
Specifikacija geometrijskih veličin izdelka - Preskusi za sprejemljivost in ponovno overjanje koordinatnih merilnih strojev (KMS) - 7. del: Koordinatni merilni stroji, opremljeni s sistemi za obdelavo slik (ISO 10360-7:2011)

Geometrical product specifications (GPS) - Acceptance and reverification tests for coordinate measuring machines (CMM) - Part 7: CMMs equipped with imaging probing systems (ISO 10360-7:2011)

Geometrische Produktspezifikation (GPS) - Annahmeprüfung und Bestätigungsprüfung für Koordinatenmessgeräte (KMG) - Teil 7: KMG mit Bildverarbeitungssystemen (ISO 10360-7:2011)

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-3a0d11818181/iso-10360-7-2011>

Spécification géométrique des produits (GPS) - Essais de réception et de vérification périodique des machines à mesurer tridimensionnelles (MMT) - Partie 7: MMT équipés de systèmes de mesurage imageurs (ISO 10360-7:2011)

Ta slovenski standard je istoveten z: EN ISO 10360-7:2011

ICS:

17.040.30	Merila	Measuring instruments
17.040.40	Specifikacija geometrijskih veličin izdelka (GPS)	Geometrical Product Specification (GPS)

SIST EN ISO 10360-7:2011

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 10360-7:2011

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 10360-7

June 2011

ICS 17.040.30

English Version

**Geometrical product specifications (GPS) - Acceptance and
reverification tests for coordinate measuring machines (CMM) -
Part 7: CMMs equipped with imaging probing systems (ISO
10360-7:2011)**

Spécification géométrique des produits (GPS) - Essais de
réception et de vérification périodique des machines à
mesurer tridimensionnelles (MMT) - Partie 7: MMT
équipées de systèmes de palpage imageurs (ISO 10360-
7:2011)

Geometrische Produktspezifikation (GPS) - Annahme- und
Bestätigungsprüfung für Koordinatenmessgeräte (KMG) -
Teil 7: KMG mit Bildverarbeitungssystemen (ISO 10360-
7:2011)

This European Standard was approved by CEN on 7 April 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....3

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[SIST EN ISO 10360-7:2011](https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011)
<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

Foreword

This document (EN ISO 10360-7:2011) has been prepared by Technical Committee ISO/TC 213 “Dimensional and geometrical product specifications and verification” in collaboration with Technical Committee CEN/TC 290 “Dimensional and geometrical product specification and verification” the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2011, and conflicting national standards shall be withdrawn at the latest by December 2011.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW Endorsement notice (standards.iteh.ai)

The text of ISO 10360-7:2011 has been approved by CEN as a EN ISO 10360-7:2011 without any modification.

[SIST EN ISO 10360-7:2011](https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011)

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 10360-7:2011](#)

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

INTERNATIONAL
STANDARDISO
10360-7First edition
2011-06-01

**Geometrical product specifications
(GPS) — Acceptance and reverification
tests for coordinate measuring machines
(CMM) —**

Part 7:

**CMMs equipped with imaging probing
systems****(standards.iteh.ai)***Spécification géométrique des produits (GPS) — Essais de réception et
de vérification périodique des machines à mesurer tridimensionnelles
(MMT) —*<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1>**Partie 7: MMT équipées de systèmes de palpée imageurs**Reference number
ISO 10360-7:2011(E)

© ISO 2011

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 10360-7:2011](https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011)

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction.....	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Symbols	7
5 Environmental and metrological requirements	8
5.1 Environmental conditions	8
5.2 Operating conditions	8
5.3 Requirements for various configuration imaging probe CMMs	8
5.3.1 General	8
5.3.2 Length measurement errors	9
5.3.3 Probing errors	10
5.3.4 Repeatability range of the length measurement error, R_B or R_U	10
5.3.5 Workpiece loading effects	10
6 Acceptance tests and reverification tests	11
6.1 General	11
6.2 Length measurement errors	11
6.2.1 General	11
6.2.2 Measuring equipment	12
6.2.3 Length measurement error, E_B or E_U	13
6.2.4 Z length measurement error, E_{BZ} or E_{UZ}	14
6.2.5 XY length measurement error, E_{BXY} or E_{UXY}	15
6.2.6 Imaging probe length measurement error, E_{BV} or E_{UV}	15
6.3 Squareness error, E_{SQ}	16
6.3.1 General	16
6.3.2 Measuring equipment	16
6.3.3 Measuring positions	17
6.3.4 Measuring procedure	17
6.3.5 Derivation of test results	18
6.4 Repeatability range of the length measurement error, R_B or R_U	19
6.5 Probing performance (P_{F2D})	19
6.5.1 Principle	19
6.5.2 Measuring equipment	19
6.5.3 Procedure	19
6.5.4 Derivation of test results	21
6.6 Probing error of the imaging probe, P_{FV2D}	21
6.6.1 Principle	21
6.6.2 Measuring equipment	21
6.6.3 Procedure	21
6.6.4 Derivation of test results	21
7 Compliance with specifications	22
7.1 Acceptance test	22
7.1.1 Acceptance criteria	22
7.1.2 Data rejection and repeated measurements	24
7.2 Reverification test	24
8 Applications	24

ISO 10360-7:2011(E)

8.1	Acceptance test	24
8.2	Reverification test.....	25
8.3	Interim check.....	25
9	Indication in product documentation and data sheets.....	26
Annex A (informative)	Interim check.....	27
Annex B (normative)	Artefacts that represent a calibrated test length.....	28
Annex C (informative)	Alternative method for checking the squareness error.....	34
Annex D (normative)	Mathematical adjustments to low CTE artefacts.....	36
Annex E (informative)	Relation to the GPS matrix model.....	38
	Bibliography.....	39

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 10360-7:2011](https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011)

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10360-7 was prepared by Technical Committee ISO/TC 213, *Dimensional and geometrical product specifications and verification*.

ISO 10360 consists of the following parts, under the general title *Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM)*:

- *Part 1: Vocabulary*
- *Part 2: CMMs used for measuring linear dimensions*
- *Part 3: CMMs with the axis of a rotary table as the fourth axis*
- *Part 4: CMMs used in scanning measuring mode*
- *Part 5: CMMs using single and multiple stylus contacting probing systems*
- *Part 6: Estimation of errors in computing of Gaussian associated features*
- *Part 7: CMMs equipped with imaging probing systems*
- *Part 9: CMMs with multiple probing systems*

The following part is under preparation:

- *Part 8: CMMs with optical distance sensors*

ISO 10360-7:2011(E)**Introduction**

This part of ISO 10360 is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO/TR 14638). It influences chain link 5 of the chains of standards on size, distance, radius, angle, form, orientation, location, run-out and datums. For more detailed information on the relation of this part of ISO 10360 to other standards and the GPS matrix model, see Annex E.

The tests of this part of ISO 10360 have two technical objectives:

- a) to test the error of indication of a calibrated test length using an imaging probing system;
- b) to test the errors in the imaging probing system.

The benefits of these tests are that the measured result has a direct traceability to the unit length, the meter, and that it gives information on how the CMM will perform on similar length measurements.

The structure of this part of ISO 10360 parallels that of ISO 10360-2, which is for CMMs equipped with contact probing systems. The testing methodology between these two parts of ISO 10360 is intentionally similar. The differences that exist may be eliminated in future revisions of either this part of ISO 10360 or ISO 10360-2.

All the definitions in Clause 3 will appear in the revision of ISO 10360-1:2000.

ITEH STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 10360-7:2011](https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011)

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) —

Part 7: CMMs equipped with imaging probing systems

1 Scope

This part of ISO 10360 specifies the acceptance tests for verifying the performance of a coordinate measuring machine (CMM) used for measuring linear dimensions as stated by the manufacturer. It also specifies the reverification tests that enable the user to periodically reverify the performance of the CMM.

The acceptance and reverification tests given in this part of ISO 10360 are applicable only to Cartesian CMMs using imaging probing systems of any type operating in the discrete-point probing mode.

This part of ISO 10360 does not explicitly apply to:

- non-Cartesian CMMs; however, parties may apply this part of ISO 10360 to non-Cartesian CMMs by mutual agreement;
<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6c1c418511b1/iso-10360-7:2011>
- CMMs using other types of optical probing; however, parties may apply this approach to other optical CMMs by mutual agreement;
- CMMs using contact probing systems (see ISO 10360-2 for contact probing systems).

This part of ISO 10360 specifies performance requirements that can be assigned by the manufacturer or the user of a CMM, the manner of execution of the acceptance and reverification tests to demonstrate the stated requirements, rules for proving conformance, and applications for which the acceptance and reverification tests can be used.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10360-1:2000, *Geometrical Product Specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 1: Vocabulary*

ISO 10360-2:2009, *Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 2: CMMs used for measuring linear dimensions*

ISO 14253-1:1998, *Geometrical Product Specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformance or non-conformance with specifications*

ISO 10360-7:2011(E)

ISO 14660-1:1999, *Geometrical Product Specifications (GPS) — Geometrical features — Part 1: General terms and definitions*

ISO/TS 23165:2006, *Geometrical product specifications (GPS) — Guidelines for the evaluation of coordinate measuring machine (CMM) test uncertainty*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10360-1, ISO 10360-2, ISO 14253-1, ISO 14660-1, ISO/TS 23165, ISO/IEC Guide 99 and the following apply.

3.1**imaging probing system**

probing system which creates measurement points through the use of an imaging system

NOTE 1 This part of ISO 10360 is primarily concerned with imaging probing systems that enable measurements in the lateral direction to the probing system axis.

NOTE 2 A video or vision probing system is an imaging probing system.

3.2**imaging probe CMM**

CMM equipped with an imaging probing system

iTeh STANDARD PREVIEW
(standards.iteh.ai)

3.3**field of view**

FOV

area viewed by the imaging probing system

SIST EN ISO 10360-7:2011

<https://standards.iteh.ai/catalog/standards/sist/365e5779-5e4e-421d-b41e-6ae1f038512a/sist-en-iso-10360-7-2011>

See Figure 1.

NOTE The measuring limits, or size, of the FOV are stated as the limits of the object space that is reproduced in the final image.

3.4**measuring window**

region of interest in the FOV that is used in the determination of the measured point(s)

See Figure 1.

NOTE Configurations of measuring windows may vary widely between various imaging probe CMMs and for different measuring applications on the same imaging probe CMM.

3.5**measuring plane (of the imaging probing system)**

two-dimensional plane defined by the FOV of an imaging probing system

3.6**coefficient of thermal expansion**

CTE

α


linear thermal expansion coefficient of a material at 20 °C

3.7**normal CTE material**

material with a CTE between $8 \times 10^{-6}/^{\circ}\text{C}$ and $13 \times 10^{-6}/^{\circ}\text{C}$

3.8**test circle**

circular material standard used for acceptance test and reverification test

**Key**

- 1 camera or other device for capturing an image of the measured object
- 2 various optical elements of the imaging probing system
- 3 measured object
- 4 FOV (object)
- 5 FOV (image)
- 6 measuring window
- 7 measured point

Figure 1 — Imaging probing system

3.9**bidirectional length measurement error**

E_B

error of indication when measuring a calibrated bidirectional test length using an imaging probe CMM with a single probing point (or equivalent) at each end of the calibrated test length

NOTE E_B is applicable only to imaging probe CMMs that are capable of three-dimensional spatial measurements, which may not always be the case.