



# INTERNATIONAL STANDARD ISO 7493:1997 TECHNICAL CORRIGENDUM 1

Published 1999-07-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

## Dental operator's stool

### TECHNICAL CORRIGENDUM 1

*Siège d'opérateur dentaire*

*RECTIFICATIF TECHNIQUE 1*

## iTeh STANDARD PREVIEW

Technical Corrigendum 1 to International Standard ISO 7493:1997 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*.

[ISO 7493:1997/Cor 1:1999](https://standards.iteh.ai/catalog/standards/sist/31006904-1338-4cdc-ac87-606003eba250/iso-7493-1997-cor-1-1999)

<https://standards.iteh.ai/catalog/standards/sist/31006904-1338-4cdc-ac87-606003eba250/iso-7493-1997-cor-1-1999>

*Page 3*

Replace the text in subclause 4.3 by the following wording:

#### **4.3 Seat height and range of adjustment**

The minimum adjustable height of the loaded seat ( $h_1$ ) and the range of adjustment of the seat height shall be given by the manufacturer with the instructions (see clause 6).

Compliance shall be checked by visual inspection.