



SLOVENSKI STANDARD

SIST ISO 18073:2010

01-september-2010

Kakovost vode - Določevanje tetra- do okta-kloriranih dioksinov in furanov - Metoda z izotopnim razredčenjem in plinsko kromatografijo visoke ločljivosti z masno selektivnim detektorjem visoke ločljivosti (HRGC/HRMS)

Water quality - Determination of tetra- to octa-chlorinated dioxins and furans - Method using isotope dilution HRGC/HRMS

iTeh STANDARD PREVIEW

(standards.iteh.ai)

Qualité de l'eau - Dosage des dioxines et furanes tétra- à octachlorés - Méthode par dilution d'isotopes HRGC/SMHR

[SIST ISO 18073:2010](https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010)

[https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-](https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010)

[693a52790dc0/sist-iso-18073-2010](https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010)

Ta slovenski standard je istoveten z: ISO 18073:2004

ICS:

13.060.50

Preiskava vode na kemične snovi

Examination of water for chemical substances

SIST ISO 18073:2010

en,fr

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 18073:2010](#)

<https://standards.iteh.ai/catalog/standards/sist/f16c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>

INTERNATIONAL STANDARD

ISO
18073

First edition
2004-04-15

Water quality — Determination of tetra- to octa-chlorinated dioxins and furans — Method using isotope dilution HRGC/HRMS

*Qualité de l'eau — Dosage des dioxines et furanes tétra- à
octachlorés — Méthode par dilution d'isotopes HRGC/SMHR*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 18073:2010

<https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>



Reference number
ISO 18073:2004(E)

© ISO 2004

ISO 18073:2004(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 18073:2010](https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010)

<https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	v
1 Scope.....	1
2 Normative references	1
3 Terms, definitions and abbreviated terms.....	2
3.1 Terms and definitions.....	2
3.2 Abbreviated terms.....	3
4 Principle	4
4.1 Spiking and extraction.....	4
4.2 Clean-up	4
4.3 Concentration	4
4.4 Identification	5
4.5 Quantification	6
4.6 Analytical quality.....	6
5 Contamination and interferences	6
6 Reagents and standards	8
7 Apparatus and materials	15
7.1 Water sample containers.....	15
7.2 Equipment for sample preparation	15
7.3 Extraction apparatus	15
7.4 Filtration apparatus.....	15
7.5 Clean-up apparatus.....	16
7.6 Concentration apparatus.....	16
7.7 Other equipment	17
8 Sample collection, preservation, storage and holding times.....	17
9 Quality assurance/quality control	18
9.1 General	18
9.2 Initial precision and recovery (IPR).....	20
9.3 Spiking	20
9.4 Method blanks	21
9.5 Quality control check sample	21
9.6 Method precision	21
10 Calibration.....	22
10.1 Operating conditions	22
10.2 Mass spectrometer (MS) resolution	23
10.3 QC limits and theoretical ion abundance ratios for standards	25
10.4 Retention time	25
10.5 Isomer specificity.....	26
10.6 Calibration by isotope dilution	27
10.7 Calibration by internal standard	28
10.8 Combined calibration	29
11 Sample preparation.....	29
11.1 General	29
11.2 Determination of suspended solids	29
11.3 Preparation of aqueous samples containing 1 % or less by mass suspended solids	30
12 Extraction and concentration	31
12.1 Extraction of samples with no visible particulates	31
12.2 Soxhlet extraction of filters and/or disks	31

ISO 18073:2004(E)

12.3	Back-extraction with acid and base	32
12.4	Macro-concentration	33
12.5	Micro-concentration and solvent exchange	34
13	Extract clean-up.....	35
13.1	General	35
13.2	Gel-permeation chromatography (GPC)	35
13.3	Silica clean-up	36
13.4	Alumina clean-up.....	37
13.5	Carbon column	37
13.6	Florisil clean-up	38
14	HRGC/HRMS analysis	38
15	System and laboratory performance.....	38
15.1	General	38
15.2	MS resolution.....	38
15.3	Calibration verification	39
15.4	GC resolution.....	39
15.5	Blank.....	39
16	Qualitative determination	40
17	Quantitative determination.....	40
17.1	Concentration calculations for isotope dilution	40
17.2	Internal standard recovery and concentration calculations.....	41
17.3	Concentration in the aqueous sample	42
17.4	Toxicity equivalents (TEQs).....	42
18	Analysis of complex samples	44
18.1	General	44
18.2	Recovery of internal standards.....	44
19	Results and reporting	45
Annex A (informative)	Application of HRGC/HRMS	46
Annex B (informative)	Interlaboratory trial data	51
Annex C (informative)	Pollution prevention and waste management.....	56
Annex D (informative)	Additional extraction and clean up techniques.....	57
Bibliography.....		61

iteh STANDARD PREVIEW

(standards.iteh.ai)

SIST ISO 18073:2010

<https://standards.iteh.ai/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18073 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 18073:2010](https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010)

<https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 18073:2010](#)

<https://standards.iteh.ai/catalog/standards/sist/f16c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>

Water quality — Determination of tetra- to octa-chlorinated dioxins and furans — Method using isotope dilution HRGC/HRMS

WARNING — Persons using this International Standard should be familiar with normal laboratory practice. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

Attention is drawn to any relevant national safety regulations. The 2,3,7,8-chloro-substituted PCDDs/PCDFs are among the most toxic of chemicals. All work with PCDDs/PCDFs requires, therefore, the utmost care; the national safety measures which correspond to those for toxic substances shall be strictly adhered to.

1 Scope

This International Standard specifies a method for the determination of tetra- to octa-chlorinated dibenzo-*p*-dioxins (PCDDs) and dibenzofurans (PCDFs) in waters and waste waters (containing less than 1 % by mass solids) using high-resolution gas chromatography/high-resolution mass spectrometry (HRGC/HRMS).

This International Standard is applicable to the seventeen 2,3,7,8-substituted PCDDs/PCDFs specified in Table 1.

SIST ISO 18073:2010

The detection limits and quantitation levels in this method are usually dependent on the level of interferences rather than instrumental limitations. The minimum levels (MLs) specified in Table 2 are the levels at which the PCDDs/PCDFs can be determined with no interferences present. The method detection limit (MDL) for 2,3,7,8-TCDD has been determined as 4,4 pg/l based on this method using a sample volume of 1 l. Lower detection limits may be achieved by using a larger sample volume.

This method is “performance based”. The analyst is permitted to modify the method to overcome interferences or lower the cost of measurements, provided that all performance criteria in this International Standard are met. The requirements for establishing method equivalency are given in 9.1.2.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes*

ISO 5667-2, *Water quality — Sampling — Part 2: Guidance on sampling techniques*

ISO 6879:1995, *Air Quality — Performance characteristics and related concepts for air quality measuring methods*.

ISO 18073:2004(E)

3 Terms, definitions and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 6879:1995 and the following apply.

3.1.1

analyte

PCDD or PCDF tested for by this method

Note See Table 1 for a list of compounds.

3.1.2

calibration standard

solution prepared from a secondary standard and/or stock solutions and used to calibrate the response of the instrument with respect to analyte concentration

3.1.3

calibration verification standard

midpoint calibration standard that is used to verify calibration

3.1.4

congener

any one of the 210 individual PCDDs/ PCDFs

3.1.5

internal standard

¹³C₁₂-labelled 2,3,7,8-PCDD/PCDF analogue added to samples prior to extraction against which the concentrations of native PCDDs and PCDFs are calculated

[SIST ISO 18073:2010](https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010)

3.1.6

keeper

high boiling-point solvent added to the sampling standard solution

3.1.7

method blank

aliquot of reagent water that is treated exactly as a sample through the complete analytical procedure including extraction, clean-up, identification and quantification including all the relevant reagents and materials

3.1.8

operational performance characteristics

influence of the physical and chemical environment and maintenance problems, for example; mains voltage, temperature, supply of certain substances, set-up time, period of unattended operation

[See ISO 6879:1995]

3.1.9

pattern

chromatographic print of any series of PCDD/PCDF isomers

3.1.10

PCDD/PCDF isomers

PCDDs or PCDFs with identical chemical compositions but different structures

3.1.11

profile

graphic representation of the sums of the isomer concentrations of the PCDDs and the PCDFs

3.1.12**recovery standard**

$^{13}\text{C}_{12}$ -labelled 2,3,7,8-chloro-substituted PCDD/PCDF, added before injection into the GC

3.1.13**spiking**

addition of $^{13}\text{C}_{12}$ -labelled PCDD/PCDF standards

3.1.14**statistical performance characteristics**

quantification, for measured values, of the possible deviations resulting from the random part of the measuring process, e.g. repeatability or instability

[See ISO 6879:1995]

3.2 Abbreviated terms

DCDPE	decachlorodiphenyl ether
GC/MS	gas chromatography/mass spectrometry
GPC	gel-permeation chromatography
HpCDD	heptachlorodibenzo- <i>p</i> -dioxin
HpCDF	heptachlorodibenzofuran
HpCDPE	heptachlorodiphenyl ether
HPLC	high-performance liquid chromatography
HRGC	high-resolution gas chromatography
HRMS	high-resolution mass spectrometry
HxCDD	hexachlorodibenzo- <i>p</i> -dioxin
HxCDF	hexachlorodibenzofuran
HxCDPE	hexachlorodiphenyl ether
MDL	method detection limit
ML	minimum level (see Table 2)
NCDPE	nonachlorodiphenyl ether
OCDD	octachlorodibenzo- <i>p</i> -dioxin
OCDF	octachlorodibenzofuran
 OCDPE	octachlorodiphenyl ether
PCDD/PCDF	polychlorinated dibenzo- <i>p</i> -dioxin/dibenzofuran
PeCDD	pentachlorodibenzo- <i>p</i> -dioxin
PeCDF	pentachlorodibenzofuran

ISO 18073:2004(E)

PTFE	polytetrafluoroethylene
SIM	selected ion monitoring
TCDD	tetrachlorodibenzo- <i>p</i> -dioxin
TCDF	tetrachlorodibenzofuran
TEF	toxic equivalent factor
TEQ	toxic equivalent

4 Principle**4.1 Spiking and extraction**

Internal standards, analogues of the 2,3,7,8-substituted PCDDs/PCDFs labelled with a stable isotope (see Table 1) in a suitable solvent such as acetone, are spiked into a 1 l aqueous sample containing less than 1 % by mass solids. A minimum of one labelled standard per homologue group is used and the sample is extracted by one of two procedures as specified in 4.1 a) or 4.1 b).

- a) Samples containing no visible particles are extracted with dichloromethane in a separatory funnel or by solid-phase extraction. The extract is concentrated for clean-up.
- b) Samples containing visible particles are vacuum filtered through a glass-fibre filter. The filter is extracted in a Soxhlet extractor using toluene and the filtrate is extracted with dichloromethane in a separatory funnel. The dichloromethane extract is concentrated and combined with the Soxhlet extract prior to clean-up.

Other solvents and extraction techniques may be substituted, provided that all the performance criteria can be met.

<https://standards.iteh.ai/catalog/standards/sist/fl6c3929-e839-4d8f-bf50-693a52790dc0/sist-iso-18073-2010>

4.2 Clean-up

After extraction, sample extracts are cleaned up to remove interfering components. Sample clean-ups may include back-extraction with acid and/or base, and gel-permeation, alumina, silica, Florisil¹⁾ or activated-carbon chromatography.

4.3 Concentration

After clean-up, the extract is concentrated to near dryness. Prior to injection, recovery standards are added to each extract, and an aliquot of the extract is injected into the gas chromatograph. The analytes are separated by the GC and detected by a high-resolution mass spectrometer. Two exact masses are monitored for each analyte.

Resolution equal to or greater than or 10 000 is recommended. High-resolution gas chromatography/high-resolution mass spectrometry at a resolution equal to or greater than 10 000 is, at present, required to achieve adequate sensitivity and selectivity, and to allow the use of all ¹³C₁₂-labelled standards. Resolution in the range of 6 000 to 10 000 should be acceptable if the absence of interferences has been documented. If a determination of PCDD/PCDF homologue group totals is required, then a resolution of 10 000 is necessary. At resolutions less than 10 000, some ¹³C₁₂-PCDFs interfere with native PCDDs of the same level of chlorination.

1) Florisil is an example of a suitable product available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO of this product.

Table 1 — PCDDs and PCDFs analyzed by this method

PCDDs/PCDFs	CAS registry	Labelled analogue	CAS registry
2,3,7,8-TCDD	1746-01-6	$^{13}\text{C}_{12}$ -2,3,7,8-TCDD	76523-40-5
–	–	$^{13}\text{C}_{12}$ -1,2,3,4-TCDD	–
Total TCDD	41903-57-5	–	–
2,3,7,8-TCDF	51207-31-9	$^{13}\text{C}_{12}$ -2,3,7,8-TCDF	89059-46-1
Total TCDF	55722-27-5	–	–
1,2,3,7,8-PeCDD	40321-76-4	$^{13}\text{C}_{12}$ -1,2,3,7,8-PeCDD	109719-79-1
Total PeCDD	36088-22-9	–	–
1,2,3,7,8-PeCDF	57117-41-6	$^{13}\text{C}_{12}$ -1,2,3,7,8-PeCDF	109719-77-9
2,3,4,7,8-PeCDF	57117-31-4	$^{13}\text{C}_{12}$ -2,3,4,7,8-PeCDF	116843-02-8
Total PeCDF	30402-15-4	–	–
1,2,3,4,7,8-HxCDD	39227-28-6	$^{13}\text{C}_{12}$ -1,2,3,4,7,8-HxCDD	109719-80-4
1,2,3,6,7,8-HxCDD	57653-85-7	$^{13}\text{C}_{12}$ -1,2,3,6,7,8-HxCDD	109719-81-5
1,2,3,7,8,9-HxCDD	19408-74-3	$^{13}\text{C}_{12}$ -1,2,3,7,8,9-HxCDD	109719-82-6
Total HxCDD	34465-46-8	–	–
1,2,3,4,7,8-HxCDF	70648-26-9	$^{13}\text{C}_{12}$ -1,2,3,4,7,8-HxCDF	114423-98-2
1,2,3,6,7,8-HxCDF	57117-44-9	$^{13}\text{C}_{12}$ -1,2,3,6,7,8-HxCDF	116843-03-9
1,2,3,7,8,9-HxCDF	72918-21-9	$^{13}\text{C}_{12}$ -1,2,3,7,8,9-HxCDF	116843-04-0
2,3,4,6,7,8-HxCDF	60851-34-5	$^{13}\text{C}_{12}$ -2,3,4,6,7,8-HxCDF	116843-05-1
Total HxCDF	55684-94-1	–	–
1,2,3,4,6,7,8-HpCDD	35822-46-9	$^{13}\text{C}_{12}$ -1,2,3,4,6,7,8-HpCDD	109719-83-7
Total HpCDD	37871-00-4	–	–
1,2,3,4,6,7,8-HpCDF	67562-39-4	$^{13}\text{C}_{12}$ -1,2,3,4,6,7,8-HpCDF	109719-84-8
1,2,3,4,7,8,9-HpCDF	55673-89-7	$^{13}\text{C}_{12}$ -1,2,3,4,7,8,9-HpCDF	109719-94-0
Total HpCDF	38998-75-3	–	–
OCDD	3268-87-9	$^{13}\text{C}_{12}$ -OCDD	114423-97-1
OCDF	39001-02-0	$^{13}\text{C}_{12}$ -OCDF	109719-78-0

4.4 Identification

An individual PCDD/PCDF is identified by comparing the GC retention time and the ion-abundance ratio of the two exact masses monitored with the respective retention time of an authentic standard and the theoretical or analytical ion-abundance ratio of the two exact masses.

The non-2,3,7,8-substituted isomers and congeners are identified by the agreement of their retention times and ion-abundance ratios with accepted values within predefined limits.

ISO 18073:2004(E)**4.5 Quantification**

Quantitative analysis is performed using selected-ion monitoring (SIM) in one of three ways.

- a) For the 2,3,7,8-substituted PCDDs/PCDFs for which labelled analogues have been added to the sample (4.1), the isotope-dilution technique is used to calibrate the GC/MS system and to determine the concentration of each compound.
- b) For the 2,3,7,8-substituted PCDDs/PCDFs for which labelled analogues are not added to the samples prior to extraction and for the labelled internal standards themselves, the internal-standard technique is used to calibrate the GC/MS system and to determine the concentration of each compound.
- c) For non-2,3,7,8-substituted isomers and for all isomers at a given level of chlorination (i.e. total TCDDs), concentrations are determined using response factors from the calibration of the PCDDs/PCDFs at the same level of chlorination.

4.6 Analytical quality

The quality of the analysis is assured through reproducible calibration and testing of the procedures for extraction, clean-up, and GC/MS operation.

5 Contamination and interferences

5.1 Where possible, purify reagents by extraction or solvent rinse.

Solvents, reagents, glassware, and other sample processing hardware can yield artefacts and/or elevated baselines causing misinterpretation of chromatograms. Specific selection of reagents and purification of solvents by distillation in all-glass systems may be required.

5.2 Clean glassware such that the method blank requirements of this International Standard are met (9.4.3). An example of a cleaning procedure is given in 5.2 a) to 5.2 c).

- a) Disassemble glassware with removable parts, particularly separatory funnels with fluoropolymer stopcocks, prior to washing with detergent. Rinse glassware with solvent and wash with a detergent solution as soon after use as is practical. Sonication of glassware in a detergent solution for approximately 30 s may aid in cleaning.
- b) After washing with detergent, rinse glassware immediately, first with methanol, then with hot tap water. The tap water rinse shall be followed by another methanol rinse, then acetone, and then dichloromethane.
- c) Immediately prior to use, pre-extract the Soxhlet apparatus with toluene for approximately 3 h. Shake separatory funnels with a dichloromethane/toluene mixture (80/20 by volume) for 2 min, drain, and then shake with pure dichloromethane for 2 min.

Proper cleaning of glassware is extremely important, because glassware not only can contaminate the samples but also can remove the analytes of interest by adsorption on the glass surface.

5.3 Demonstrate that all materials used in the analysis are free from interferences by running reference-matrix method blanks initially and with each sample batch (samples, up to a maximum of 20, processed through the extraction procedure on a given 12-h shift).

5.4 The reference matrix shall simulate, as closely as possible, the sample matrix under test. Ideally, the reference matrix shall not contain the PCDDs/PCDFs in detectable amounts, but shall contain potential interferences in the concentrations expected to be found in the samples to be analyzed.

Table 2 — Suggested quantitation relationships for individual PCDDs/PCDFs

PCDDs/PCDFs	Reference for retention time and quantitation	Relative retention time ^a	Minimum level ^{a,b} in water pg/l	Minimum level ^{a,b} in extract pg/μl
2,3,7,8-TCDF	¹³ C ₁₂ -2,3,7,8-TCDF	0,999 to 1,003	10	0,5
2,3,7,8-TCDD	¹³ C ₁₂ -2,3,7,8-TCDD	0,999 to 1,002	10	0,5
1,2,3,7,8-PeCDF	¹³ C ₁₂ -1,2,3,7,8-PeCDF	0,999 to 1,002	50	2,5
2,3,4,7,8-PeCDF	¹³ C ₁₂ -2,3,4,7,8-PeCDF	0,999 to 1,002	50	2,5
1,2,3,7,8-PeCDD	¹³ C ₁₂ -1,2,3,7,8-PeCDD	0,999 to 1,002	50	2,5
¹³ C ₁₂ -2,3,7,8-TCDF	¹³ C ₁₂ -1,2,3,4-TCDD	0,923 to 1,103	50	2,5
¹³ C ₁₂ -2,3,7,8-TCDD	¹³ C ₁₂ -1,2,3,4-TCDD	0,976 to 1,043	50	2,5
¹³ C ₁₂ -1,2,3,7,8-PeCDF	¹³ C ₁₂ -1,2,3,4-TCDD	1,000 to 1,425	50	2,5
¹³ C ₁₂ -2,3,4,7,8-PeCDF	¹³ C ₁₂ -1,2,3,4-TCDD	1,011 to 1,526	50	2,5
¹³ C ₁₂ -1,2,3,7,8-PeCDD	¹³ C ₁₂ -1,2,3,4-TCDD	1,000 to 1,567	50	2,5
1,2,3,4,7,8-HxCDF	¹³ C ₁₂ -1,2,3,4,7,8-HxCDF	0,999 to 1,001	50	2,5
1,2,3,6,7,8-HxCDF	¹³ C ₁₂ -1,2,3,6,7,8-HxCDF	0,997 to 1,005	50	2,5
1,2,3,7,8,9-HxCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDF	0,999 to 1,001	50	2,5
2,3,4,6,7,8-HxCDF	¹³ C ₁₂ -2,3,4,6,7,8-HxCDF	0,999 to 1,001	50	2,5
1,2,3,4,7,8-HxCDD	¹³ C ₁₂ -1,2,3,4,7,8-HxCDD	0,999 to 1,001	50	2,5
1,2,3,6,7,8-HxCDD	¹³ C ₁₂ -1,2,3,6,7,8-HxCDD	0,998 to 1,004	50	2,5
1,2,3,7,8,9-HxCDD	See note ^c	1,000 to 1,019	50	2,5
1,2,3,4,6,7,8-HpCDF	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF	0,999 to 1,001	50	2,5
1,2,3,4,7,8,9-HpCDF	¹³ C ₁₂ -1,2,3,4,7,8,9-HpCDF	0,999 to 1,001	50	2,5
1,2,3,4,6,7,8-HpCDD	¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD	0,999 to 1,001	50	2,5
OCDD	¹³ C ₁₂ -OCDD	0,999 to 1,001	100	5,0
OCDF	¹³ C ₁₂ -OCDD	0,999 to 1,008	100	5,0
¹³ C ₁₂ -1,2,3,4,7,8-HxCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	0,944 to 0,970	—	—
¹³ C ₁₂ -1,2,3,6,7,8-HxCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	0,949 to 0,975	—	—
¹³ C ₁₂ -1,2,3,7,8,9-HxCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	0,977 to 1,047	—	—
¹³ C ₁₂ -2,3,4,6,7,8-HxCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	0,959 to 1,021	—	—
¹³ C ₁₂ -1,2,3,4,7,8-HxCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	0,977 to 1,000	—	—
¹³ C ₁₂ -1,2,3,6,7,8-HxCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	0,981 to 1,003	—	—
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	1,043 to 1,085	—	—
¹³ C ₁₂ -1,2,3,4,7,8,9-HpCDF	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	1,057 to 1,151	—	—
¹³ C ₁₂ -1,2,3,4,6,7,8-HpCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	1,086 to 1,110	—	—
¹³ C ₁₂ -OCDD	¹³ C ₁₂ -1,2,3,7,8,9-HxCDD	1,032 to 1,311	—	—

^a Minimum levels and relative retention times are given for guidance only.

^b The minimum level (ML) for each analyte is defined as the level for which the entire analytical system shall give a recognizable signal and acceptable calibration point. It is equivalent to the concentration of the lowest calibration standard, assuming that all method-specific sample masses, volumes and clean-up procedures have been used, i.e. based on 1 l of sample.

^c The retention time reference for 1,2,3,7,8,9-HxCDD is ¹³C₁₂-1,2,3,6,7,8-HxCDD; 1,2,3,7,8,9-HxCDD is quantified using the averaged responses for ¹³C₁₂-1,2,3,4,7,8-HxCDD and ¹³C₁₂-1,2,3,6,7,8-HxCDD.