
**Surface chemical analysis —
Secondary-ion mass spectrometry —
Determination of relative sensitivity
factors from ion-implanted reference
materials**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

*Analyse chimique des surfaces — Spectrométrie de masse des ions
secondaires — Détermination des facteurs de sensibilité relative à
l'aide de matériaux de référence à ions implantés*

ISO 18114:2003

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>



Reference number
ISO 18114:2003(E)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 18114:2003

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references.....	1
3 Terms and definitions	1
4 Symbols and abbreviated terms	1
5 Principle	2
6 Apparatus.....	2
7 Ion-implanted reference materials	2
8 Procedure.....	2
9 Test report.....	3
Bibliography.....	4

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 18114:2003

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18114 was prepared by Technical Committee ISO/TC 201, *Surface chemical analysis*, Subcommittee SC 6, *Secondary ion mass spectrometry*.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 18114:2003](https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003)

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>

Introduction

Ion-implanted materials are commonly used in secondary-ion mass spectrometry for the calibration of instruments. This international Standard was prepared to provide a uniform method for determining the relative sensitivity factor of an element in a specified matrix from an ion-implanted reference material, and to show how the concentration of the element in a different sample of the same matrix material can be determined.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 18114:2003

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

ISO 18114:2003

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>

Surface chemical analysis — Secondary-ion mass spectrometry — Determination of relative sensitivity factors from ion-implanted reference materials

1 Scope

This International Standard specifies a method of determining relative sensitivity factors (RSFs) for secondary-ion mass spectrometry (SIMS) from ion-implanted reference materials.

The method is applicable to specimens in which the matrix is of uniform chemical composition, and in which the peak concentration of the implanted species does not exceed one atomic percent.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 18115, *Surface chemical analysis — Vocabulary*

[ISO 18114:2003](https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003)

<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18115 apply.

4 Symbols and abbreviated terms

$C_i^{A,M}$	the atomic concentration of the analyte element A in the matrix M at cycle i of a depth profile, expressed in atoms per unit volume
d	the depth over which the depth profile is integrated, expressed in length units
$I_i^{A_j}$	the detected count rates of the analyte ion of isotope A_j at measurement cycle i , expressed in counts/s
$I_i^{M_k}$	the detected count rates of the reference isotope M_k at measurement cycle i , expressed in counts/s
I_{BG}	the mean background count rate of species A_j , expressed in counts/s
N^{A_j}	the fractional isotopic abundance of the analyte isotope A_j in the unknown sample
n	the number of cycles over which the depth profile is integrated
Φ	the implanted fluence of isotope A_j , expressed in atoms per unit area
RSF	the relative sensitivity factor, expressed in atoms per unit volume
SIMS	secondary-ion mass spectrometry

5 Principle

An isotopic relative sensitivity factor (RSF) for a particular element-matrix combination is derived from a SIMS depth profile of an ion-implanted external standard. This RSF can then be used to quantify the concentration of the same element as a function of depth in a different sample of the same matrix material, using the equation:

$$C_i^{A,M} = \frac{\text{RSF}^{A_j, M_k} \times I_i^{A_j}}{I_i^{M_k} \times N^{A_j}} \quad (1)$$

This procedure only applies to samples and reference materials in which the detected count rate of the analyte ion is directly proportional to its concentration. In practice, the upper limit of analyte concentration for which this proportionality holds is generally assumed to be one atomic percent. The measurement conditions used for the analysis of reference materials and unknown samples shall be the same.

6 Apparatus

The procedure described here can be used to determine an RSF from data obtained with any SIMS instrument that is capable of obtaining depth profiles. Follow the instrument manufacturer's instructions or local documented procedures for setting up the instrument to obtain the best quality depth profile data.

iTeh STANDARD PREVIEW

7 Ion-implanted reference materials (standards.iteh.ai)

Ion-implanted reference materials for this procedure shall have peak concentrations below one atomic percent, but at least a factor of 100 greater than the background intensity or detection limit of the analyte element in the SIMS instrument. The depth of the peak in the depth distribution of the analyte shall be at least a factor of two below the depth of the onset of steady-state sputtering conditions as indicated by the stability of a matrix reference signal. Use certified reference materials (CRMs) or secondary materials derived from CRMs, when available. If no CRM is available, use reference materials for which the implanted fluence has been measured by an independent method such as Rutherford backscattering spectrometry or neutron activation analysis, if possible.

8 Procedure

Determine the count rates of the ions of analyte isotope A_j , $I_i^{A_j}$, and reference isotope M_k , $I_i^{M_k}$, at each measurement cycle i of the depth profile of the implanted reference material. Calculate the relative sensitivity factor for isotopic species A_j of element A in matrix M with isotopic reference species M_k from the equation:

$$\text{RSF}^{A_j, M_k} = \frac{\Phi \times n}{\sum_{i=1}^n \left[\frac{(I_i^{A_j} - I_{\text{BG}})}{I_i^{M_k}} \right] \times d} \quad (2)$$

The depth d is typically calculated by measuring the depth of the sputtered crater in the sample with a calibrated stylus profilometer and assuming that a constant depth is removed over each measurement cycle.

If the variation of the reference species count rate is smaller than a specified tolerance level, it may be regarded as constant and its value I^{M_k} need be measured only once during the depth profile. In this case, the equation used to calculate the RSF becomes:

$$\text{RSF}^{A_j, M_k} = \frac{\Phi \times n \times I^{M_k}}{\sum_{i=1}^n \left(I_i^{A_j} - I_{\text{BG}} \right) \times d} \quad (3)$$

9 Test report

The following information shall be recorded when a relative sensitivity factor is determined by this method:

- a) the relative sensitivity factor;
- b) the model of SIMS instrument used;
- c) the primary-beam species and energy;
- d) the secondary-ion species and polarity;
- e) the reference ion isotope and species;
- f) any special spectrometer conditions (e.g. high mass resolution, kinetic energy filtering);
- g) any special analytical conditions (e.g. oxygen flooding of surface).

ISO 18114:2003
<https://standards.iteh.ai/catalog/standards/sist/7189e51c-b825-4079-aa66-f696ca55f3e9/iso-18114-2003>