

### SLOVENSKI STANDARD SIST ISO 18653:2005

01-januar-2005



Gears -- Evaluation of instruments for the measurement of individual gears

Engrenages -- Évaluation des instruments de mesure des engrenages individuels

# Ta slovenski standard je istoveten z: ISO 18653:2003

	https://standard	SIST ISO 18653:2005 Is.iteh.ai/catalog/standards/sist/80c254ff-2c47-49c6-a3a9-	
<u>ICS:</u>	*	3e2274182911/sist-iso-18653-2005	
21.200	Gonila	Gears	
SIST ISO 1	18653:2005	en	



### iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 18653:2005</u> https://standards.iteh.ai/catalog/standards/sist/80c254ff-2c47-49c6-a3a9-3e2274182911/sist-iso-18653-2005



## INTERNATIONAL STANDARD

ISO 18653

First edition 2003-12-01

# Gears — Evaluation of instruments for the measurement of individual gears

Engrenages — Évaluation des instruments de mesure des engrenages individuels

## iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 18653:2005</u> https://standards.iteh.ai/catalog/standards/sist/80c254ff-2c47-49c6-a3a9-3e2274182911/sist-iso-18653-2005



Reference number ISO 18653:2003(E)

#### PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

### iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 18653:2005</u> https://standards.iteh.ai/catalog/standards/sist/80c254ff-2c47-49c6-a3a9-3e2274182911/sist-iso-18653-2005

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.org Web www.iso.org Published in Switzerland

### Contents

Forew	ord	iv
1	Scope	1
2	Normative references	1
3	Terms, definitions and symbols	1
4 4.1 4.2 4.3 4.4 4.5 4.6	Application General Traceability Artifacts Measurement uncertainty Sources of uncertainty Evaluation interval	4 4 4 5 5
5 5.1 5.2 5.3 5.4	Condition of the measurement system System characteristics Suitability for calibration Table load considerations Tooling and gauges	6 6 6 6
6	Environment	7
7 7.1 7.2 7.3 7.4 7.5 7.6 7.7	ArtifactsArtifact size and geometry Involute artifactsSIST ISO 18653:2005 Helix artifactsSIST iso 18653:2005 Pitch artifacts	7 8 9 . 10 . 11 . 11
8 8.1 8.2 8.3 8.4	Method for estimating measurement uncertainty Methods Comparator method Calculation of $U_{95}$ measurement uncertainty Procedure	. 13 . 14 . 14 . 14 . 15
Annex	A (normative) Artifact calibration certificate requirements	. 17
Bibliog	graphy	. 19

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18653 was prepared by Technical Committee ISO/TC 60, Gears.

### iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 18653:2005</u> https://standards.iteh.ai/catalog/standards/sist/80c254ff-2c47-49c6-a3a9-3e2274182911/sist-iso-18653-2005

# Gears — Evaluation of instruments for the measurement of individual gears

#### 1 Scope

This International Standard specifies methods for the evaluation of measuring instruments used for gear measurements of involute, helix, pitch and runout. It is applicable both to instruments that measure runout directly and to those that compute it from index measurements. It also gives recommendations for the evaluation of tooth thickness measuring instruments and, of necessity, includes the estimation of measurement uncertainty with the use of calibrated gear artifacts. It does not address the calibration of artifacts by laboratories accredited in accordance with ISO/IEC 17025; nor are its requirements intended as an acceptance specification of product gears (see ISO 1328-1, ISO 1328-2, ISO/TR 10064-1 and ISO/TR 10064-2). The estimation of product gear measurement uncertainty is beyond its scope (see ISO/TR 10064-5 for recommendations).

### 2 Normative references STANDARD PREVIEW

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies SO 18653-2005

ISO 1328-1, Cylindrical gears — ISO system of accuracy — Part 1: Definitions and allowable values of deviations relevant to corresponding flanks of gear teeth

ISO/TR 10064-3, Cylindrical gears — Code of inspection practice — Part 3: Recommendations relative to gear blanks, shaft centre distance and parallelism of axes

ISO/TR 10064-5<sup>1</sup>), Cylindrical gears — Code of inspection practice — Part 5: Recommendations relative to evaluation of gear measuring instruments

ISO 14253-1, Geometrical Product Specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformance or non-conformance with specifications

ISO/IEC 17025, General requirements for the competence of testing and calibration laboratories

#### 3 Terms, definitions and symbols

For the purposes of this document, the following terms, definitions and symbols (see Table 1) apply.

NOTE 1 The definitions, when applicable, conform to ISO 122-1, ISO 1328-1, ISO 1328-2 and ISO/TR 10064-1.

NOTE 2 The terms, definitions and symbols used in this document may differ from those used in other International Standards. The user needs to be certain of fully understanding them, as used here.

<sup>1)</sup> Under preparation.

#### 3.1

#### accuracy

closeness of agreement between a measured value and an accepted reference (or calibrated) value

#### 3.2

#### artifact

object of specific shape used to determine the accuracy of measuring devices

See Clause 7.

#### 3.3

#### bias

difference between the observed average of measurements and the calibration value

See Figure 1.

NOTE Bias can be affected by systematic errors such as linearity or gain and can be different throughout the operating range of the measurement system.



#### Key

- 1 calibration value
- 2 observed average
- 3 bias

#### Figure 1 — Bias

#### 3.4

#### calibration

set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system and the corresponding values realized by standards

#### 3.5

#### gain

magnification factor between the input and the output

#### 3.6

#### helix artifact

artifact having a calibrated helix form

#### 3.7

#### involute artifact

calibrated artifact having an involute form determined by a specific base circle

#### 3.8

#### measurand

particular quantity subject to measurement

#### 3.9

#### pitch and runout artifact

artifact with calibrated index features for pitch or runout or both

#### 3.10

#### repeatability (of measurement results)

closeness of the agreement between results of successive measurements of the same measurand carried out under the same conditions of measurement

#### 3.11

#### reproducibility (of measurement results)

closeness of the agreement between results of measurements of the same measurand carried out under changed conditions of measurement

NOTE 1 A valid statement of reproducibility requires specification of the conditions changed.

NOTE 2 The changed conditions may include

- principle of measurement,
- method of measurement,
- observer.
- iTeh STANDARD PREVIEW (standards.iteh.ai)
- measuring instrument,

SIST ISO 18653:2005

- reference standard, https://standards.iteh.ai/catalog/standards/sist/80c254ff-2c47-49c6-a3a9-3e2274182911/sist-iso-18653-2005
- location.
- conditions of use, and
- time.

NOTE 3 Reproducibility may be expressed quantitatively in terms of dispersion characteristics of the results.

#### 3.12

#### uncertainty (of measurement results)

parameter associated with the result of a measurement that characterizes the dispersion of the values that could be reasonably attributed to the measurand

NOTE 1 The parameter can be, for example, a standard deviation (or a given multiple of it), or the half-width of an interval having a stated level of confidence.

NOTE 2 Uncertainty of measurement comprises, in general, many components. Some of these components can be evaluated from the statistical distribution of the results of a series of measurements and can be characterized by experimental standard deviations. The other components, which also can be characterized by standard deviations, are evaluated from assumed probability distributions based on experience or other information.

It is understood that the result of the measurement is the best estimate of the value of the measurand, and NOTE 3 that all components of uncertainty, including those arising from systematic effects, such as components associated with corrections and reference standards, contribute to the dispersion.

Symbol	Definition	Unit	Where first used
Ε	Bias	μm	Eq. 1
k	Coverage factor	_	Eq. 1
п	Number of measurements	_	Eq. 2
U <sub>95</sub>	Measuring uncertainty	μm	Eq. 1
U <sub>95c</sub>	Uncertainty estimation	μm	Eq. 4
$U_{95(cal)}$	Measurement uncertainty stated in reference artifact calibration document	μm	Eq. 3
u <sub>m</sub>	Standard uncertainty	μm	Eq. 1
u <sub>n</sub>	Reference artifact calibration uncertainty	μm	Eq. 1
$u_{g}$	Geometry similarity influence	μm	Eq. 1
u <sub>w</sub>	Workpiece characteristic influence	μm	Eq. 1
X <sub>i</sub>	Individual measured value of parameter calibrated	μm	Eq. 2
$\overline{X}$	Mean of measured values	μm	Eq. 2

#### Table 1 — Symbols

#### iTeh STANDARD PREVIEW Application 4 (standards.iteh.ai)

#### 4.1 General

The purpose of the tests prescribed in this standard is to estimate measurement uncertainty. It has been assumed that the gear-measuring instrument has been installed on site and a series of acceptance tests have been completed successfully. Prescribed tests 4 may serve as a fifter m<sup>5</sup> checks to verify the measurement process.

The measurement and evaluation procedures may be used as part of acceptance tests for a new gearmeasuring instrument, with prior agreement between customer and supplier. In this situation it is recommended that a series of traceably calibrated gear artifacts be used to verify the measurement uncertainty at specific points throughout the working volume of the instrument. These measurements should include provision for testing the machine with table loads that represent the weight of product gears being tested.

#### 4.2 Traceability

The term traceability implies an unbroken calibration chain from measurements taken on shop floor inspection instruments to the primary artifacts at a national laboratory, see Figure 2. Traceability is transferred by calibrated gear artifacts. The primary laboratory has the lowest uncertainty, and uncertainty increases at each level as the traceability chain is transferred to shop floor measuring instruments. Minimizing the steps from the primary laboratory to a shop floor measuring instrument will reduce the measurement uncertainty.

#### 4.3 Artifacts

The gear artifacts used for these tests shall be of similar size and geometry to product gears inspected on the measuring instrument. Artifacts shall be used to evaluate the accuracy of each parameter inspected: helix (lead), profile, pitch, runout and tooth thickness. Specific artifact requirements are given in Clause 7.

#### 4.4 Measurement uncertainty

Conventional practice recommends the uncertainty of a measurement process be less than 10 % of the parameter tolerance measured, to ensure that the reliable interpretation of the measurement results is possible. However, this is not technically achievable when inspecting high accuracy gears.

For example, in gears with tolerances better than 10  $\mu$ m, the best achievable uncertainty may be only 20 to 30 %. To determine the uncertainty, see Clause 8 and ISO/TR 10064-5. Recommendations for allowable measurement uncertainty are made in ISO/TR 10064-5.



Figure 2 — Hierarchy of calibration

#### 4.5 Sources of uncertainty

The verification of measurement uncertainty shall include, but not be limited to, the assessment of the principal contributions to uncertainty in gear measurement as follows:

- artifact data;
- calibration data;
- repeatability of the instrument;
- reproducibility of the instrument;
- probe system filtering, damping and dynamic response, and accuracy;
- environmental influence, including temperature, vibration;
- mechanical alignment;