
Laboratorijska steklovina - Instrumenti za volumetrična merjenja - Metode za preskušanje zmogljivosti in uporaba (ISO 4787:2010)

Laboratory glassware - Volumetric instruments - Methods for testing of capacity and for use (ISO 4787:2010)

Laborgeräte aus Glas - Volumenmessgeräte - Prüfverfahren und Anwendung (ISO 4787:2010)

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Verrerie de laboratoire - Instruments volumétriques - Méthodes de vérification de la capacité et d'utilisation (ISO 4787:2010)

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

Ta slovenski standard je istoveten z: EN ISO 4787:2010

ICS:

| | | |
|-----------|--|---|
| 17.060 | Merjenje prostornine, mase, gostote, viskoznosti | Measurement of volume, mass, density, viscosity |
| 71.040.20 | Laboratorijska posoda in aparati | Laboratory ware and related apparatus |

SIST EN ISO 4787:2010

en,fr

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 4787:2010

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 4787

April 2010

ICS 17.060

English Version

Laboratory glassware - Volumetric instruments - Methods for testing of capacity and for use (ISO 4787:2010)

Verrerie de laboratoire - Instruments volumétriques -
Méthodes de vérification de la capacité et d'utilisation (ISO
4787:2010)

Laborgeräte aus Glas - Volumenmessgeräte -
Prüfverfahren und Anwendung (ISO 4787:2010)

This European Standard was approved by CEN on 15 March 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

[SIST EN ISO 4787:2010](https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010)

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....3

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

SIST EN ISO 4787:2010

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

Foreword

This document (EN ISO 4787:2010) has been prepared by Technical Committee ISO/TC 48 "Laboratory equipment" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2010, and conflicting national standards shall be withdrawn at the latest by October 2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of ISO 4787:2010 has been approved by CEN as a EN ISO 4787:2010 without any modification.

[SIST EN ISO 4787:2010](https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010)
<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN ISO 4787:2010

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

INTERNATIONAL STANDARD

**ISO
4787**

Second edition
2010-04-15

Laboratory glassware — Volumetric instruments — Methods for testing of capacity and for use

*Verrerie de laboratoire — Instruments volumétriques — Méthodes de
vérification de la capacité et d'utilisation*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 4787:2010](https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010)

[https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-
25eabb85b4eb/sist-en-iso-4787-2010](https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010)



Reference number
ISO 4787:2010(E)

© ISO 2010

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 4787:2010](https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010)

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

| | |
|---|-----------|
| Foreword | iv |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 2 |
| 4 Summary of method | 2 |
| 5 Volume and reference temperature | 2 |
| 5.1 Unit of volume..... | 2 |
| 5.2 Reference temperature | 2 |
| 6 Apparatus and calibration liquid | 2 |
| 7 Factors affecting the accuracy of volumetric instruments | 3 |
| 7.1 General | 3 |
| 7.2 Temperature | 3 |
| 7.3 Cleanliness of glass surface | 3 |
| 7.4 Quality of used volumetric instruments..... | 4 |
| 7.5 Delivery time and waiting time | 4 |
| 8 Setting the meniscus | 4 |
| 8.1 General | 4 |
| 8.2 Meniscus of transparent liquids | 4 |
| 8.3 Meniscus of opaque liquids | 5 |
| 9 Calibration procedure | 5 |
| 9.1 General | 5 |
| 9.2 Test room | 5 |
| 9.3 Filling and delivery | 6 |
| 9.4 Weighing..... | 7 |
| 9.5 Evaluation..... | 7 |
| 10 Use | 7 |
| 10.1 General | 7 |
| 10.2 Volumetric flasks (see ISO 1042) | 8 |
| 10.3 Measuring cylinders (see ISO 4788) | 8 |
| 10.4 Burettes (see ISO 385)..... | 8 |
| 10.5 Pipettes..... | 9 |
| Annex A (informative) Cleaning of volumetric glassware | 10 |
| Annex B (normative) Calculation of volume | 11 |
| Bibliography..... | 21 |

ISO 4787:2010(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 4787 was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, Subcommittee SC 6, *Laboratory and volumetric ware*.

This second edition cancels and replaces the first edition (ISO 4787:1984), which has been technically revised to incorporate the following changes:

- iTeh STANDARD PREVIEW**
(standards.iteh.ai)
- a) the potassium dichromate cleaning method in Annex A has been deleted;
 - b) new tables for calculation of test results have been added to Annex B;
 - c) the description of the test (calibration) methods has been modified to be more precise;
 - d) test methods have been separated from recommendations for use.

Laboratory glassware — Volumetric instruments — Methods for testing of capacity and for use

1 Scope

This International Standard provides methods for the testing, calibration and use of volumetric instruments made from glass in order to obtain the best accuracy in use.

NOTE Testing is the process by which the conformity of the individual volumetric instrument with the appropriate standard is determined, culminating in the determination of its error of measurement at one or more points.

The International Standards for the individual volumetric instruments include clauses on the definition of capacity; these clauses describe the method of manipulation in sufficient detail to define the capacity without ambiguity. This International Standard contains supplementary information.

The procedures are applicable to volumetric instruments with nominal capacities in the range of 0,1 ml to 10 000 ml. These include: single-volume pipettes (see ISO 648) without subdivisions; graduated measuring pipettes and dilution pipettes, with partial or complete subdivisions (see ISO 835); burettes (see ISO 385); volumetric flasks (see ISO 1042); and graduated measuring cylinders (see ISO 4788). The procedures are not recommended for testing of volumetric instruments with capacities below 0,1 ml such as micro-glassware.

This International Standard does not deal specifically with pyknometers as specified in ISO 3507. However, the procedures specified below for the determination of volume of glassware can, for the most part, also be followed for the calibration of pyknometers.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 385, *Laboratory glassware — Burettes*

ISO 648, *Laboratory glassware — Single-volume pipettes*

ISO 835, *Laboratory glassware — Graduated pipettes*

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 4788, *Laboratory glassware — Graduated measuring cylinders*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

ISO 4787:2010(E)

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 99 apply.

4 Summary of method

The general procedure is based upon a determination of volume of water, either contained in or delivered by the volumetric instrument. This volume of water is based upon knowledge of its mass under consideration of buoyancy and its tabulated density (gravimetric method).

5 Volume and reference temperature

5.1 Unit of volume

The unit of volume shall be the millilitre (ml), which is equivalent to one cubic centimetre (cm³).

5.2 Reference temperature

The standard reference temperature, i.e. the temperature at which the volumetric instrument is intended to contain or deliver its volume (capacity), shall be 20 °C.

When the volumetric instrument is required for use in a country which has adopted a standard reference temperature of 27 °C (the alternative recommended in ISO 384 for tropical use), this figure shall be substituted for 20 °C.

PRE-STANDARD PREVIEW
(standards.iteh.ai)

6 Apparatus and calibration liquid

<https://standards.iteh.ai/catalog/standards/sist/724af10b-94ae-48fb-844b-25eabb85b4eb/sist-en-iso-4787-2010>

6.1 Balance, with a resolution and standard deviation appropriate to the selected volume of the apparatus under test (see Table 1).

The resolution of the display, the standard deviation and the linearity of the balance will be a limiting factor in the accuracy of the measurements. The balance shall be calibrated with adequate accuracy (see 9.4).

Table 1 — Recommended balance

| Selected volume under test ^a | Resolution | Standard deviation (repeatability) | Linearity |
|---|------------|------------------------------------|-----------|
| V | mg | mg | mg |
| 100 µl < V ≤ 10 ml | 0,1 | 0,2 | 0,2 |
| 10 ml < V < 1 000 ml | 1 | 1 | 2 |
| 1 000 ml ≤ V ≤ 2 000 ml | 10 | 10 | 20 |
| V > 2 000 ml | 100 | 100 | 200 |

^a For practical purposes, the nominal volume may be used to choose the balance.

6.2 Thermometer, to measure the temperature of the calibration liquid (water) with a measurement error of maximum 0,2 °C for liquid volumes < 1 000 ml and with a measurement error of maximum 0,1 °C for liquid volumes ≥ 1 000 ml.

6.3 Hygrometer, to measure the humidity in the test room with a measurement error of maximum 5 % within the humidity range of 35 % to 85 %.