

---

---

**Animal feeding stuffs — Determination of  
potassium and sodium contents —  
Methods using flame-emission  
spectrometry**

*Aliments des animaux — Détermination des teneurs en potassium et  
sodium — Méthodes par spectrométrie d'émission de flamme*

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[ISO 7485:2000](https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5ece02a/iso-7485-2000)

<https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5ece02a/iso-7485-2000>



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 7485:2000](#)

<https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5ece02a/iso-7485-2000>

© ISO 2000

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

Printed in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 7485 was prepared by Technical Committee ISO/TC 34, *Agricultural food products*, Subcommittee SC 10, *Animal feeding stuffs*.

Annex A of this International Standard is for information only.

ITIH STANDARD PREVIEW  
(standards.iteh.ai)

[ISO 7485:2000](https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5ece02a/iso-7485-2000)

<https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5ece02a/iso-7485-2000>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 7485:2000

<https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5ece02a/iso-7485-2000>

# Animal feeding stuffs — Determination of potassium and sodium contents — Methods using flame-emission spectrometry

## 1 Scope

This International Standard specifies a calibration method and a standard addition method for the determination of potassium and sodium contents of animal feeding stuffs by flame-emission spectrometry.

The standard addition method applies if it is not known whether the emissions measured for potassium or sodium are subject to matrix effects. The limit of determination for potassium and sodium is 0,04 g/kg.

The calibration method applies if such matrix effects have been shown to be absent.

## 2 Normative reference

The following normative document contains provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent edition of the normative document indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 6498:1998, *Animal feeding stuffs — Preparation of test samples*.

## 3 Principle

The test portion is incinerated to destroy any organic matter. The potassium and sodium are dissolved by treatment with hydrochloric acid. The potassium and sodium contents are determined by flame-emission spectrometry at wavelengths of 766 nm and 589 nm respectively, by means of a calibration curve (calibration method) or a standard addition graph (standard addition method).

## 4 Reagents

Use only reagents of recognized analytical quality.

- 4.1 **Water**, double distilled, or double deionized.
- 4.2 **Potassium chloride** (KCl), dried for 2 h at  $(110 \pm 2)$  °C.
- 4.3 **Sodium chloride** (NaCl), dried for 2 h at  $(110 \pm 2)$  °C.
- 4.4 **Concentrated hydrochloric acid**,  $c(\text{HCl}) = 12 \text{ mol/l}$  ( $\rho_{20}(\text{HCl}) = 1,19 \text{ g/ml}$ ).
- 4.5 **Hydrochloric acid**,  $c(\text{HCl}) = 6 \text{ mol/l}$ .
- 4.6 **Dilute hydrochloric acid**,  $c(\text{HCl}) = 0,1 \text{ mol/l}$ .

**4.7 Ionization and release buffer solution.**

Dissolve 50 g of caesium chloride and 250 g of aluminium nitrate nonahydrate  $[Al(NO_3)_3 \cdot 9H_2O]$  in water. Dilute to 1 litre with water and mix. Store in a plastic bottle, e.g. polyethene or polypropylene.

**4.8 Potassium standard solutions.**

**4.8.1 Potassium stock solution,  $\rho(K) = 1 \text{ mg/ml}$ .**

Weigh 1,906 g of potassium chloride (4.2) and transfer to a 1 000 ml volumetric flask with approximately 250 ml of dilute hydrochloric acid (4.6). Dissolve and dilute to the mark with the hydrochloric acid (4.6).

Store in a plastic bottle, e.g. polyethene or polypropylene. At room temperature the solution is stable for up to 6 months.

**4.8.2 Potassium intermediate solution,  $\rho(K) = 100 \text{ }\mu\text{g/ml}$ .**

Transfer 10 ml of the potassium stock solution (4.8.1) to a 100 ml volumetric flask. Dilute to the mark with dilute hydrochloric acid (4.6) and mix.

Store in a plastic bottle, e.g. polyethene or polypropylene. At room temperature the solution is stable for 1 month.

**4.8.3 Potassium calibration solution,  $\rho(K) = 10 \text{ }\mu\text{g/ml}$ .**

Transfer 10 ml of the potassium intermediate solution (4.8.2) to a 100 ml volumetric flask. Add 10 ml of the ionization and release buffer solution (4.7). Dilute to the mark with dilute hydrochloric acid (4.6) and mix.

Prepare fresh on the day of use.

**4.9 Sodium standard solutions.**

ISO 7485:2000

[https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-](https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5eccc02a/iso-7485-2000)

**4.9.1 Sodium stock solution,  $\rho(Na) = 1 \text{ mg/ml}$ .**

[849b-cae4f5eccc02a/iso-7485-2000](https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae4f5eccc02a/iso-7485-2000)

Weigh 2,542 g of sodium chloride (4.3) and transfer to a 1 000 ml volumetric flask with approximately 250 ml of dilute hydrochloric acid (4.6). Dissolve and dilute to the mark with the hydrochloric acid (4.6). Mix well.

Store in a plastics bottle. At room temperature the solution is stable for up to 6 months.

**4.9.2 Sodium intermediate solution,  $\rho(Na) = 100 \text{ }\mu\text{g/ml}$ .**

Transfer 10 ml of the sodium stock solution (4.9.1) to a 100 ml volumetric flask. Dilute to the mark with dilute hydrochloric acid (4.6) and mix.

Store in a plastics bottle. At room temperature the solution is stable for 1 month.

**4.9.3 Sodium calibration solution,  $\rho(Na) = 10 \text{ }\mu\text{g/ml}$ .**

Transfer 10 ml of the sodium intermediate solution (4.9.2) to a 100 ml volumetric flask. Add 10 ml of the ionization and release buffer solution (4.7). Dilute to the mark with the hydrochloric acid (4.6) and mix.

Prepare fresh on the day of use.

**4.10 Biological Certified Reference Materials.**

For example: V8-Rye flour or V10-Hay (powdered), from the International Atomic Energy Agency (IAEA), Austria; SRM 157-Wheat flour, SRM 1568a-Rice flour or SRM 1548-Total diet, from the National Institute of Standards and Technology (NIST), USA.

## 5 Apparatus

Usual laboratory equipment and, in particular, the following.

- 5.1 **Analytical balance**, capable of weighing to the nearest 1 mg.
- 5.2 **Muffle furnace**, electrically heated, capable of being maintained at  $(550 \pm 20)$  °C.
- 5.3 **Ashing dishes**, of platinum or silica.
- 5.4 **Flame photometer** with an air-butane, air-propane or air-acetylene flame in a ratio allowing complete combustion; or an **atomic absorption spectrometer** operating in emission mode and equipped to determine potassium and sodium with an air-acetylene flame.
- 5.5 **Sand bath or hot plate**, capable of being maintained at approximately 150 °C.
- 5.6 **TD (to deliver) precision pipettes or microburette** of 5 ml capacity.
- 5.7 **TC (to contain) precision pipettes**, of various capacities.
- 5.8 **Beakers**, of 250 ml capacity.
- 5.9 **Volumetric flasks**, of various capacities.

## 6 Sampling

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 6497 [1].

[ISO 7485:2000](https://standards.iteh.ai/catalog/standards/sist/b6818738-2168-49d8-849b-cae45eccc02a/iso-7485-2000)

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

## 7 Preparation of test sample

Prepare the test sample in accordance with ISO 6498.

## 8 Procedure

### 8.1 Test portion

#### 8.1.1 Samples containing organic matter

Depending on the expected content, weigh 1 g to 5 g of the prepared test sample (clause 7) to the nearest 1 mg into an ashing dish (5.3). Proceed in accordance with 8.2.1.

#### 8.1.2 Samples not containing organic matter

Depending on the expected content of potassium and sodium, weigh 1 g to 5 g of the prepared test sample to the nearest 1 mg into a beaker (5.8). Proceed in accordance with 8.2.2.

## 8.2 Preparation of test solution

### 8.2.1 Incineration and rendering silica insoluble

Place the incineration dish containing the test portion (8.1.1) in the cold muffle furnace (5.2). Close the furnace and gradually increase the temperature to 550 °C and ash for 3 h at this temperature.

If many carbon particles are present, allow to cool down, moisten the contents of the dish with 2 ml of water and dry the dish on the sand bath or hot plate (5.5). Ash for another 2 h at 550 °C in the muffle furnace.

Allow to cool down, moisten the ash with a few drops of water and transfer it to a 250 ml beaker (5.8). Rinse the incineration dish with a total of approximately 5 ml of concentrated hydrochloric acid (4.4) and then with a few millilitres of water, collecting the rinsings in the beaker.

**CAUTION — Take care during the addition of hydrochloric acid, as an effervescent reaction may occur with some samples.**

Put a watch glass on the beaker to reduce spattering and slowly evaporate the contents of the beaker to near dryness on the sand bath or hot plate (5.5). Allow to cool to room temperature.

Proceed in accordance with 8.2.3.

### 8.2.2 Dissolution of potassium and sodium

Add to the beaker containing the test portion (8.1.2), slowly and carefully, a sufficient quantity (15 ml to 30 ml) of concentrated hydrochloric acid (4.4). Put a watch glass on the beaker to reduce spattering and slowly evaporate the contents of the beaker to dryness on the sand bath or hot plate (5.5). Allow to cool to room temperature.

### 8.2.3 Re-dissolution of potassium and sodium

Treat the residue (8.2.1 or 8.2.2) with 5 ml of hydrochloric acid (4.5) and 45 ml of water. Bring to the boil and allow to cool to room temperature. Transfer quantitatively into a 100 ml volumetric flask. Dilute to the mark with water and mix. Allow particles to settle for at least 4 h. Filter if the solution is not clear.

**CAUTION — Perform the above operations in an effective fume-removal device.**

## 8.3 Blank test solution

Prepare a solution in the same way as for the test solution, carrying out all the operations specified in 8.2 but omitting the test portion.

## 8.4 Selection of method

If it is not known whether the emissions measured for potassium or sodium are subject to matrix effects, proceed in accordance with 8.6.

If such matrix effects have been shown to be absent, proceed in accordance with 8.5.

## 8.5 Calibration method

### 8.5.1 Test solutions

Transfer to a 100 ml volumetric flask an aliquot portion of the test solution (8.2.3) containing a maximum of 1 mg of potassium or sodium. Add 10 ml of the ionization and release buffer solution (4.7), dilute to the mark with water and mix.

Prepare the same dilution of the blank test solution (8.3) in the same way.



## 8.5.2 Calibration solutions

Transfer to six 100 ml volumetric flasks 0 ml, 2,0 ml, 4,0 ml, 6,0 ml, 8,0 ml and 10,0 ml of the potassium intermediate solution (4.8.2) or sodium intermediate solution (4.9.2). Add 10 ml of the ionization and release buffer solution (4.7) to each flask. Dilute to the mark with hydrochloric acid (4.6) and mix. The potassium or sodium contents of these solutions are 0 µg/ml, 2,0 µg/ml, 4,0 µg/ml, 6,0 µg/ml, 8,0 µg/ml and 10,0 µg/ml respectively.

Prepare fresh on the day of use.

## 8.5.3 Spectrometric measurements

### 8.5.3.1 Adjustment of the instrument

Adjust the instrument in accordance with the manufacturer's instructions.

Light the flame and allow some minutes for flame stabilization. Adjust the burner position, the gas flows and the wavelengths (theoretically 766 nm for potassium and 589 nm for sodium) or select the filters so that maximum reading is obtained when the calibration solution of potassium (4.8.3) or sodium (4.9.3) is atomized in the flame.

### 8.5.3.2 Test solution measurements

Measure the emissions of the calibration solutions (8.5.2) and then the emission of the test solution (8.5.1).

Correct the emission of the test solution for the emission of the diluted blank test solution (8.5.1), if the latter differs from the emission of the zero concentration calibration solution.

## 8.5.4 Calibration graph

Draw the calibration graph by plotting the emissions of the calibration solutions against their respective potassium or sodium contents in micrograms per millilitre.

## 8.6 Standard addition method

### 8.6.1 Test solutions

Carry out the dilutions according to the expected potassium or sodium content.

Transfer to each of three volumetric flasks of suitable capacity (flasks A, B and C) the same aliquot portion of the test solution (8.2.3). Add to flasks B and C aliquot portions of the potassium intermediate solution (4.8.2) or sodium intermediate solution (4.9.2) corresponding to a content of 2,5 µg/ml for flask B and 5 µg/ml for flask C. Add to each flask an aliquot portion of the ionization and release buffer solution (4.7) corresponding to 10 ml per 100 ml of the final solution.

Dilute the contents of the three flasks (A, B and C) to the mark with water and mix.

The potassium or sodium content of the solution in flask A shall be smaller than 5 µg/ml.

Prepare the same dilution of the blank test solution (8.3) in the same way as for the solution in flask A.

### 8.6.2 Spectrometric measurements

#### 8.6.2.1 Adjustment of the instrument

Adjust the instrument in accordance with 8.5.3.1.