



**SLOVENSKI STANDARD**  
**SIST EN 12470-1:2000/kprA1:2009**  
**01-april-2009**

---

**Klinični termometri - 1. del: Zaprti stekleni termometri s tekočimi kovinami**

Clinical thermometers - Part 1: Metallic liquid-in-glass thermometers with maximum device

Medizinische Thermometer - Teil 1: Mit metallischer Flüssigkeit gefüllte Glasthermometer mit Maximumvorrichtung

Thermomètres médicaux - Partie 1: Thermomètres à dilatation de liquide métallique dans une gaine de verre, avec dispositif à maximum

**Ta slovenski standard je istoveten z: EN 12470-1:2000/prA1**

---

**ICS:**

11.040.55      Diagnostična oprema      Diagnostic equipment

**SIST EN 12470-1:2000/kprA1:2009      en,fr**



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**FINAL DRAFT**  
**EN 12470-1:2000**

**prA1**

January 2009

---

ICS 17.200.20

English Version

## Clinical thermometers - Part 1: Metallic liquid-in-glass thermometers with maximum device

Thermomètres médicaux - Partie 1: Thermomètres à  
dilatation de liquide métallique dans une gaine de verre,  
avec dispositif à maximum

Medizinische Thermometer - Teil 1: Mit metallischer  
Flüssigkeit gefüllte Glasthermometer mit  
Maximumvorrichtung

This draft amendment is submitted to CEN members for unique acceptance procedure. It has been drawn up by the Technical Committee CEN/TC 205.

This draft amendment A1, if approved, will modify the European Standard EN 12470-1:2000. If this draft becomes an amendment, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for inclusion of this amendment into the relevant national standard without any alteration.

This draft amendment was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

**Warning** : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

Page

Foreword.....	3
1 <b>Modification to 8.1, General</b> .....	4
2 <b>Modification to 8.2, Marking</b> .....	4
3 <b>Modification to 8.3, Instructions for use</b> .....	4
4 <b>Modification to Annex ZA</b> .....	4

## Foreword

This document (EN 12470-1:2000/prA1:2009) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by DIN.

This document is currently submitted to the Unique Acceptance Procedure.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

This European Standard applies to clinical thermometers which are used for measuring the body temperature of humans.

EN 12470 consists of the following Parts under the general title “Clinical thermometers”:

Part 1: Metallic liquid-in-glass thermometers with maximum device

Part 2: Phase change-type (dot matrix) thermometers

Part 3: Performance of compact electrical thermometers (non-predictive and predictive) with maximum device

Part 4: Performance of electrical thermometers for continuous measurement

Part 5: Performance of infra-red ear thermometers (with maximum device)

**EN 12470-1:2000/prA1:2009 (E)****1 Modification to 8.1, General**

*Add a new sentence to read as follows:*

"The manufacturer shall state the date of issue or the latest revision of the instruction for use."

**2 Modification to 8.2, Marking**

*Add the following new item f):*

"f) the name and address of the authorised representative where the manufacturer does not have a registered place of business in the community."

**3 Modification to 8.3, Instructions for use**

*Add the following new item f):*

"f) the name and address of the authorised representative where the manufacturer does not have a registered place of business in the community."

**4 Modification to Annex ZA**

*Replace the existing Annex ZA with the following:*

"

**Annex ZA**  
(informative)

**Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC**

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Communities under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC

Clause(s)/sub-clause(s) of this EN	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
4	10.3	
5	1, 2	
6	1, 2, 3, 4, 5	
6.1	10.2	
6.1.2.7	9.2, 12.7.1	
6.2	7.1	
6.3.3	9.2	
6.4	10.1	
6.5	10.1	
6.6	10.1	
7	1, 2, 3, 4, 5	
7.2, 7.3, 7.4	7.1	
7.5, 7.6, 7.7	10.1	
7.8	9.2, 12.7.1	
7.9	10.1	
8	13.1	
8.1	13.2, 13.6 q)	
8.2	13.2	
8.2 a), f)	13.3 a)	
8.2 b)	13.3 b)	
8.2 c)	13.3 d)	
8.2 d)	10.3	
8.3	13.1 i), j), k), 13.6 a)	
8.3 e)	8.1, 13.6 h)	
8.3.f)	13.3 a)	
	1, Dash 1 and dash 2	The modification of this Essential Requirement is not addressed in this European Standard
	6a	This Essential Requirement is not addressed in this European Standard
	7.1, Dash 3	The modification of this Essential Requirement is not addressed in this European Standard
	7.4	The modification of this Essential Requirement is not addressed in this European Standard
	7.5	The modification of this Essential Requirement is not addressed in this European Standard
	12.1a	This Essential Requirement is not addressed in this European Standard
	13.3 f)	This Essential Requirement is not addressed in this European Standard
	13.6 h) 2 <sup>nd</sup> para	The modification of this Essential Requirement is not addressed in this European Standard

**EN 12470-1:2000/prA1:2009 (E)**

**WARNING** — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard."