
Krma - Določevanje ohratoksina A v krmi z imunoafinitetnim kolonskim čiščenjem in s tekočinsko kromatografijo visoke ločljivosti s fluorescentno detekcijo

Animal feeding stuffs - Determination of Ochratoxin A in animal feed by immunoaffinity column clean-up and High Performance Liquid Chromatography with fluorescence detection

Futtermittel - Bestimmung von Ochratoxin A in Tierfutter mit Reinigung an einer Immunoaffinitätssäule und Hochleistungs-Flüssig-Chromatographie mit Fluoreszenzdetektion

Aliments pour animaux - Dosage de l'ochratoxine A dans les aliments pour animaux par purification sur colonne d'immuno-affinité et chromatographie liquide à haute performance avec détection par fluorescence

Ta slovenski standard je istoveten z: EN 16007:2011

ICS:

65.120 Krmila Animal feeding stuffs

SIST EN 16007:2011 en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 16007:2011

<https://standards.iteh.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-4f07d59b84e7/sist-en-16007-2011>

EUROPEAN STANDARD

EN 16007

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2011

ICS 65.120

English Version

Animal feeding stuffs - Determination of Ochratoxin A in animal feed by immunoaffinity column clean-up and High Performance Liquid Chromatography with fluorescence detection

Aliments pour animaux - Dosage de l'ochratoxine A dans les aliments pour animaux par purification sur colonne d'immuno-affinité et chromatographie liquide à haute performance avec détection par fluorescence

Futtermittel - Bestimmung von Ochratoxin A in Tierfutter durch Reinigung an einer Immunoaffinitätssäule und Hochleistungs-Flüssig-Chromatographie mit Fluoreszenzdetektion

This European Standard was approved by CEN on 25 June 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents and materials	4
5 Apparatus	7
6 Sample preparation	8
7 Measurements.....	10
8 Determination of concentrations	11
9 Precision	11
10 Test report	12
Annex A (informative) Example of a chromatogram.....	14
Annex B (informative) Results of the interlaboratory study	15
Bibliography	18

SIST EN 16007:2011

<https://standards.itech.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-4f07d59b84e7/sist-en-16007-2011>

Foreword

This document (EN 16007:2011) has been prepared by Technical Committee CEN/TC 327 “Animal feeding stuffs”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2012, and conflicting national standards shall be withdrawn at the latest by February 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

WARNING — The use of this protocol involves hazardous materials, operations and equipment. This protocol does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this protocol to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

[SIST EN 16007:2011](https://standards.iteh.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-4f07d59b84e7/sist-en-16007-2011)

<https://standards.iteh.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-4f07d59b84e7/sist-en-16007-2011>

EN 16007:2011 (E)**1 Scope**

This European Standard specifies a method for the determination of Ochratoxin A (OTA) in cereal based animal feed using immunoaffinity for clean-up followed by liquid-chromatography with fluorescence detection.

NOTE The validated mass fraction range was 39 µg/kg to 338 µg/kg OTA.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 1042, *Laboratory glassware — One-mark volumetric flasks (ISO 1042:1998)*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

3 Principle

OTA is extracted from the test material with a mixture of methanol – 3% aqueous sodium bicarbonate solution. The extract is filtered, diluted with PBS and purified using immunoaffinity columns (IAC). The purified OTA is eluted from the IAC using first methanol and then water, brought to a defined volume with water and quantified by HPLC with fluorescence detection.

(standards.iteh.ai)

4 Reagents and materials

SIST EN 16007:2011

<https://standards.iteh.ai/catalog/standards/sist/5409891-78b2-4f20-9cb8->

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade. Solvents shall be of HPLC or better quality.

4.1 Methanol, CH₃OH, technical grade.

4.2 Methanol, CH₃OH, HPLC grade.

4.3 Water, water (EN ISO 3696 grade 1 (HPLC grade)) and water (EN ISO 3696 grade 3), or equivalent.

4.4 Potassium chloride, KCl.

4.5 Sodium chloride, NaCl.

4.6 Disodium hydrogenphosphate dodecahydrate, Na₂HPO₄*12 H₂O .

4.7 Acetonitrile, CH₃CN, HPLC grade.

4.8 Glacial acetic acid, CH₃COOH, 96% minimum.

4.9 Solution of acetonitrile/ glacial acetic acid, acetonitrile (4.7) and glacial acetic acid (4.8) in proportion of 99/1 (v/v).

4.10 Toluene, C₆H₅CH₃, analytical grade.

4.11 Toluene/Acetic acid solution, Toluene (4.10) / Glacial acetic acid (4.8) in proportion of 99/1 (v/v).

4.12 Sodium hydrogen carbonate, NaHCO_3 , minimum 99% purity.

4.13 PBS concentrate, Phosphate buffered saline concentrate.

Dissolve the following in 1 800 ml of water (EN ISO 3696 grade 1):

- 4 g KCl (4.4);
- 160 g NaCl (4.5);
- 72 g $\text{Na}_2\text{HPO}_4 \cdot 12 \text{H}_2\text{O}$ (4.6).

4.14 PBS Ready to use.

Dilute 100 ml of PBS concentrate (4.13) to 1 000 ml with water (EN ISO 3696 grade 1). Adjust to pH 7,4 with 1 mol/l HCl and make up to 2 000 ml with water (EN ISO 3696 grade 1),

or

PBS tablets, Phosphate buffered saline tablets,

One tablet dissolved in 200 ml of water (EN ISO 3696 grade 1) yields 0,01 mol/l phosphate buffer, 0,002 7 mol/l potassium chloride and 0,137 mol/l sodium chloride, pH 7,4, at 25° C (e.g. Sigma P4417).

4.15 3% aqueous sodium hydrogen carbonate solution.

Add 30 g of sodium hydrogen carbonate (4.12) to 1 000 ml of water (EN ISO 3696 grade 3).

4.16 Extraction solvent, methanol / 3% aqueous sodium hydrogen carbonate solution in proportion of 50/50 (v/v).

Add 500 ml of methanol (4.1) to 500 ml of 3% aqueous sodium hydrogen carbonate solution (4.15). Mix well.

4.17 Aqueous solution of glacial acetic acid.

Add 30 ml of glacial acetic acid (4.8) to 870 ml of water (EN ISO 3696 grade 1) and filter.

4.18 HPLC mobile phase, acetonitrile / methanol / aqueous solution of glacial acetic acid in proportion of 35/35/30 (v/v/v).

Mix 1 050 ml of methanol (4.2) with 1 050 ml of acetonitrile (4.77) and with 900 ml of aqueous solution of glacial acetic acid (4.1717). Mix well and degas.

4.19 OTA, ochratoxin A.

OTA in pure form as crystals, dried film, powder or in solution.

4.20 OTA Stock Solution.

Prepare from the OTA (4.19) a solution of 20 $\mu\text{g}/\text{ml}$ in toluene/acetic acid solution (4.11).

To determine the exact concentration, record the absorption curve of this solution between a wavelength of 300 nm and 370 nm in a 1 cm quartz cell with toluene/acetic acid solution (4.11) as reference using the UV-

EN 16007:2011 (E)

spectrophotometer (5.21). Identify the wavelength for maximum absorption. Calculate the mass concentration of OTA, ρ_{ota} , in micrograms per millilitre using Equation (1):

$$\rho_{ota} = \frac{A_{max} \times M \times 100}{\epsilon \times b} \quad (1)$$

where

A_{max} is the absorption determined at the maximum of the absorption curve (here: at 333 nm);

M is the molar mass, in grams per mol, of OTA ($M = 403,8$ g/mol);

ϵ is the molar absorption coefficient, in square metres per mol, of OTA in toluene/acetic acid solution (4.11), (here: 544 m²/mol);

b is the optical path length, in centimetres, of the quartz cell.

4.21 Nitrogen.**4.22 OTA diluted stock solution for calibration.**

Prepare 10 ml of OTA diluted stock solution for calibration by diluting 100 times the OTA stock standard solution (4.20) with mobile phase (4.18). For this pipette 100 μ l of the OTA stock standard solution into a volumetric cylinder, evaporate the solution under a slight stream of nitrogen (4.21) at 40 °C and take up with mobile phase (4.18).

NOTE To ease complete dissolution of the OTA in mobile phase, first fill the volumetric cylinder to approximately 1/3, let the OTA dissolve, and then fill up to the mark and shake.

[SIST EN 16007:2011](https://standards.iteh.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-4f07d59b84e7/sist-en-16007-2011)

4.23 Calibration solutions.

<https://standards.iteh.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-4f07d59b84e7/sist-en-16007-2011>

From the OTA diluted stock solution for calibration (4.22) prepare six levels of calibration solutions by adding the volumes of diluted stock solution listed below to a volumetric flask of the indicated volume and make up to the mark with mobile phase (4.18):

Table 1 — recommended calibration solutions (4.23) for the determination of OTA

Calibrant	OTA diluted stock solution for calibration (4.22) (μ l)	Volumetric flask (5.5) (ml)	Concentration OTA (ng/ml)
1	50	10,0	1
2	250	10,0	5
3	500	10,0	10
4	750	10,0	15
5	1 000	10,0	20
6	1 250	10,0	25

4.24 IAC with antibodies specific to OTA.

The IAC contains antibodies raised against OTA. The column shall have a total capacity of not less than 100 ng of OTA. OTA recovery shall not be less than 85 % when 5 ng OTA is applied in 50 ml of a mixture of 4 parts per volume of extraction solvent (4.16) and 96 parts per volume of PBS Ready to use (4.14).

4.25 Spiking solution.

SIST EN 16007:2011

<https://standards.iteh.ai/catalog/standards/sist/f5409891-78b2-4f20-9cb8-497d9989c7/sist-en-16007-2011>

OTA in a solution of acetonitrile/ glacial acetic acid in proportion of 99/1 (v/v). The OTA concentration of the spiking solution will depend on the spiking level required. The spiking volume should be approximately 500 μ l. The solution can be made from the OTA Stock Solution (4.20) in a similar manner as the OTA diluted stock solution for calibration (4.22).

5 Apparatus

Usual laboratory equipment and, in particular, the following:

5.1 Common laboratory glassware, such as graduated cylinders, beakers, volumetric pipettes

5.2 Analytical balance, capable of weighing to 0,1 mg

5.3 Horizontal or vertical shaker

5.4 Automated SPE Vacuum System

5.5 Volumetric flasks (class A, EN ISO 1042), 5 ml, 10 ml, 100 ml

5.6 Filter paper pre-folded, 30 μ m particle retention

5.7 Screw-cap flasks, 250 ml and 500 ml

EN 16007:2011 (E)**5.8 Glass funnels**, 9 cm ID**5.9 Reservoirs**, polypropylene, suitable for attachment to top of IAC, 50 ml to 75 ml size**5.10 Plastic syringes**, 5 ml**5.11 Calibrated displacement micropipette**, 100 µl, 500 µl and 1 000 µl (variable volume)**5.12 Solvent vacuum filtration system**, suitable for 47 mm filter**5.13 Glass microfibre filter**, 1,6 µm particle retention**5.14 Vortex mixer****5.15 HPLC syringe filter cartridges**, Nylon with 0,45 µm pore size**5.16 Ultrasonic bath****5.17 Amber glass vials**, ca 2 ml capacity and crimp caps or equivalent**5.18 HPLC system**

The HPLC system comprises the following:

- pump, pulse free, flow capacity 0,5 ml/min to 1,5 ml/min;
- injector system, manual or autosampler, with loop suitable for 100 µl to 300 µl injections;
- fluorescence detector, suitable for measurements with excitation wavelengths 333 nm and emission at 467 nm;
- integrator, or PC workstation.

5.19 Analytical reversed phase HPLC column, C18 RP-column suitable to allow a sufficient separation of OTA from other interfering components; Fully End Capped with column dimensions preferably 250 mm X 4,6 mm ID stationary phase with particle size 5 µm.**5.20 Pre-column (guard-column)**, with preferably the same stationary phase material as the analytical column and internal diameter of 4,0 mm, stationary phase with particle size 5 µm.**5.21 UV-Spectrophotometer.****6 Sample preparation****6.1 Extraction of OTA**

- Weigh 25,0 g, to the nearest 0,1 g, of the test sample into a large enough container with lid, e.g. 500 ml screw-cap flask (5.67),

NOTE 1 The materials for the interlaboratory study [1] were milled to a particle size of ca. 0,5 mm.

- add 200,0 ml of extraction solvent (4.16), cap, and shake vigorously by hand for a few seconds; so that the material disperses evenly (check visually),
- put on a shaker (5.3) for 40 min. Choose a speed such that the material is mixed well without collecting in the top of the flask. A possible alternative is to ultrasonificate for 15 minutes; in this case shortly shake the extract by hand for a few seconds,
- allow the extracted sample to settle after extraction,
- filter the extract through folded filter paper (5.6) and collect the filtered extract in a screw cap flask of 250 ml (5.7), and
- transfer approx. 10 ml of the filtered extract into the vacuum system and filter through glass microfibre filter (5.13) by applying a slight vacuum (5.12). Discard this volume and filter again the remaining extract in another screw cap flask of 250 ml (5.7) to obtain a clear extract for further analysis. Proceed immediately with the IAC clean-up procedure (6.2).

NOTE 2 Do not apply a strong vacuum in the beginning of the filtration process as this can lead to turbid filtered extract after filtration.

6.2 Clean up and test solution

- Take one IAC (4.24) per extract;
- attach a reservoir (5.9); do not empty storage solution from column;
- transfer 4,00 ml of double filtered extract into a 100 ml volumetric flask (5.5), fill to volume with PBS ready for use (4.14) and shake;
- apply 50,0 ml of the diluted extract to the reservoir on the IAC (see previous point); this is equivalent to 2,00 ml of the double filtered and undiluted extract;
- open the IAC outlet;
- draw extract through the IAC by gravity at a steady flow rate until all extract has passed the immunoaffinity column and the last solvent portion reaches the lower frit of the IAC;
- after the extract has passed completely, wash the IAC with 10 ml of water (EN ISO 3696 grade 1); make sure that the water is pH neutral;
- pass air through the IAC (f.i., using a properly fitted large syringe) in order to expel excess water;
- place a 5 ml volumetric flask (5.5) underneath the IAC and add 0,75 ml of methanol (4.2) to the IAC;
- collect the eluate in the 5 ml volumetric flask;
- after the last drops of methanol have passed through the IAC, allow the methanol to remain on the IAC for approximately 1 min; then add a further 0,75 ml of methanol (4.2) and continue to collect the eluate, followed by 0,50 ml of water (EN ISO 3696 grade 1);
- carefully pass air through the IAC in order to collect most of the applied methanol and water;
- immediately add 1,5 ml of glacial acetic acid solution (4.17) into the volumetric flask; and
- fill the volumetric flask up to the mark with water (EN ISO 3696 grade 1) and shake using a vortex. In case of turbid samples, filter test solutions through a HPLC syringe filter (5.15) with a plastic syringe (5.10). This can be used directly as test solution.