
**Ophthalmic implants — Irrigating
solutions for ophthalmic surgery**

*Implants ophtalmiques — Solutions d'irrigation pour la chirurgie
ophtalmique*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 16671:2003](https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003)

[https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-
ca905ff7c658/iso-16671-2003](https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 16671:2003](https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003)

<https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003>

© ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Intended performance	2
5 Design attributes	2
6 Design evaluation	4
7 Sterilization	5
8 Product stability	6
9 Packaging	6
10 Information supplied by the manufacturer	6
Annex A (informative) Example of a suitable method for pH measurement and buffer capacity determination	8
Annex B (normative) Particulate contamination: visible particulates	9
Annex C (informative) Light obscuration test method for particulate contamination: subvisible particles	10
Annex D (informative) Microscopic test method for particulate contamination: subvisible particles	12
Annex E (normative) Intraocular irrigation test	17
Annex F (informative) Clinical investigation	18
Bibliography	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 16671 was prepared by Technical Committee ISO/TC 172, *Optics and optical instruments*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 16671:2003](https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003)

<https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003>

Ophthalmic implants — Irrigating solutions for ophthalmic surgery

1 Scope

This International Standard defines requirements with regards to safety for the intended performance, design attributes, preclinical and clinical evaluation, sterilization, product packaging, product labelling and the information supplied by the manufacturer.

This International Standard applies to ophthalmic irrigating solutions (OISs), used during ophthalmic surgery. These solutions do not provide any primary immunological, pharmacological or metabolic function.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1:—¹⁾, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

ISO 10993-2:1992, *Biological evaluation of medical devices — Part 2: Animal welfare requirements*

ISO 11607:2002, *Packaging for terminally sterilized medical devices*

ISO 13408-1:1998, *Aseptic processing of health care products — Part 1: General requirements*

ISO 14155-1, *Clinical investigation of medical devices for human subjects — Part 1: General requirements*

ISO 14155-2, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans*

ISO 14630:1997, *Non-active surgical implants — General requirements*

ISO 14971-1:1998, *Medical devices — Risk management — Part 1: Application of risk analysis*

ISO 15223:2000, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

EN 868-1:1997, *Packaging materials and systems for medical devices which are to be sterilized — Part 1: General requirements and test methods*

EN 1041:1998, *Information supplied by the manufacturer with medical devices*

EN 12442-1:2000, *Animal tissues and their derivatives utilized in the manufacture of medical devices — Part 1: Analysis and management of risk*

1) To be published. (Revision of ISO 10993-1:1997)

3 Terms and definitions

For the purposes of this International Standard, the following terms and definitions apply.

3.1 delivery system
sealed container in which the product is supplied and any additional components provided to introduce the product into the eye

3.2 ophthalmic irrigating solution (OIS)
aqueous solution that is physiologically compatible with the intraocular environment and functions solely by mechanical means

NOTE It does not provide any primary immunological, pharmacological or metabolic function.

4 Intended performance

The intended performance of the product shall be defined by the manufacturer. The intended performance shall include the general requirements for the intended performance of non-active surgical implants outlined in ISO 14630.

The extent to which the intended performance has been achieved shall be determined, taking into account published standards, published clinical and scientific literature, validated test results, preclinical evaluation and clinical trials.

ITEH STANDARD PREVIEW
(standards.iteh.ai)

5 Design attributes

[ISO 16671:2003](https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003)

5.1 General <https://standards.iteh.ai/catalog/standards/sist/b30b6569-ca09-4ae9-93bc-ca905ff7c658/iso-16671-2003>

The general requirements for non-active surgical implants outlined in ISO 14630 shall apply.

5.2 Concentration of the components

The concentration of each component material in the finished product shall be determined and documented, and the concentration of each component shall be expressed as weight of material per unit volume of solution. Since the testing methodology may affect the actual concentration reported, the standard physical or chemical techniques utilized shall be described and documented. Wherever possible, components shall comply with stated compendial standards.

5.3 Water used

The purity of the water used shall be Water for Injections (e.g. see [3]).

5.4 Characterization of the finished product

5.4.1 General

The manufacturer shall describe and document the physical characteristics that affect the performance of the OIS efficacy in ophthalmic surgery.

NOTE These physical properties should be measured at the conditions expected and relevant at the time of use.

5.4.2 pH and buffering capacity

The pH of the finished product shall be determined and documented with a calibrated pH meter at $25\text{ °C} \pm 2\text{ °C}$.

NOTE The pH of the product should be close to that of the aqueous humour (pH 7,38) in order to prevent damage to the corneal endothelial cells. *In vitro* studies have shown that the pH range tolerated by the endothelium narrows as exposure time increases.

A suitable method shall be used to determine buffering capacity. An example of a suitable method is given in Annex A. The products shall be classified as in Table 1.

Table 1 — Classification according to pH and buffering capacity

Group	Base buffering capacity (mol/l per pH)	Acid buffering capacity (mol/l per pH)	pH range
Essentially unbuffered	< 0,000 5	< 0,004	6,5 to 8,5
Moderately buffered	0,000 5 to 0,005	0,004 to 0,04	6,7 to 8,2
Buffered	> 0,005	> 0,04	7,2 to 7,6

5.4.3 Chemical and biological contaminants

Potentially hazardous chemical or biological contaminants and impurities shall be determined by a risk analysis. For raw materials of biological origin, these contaminants may include proteins, nucleic acids or other biological materials. Contaminants of the finished product that are potentially hazardous either to the tissues of the eye or systemically, shall be identified, whenever possible, and their concentrations in the finished product reported.

ISO 16671:2003

<https://standards.iteh.ai/catalog/standards/sist/b30b6569-ea09-4ae9-93bc-4a09517c9583/iso-16671-2003>

Contaminants shall be determined using standard analytical methods when available, and all methods shall be described. Limits for identified contaminants shall be set and included. Testing for the biological effects of these contaminants during evaluation of biological safety may be required if the risk analysis determines it necessary.

5.4.4 Osmolality

The manufacturer shall determine and document the osmolality range of the OIS. Osmolality of the finished product shall be not less than 200 mosm/kg or greater than 400 mosm/kg. Osmolality shall be determined using either a vapour pressure osmometer or a cryoscopic osmometer.

5.4.5 Spectral transmittance

The spectral transmittance of the finished product shall be recorded over the range 300 nm to 1 100 nm. Results shall be presented graphically, plotting % transmission against wavelength.

5.4.6 Particulates

5.4.6.1 General

There is a potential for adverse events to take place as a result of particles of certain sizes and characteristics in the finished product.

Particulate contamination of OISs consists of extraneous, mobile undissolved particles other than gas bubbles, unintentionally present in the solutions.

A risk assessment shall evaluate the potential for contamination by, or formation of, particulates in the product during manufacture as well as the conditions expected during transport and storage, and during use of the product and the hazards associated with these.

In multi-component products (i.e. where there are two or more separate parts of a product that have to be mixed prior to use) tests shall be applied to the mixed product.

5.4.6.2 Visible particles

The OIS shall be essentially free of visible particles. The method described in Annex B shall be used to determine this.

5.4.6.3 Subvisible particles

Either the light obscuration test method given in Annex C or the microscopic test method given in Annex D shall be used to determine the subvisible particulate level for OISs with the corresponding limits of each method as given below.

The following limits shall apply for the light obscuration test method described in Annex C:

- No more than 50 particles equal to or greater than 10 µm per ml of OIS;
- No more than 5 particles equal to or greater than 25 µm per ml of OIS;
- No more than 2 particles equal to or greater than 50 µm per ml of OIS.

The following limits shall apply for the microscopic test method of Annex D:

- No more than 25 particles equal to or greater than 10 µm per ml of OIS;
- No more than 2,5 particles equal to or greater than 25 µm per ml of OIS;
- No more than 1 particle equal to or greater than 50 µm per ml of OIS.

NOTE The light obscuration test method of Annex C is based on light blockage. Amorphous, semiliquid or otherwise morphologically indistinct materials contribute to light obscuration and therefore contribute to the particle count. In the microscopic test method of Annex D, amorphous, semiliquid or otherwise morphologically indistinct materials appear as stain or discolouration on the surface of the membrane filter, and are not counted as particles. To compensate for this difference, the limits for the microscopic test method are one half those for the light obscuration method.

6 Design evaluation

6.1 General

The OIS shall be evaluated to demonstrate that the intended performance is achieved. The requirements for evaluation of non-active implants outlined in ISO 14630 shall apply.

6.2 Preclinical evaluation of biological safety

6.2.1 General

The procedure for evaluation of biological safety of an OIS shall commence with an assessment of risk, carried out and documented in accordance with ISO 14971-1. The results of the risk analysis shall determine the tests required to evaluate the biological safety of the OIS.

For OISs containing material of animal origin the risk analysis and management requirements outlined in EN 12442-1 shall apply.

NOTE 1 During the risk assessment the manufacturer should take into account the interaction with other ophthalmic products.

For all OISs the requirements for evaluation of biological safety specified in ISO 10993-1 shall apply.

NOTE 2 Based upon the typical clinical applications, OISs are categorized as "Implant devices, tissue/bone". The tests for this and other categories of devices identified in Table 1 of ISO 10993-1:1997 are for guidance only. They do not represent maximum or minimum test requirements.

NOTE 3 It may be possible to combine biocompatibility tests, thereby reducing the number of animals required for testing. Multiple tests can be conducted simultaneously in a single animal provided that the test animal is not subjected to undue pain or distress.

In addition to the biocompatibility tests identified in ISO 10993-1 and by the risk analysis, all of the following tests shall be considered in the selection of tests to evaluate the biological safety of an OIS.

6.2.2 Bacterial endotoxins test

The OIS shall be evaluated for the presence of bacterial endotoxins using the Limulus Amebocyte Lysate (LAL) test, in accordance with USP 24 [5] or an equivalent, validated test procedure. Any product that exceeds a bacterial endotoxin limit of 0,5 endotoxin units (EU) per ml fails the test.

6.2.3 Intraocular irritation and inflammation

If the risk assessment indicates that it is necessary to undertake tests for intraocular irritation, inflammation, intraocular pressure and other local events, such tests shall be conducted in a suitable animal model in accordance with Annex E. The choice of animal species shall be justified and documented. The animal welfare requirements as described in ISO 10993-2 shall apply.

The animal testing shall mirror the intended clinical use as closely as possible.

The study design should assess the intraoperative and postoperative ocular irritation and inflammation of the ophthalmic surgery, with comparative use of the OIS under evaluation and a control OIS that has already been proven to be non-irritating and non-inflammatory in clinical use for five years. The volume of OIS used shall simulate the intended use, accounting for ocular volume differences between the human eye and that of the animal model.

The post-surgical irritation and inflammation shall be monitored and graded in accordance with Annex E. Based upon the risk management plan, appropriate evaluation at appropriate times may include corneal pachymetry and slit lamp biomicroscopy. All adverse effects shall be documented.

The OIS shall show ocular irritation and inflammation results less than or equal to the control OIS, or it shall be excluded from clinical use.

6.3 Clinical evaluation

If clinical evaluation and risk assessment identify the need for a clinical investigation, Annex F shall be considered. In addition, the general requirements concerning clinical investigations of medical devices for human subjects specified in ISO 14155-1 and ISO 14155-2 shall apply.

7 Sterilization

Whenever possible, the product shall be terminally sterilized in its final container. The requirements for sterilization of non-active surgical implants outlined in ISO 14630 shall apply.

Ethylene oxide shall not be used to sterilize the OIS solution and, unless justified, not to sterilize the primary container either. In case of justification and use for the latter, ethylene oxide and related contaminants may diffuse into the solution, for which the following limits shall then apply:

- ethylene oxide: less than 20 µg/ml
- ethylene chlorohydrin: less than 100 µg/ml

NOTE 1 It has been found that the requirements determining acceptable limits for ethylene oxide residuals specified in ISO 10993-7 are inadequate for devices in contact with highly sensitive tissues, such as those of the eye. For this case AAMI TIR No. 19 provides additional guidance to the application of ISO 10993-7.

For OISs that are not terminally sterilized, but aseptically processed, ISO 13408-1 shall apply. The process shall be demonstrated to comply with a contamination rate limit of 10^{-3} by a validated media fill study.

NOTE 2 ISO 13408-1 specifies the general requirements for, and offers guidance on, processes, programmes and procedures for the validation and control of aseptically processed health care products. ISO 13408-1 particularly applies, but is not limited to, the processing of aqueous solutions, and is thus relevant to the preparation of OISs. Future parts of that International Standard will address specialized processes, such as filtration and lyophilization.

8 Product stability

The manufacturer shall define and state the shelf life of the product. Real time or accelerated shelf life testing shall be performed to demonstrate that the finished product remains within specifications for a period of the labelled shelf life under expected conditions of transport and storage. Real time testing shall reflect normal storage temperature and temperature fluctuations and the relative humidity shall be controlled within $60\% \pm 20\%$. In accelerated testing the temperature shall not exceed 45 °C and the relative humidity shall be at least 40 %. The parameters that shall be followed during shelf life studies are the pH, osmolality, particulate levels, sterility (using EP, JP or USP [5-8]), colour and clarity, plus any other factors identified by risk analysis as crucial to safe use of the product.

The established shelf life of the OIS shall be re-validated if a risk assessment identifies any change in manufacture that may affect the stability of the product.

NOTE Changes in the composition of the product, source materials, material suppliers, manufacturing conditions, including the sterilization process, package design or package materials may affect the shelf life of the product.

9 Packaging

9.1 Protection from damage during storage and transport

The packaging requirements for medical devices outlined in ISO 11607 and ISO 14630 shall apply.

9.2 Maintenance of sterility in transit

OISs shall be packaged in such a way that they remain sterile or below the contamination rate specified, as applicable, within the limits specified for conditions of transport, storage and handling. The sterile packaging requirements outlined in EN 868-1 shall apply.

10 Information supplied by the manufacturer

The general requirements for information provided by the manufacturer of medical devices specified in EN 1041 shall apply together with the following particular requirements. Symbols may be used instead of text, where appropriate. When symbols are used the requirements of ISO 15223 shall apply.

The labelling shall contain information on whether the OIS is buffered and if so give information on the buffer type and capacity.

If the product is vulnerable to damage by exposure to environmental elements, there shall be clear warning signs on the shipping container.

NOTE The batch number and expiration date may be provided on a self-adhesive label.

A package insert shall be included within the storage container, provided in such a way that it can be removed and read without damaging the sterile barrier.

The minimum information required on the storage container, package insert, sterile barrier and primary container is listed in Table 2.

Table 2 — Information supplied by the manufacturer

Information	Storage container	Package insert	Sterile barrier (if present)	Primary container
Name of the manufacturer or authorized representative	x	x	x	x
Address of the manufacturer or authorized representative	x	x		
Trade name of product	x	x	x	x
Brief description of the chemical composition of the product and the volume supplied	x	x		
Conditions for storage	x	x		
Indications for use		x		
Contra-indications for use		x		
Warnings, precautions and known interactions with other ocular products		x		
Statement that the contents are for single use only	x	x	x	x
Statement "Sterile" and the method(s) of sterilization of the product and primary container	x	x	x	x
Statement "Do not use if sterile barrier is breached"		x	x	x
Expiration date	x		x	x
Information on whether the OIS is buffered and if so information on the buffer type and capacity		x		
Buffering category (see Table 1)		x		
Batch number preceded by the word "LOT"	x		x	x
NOTE The batch number, expiration date and sterility data need not be provided on the sterile barrier if that is transparent and the required information can be read directly from the primary container without breaching the seal.				