

# SLOVENSKI STANDARD SIST ISO 14507:2019

01-oktober-2019

# Kakovost tal - Priprava vzorcev za določevanje organskih onesnaževal

Soil quality - Pretreatment of samples for determination of organic contaminants

Qualité du sol - Prétraitement des échantillons pour la détermination des contaminants organiques

(standards.iteh.ai)

Ta slovenski standard je istoveten z: ISO 14507:2003

SIST ISO 14507:2019

https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-

22e884277cda/sist-iso-14507-2019

ICS:

13.080.01 Kakovost tal in pedologija na Soil quality and pedology in

splošno general

SIST ISO 14507:2019 en

**SIST ISO 14507:2019** 

# iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 14507:2019

https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019 SIST ISO 14507:2019

# INTERNATIONAL STANDARD

ISO 14507

First edition 2003-03-15

# Soil quality — Pretreatment of samples for determination of organic contaminants

Qualité du sol — Prétraitement des échantillons pour la détermination des contaminants organiques

# iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 14507:2019

https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019



Reference number ISO 14507:2003(E)

### ISO 14507:2003(E)

#### PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

# iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 14507:2019

https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019

#### © ISO 2003

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

# **Contents** Page

Forev	vord	iv
Introd	duction	v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4 4.1 4.2 4.3	Principle Sampling Volatile compounds Moderately volatile compounds	2 2
5	Reagents	3
6	Apparatus	3
7	Preservation and storage	4
8 8.1 8.2 8.3 8.4	Procedure  General	4 5 5
9	https://standards.iteh.a/catalog/standards/sist/520c72e7-d27b-4001-8537- <b>Test report</b>	8
-	x A (informative) Information concerning vapour pressure, boiling and melting points of volatile organic compounds	
Anne	x B (informative) Validation	12
Biblio	ography	14

ISO 14507:2003(E)

# **Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 14507 was prepared by Technical Committee ISO/TC 190, Soil quality, Subcommittee SC 3, Chemical methods and soil characteristics.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 14507:2019</u> https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019

## Introduction

The properties of organic micropollutants can differ greatly according to chemical species:

- they can range from non-volatile to very volatile compounds (low to high vapour pressure);
- they can be labile or reactive at ambient or elevated temperatures;
- they can be biodegradable or UV-degradable;
- they can have considerably different solubilities in water;
- they require different analytical procedures.

Because of these differences, a general pretreatment procedure cannot be proposed. The goal of a pretreatment procedure is to prepare a test sample in which the concentration of the contaminant is equal to the concentration in the original soil, provided, however, that this procedure does not alter the chemical species to be analysed. For instance, if the sample contains only small particles and the contaminant is homogeneously distributed, it is not necessary to grind the sample. In this International Standard, the size 2 mm is used to distinguish between small and large soil particles.

11 en STANDARD PREV

Consistency among the following aspects is important: (standards.iteh.ai)

— soil diversity;

SIST ISO 14507:2019

- the aim of the analysis (including its accuracy) irds/sist/520c72e7-d27b-4001-8537-
  - 22e884277cda/sist-iso-14507-2019
- the nature of the chemical species to be analysed.

The particle size distribution of the sample in relation to the mass of sample taken for analysis is also important to pretreatment. For the analysis of organic contaminants, the sample mass taken in most cases is about 20 g.

With such a sample mass, and provided that the contaminant is homogeneously distributed and the particles in the sample are smaller than about 2 mm, further grinding of the sample is not necessary. If the sample contains large particles or if the contaminant is heterogeneously distributed (as occurs for instance with tar particles), it is not possible to take a representative test sample of about 20 g without grinding the sample. To improve the homogeneity, samples are ground to a size smaller than 1 mm. Prior to analysis, very often no information about the distribution of the contaminant in the soil is known.

Some analytical procedures start with a field-moist sample. Drying of the sample gives lower extraction results. If the sample is not dried, grinding is not possible.

In a situation in which accurate results are needed, the best available pretreatment procedure should be used. If it is necessary to establish whether the concentration is above a certain limit and it is already known that the soil is heavily polluted, the simplest pretreatment procedure may meet the need. In this case however, the result may not be presented as a representative value for the whole sample.

The choice depends above all on the volatility of the organic compounds under analysis. It also depends on the soil particle size distribution, the heterogeneity of the sample and the analytical procedure.

**SIST ISO 14507:2019** 

# iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 14507:2019

https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019

# Soil quality — Pretreatment of samples for determination of organic contaminants

### 1 Scope

This International Standard specifies three methods for the pretreatment of soil samples in the laboratory prior to the determination of organic contaminants:

- if volatile organic compounds are to be measured;
- if moderately volatile to non-volatile organic compounds are to be measured, if the result of the subsequent analysis must be accurate and reproducible, and if the sample contains particles larger than 2 mm and/or the contaminant is heterogeneously distributed;
- if non-volatile organic compounds are to be measured and the extraction procedure prescribes a field-moist sample, or if the largest particles of the sample are smaller than 2 mm and the contaminant is homogeneously distributed. This procedure is also applicable if reduced accuracy and repeatability are acceptable.

The pretreatment described in this International Standard is used in combination with an extraction procedure in which the contaminant is available for the extraction liquid.

NOTE For the pretreatment of soil samples for the purposes of determining non-volatile inorganic compounds and physico-chemical soil characteristics, see ISO 11464.

#### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10381-1, Soil quality — Sampling — Part 1: Guidance on the design of sampling programmes

ISO 11074-2, Soil quality — Vocabulary — Part 2: Terms and definitions relating to sampling

ISO 11465:1993, Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method

### 3 Terms and definitions

For the purposes of this document, the terms and definitions in ISO 11074-2 and the following apply:

#### 3.1

#### volatile compound

organic compound having a boiling point below 300 °C (at a pressure of 101 kPa)

NOTE 1 This includes volatile aromatic and volatile halogenated hydrocarbons as determined in accordance with ISO 15009. Some mono- and dichlorophenols, for instance, and naphthalene also belong to this group.

### ISO 14507:2003(E)

NOTE 2 The selection of the categories for volatile and moderately volatile compounds can be related in principle to the vapour pressure. However, as the vapour pressure of only a small number of compounds is known, and in view of the relationship between vapour pressure and boiling point, the boiling point has been chosen as the criterion for distinction. See Annex A.

#### 3.2

#### moderately volatile compound

organic compound having a boiling point above 300 °C (at a pressure of 101 kPa)

NOTE 1 This definition includes:

- a) mineral oil (see ISO/TR 11046);
- b) most polycyclic aromatic hydrocarbons (PAH) (see ISO 13877);
- c) polychlorobiphenyls (PCB) (see ISO 10382);
- d) organochlorine pesticides (see ISO 10382).

NOTE 2 In ISO/TR 11046, mineral oil is defined as the group of hydrocarbons which, on chromatography, have retention times lying between that of n-decane ( $C_{10}H_{22}$ ) and n-tetracontane ( $C_{40}H_{82}$ ). As in this International Standard the limit for volatile compounds lies at a boiling point of 300 °C [approximately hexadecane ( $C_{16}$ )], this means that mineral oil should be considered as a volatile compound, as part of the compounds to be determined fall within the boiling point range of volatile compounds. However for practical reasons, it has been decided that the pretreatment for moderately volatile compounds should be specified for the determination of mineral oil. As a result of cryogenic crushing, an improvement in the extraction yield occurs for compounds with a boiling point above 300 °C. The possible losses for the lower-boiling hydrocarbons ( $C_{10}$  to  $C_{16}$ ) are assumed to be low, due to the retaining effect of the higher-boiling hydrocarbons present in mineral oil and to be compensated by the higher extraction yield of the other hydrocarbons present. As the total yield is used to determine the mineral oil as a group parameter, it is assumed that pretreatment using the method for moderately volatile compounds gives the best results at present.

### 4 Principle

<u>SIST ISO 14507:2019</u> https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019

# 4.1 Sampling

Sampling shall be carried out in accordance with ISO 10381-1. The samples taken shall be kept cool and processed as soon as possible. The method of pretreatment depends on the volatility of the organic compound(s) or group(s) of organic compounds to be determined.

When taking a subsample, the quantity of soil should be removed from the sample container in such a way that this quantity:

- a) is a subsample over the full depth of the sample, and
- b) can be taken quickly.

### 4.2 Volatile compounds

For determination of volatile organic compounds, core test samples are taken from the sample and extracted according to the relevant specific analytical procedure. If composite samples are required, extracts of individual samples are mixed. It is not possible to obtain composite samples without severe losses of volatiles. The procedure is described in 8.2.

### 4.3 Moderately volatile compounds

Samples are chemically dried at a low temperature (– 196 °C, liquid nitrogen). The dried samples are cooled with liquid nitrogen and ground with a cross-beater mill with a sieve of 1 mm mesh size (cryogenic crushing). After grinding, suitable test portions are processed according to the relevant specific analytical procedures. Composite samples can be prepared by mixing the ground samples. This procedure is described in 8.3.

If the extraction procedure requires a field-moist sample, drying and grinding are not possible.

If the original sample only contains a small fraction of particles greater than 2 mm and the distribution of contaminants is likely to be homogeneous, grinding may be omitted. In these two cases, suitable test portions are taken directly after mixing the sample. This procedure is described in 8.4.

NOTE 1 To distinguish the volatile organic compounds from the moderately volatile organic compounds, boiling points are used instead of the vapour pressure at ambient temperature. This is explained in Annex A, which also gives boiling points and vapour pressures of compounds regularly determined in soil investigations.

NOTE 2 For some specific components in the group of moderately volatile organic compounds, freeze-drying may give good results. In this International Standard freeze-drying is not described.

# 5 Reagents

Use only reagents of recognized analytical quality. Check samples of each batch of the reagents for the presence of contaminating compounds.

5.1 Sodium sulfate, anhydrous.

Heat the sodium sulfate before use for at least 6 h at about 550 °C to remove crystalline water and organic materials. After heating, allow to cool in a desiccator and store in a closed container.

5.2 Magnesium silicate (talcum powder).

# Teh STANDARD PREVIEW

5.3 Sand or gravel.

(standards.iteh.ai)

Before use, wash the sand or gravel at least twice with an equivalent mass of demineralized water and then heat for 6 h at about 550 °C to remove organic materials 2010

**5.4 Liquid nitrogen.** https://standards.iteh.ai/catalog/standards/sist/520c72e7-d27b-4001-8537-22e884277cda/sist-iso-14507-2019

For safety precautions, see warnings in 8.3.2.

# 6 Apparatus

Usual laboratory glassware and equipment, and in particular:

- **6.1 Glass containers** with a volume of 750 ml to 1 000 ml, with a wide neck and screw cap with a polytetrafluoroethylene (PTFE) inlay.
- **6.2 Polyethylene containers** with a volume of 750 ml to 1 000 ml, with a wide neck and screw cap. To prevent a significant headspace being formed, do not use larger containers.
- **6.3** Refrigerator, capable of maintaining a temperature below 10 °C.
- **6.4 Dewar vessel(s)**, capable of holding at least one polyethylene container of about 750 ml.
- **6.5** Cross-beater mill or mill with comparable qualities, with a sieve of mesh size 1 mm and accessories. A cross-beater mill as used in most soil laboratories is suitable for milling soil samples cooled with liquid nitrogen.
- **6.6 Gloves**, suitable for working at low temperatures.
- **6.7** Oven, suitable for heating to about 550 °C.