



# SLOVENSKI STANDARD SIST ISO 10382:2019

01-oktober-2019

---

## Kakovost tal - Določevanje organoklornih pesticidov in polikloriranih bifenilov - Plinska kromatografija z detektorjem z zajemom elektronov (ECD)

Soil quality - Determination of organochlorine pesticides and polychlorinated biphenyls -  
Gas-chromatographic method with electron capture detection

### iTeh STANDARD PREVIEW

Qualité du sol - Dosage des pesticides organochlorés et des biphényles polychlorés -  
Méthode par chromatographie en phase gazeuse avec détection par capture d'électrons

[SIST ISO 10382:2019](http://standards.iTeh.si/catalog/stanards/SIST/ISO/10382-2019-4ccc-b8a8-3c373685d053/sist-iso-10382-2019)

Ta slovenski standard je istoveten z: **ISO 10382:2002**

#### ICS:

13.080.10	Kemijske značilnosti tal	Chemical characteristics of soils
71.040.50	Fizikalnokemijske analitske metode	Physicochemical methods of analysis

**SIST ISO 10382:2019**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST ISO 10382:2019

<https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019>

# INTERNATIONAL STANDARD

**ISO  
10382**

First edition  
2002-10-15

---

---

## **Soil quality — Determination of organochlorine pesticides and polychlorinated biphenyls — Gas- chromatographic method with electron capture detection**

**iTeh STANDARD PREVIEW**  
*Qualité du sol — Dosage des pesticides organochlorés et des biphényles  
polychlorés — Méthode par chromatographie en phase gazeuse avec  
détection par capture d'électrons*  
(standards.iteh.ai)

SIST ISO 10382:2019

<https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019>



Reference number  
ISO 10382:2002(E)

© ISO 2002

## ISO 10382:2002(E)

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 10382:2019](https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019)

<https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019>

© ISO 2002

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.ch](mailto:copyright@iso.ch)  
Web [www.iso.ch](http://www.iso.ch)

Printed in Switzerland

## Contents

	Page
Foreword .....	iv
1 Scope .....	1
2 Normative references .....	1
3 Principle .....	1
4 Reagents .....	2
5 Apparatus .....	4
6 Preparation of standard solutions of PCB and OCP .....	7
7 Sampling and preservation of samples .....	8
7.1 Sampling .....	8
7.2 Sample preservation and pretreatment .....	8
8 Procedure .....	8
8.1 Blank .....	8
8.2 Extraction and concentration .....	8
8.3 Clean-up of the extract .....	9
8.4 Column-chromatographic separation of PCBs and non-polar OCPs from several polar OCPs .....	9
8.5 Gas chromatographic analysis .....	10
9 Test report .....	14
10 Accuracy .....	14
Annex A (informative) Table of retention times of polychlorinated biphenyls and organochlorine pesticides for two different capillary columns .....	15
Annex B (informative) Scheme for the preparation of standard solutions including injection standards .....	16
Annex C (informative) Results of an interlaboratory trial carried out in the Netherlands .....	17
Annex D (informative) Clean-up to remove elemental sulfur and some other organic sulfur compounds .....	20
Bibliography .....	22

**ISO 10382:2002(E)****Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10382 was prepared by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

Annexes A, B, C and D of this International Standard are for information only.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST ISO 10382:2019](https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019)

<https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019>

# Soil quality — Determination of organochlorine pesticides and polychlorinated biphenyls — Gas-chromatographic method with electron capture detection

## 1 Scope

This International Standard specifies a method for quantitative determination of seven polychlorinated biphenyls and seventeen organochlorine pesticides in soil.

This International Standard is applicable to all types of soil.

Under the conditions specified in this International Standard, limits of detection of 0,1 µg/kg to 4 µg/kg (expressed as dry matter) can be achieved.

## 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 10381-1, *Soil quality — Sampling — Part 1: Guidance on the design of sampling programmes*

ISO 10381-2, *Soil quality — Sampling — Part 2: Guidance on sampling techniques*

ISO 11465:1993, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*

ISO 14507, *Soil quality — Pretreatment of samples for the determination of organic contaminants*

## 3 Principle

After pretreatment, the soil test sample is extracted with a hydrocarbon solvent.

The extract is concentrated; polar compounds are removed by passing the concentrated extract through a column filled with aluminium oxide. The eluate is concentrated.

Elemental sulfur is removed from the concentrated extract, if necessary, by treatment with tetrabutylammonium sulfite reagent.

The extract is analysed by gas chromatography. The various compounds are separated using a capillary column with an immobile phase of low polarity. Detection occurs with an electron-capture detector (ECD).

Polychlorinated biphenyls (PCBs) and organochlorine pesticides (OCPs) are assigned and quantified by comparison of relative retention times and relative peak heights (or peak areas) with respect to injection standards

**ISO 10382:2002(E)**

added, with the corresponding variables of an external standard solution. The efficiency of the procedure depends on the composition of the soil that is investigated. The described procedure does not take account of incomplete extraction due to the structure and composition of the soil sample.

The limit of detection is dependent on the determinands, the equipment used, the quality of chemicals used for extraction of the soil sample, and the clean-up of the extract.

NOTE 1 For confirmation of the identity of detected compounds and the concentrations found, further investigation is necessary. Confirmation can be carried out by repeating the gas chromatographic analysis using a column of different polarity and/or using gas chromatography/mass spectrometry (GC/MS).

NOTE 2 Other non-volatile organochlorine compounds, e.g. some chlorobenzenes, can also be identified and quantified by this method.

**4 Reagents**

All reagents shall be of recognized analytical grade. The purity of the reagents used shall be checked by running a blank determination as described in 8.1.

**4.1 Petroleum ether**, boiling range 40 °C to 60 °C.

**4.2 Acetone**.

**4.3 n-Hexane**.

**4.4 Diethyl ether**.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

Diethyl ether can contain peroxides which may oxidize some of the determinands. Check for the absence of peroxides, e.g. by shaking with a freshly prepared 10 % (mass fraction) KI solution.

**4.5 Anhydrous sodium sulfate**, heated for at least 6 h to 550 °C ± 20 °C, cooled to about 200 °C in a furnace and then to ambient temperature in a desiccator containing magnesium perchlorate or a suitable alternative.

The anhydrous sodium sulfate shall be kept carefully sealed.

**4.6 Aluminium oxide**, basic or neutral, areic mass 200 m<sup>2</sup>/g, activity Super I according to Brockmann.

**4.7 Aluminium oxide**, deactivated with 10 % water.

To 90 g of aluminium oxide (4.6) add 10 g of water. Shake until all lumps have disappeared. Allow the aluminium oxide to condition before use for approximately 16 h, sealed from the air.

**4.8 Silica gel**, particle size 60 µm to 200 µm, deactivated with 5 % water.

Heat 95 g of silica gel for at least 24 h in an oven at 150 °C. Then allow to cool in a desiccator and add 5 g of water. Shake until all lumps have disappeared. Allow the silica gel to condition before use for approx. 16 h, sealed from the air.

For each new batch of aluminium oxide or silica gel, the elution pattern should be checked against a standard solution of PCB and OCP. If necessary, the deactivation of the adsorbent should be adjusted (see 8.4).



## 4.9 Standards.

### 4.9.1 Polychlorinated biphenyls.

PCB- 28: 2,4,4'-trichlorobiphenyl	CAS number <sup>1)</sup> : 7012-37-5
PCB- 52: 2,2',5,5'-tetrachlorobiphenyl	CAS number: 35693-99-3
PCB-101: 2,2',4,5,5'-pentachlorobiphenyl	CAS number: 37680-73-2
PCB-118: 2,3',4,4',5-pentachlorobiphenyl	CAS number: 31508-00-6
PCB-138: 2,2',3,4,4',5'-hexachlorobiphenyl	CAS number: 35065-28-2
PCB-153: 2,2',4,4',5,5'-hexachlorobiphenyl	CAS number: 35065-27-1
PCB-180: 2,2',3,4,4',5,5'-heptachlorobiphenyl	CAS number: 35065-29-3

NOTE The numbers 28, 52, etc. correspond with the sequential numbers of chlorobiphenyls according to the IUPAC rules for the nomenclature of organic compounds.

### 4.9.2 Organochlorine pesticides.

Hexachlorobenzene (HCB)	CAS number: 118-74-1
$\alpha$ -Hexachlorocyclohexane ( $\alpha$ -HCH)	CAS number: 319-84-6
$\beta$ -Hexachlorocyclohexane ( $\beta$ -HCH)	CAS number: 319-85-7
$\gamma$ -Hexachlorocyclohexane ( $\gamma$ -HCH)	CAS number: 58-89-9
Aldrin	CAS number: 309-00-2
Dieldrin	CAS number: 60-57-1
Endrin	CAS number: 72-20-8
Heptachlor	CAS number: 76-44-8
Heptachloro epoxide (exo-, cis- or a-isomer)	CAS number: 28044-83-9
Heptachloro epoxide (endo-, trans- or b-isomer)	CAS number: 1024-57-3
$\alpha$ -Endosulfan	CAS number: 959-98-7
<i>p,p'</i> -DDE	CAS number: 72-55-9
<i>o,p'</i> -DDD	CAS number: 53-19-0
<i>o,p'</i> -DDT	CAS number: 784-02-6
<i>p,p'</i> -DDD	CAS number: 72-54-8
<i>o,p'</i> -DDE	CAS number: 3424-82-6
<i>p,p'</i> -DDT	CAS number: 50-29-3

1) Registration used by the Chemical Abstracts Service.

**ISO 10382:2002(E)****4.9.3 Injection standards.**

PCB-155: 2,2',4,4',6,6'-hexachlorobiphenyl CAS number: 33979-03-2

Select a second injection standard, not interfering with the analytes, from the following substances:

PCB-143: 2,2',3,4,5,6'-hexachlorobiphenyl CAS number: 68194-15-0

PCB-207: 2,2',3,3',4,4',5,6,6'-nonachlorobiphenyl CAS number: 52663-79-3

Mirex CAS number: 2385-85-5

**4.10 Tetrabutylammonium reagent (TBA sulfite reagent).**

Saturate a solution of tetrabutylammonium hydrogen sulfate in a mixture of equal volumes of water and 2-propanol,  $c[(C_4H_9)_4NHSO_4] = 0,1 \text{ mol/l}$ , with sodium sulfite.

NOTE 25 g of sodium sulfite is normally sufficient for 100 ml of solution.

**4.11 n-Heptane.****5 Apparatus****5.1 Customary laboratory glassware**

All glassware to be used shall be thoroughly cleaned, preferably in a dishwasher using a customary cleaning procedure, followed by rinsing with acetone and a subsequent rinsing with petroleum ether or hexane.

**5.2 Glass sample bottles**, of nominal capacity 1 l, with screw top and polytetrafluoroethylene seal (PTFE).

**5.3 Shaking device**, with horizontal movement (200 to 300 strokes per minute).

**5.4 Water bath**, capable of being heated to 100 °C.

**5.5 Shaking funnels**, with a capacity of 2 l.

**5.6 Conical flasks**, with a capacity of 500 ml.

**5.7 Evaporator**, Kuderna Danish (see Figure 1).

Other evaporators, e.g. a rotary evaporator, may be used if found to be equally suitable.

**5.8 Quartz wool or silanized glass wool**, rinsed with petroleum ether or hexane.

**WARNING Working with quartz wool imposes a risk to health through the release of fine quartz particles. Prevent inhalation of these by using a fume cupboard and wearing a dust mask.**

**5.9 Boiling chips**, of glass or porcelain beads, rinsed with petroleum ether or hexane.

**5.10 Calibrated test tubes**, with a capacity of 15 ml and ground glass stoppers.

**5.11 Chromatography tubes** (see Figure 2).

**5.12 Gas chromatograph**, equipped with a non-discriminating injection system, capillary column and electron-capture detector (ECD) based on  $^{63}\text{Ni}$ .

NOTE 1 Working with an encapsulated radioactive source such as that present in an ECD requires a licence in accordance with the appropriate national regulations.

NOTE 2 Gas chromatographs equipped with two detectors and with facilities for connecting two capillary columns to the same injection system are very well suited for this analysis; with such apparatus the confirmatory analysis can be performed simultaneously.

**5.13 Capillary column**, of fused silica, with a length of 50 m and an internal diameter of about 0,25 mm coated with a film of cross-linked polysiloxane.

Other columns can also be used, although in some cases unsatisfactory separation is obtained. A column coated with a moderate polar phase, e.g. CP-Sil 19, OV 1701 etc., shall be used to confirm the result obtained.

NOTE The retention times for PCB and OCP on capillary columns coated with CP-Sil 8 and CP-Sil 19 are given in annex A.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 10382:2019](https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019)

<https://standards.iteh.ai/catalog/standards/sist/6f66a718-a0a7-4ccc-b8a8-3c373685d053/sist-iso-10382-2019>