



# SLOVENSKI STANDARD

## SIST EN ISO 12870:2012

01-september-2012

**Nadomešča:**  
**SIST EN ISO 12870:2009**

---

**Očesna optika - Okviri očal - Zahteve in preskusne metode (ISO 12870:2012)**

Ophthalmic optics - Spectacle frames - Requirements and test methods (ISO 12870:2012)

Augenoptik - Brillenfassungen - Anforderungen und Prüfverfahren (ISO 12870:2012)

Optique ophtalmique - Montures de lunettes - Exigences et méthodes d'essai (ISO 12870:2012)

[SIST EN ISO 12870:2012](https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-6551c65/sist-en-iso-12870-2012)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-6551c65/sist-en-iso-12870-2012>

**Ta slovenski standard je istoveten z: EN ISO 12870:2012**

---

**ICS:**

11.040.70      Oftalmološka oprema      Ophthalmic equipment

**SIST EN ISO 12870:2012**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 12870:2012](#)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>

EUROPEAN STANDARD

EN ISO 12870

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2012

ICS 11.040.70

Supersedes EN ISO 12870:2009

English Version

## Ophthalmic optics - Spectacle frames - Requirements and test methods (ISO 12870:2012)

Optique ophtalmique - Montures de lunettes - Exigences et méthodes d'essai (ISO 12870:2012)

Augenoptik - Brillenfassungen - Anforderungen und Prüfverfahren (ISO 12870:2012)

This European Standard was approved by CEN on 23 March 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST EN ISO 12870:2012](https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

<b>Contents</b>	<b>Page</b>
Foreword.....	<b>3</b>
<b>Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EC Directive 93/42/EEC.....</b>	<b>4</b>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 12870:2012](https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012)  
<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>

## Foreword

This document (EN ISO 12870:2012) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2012, and conflicting national standards shall be withdrawn at the latest by October 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 12870:2009.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN ISO 12870:2012](https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7e5f/sist-en-iso-12870-2012)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7e5f/sist-en-iso-12870-2012>

### Endorsement notice

The text of ISO 12870:2012 has been approved by CEN as a EN ISO 12870:2012 without any modification.

## Annex ZA (informative)

### Relationship between this European Standard and the Essential Requirements of EC Directive 93/42/EEC

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

**Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC**

Clauses/subclauses of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/notes
4.1, 4.2, 4.5, 4.6, 4.7, 4.8, 4.9, 9, 10	1	The requirements declared as optional in Table 1, but cited in the list given in footnote <sup>a</sup> , only provide presumption of conformity to their corresponding Essential Requirements if they are complied with. See below in the present table.
4.1, 4.2, 4.5, 4.6, 4.7, 4.8, 4.9, 9, 10	2	The requirements declared as optional in Table 1, but cited in the list given in footnote <sup>a</sup> , only provide presumption of conformity to their corresponding Essential Requirements if they are complied with. See below in the present table.
4	3	Testing in accordance with Clauses 5, 6, 7 and 8.
4.6 to 4.9	4	Testing in accordance with 8.2 to 8.6.
4.2.1, 4.2.2, 4.2.3	6	Testing in accordance with 8.8. The requirement of 4.2.3 (i.e. 0,5 µg/cm <sup>2</sup> /week) is the requirement set forth by Directive 94/27/EEC. The test in 8.8 makes reference to EN 16128 and EN 12472. See also explanations in Annex D.
4.2.4	6 a)	—
4.2.1, 4.2.2, 4.2.3	7.1	Testing in accordance with 8.8. The requirement of 4.2.3 (i.e. 0,5 µg/cm <sup>2</sup> /week) is the requirement set forth by Directive 94/27/EEC. The test in 8.8 makes reference to EN 16128 and EN 12472. See also explanations in Annex D.
4.6 to 4.9	7.1	Testing in accordance with 8.2 to 8.6.

Table ZA.1 (end)

Clauses/subclauses of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/notes
4.2.1, 4.2.2, 4.2.3	7.2	Testing in accordance with 8.8. The requirement of 4.2.3 (i.e. 0,5 µg/cm <sup>2</sup> /week) is the requirement set forth by Directive 94/27/EEC. The test in 8.8 makes reference to EN 16128 and EN 12472. See also explanations in Annex D.
4.6 to 4.9	7.3	Testing in accordance with 8.2 to 8.6.
4.2.2, 4.2.3	7.5	Testing in accordance with 8.8. Essential Requirement 7.5 is only partly addressed in ISO 12870. To the extent that it is covered in ISO 12870, testing is carried out in accordance with 8.8. The requirement of 4.2.3 (i.e. 0,5 µg/cm <sup>2</sup> /week) is the requirement set forth by Directive 94/27/EEC. The test in 8.8 makes reference to EN 16128 and EN 12472. See also explanations in Annex D.
4.8	9.1	Testing in accordance with 8.4 and 8.5.
4.9	9.3	Testing in accordance with 8.6.
9, 10	13.1	—
9, 10	13.3	The statement in 10.4 is true for the countries of the Community [cf. ER 13.3 a)].

**WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 12870:2012](#)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>



# INTERNATIONAL STANDARD

**ISO**  
**12870**

Third edition  
2012-04-01

---

---

## Ophthalmic optics — Spectacle frames — Requirements and test methods

*Optique ophtalmique — Montures de lunettes — Exigences et  
méthodes d'essai*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 12870:2012](https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>



Reference number  
ISO 12870:2012(E)

© ISO 2012

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 12870:2012

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

<b>Contents</b>	<b>Page</b>
Foreword .....	iv
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Requirements</b> .....	<b>2</b>
<b>4.1 General</b> .....	<b>2</b>
<b>4.2 Physiological compatibility</b> .....	<b>3</b>
<b>4.3 Measurement system</b> .....	<b>4</b>
<b>4.4 Dimensional tolerances on nominal size</b> .....	<b>4</b>
<b>4.5 Tolerance on screw threads</b> .....	<b>4</b>
<b>4.6 Dimensional stability at elevated temperature</b> .....	<b>4</b>
<b>4.7 Resistance to perspiration</b> .....	<b>4</b>
<b>4.8 Mechanical stability</b> .....	<b>5</b>
<b>4.9 Resistance to ignition</b> .....	<b>6</b>
<b>4.10 Resistance to optical radiation</b> .....	<b>6</b>
<b>5 Selection of test samples</b> .....	<b>6</b>
<b>5.1 General</b> .....	<b>6</b>
<b>5.2 Testing for nickel release</b> .....	<b>6</b>
<b>5.3 Change in spectacle frame model</b> .....	<b>6</b>
<b>6 Preparation and conditioning of test samples</b> .....	<b>6</b>
<b>6.1 Test lenses</b> .....	<b>6</b>
<b>6.2 Sample conditioning and test conditions</b> .....	<b>7</b>
<b>7 Testing, inspection and compliance</b> .....	<b>7</b>
<b>7.1 Testing</b> .....	<b>7</b>
<b>7.2 Inspection and examination</b> .....	<b>8</b>
<b>7.3 Compliance</b> .....	<b>8</b>
<b>8 Test methods</b> .....	<b>9</b>
<b>8.1 General</b> .....	<b>9</b>
<b>8.2 Test for dimensional stability at elevated temperature</b> .....	<b>9</b>
<b>8.3 Test for resistance to perspiration</b> .....	<b>10</b>
<b>8.4 Bridge deformation and lens retention test</b> .....	<b>11</b>
<b>8.5 Endurance test</b> .....	<b>13</b>
<b>8.6 Test for resistance to ignition</b> .....	<b>14</b>
<b>8.7 Test for resistance to optical radiation</b> .....	<b>15</b>
<b>8.8 Nickel release</b> .....	<b>16</b>
<b>9 Marking</b> .....	<b>18</b>
<b>10 Additional information to be supplied by the manufacturer or other person placing the product on the market</b> .....	<b>19</b>
<b>11 Reference to ISO 12870</b> .....	<b>20</b>
<b>Annex A (informative) Recommendations for the design of spectacle frames</b> .....	<b>21</b>
<b>Annex B (informative) Examples of layout of test equipment</b> .....	<b>23</b>
<b>Annex C (informative) Examples of locations for cutting metal spectacle frames before testing for nickel release</b> .....	<b>26</b>
<b>Annex D (informative) European requirements and legislation on nickel release</b> .....	<b>27</b>
<b>Bibliography</b> .....	<b>28</b>

**ISO 12870:2012(E)****Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 12870 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This third edition cancels and replaces the second edition (ISO 12870:2004) which has been technically revised.

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

[SIST EN ISO 12870:2012](https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012)

<https://standards.iteh.ai/catalog/standards/sist/225923a1-5b2e-4d5c-924a-23a5da5e7c65/sist-en-iso-12870-2012>

# Ophthalmic optics — Spectacle frames — Requirements and test methods

## 1 Scope

This International Standard specifies fundamental requirements for unglazed spectacle frames designed for use with all prescription lenses. It is applicable to frames at the point of sale by the manufacturer or supplier to the retailer.

This International Standard is applicable to all spectacle frame types, including rimless mounts, semi-rimless mounts and folding spectacle frames. It is also applicable to spectacle frames made from natural organic materials.

NOTE See Annex A for recommendations on the design of spectacle frames.

This International Standard is not applicable to complete custom-made spectacle frames or to products designed specifically to provide personal eye protection.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 105-A02, *Textiles — Tests for colour fastness — Part A02: Grey scale for assessing change in colour*

ISO 105-B02, *Textiles — Tests for colour fastness — Part B02: Colour fastness to artificial light: Xenon arc fading lamp test*

ISO 3160-1, *Watch cases and accessories — Gold alloy coverings — Part 1: General requirements*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 7998, *Ophthalmic optics — Spectacle frames — Lists of equivalent terms and vocabulary*

ISO 8596, *Ophthalmic optics — Visual acuity testing — Standard optotype and its presentation*

ISO 8624:2011, *Ophthalmic optics — Spectacle frames — Measuring system and terminology*

ISO 11380, *Optics and optical instruments — Ophthalmic optics — Formers*

ISO 11381, *Optics and optical instruments — Ophthalmic optics — Screw threads*

ISO/TS 24348, *Ophthalmic optics — Spectacle frames — Method for the simulation of wear and detection of nickel release from coated metal and combination spectacle frames*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7998 and ISO 8624 and the following apply.

### 3.1

#### spectacle frame model

spectacle frame produced to a common design, using the same materials (but not necessarily the same pigmentation) and surface treatment