
Tekoči naftni proizvodi - Goriva na osnovi srednjih destilatov, metilnih estrov maščobnih kislin (FAME) in mešanic - Ugotavljanje oksidacijske stabilnosti z oksidacijsko metodo rapidne male skale

Liquid petroleum products - Middle distillates and fatty acid methyl ester (FAME) fuels and blends - Determination of oxidation stability by rapid small scale oxidation method

Flüssige Mineralölerzeugnisse - Mitteldestillat- und Fettsäuremethylesterkraftstoffe und Mischungen - Bestimmung der Oxidationsstabilität mit beschleunigtem Verfahren und kleiner Probenmenge

Produits pétroliers liquides - Carburants et mélanges des distillats moyens et des esters méthyliques d'acides gras (EMAG) - Détermination de la stabilité à l'oxydation par méthode d'oxydation accélérée petite échelle

Ta slovenski standard je istoveten z: EN 16091:2011

ICS:

75.160.20 Tekoča goriva Liquid fuels

SIST EN 16091:2012 **en,de**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 16091:2012

<https://standards.iteh.ai/catalog/standards/sist/b1694fe5-beea-425e-af4c-6fd31223bb16/sist-en-16091-2012>

EUROPEAN STANDARD

EN 16091

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2011

ICS 75.160.20

English Version

Liquid petroleum products - Middle distillates and fatty acid methyl ester (FAME) fuels and blends - Determination of oxidation stability by rapid small scale oxidation method

Produits pétroliers liquides - Distillats moyens, esters méthyliques d'acides gras (EMAG) et leurs mélanges - Détermination de la stabilité à l'oxydation par méthode d'oxydation accélérée à petite échelle

Flüssige Mineralölerzeugnisse - Mitteldestillat- und Fettsäuremethylesterkraftstoffe und Mischungen - Bestimmung der Oxidationsstabilität mit beschleunigtem Verfahren und kleiner Probenmenge

This European Standard was approved by CEN on 15 October 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Principle.....	4
5 Reagents and materials	4
6 Apparatus	5
6.1 Automatically controlled oxidation tester, comprising an oxidation reaction vessel containing:	5
6.2 Pipette, capable of delivering 5,0 ml \pm 0,1 ml.	5
6.3 O-ring seals, see A.6.	5
7 Sampling and sample handling.....	5
8 Performance check of the apparatus.....	5
9 Preparation of the apparatus.....	6
10 Procedure	6
11 Expression of results	7
12 Precision.....	7
12.1 General.....	7
12.2 Repeatability, <i>r</i>	7
12.3 Reproducibility, <i>R</i>	7
13 Test report	7
Annex A (normative) Apparatus description.....	8
A.1 General requirements.....	8
A.2 Assembly of apparatus	8
A.3 Reaction vessel.....	9
A.4 Screw cap	9
A.5 Electric heater	9
A.6 O-ring seals	9
A.7 Valves.....	9
A.8 Pressure sensor.....	9
A.9 Temperature sensor	9
A.10 Connecting pipes.....	9
A.11 Cooling fan	10
Annex B (informative) Calibration process.....	11
Annex C (informative) Determination of induction period	12
Bibliography	13

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 16091:2012

<https://standards.iteh.ai/catalog/standards/sist/b1694fc5-bcca-425e-a44c-6fd31225bb16/sist-en-16091-2012>

Foreword

This document (EN 16091:2011) has been prepared by Technical Committee CEN/TC 19 "Gaseous and Liquid fuels, lubricants and related products of petroleum, synthetic and biological origin", the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2012, and conflicting national standards shall be withdrawn at the latest by May 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 16091:2012](#)

<https://standards.iteh.ai/catalog/standards/sist/b1694fe5-beea-425e-af4c-6fd31223bb16/sist-en-16091-2012>

EN 16091:2011 (E)**1 Scope**

This European Standard specifies a method for the determination of the oxidation stability of middle distillate fuels, fatty acid methyl ester (FAME) fuel and blends thereof, under accelerated conditions, by measuring the induction period to the specified breakpoint in a reaction vessel charged with the sample and oxygen.

NOTE 1 For the purposes of this European Standard, the term “% (V/V)” is used to represent the volume fraction (φ).

NOTE 2 The induction period is used as an indication for the resistance of middle distillates, fatty acid methyl ester (FAME) fuels and blends thereof against oxidation. It should be recognized, however, that this correlation can vary markedly under different conditions with different FAMEs and diesel fuel blends.

NOTE 3 The presence of ignition improvers may lead to lower oxidation stability results determined by this method. It has for instance been observed that the addition of 2-ethyl hexyl nitrate (2EHN) can reduce the measured oxidation stability values.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3170, *Petroleum liquids — Manual sampling (ISO 3170:2004)*

EN ISO 3171, *Petroleum liquids — Automatic pipeline sampling (ISO 3171:1988)*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

3 Terms and definitions

[SIST EN 16091:2012](#)

<https://standards.iteh.ai/catalog/standards/sist/b1694fe5-beea-425e-af4c->

For the purposes of this document, the following terms and definitions apply.

3.1**breakpoint**

point in the pressure-time curve, where the pressure in the vessel has dropped 10 % below the maximum pressure achieved in the current test run

3.2**induction period**

time elapsed between starting the heating procedure of the reaction vessel, charged with sample and oxygen, and the breakpoint at 140 °C

4 Principle

At ambient temperature, a known volume of a sample is placed in a reaction vessel charged with oxygen to a pressure of 700 kPa \pm 5 kPa. The reaction vessel is heated to 140 °C. The pressure in the vessel drops as the oxygen is consumed during the oxidation of the sample. The pressure in the vessel is recorded at intervals of 1s until the breakpoint is reached. The elapsed time from start to the breakpoint is the induction period at the test temperature of 140 °C \pm 0,5 °C.

5 Reagents and materials

5.1 Cleaning solvent, for the removal of oxidation residues from the test vessel, of suitable purity to leave no residue on the apparatus.

NOTE Commercially available ethanol of approx. 95 % (V/V) purity was found to be suitable.

5.2 Oxygen, extra dry (< 5 mg/kg water), commercially available, with a purity of not less than 99,6 %.

5.3 Cleaning tissues, lint-free for cleaning sensitive galvanic coated surfaces without scratching.

5.4 Verification fluid, hydrocarbon fuel (B0).

NOTE Any hydrocarbon fuel with sufficient stability and a known induction period may be used as verification fluid. In general, verification fluid with a certified induction period is available from the manufacturer of the apparatus.

6 Apparatus

6.1 Automatically controlled oxidation tester, comprising an oxidation reaction vessel containing:

- test sample cup capable of being rapidly heated;
- pressure sensor capable of measuring pressures of 1 kPa up to 2 000 kPa with a resolution of 1 kPa;
- temperature sensor with a resolution of 0,1 °C;
- pressure and temperature recording of 1 second.

The oxidation reaction vessel shall be fitted with filling and relief valves, meant to release the pressure, and a fan to cool the vessel from test to ambient temperature.

The requirements for the apparatus are described in detail in Annex A.

6.2 Pipette, capable of delivering 5,0 ml ± 0,1 ml.

6.3 O-ring seals, see A.6. standards.iteh.ai/catalog/standards/sist/b1694fe5-beea-425e-af4c-6fd31223bb16/sist-en-16091-2012

7 Sampling and sample handling

Unless otherwise specified, samples shall be taken as described in EN ISO 3170 or EN ISO 3171 and/or in accordance with the requirements of national standards or regulations for the sampling of the product under test.

Collect and store samples in an opaque container to minimize exposure to light.

8 Performance check of the apparatus

8.1 Verify the correct operation of the equipment by following the procedure in section 10 using the verification fluid (5.4).

8.2 If the induction period of the verification fluid does not meet specified time, repeat 8.1. If it fails a second time, refer to the manufacturer's instructions.

8.3 Pressure and temperature are critical parameters. Proper calibration of the respective sensors (see 6.1) is therefore important. Recalibrate the tester at least every 12 months for correct temperature and pressure readings.

NOTE In view of general good laboratory practice and harmonization of test methods, calibration as in Annex B is strongly recommended. As the calibration procedure for a pressure sensor is complicated and difficult to be executed by an untrained lab technician, one may (in most cases sufficiently) rely on the verification fluid.

EN 16091:2011 (E)**9 Preparation of the apparatus**

9.1 Remove the previous test portion by means of a pipette or similar device.

9.2 Remove the used "O-ring" seal and discard.

NOTE To avoid contamination of the new test, it is necessary to discard the used "O-ring" seal, because it may be soaked with oxidation products from the previous test.

9.3 Wipe the sample cup, the seal groove and the cover with lint-free cleaning tissue (5.3) soaked with cleaning solvent (5.1) until free of gum or other oxidation residues.

9.4 Allow the test sample cup and cover to dry in air and visually inspect for cleanliness.

NOTE Compressed air is generally unsuitable to accelerate the evaporation of solvent because it can contain traces of oil that may influence the next test.

9.5 Insert a new "O-ring" seal (6.3).

10 Procedure

10.1 Bring the reaction vessel and the sample to be tested to room temperature.

10.2 Using a pipette (6.2), place $5 \text{ ml} \pm 0,1 \text{ ml}$ of the sample into the reaction vessel sample cup.

10.3 Cover the sample-cup with the screw cap (see A.2) and close the reaction vessel.

10.4 Pre-flush the reaction vessel with oxygen. Introduce oxygen (5.2) into the vessel until a pressure of $700 \text{ kPa} \pm 5 \text{ kPa}$ is attained and stabilized over 20 sec.

10.5 Start the heater with no delay between charging with oxygen and starting the test.

10.6 The test temperature of $140 \text{ }^\circ\text{C} \pm 0,5 \text{ }^\circ\text{C}$ shall be reached after $270 \text{ s} \pm 30 \text{ s}$.

10.7 If, during the initial 5 min of the test, a steady drop in pressure is observed, discontinue the test and discard the test specimen.

The leakage rate at $140 \text{ }^\circ\text{C}$ with empty sample cup shall not exceed a value of 2 kPa/h . If the leakage rate shows an increase, check the following components:

- O-ring for damage or residues of samples,
- surface of sample cup for damage, and
- sample cup for sample residues.

Contact the manufacturer to resolve leakage problems from other parts of the instrument.

10.8 The apparatus automatically records the temperature (to the nearest $0,1 \text{ }^\circ\text{C}$) and pressure (to the nearest 1 kPa) of the oxidation vessel continuously.

10.9 The apparatus terminates the test when the pressure readings show a 10 % drop from the maximum observed pressure (breakpoint) and records the time elapsed from the start of the test to the breakpoint (induction period). An example is given in Annex C.

10.10 The apparatus automatically records the average temperature to the nearest $0,1 \text{ }^\circ\text{C}$ as the temperature of the test.

NOTE The apparatus will automatically switch on the fan to cool the reaction vessel to approximately room temperature. When the apparatus has cooled, the pressure is slowly released, at a rate not exceeding 345 kPa/min.

10.11 When the pressure release has finished, open the apparatus and clean according to Clause 9.

11 Expression of results

Report the induction period expressed in min., rounded to the nearest 0,01 min.

12 Precision

12.1 General

The precision statements were developed in Round Robin tests performed in 2008 and 2009, in accordance with EN ISO 4259 [1]. This method was tested for FAME, B0, B5, B7, B10, and B30 samples, the induction period measured being between 22 min and 215 min.

12.2 Repeatability, r

The difference between two test results, obtained by the same operator with the same apparatus under constant operating conditions on identical test material, would in the long run, in the normal and correct operation of the test method, exceed the values calculated from the following equation only in one case in twenty:

$$r = 0,0288X + 0,4965 \quad (\text{standards.iteh.ai}) \quad (1)$$

where

X represents the mean of the two results expressed in min, rounded to the nearest 0,01 min.

12.3 Reproducibility, R

The difference between two single and independent test results, obtained by different operators working in different laboratories on identical test material, would in the long term, in the normal and correct operation of the test method, exceed the values calculated from the following equation only in one case in twenty:

$$R = 0,0863X + 1,3772 \quad (2)$$

where

X represents the mean of the two results expressed in min, rounded to the nearest 0,01 min.

13 Test report

The test report shall contain at least the following information:

- a reference to this European Standard;
- the type and complete identification of the product tested;
- the result of the test (see 11);
- any deviation, by agreement or otherwise, from the procedure specified;
- the date of the test.