



INTERNATIONAL STANDARD ISO 407:1991 TECHNICAL CORRIGENDUM 1

Published 1999-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Small medical gas cylinders — Pin-index yoke-type valve connections

TECHNICAL CORRIGENDUM 1

Petites bouteilles à gaz médicaux — Raccords de robinets du type à étrier avec ergots de sécurité

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 407:1991 was prepared by Technical Committee ISO/TC 58, Gas cylinders, Subcommittee SC 2, Cylinder fittings.

ISO 407:1991/Cor 1:1999

<https://standards.iteh.ai/catalog/standards/sist/0d211d38-49d2-44ad-ac9f-e50bb401750e/iso-407-1991-cor-1-1999>

Page 7

Figure 5, dimension q , should read $24^{+0,5}_0$ (instead of 24 max.).

Page 8

Figure 6, dimension q_1 , should read $24^{+0,5}_0$ (instead of 24 max.).