

SLOVENSKI STANDARD SIST ISO 18287:2019

01-oktober-2019

Kakovost tal - Določevanje policikličnih aromatskih ogljikovodikov (PAH) - Plinska kromatografija z masno selektivnim detektorjem (GC-MS)

Soil quality - Determination of polycyclic aromatic hydrocarbons (PAH) - Gas chromatographic method with mass spectrometric detection (GC-MS)

iTeh STANDARD PREVIEW

Qualité du sol - Dosage des hydrocarbures aromatiques polycycliques (HAP) - Méthode par chromatographie en phase gazeuse avec détection par spectrométrie de masse (CG-SM)

SIST ISO 18287:2019

https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-

Ta slovenski standard je istoveten z. 16664/s SO-18287:2006

ICS:

13.080.10 Kemijske značilnosti tal Chemical characteristics of

soils

71.040.50 Fizikalnokemijske analitske Physicochemical methods of

metode analysis

SIST ISO 18287:2019 en

SIST ISO 18287:2019

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 18287:2019

https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-7763ba1f6fe4/sist-iso-18287-2019

SIST ISO 18287:2019

INTERNATIONAL STANDARD

ISO 18287

First edition 2006-01-15

Soil quality — Determination of polycyclic aromatic hydrocarbons (PAH) — Gas chromatographic method with mass spectrometric detection (GC-MS)

Qualité du sol — Dosage des hydrocarbures aromatiques polycycliques

Teh ST (HAP) — Méthode par chromatographie en phase gazeuse avec
détection par spectrométrie de masse (CG-SM)

(standards.iteh.ai)

SIST ISO 18287:2019

https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-7763ba1f6fe4/sist-iso-18287-2019



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 18287:2019</u> https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-7763ba1f6fe4/sist-iso-18287-2019

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Cont	tents	Page		
Forewo	ord	iv		
Introdu	uction	v		
1	Scope	1		
2	Normative references	1		
3	Principle	2		
4	Reagents	2		
5	Apparatus	4		
6	Sampling, preservation and pretreatment	5		
7 7.1 7.2 7.3	Procedure	5 6 6		
8	Evaluation			
9	Performance characteristics ANDARD PREVIEW	9		
10	Quality assurance (standards.itch.ai)	9		
11	Test report	9		
Annex	Annex A (informative) Examples of typical GC-MS chromatograms and instrument conditions10			
Annex	Annex B (informative) Results of interlaboratory comparisons13			
Bibliog	Bibliography17			

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18287 was prepared by Technical Committee ISO/TC 190, Soil quality, Subcommittee SC 3, Chemical methods and soil characteristics.

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST ISO 18287:2019</u> https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-7763ba1f6fe4/sist-iso-18287-2019

Introduction

This International Standard is principally based on the extraction method described in ISO 13877. It is modified for the use of gas-chromatography with mass spectrometric detection and is applicable for different PAH pollution levels of soils.

Two alternative extraction methods, A and B, are described in this International Standard.

Method A (two-step method): Extraction of the field-moist soil sample with acetone and petroleum ether, followed by the removal of acetone by washing the extract with water as prescribed in ISO 13877.

Method B (one-step method or on-line method): Extraction of the field-moist soil sample with a mixture of acetone, petroleum ether and water in the presence of sodium chloride. This method is preferred for soil samples with a high content of organic matrix.

Experience has shown that these two methods are applicable with comparable results to less as well as highly polluted soils

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 18287:2019 https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-7763ba1f6fe4/sist-iso-18287-2019 **SIST ISO 18287:2019**

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST ISO 18287:2019

https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-7763ba1f6fe4/sist-iso-18287-2019

Soil quality — Determination of polycyclic aromatic hydrocarbons (PAH) — Gas chromatographic method with mass spectrometric detection (GC-MS)

SAFETY PRECAUTIONS — Certain PAH are highly carcinogenic and must be handled with extreme care. Contact of solid materials, solvent extracts and solutions of standard PAH with the body must not be allowed to occur. PAH may co-distil with solvent and become deposited outside of stoppered bottles; all containers containing solutions of PAH in solvent must therefore always be handled using gloves which are solvent resistant and preferably disposable. PAH contamination of vessels may detected by irradiation with 366 nm UV light. Vessels containing PAH solutions should be stored standing in beakers to contain any spillage in the case of breakage.

Solid PAH are most dangerous and give rise to a dust hazard due to their crystals becoming electrostatically charged. These materials must only be handled where proper facilities are available (e.g. adequate fume hoods, protective clothing, dust masks). It is strongly advised that standard solutions be prepared centrally in suitably equipped laboratories or are purchased from suppliers specialized in their preparation.

Solvent solutions containing PAH must be disposed of in a manner approved for the disposal of toxic wastes.

(standards.iteh.ai)

1 Scope

<u>SIST ISO 18287:2019</u>

https://standards.iteh.ai/catalog/standards/sist/dbb7c859-351c-48dd-a074-

This International Standard specifies the quantitative determination of 16 polycyclic aromatic hydrocarbons (PAH) according to the priority list of the Environmental Protection Agency, USA (EPA, 1982). This International Standard is applicable to all types of soil (field-moist or chemically dried samples), covering a wide range of PAH contamination levels.

Under the conditions specified in this International Standard, a lower limit of application of 0,01 mg/kg (expressed as dry matter) can be ensured for each individual PAH.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10381-1, Soil quality — Sampling — Part 1: Guidance on the design of sampling programmes

ISO 11465, Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method

ISO 14507:2003, Soil quality — Pretreatment of samples for determination of organic contaminants

ISO 10381-8, Soil quality — Sampling — Part 8: Guidance on sampling of stockpiles

ISO 16720, Soil quality —Pretreatment of samples by freeze-drying for subsequent analysis

3 Principle

The extraction is carried out using acetone and petroleum ether. Acetone is an efficient extractant, in particular because it is able to break down soil aggregates. Petroleum ether increases the efficiency of the extraction and is necessary as solvent in the subsequent concentration procedure. An extraction method for PAH should at least use 50 ml of acetone and 50 ml of petroleum ether (for a chemically dried sample). If a wet sample is used, the amount of acetone should be increased to at least 100 ml. Different ways of extraction can lead to the same results provided that the above recommendations are observed. If the sample contains a large amount of water, or if water has been added, sodium chloride (NaCl) should be added to improve the efficiency of the extraction.

Two alternative extraction methods, A and B, are described in this International Standard.

Method A (two-step method): A field moist soil sample is extracted two times with acetone, then petroleum ether is added to the acetone extract. The extract is washed two times with water. The organic layer is dried with anhydrous sodium sulfate.

Method B (one-step method or on-line method): A field-moist soil sample is extracted with a fixed ratio of a mixture of acetone, petroleum ether, water and sodium chloride. An aliquot of the organic layer is dried with anhydrous sodium sulfate.

If necessary, a clean-up step using adsorption chromatography on silica gel, as well as a concentration step, may be included.

The extract is then analysed by capillary gas-chromatography. The identification and the quantification of the PAH is made with mass spectrometric detection, using appropriate deuterated PAH as internal standards.

(standards.iteh.ai)

4 Reagents

SIST ISO 18287:2019

All reagents used shall be of recognized analytical grade and free from PAH: Asblank determination shall be carried out to ensure that the reagents do not contain PAH: detectable concentrations.

- **4.1 Water**, use only distilled water or water of equivalent purity.
- **4.2 Acetone**, for residue analysis.
- **4.3** Petroleum ether, for residue analysis (boiling range 40 °C to 60 °C).
- **4.4 Cyclohexane**, for residue analysis.
- **4.5 Isooctane**, for residue analysis.
- 4.6 Sodium chloride, anhydrous.
- **4.7 Sodium sulfate**, anhydrous, for residue analysis.
- 4.8 Magnesium perchlorate or suitable drying agent.
- **4.9** Silica gel **60**, for column chromatography, particle size 63 μm to 200 μm.
- **4.10** Silica gel 60, water content: $w(H_2O) = 10 \%$ (mass fraction).

Use silica gel 60 (4.9), heated for at least 3 h at 450 °C, cooled down in a desiccator (5.13) and stored with magnesium perchlorate or a suitable drying agent (4.8). Before use, heat for at least 5 h at 130 °C in a drying oven (5.14). Then allow to cool in a desiccator and add 10 % water (mass fraction) in a flask. Shake intensively by hand for 5 min until all lumps have disappeared and then for 2 h in a shaking machine (5.3).

Store the desactivated silica gel in the absence of air and use it for a maximum of one week.

4.11 Reference substances, internal standards

Choose as internal standards substances whose physical and chemical properties (such as extraction behaviour, retention time) are similar to those of the compounds to be analysed. Deuterated PAH should be used as internal standards for the GC-MS method for the evaluation of results. Verify the stability of the internal standards regularly. Table 1 lists native and deuterated PAH.¹⁾

Table 1 — Native PAH and deuterated PAH

PAH reference substances	CAS No.	Deuterated PAH internal standard substances
Naphthalene	91-20-3	Naphthalene-d ₈
Acenaphthene	83-32-9	Acenaphthene-d ₁₀
Acenaphthylene	208-96-8	
Fluorene	86-73-7	
Anthracene	120-12-7	
Phenanthrene	85-01-8	Phenanthrene-d ₈
Fluoranthene	206-44-0	
Pyrene iTeh ST	AND A129-00-0 PREV	EW
Benz[a]anthracene	tandarűs iteh ai)	Benz[a]anthracene-d ₁₂
Chrysene	218-01-9	
Benzo[b]fluoranthene	<u>SIST ISO 205-95-2019</u>	Benzo[e]pyrene-d ₁₂
Benzo[k]fluoranthene	763ba1f6fe4/si20750959287-2019	- 10dd-d071
Benzo[a]pyrene	50-32-8	
Indeno[1,2,3-cd]pyrene	193-39-5	Perylene-d ₁₂
Dibenz[ah]anthracene	53-70-3	
Benzo[ghi]perylene	191-24-2	

4.12 Standard solutions for GC-MS

If a commercially available certified standard stock solution of the relevant PAH is used, the calibration solutions with different levels of PAH concentrations are prepared by diluting an appropriate volume of stock solution with cyclohexane (4.4) in volumetric flasks (5.8).

4.12.1 Single-substance stock solution

If no certified standard solutions are used, prepare single-substance stock solutions by weighing approximately 10 mg of each of the standards "native PAH" and "deuterated PAH" (see Table 1) to the

¹⁾ Certified solutions of PAH and single solid PAH substances with certified purity are available from a limited number of suppliers, e.g. Institute for Reference Materials and Measurements (IRMM), B-2440 Geel, Belgium. National Institute of Science and Technology, Office of Standard Ref. Data, Washington D.C. 20234 USA, or from other commercial providers. This information is given for the convenience of users of this International Standard and does not constitute a endorsement by ISO of these suppliers.