



SLOVENSKI STANDARD SIST EN ISO 5814:2013

01-januar-2013

Nadomešča:
SIST EN 25814:1996

Kakovost vode - Določevanje raztopljenega kisika - Elektrokemijska metoda (ISO 5814:2012)

Water quality - Determination of dissolved oxygen - Electrochemical probe method (ISO 5814:2012)

Wasserbeschaffenheit - Bestimmung des gelösten Sauerstoffs - Elektrochemisches Verfahren (ISO 5814:2012)

Qualité de l'eau - Dosage de l'oxygène dissous - Méthode électrochimique à la sonde (ISO 5814:2012)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

Ta slovenski standard je istoveten z: EN ISO 5814:2012

ICS:

13.060.50	Preiskava vode na kemične snovi	Examination of water for chemical substances
-----------	---------------------------------	--

SIST EN ISO 5814:2013

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 5814:2013](#)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 5814

October 2012

ICS 13.060.50

Supersedes EN 25814:1992

English Version

Water quality - Determination of dissolved oxygen - Electrochemical probe method (ISO 5814:2012)

Qualité de l'eau - Dosage de l'oxygène dissous - Méthode
électrochimique à la sonde (ISO 5814:2012)

Wasserbeschaffenheit - Bestimmung des gelösten
Sauerstoffs - Elektrochemisches Verfahren (ISO
5814:2012)

This European Standard was approved by CEN on 4 August 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....3

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[SIST EN ISO 5814:2013](https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

Foreword

This document (EN ISO 5814:2012) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2013, and conflicting national standards shall be withdrawn at the latest by April 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 25814:1992.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW
(standards.iteh.ai)
Endorsement notice

The text of ISO 5814:2012 has been approved by CEN as a EN ISO 5814:2012 without any modification.

[SIST EN ISO 5814:2013](https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 5814:2013](#)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

INTERNATIONAL STANDARD

**ISO
5814**

Third edition
2012-10-15

Water quality — Determination of dissolved oxygen — Electrochemical probe method

*Qualité de l'eau — Dosage de l'oxygène dissous — Méthode
électrochimique à la sonde*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 5814:2013](https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>



Reference number
ISO 5814:2012(E)

© ISO 2012

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 5814:2013](https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents	Page
Foreword	iv
1 Scope	1
2 Normative references	1
3 Principle	1
4 Interferences	2
5 Reagents	2
6 Apparatus	2
7 Sampling and analysis procedure	3
7.1 Sampling	3
7.2 Measuring technique and precautions to be taken	3
7.3 Calibration	4
7.4 Determination	4
8 Calculation and expression of results	5
8.1 Dissolved oxygen concentration	5
8.2 Dissolved oxygen expressed as percentage saturation	5
9 Test report	5
Annex A (informative) Physicochemical data for oxygen in water	6
Annex B (informative) Performance data	12
Bibliography	14

[SIST EN ISO 5814:2013](https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013)

<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

ISO 5814:2012(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 5814 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

This third edition cancels and replaces the second edition (ISO 5814:1990), which has been technically revised.

The main changes compared to the second edition are:

- a) a calibration procedure using water-saturated air is specified;
- b) the calibration procedure using air-saturated water is omitted.

ITIH STANDARD PREVIEW
(standards.iteh.ai)
SIST EN ISO 5814:2013
<https://standards.iteh.ai/catalog/standards/sist/48e9ff8b-26d4-48e0-b256-5d7de64d4d73/sist-en-iso-5814-2013>

Water quality — Determination of dissolved oxygen — Electrochemical probe method

WARNING — Persons using this International Standard should be familiar with normal laboratory practice. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this International Standard be carried out by suitably trained staff.

1 Scope

This International Standard specifies an electrochemical method for the determination of dissolved oxygen in water by means of an electrochemical cell which is isolated from the sample by a gas permeable membrane.

Measurement can be made either as a concentration of oxygen in milligrams per litre, percentage saturation (% dissolved oxygen) or both. The method measures oxygen in water corresponding to 1 % to 100 % saturation. However, most instruments permit measurement of values higher than 100 %, i.e. supersaturation.

NOTE Supersaturation is possible when the partial pressure of oxygen is higher than in air. Especially when strong algal growth is present, supersaturation of up to 200 % and above can occur.

The method measures oxygen in water with a saturation higher than 100 %, when special arrangements to prevent the outgassing of oxygen during the handling and measurement of the sample are made.

The method is suitable for measurements made in the field and for continuous monitoring of dissolved oxygen, as well as measurements made in the laboratory. It is the preferred method for highly coloured and turbid waters, and also for analysis of waters not suitable for the Winkler titration method because of iron- and iodine-fixing substances, which can interfere in the iodometric method specified in ISO 5813^[1].

The method is suitable for drinking waters, natural waters, waste waters, and saline waters. If used for saline waters, such as sea or estuarine waters, a correction for salinity is essential.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Principle

Immersion of a probe, consisting of a cell enclosed by a selective membrane and containing the electrolyte and at least two metallic electrodes, in the water to be analysed.

NOTE The membrane is effectively impermeable to water and ionic dissolved matter, but is permeable to oxygen and a certain number of other gases.

One of the electrodes is made of a noble metal like gold or platinum. Oxygen is reduced at its surface by an electrochemical process. In order to make this process possible, a suitable electrochemical potential is established at this electrode. For polarographic probes, this is achieved by applying an external voltage related to a second electrode. Galvanic probes are able to build up the potential by themselves.