



SLOVENSKI STANDARD SIST ISO 7981-1:2007

01-februar-2007

Water quality -- Determination of polycyclic aromatic hydrocarbons (PAH) -- Part 1:
Determination of six PAH by high-performance thin-layer chromatography with
fluorescence detection after liquid-liquid extraction

STANDARD PREVIEW
(standards.iteh.ai)

Qualité de l'eau -- Détermination des hydrocarbures aromatiques polycycliques (HAP) --
Partie 1: Dosage de six HAP par chromatographie de haute performance sur couche
mince avec détection fluorimétrique à la suite d'une extraction liquide-liquide

Ta slovenski standard je istoveten z: ISO 7981-1:2005

ICS:

13.060.50 Examination of water for
chemical substances

SIST ISO 7981-1:2007 en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 7981-1:2007

<https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007>

INTERNATIONAL
STANDARDISO
7981-1First edition
2005-06-15

**Water quality — Determination of
polycyclic aromatic hydrocarbons
(PAH) —**

Part 1:

**Determination of six PAH by high-
performance thin-layer chromatography
with fluorescence detection after
liquid-liquid extraction**

*Qualité de l'eau — Détermination des hydrocarbures aromatiques
polycycliques (HAP) —*

<https://standards.iteh.org/standards/ISO/7981-1>

751b5ae05f08/sist-iso-7981-1-2007

*Partie 1. Dosage de six HAP par chromatographie de haute
performance sur couche mince avec détection fluorimétrique à la suite
d'une extraction liquide-liquide*

Reference number
ISO 7981-1:2005(E)

© ISO 2005

ISO 7981-1:2005(E)**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 7981-1:2007](https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007)

<https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007>

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction	v
1 Scope	1
2 Principle	1
3 Interferences	2
4 Reagents	3
5 Apparatus	4
6 Sampling	5
7 Procedure	6
8 Evaluation	8
9 Calibration	11
10 Determination of the recovery	13
11 Blank measurements	13
12 Calculation	13
13 Expression of results	14
14 Test report	15
15 Accuracy	15
Annex A (informative) R_f values of the six PAH under various chromatographic conditions	16
Annex B (informative) Spectroscopic identification	17
Annex C (informative) Examples for the construction of special apparatus	18
Annex D (informative) Accuracy	21
Bibliography	22

ISO 7981-1:2005(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 7981-1 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

ISO 7981 consists of the following parts, under the general title *Water quality — Determination of polycyclic aromatic hydrocarbons (PAH)*:

- *Part 1: Determination of six PAH by high-performance thin-layer chromatography with fluorescence detection after liquid-liquid extraction*
[SIST ISO 7981-1:2007](https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007)
<https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007>
- *Part 2: Determination of six PAH by high-performance liquid chromatography with fluorescence detection after liquid-liquid extraction*

Introduction

Polycyclic aromatic hydrocarbons (PAH) are present in nearly all types of waters. These substances are adsorbed on solids (sediments, suspended matter) as well as dissolved in the liquid phase.

Some PAH are known or suspected to cause cancer. The maximum acceptable levels of PAH in waters intended for human consumption are given in European Legislation [1] [2] [3] [4].

The sum of the mass concentrations of the six PAH specified in this part of ISO 7981 normally is about 0,01 µg/l to 0,05 µg/l in ground water and up to 1 µg/l in surface water.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 7981-1:2007](https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007)

<https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 7981-1:2007

<https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007>

Water quality — Determination of polycyclic aromatic hydrocarbons (PAH) —

Part 1: Determination of six PAH by high-performance thin-layer chromatography with fluorescence detection after liquid-liquid extraction

WARNING — Some substances being measured are presumed to be carcinogenic. Acetonitrile and hexane are harmful.

Persons using this part of ISO 7981 should be familiar with normal laboratory practise. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user of this part of ISO 7981 to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this part of ISO 7981 be carried out by suitably trained staff.

1 Scope

SIST ISO 7981-1:2007

[https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-](https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007)

[751b5ae05f08/sist-iso-7981-1-2007](https://standards.iteh.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ae05f08/sist-iso-7981-1-2007)

This part of ISO 7981 specifies the determination of six selected PAH in drinking water by high-performance thin-layer chromatography with fluorescence detection after liquid-liquid extraction. The six PAH are: fluoranthene, benzo[*b*]fluoranthene, benzo[*a*]pyrene, benzo[*k*]fluoranthene, indeno[1,2,3-*cd*]pyrene, and benzo[*ghi*]perylene (see Table 1).

A screening method (method A) is described to exclude those samples containing less than 20 % of the limit values given in References [1], [2], [3] and [4].

A quantitative method (method B) is also described, with a working range of 40 ng/l to 240 ng/l (sum of 6 PAH). Higher concentrations can be determined by using a smaller aliquot of the sample.

With some modifications, this method is also applicable for the analysis of ground waters and moderately polluted surface waters.

2 Principle

Since PAH can to a large extent be adsorbed on particulate matter, the whole test sample is analysed.

NOTE For the analysis of surface water, a differentiation between dissolved and undissolved PAH may be desirable, but this is not relevant for drinking water.

PAH are extracted from the water sample by liquid-liquid extraction. The extract is evaporated to dryness and the residue is taken up in a solvent and analysed.

Extracts of surface waters and other contaminated water samples should be cleaned prior to analysis (7.4).

ISO 7981-1:2005(E)

PAH are then separated by high-performance thin-layer chromatography (HPTLC) on appropriate stationary phases and detected either visually or by *in situ* fluorescence measurement at constant or differing wavelength combinations.

Table 1 — Polycyclic aromatic hydrocarbons determinable by this method

Name	Chemical formula	Molar mass g/mol	Carbon fraction	CAS-number	Structure
Fluoranthene	C ₁₆ H ₁₀	202,26	95,0	206-44-0	
Benzo[b]fluoranthene	C ₂₀ H ₁₂	252,32	95,2	205-99-2	
Benzo[a]pyrene	C ₂₀ H ₁₂	252,32	95,2	50-32-8	
Benzo[k]fluoranthene	C ₂₀ H ₁₂	252,32	95,2	207-08-9	
Indeno[1,2,3-cd]pyrene	C ₂₂ H ₁₂	276,34	95,6	193-39-5	
Benzo[ghi]perylene	C ₂₂ H ₁₂	276,34	95,6	191-24-2	

3 Interferences

3.1 Interferences with screening method (method A)

Other compounds and/or impurities can interfere in the screening method, thus the use of spectroscopic identification is recommended in order to lower the occurrence of false positives in screening test samples.

It may be necessary to clean coloured extracts or test samples known to contain other organic substances on silica prior to analysis (7.4).

3.2 Interferences with sampling and extraction

Use sampling containers made of materials (preferably of glass or steel) that do not affect the test sample during the contact time. Avoid plastics and other organic materials during sampling, sample storage or extraction.

If automatic samplers are used, avoid the use of silicone or rubber material for the tubes. If present, make sure that the tubes are as short as possible. Rinse the sampling line with the water to be sampled before the test sample is taken. ISO 5667-2 and ISO 5667-3 can be used for guidance.

Keep the samples from direct sunlight and prolonged exposure to light.

During storage of the test sample, losses of PAH can occur due to adsorption on the walls of the containers. The extent of the losses depends on the storage time.

3.3 Interferences with HPTLC

Substances that exhibit either fluorescence or quenching and co-elute with the PAH to be determined can interfere with the determination. These interferences can lead to incompletely resolved signals and can, depending on their magnitude, affect the accuracy and precision of the analytical results. Band overlap will make an interpretation of the result impossible. Unsymmetrical bands and bands broader than the corresponding bands of the reference substance suggest interferences.

The identity and purity of the bands can be checked by recording the excitation and emission spectra.

4 Reagents

Use only reagents of recognized analytical grade (e.g. "for residue analysis" or "for HPLC analysis") as far as available, and only distilled water or water of equivalent purity showing the lowest possible fluorescence.

Monitor the blank to guarantee that the reagents do not contain PAH in detectable concentrations (see Clause 11).

STANDARD PREVIEW
(standards.itech.ai)

4.1 Solvents

SIST ISO 7981-1:2007

4.1.1 Extraction and clean-up solvents

<https://standards.itech.ai/catalog/standards/sist/391d9124-4f6c-41c3-981b-751b5ac05f08/sist-iso-7981-1-2007>

4.1.1.1 Cyclohexane, C₆H₁₂

4.1.1.2 Hexane, C₆H₁₄

4.1.1.3 Dichloromethane, CH₂Cl₂

Other volatile solvents may be used as well, if it is proved that the recovery is equivalent or better.

NOTE Dichloromethane often contains stabilizers, e.g. ethanol or amylene. Stabilizers can influence the elution strength of the eluent. Without stabilizer, free radicals might develop. This can lead to degradation of PAH. The presence of hydrogen chloride indicates the presence of radicals. Hydrogen chloride can be determined by extracting dichloromethane with water and measuring the pH value.

4.1.2 HPTLC solvents

4.1.2.1 Methanol, CH₃OH

4.1.2.2 Acetonitrile, CH₃CN

4.1.2.3 2,4-dimethyl-3-oxypentane (diisopropyl ether), C₆H₁₄O

4.1.2.4 2-propanol (isopropanol), C₃H₇OH

4.2 Sodium thiosulfate pentahydrate, Na₂S₂O₃·5H₂O

4.3 Sodium chloride, NaCl

ISO 7981-1:2005(E)

4.4 Sodium sulfate, Na₂SO₄, anhydrous, precleaned by heating to 500 °C.

4.5 Nitrogen, having a purity (volume fraction) of at least 99,999 %.

4.6 Silica, with an average particle size of approximately 40 µm and stored in a desiccator to ensure maximum activity.

NOTE Prepacked silica cartridges are commercially available.

4.7 Molecular sieve beads, pore size 0,4 nm.

4.8 Caffeine, C₈H₁₀N₄O₂

4.9 Liquid paraffin

4.10 Reference substances (see Table 1)

Because of the dangerous nature of the substances to be used, it is highly recommended to use commercially available, preferably certified, standard solutions. Avoid skin contact.

4.11 Single-substance stock solutions, of those listed in Table 1, diluted in cyclohexane (4.1.1.1) or methanol (4.1.2.1) to a mass concentration of, for example, 10 µg/ml.

4.12 Multiple-substance stock solution, preferably certified, diluted in cyclohexane (4.1.1.1) or methanol (4.1.2.1) to a mass concentration of, for example, 10 µg/ml for fluoranthene and 2 µg/ml for the other reference substances (4.10).

4.13 Calibration solutions

Transfer 20 µl, 40 µl, 60 µl, 80 µl, 100 µl and 120 µl of the stock solution (4.12) into a graduated 10 ml flask (5.14) and make up to volume with cyclohexane (4.1.1.1) or methanol (4.1.2.1).

1 ml of this reference solution contains 20 ng, 40 ng, 60 ng, 80 ng, 100 ng and 120 ng of fluoranthene and 4 ng, 8 ng, 12 ng, 16 ng, 20 ng and 24 ng of the other reference substances (4.10).

NOTE The solutions 4.11 to 4.13 are stable for at least one year when stored in the dark at room temperature and protected from evaporation.

5 Apparatus

Use standard laboratory apparatus, cleaned to eliminate all interferences.

Clean all glassware, for example by rinsing with detergent and hot water, and dry for about 15 min to 30 min at about 120 °C. After cooling, rinse with acetone, seal the glassware and store in a clean environment.

Glassware that has been in contact with waste water samples or samples with high PAH concentrations shall not be re-used for drinking water analysis.

5.1 Brown glass bottles, narrow-necked, flat-bottomed, nominal capacity 1 000 ml, with solid glass stopper.

5.2 Magnetic stirrer with stirring rods, PTFE-coated, kept under cyclohexane, with a maximum rotational frequency of 1 000 min⁻¹.

5.3 Measuring cylinders, nominal capacities 10 ml, 25 ml and 1 000 ml.

5.4 Separating funnel, nominal capacity 1 000 ml, with PTFE stopcock, kept under cyclohexane, and glass stopper, e.g. a Squibb funnel.