



**SLOVENSKI STANDARD**  
**SIST EN 1060-3:2000+A2:2010**  
**01-marec-2010**

---

**Neinvazivni sfigmomanometri - 3. del: Dodatne zahteve za avtomatizirane sisteme za merjenje krvnega tlaka**

Non-invasive sphygmomanometers - Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems

Nichtinvasive Blutdruckmessgeräte - Teil 3: Ergänzende Anforderungen für elektromechanische Blutdruckmesssysteme

Tensiomètres non invasifs - Partie 3: Exigences complémentaires concernant les systèmes électromécaniques de mesure de la pression sanguine

<https://standards.iteh.ai/catalog/standards/sist/014f421d-bf1e-4c35-b63f-ec44a11ab831/sist-en-1060-3-2000a2-2010>

**Ta slovenski standard je istoveten z: EN 1060-3:1997+A2:2009**

---

**ICS:**

11.040.55      Diagnostična oprema      Diagnostic equipment

**SIST EN 1060-3:2000+A2:2010**      **en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 1060-3:2000+A2:2010

<https://standards.iteh.ai/catalog/standards/sist/014f421d-bf1e-4c35-b63f-ec44a11ab831/sist-en-1060-3-2000a2-2010>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN 1060-3:1997+A2**

November 2009

ICS 11.040.55

Supersedes EN 1060-3:1997

English Version

## Non-invasive sphygmomanometers - Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems

Tensiomètres non invasifs - Partie 3: Exigences complémentaires concernant les systèmes électromécaniques de mesure de la pression sanguine

Nichtinvasive Blutdruckmessgeräte - Teil 3: Ergänzende Anforderungen für elektromechanische Blutdruckmesssysteme

This European Standard was approved by CEN on 27 January 1997 and includes Amendment 1 approved by CEN on 24 November 2005 and Amendment 2 approved by CEN on 17 October 2009.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

## Contents

	Page
Foreword.....	3
1 Scope .....	4
2 Normative references .....	4
3 Definitions .....	4
4 Cuff.....	5
5 Display .....	6
6 Units .....	6
7 Requirements .....	6
8 Test methods.....	9
9 Information supplied by the manufacturer .....	19
Annex A (informative) Bibliography .....	22
Annex ZA (informative) <b>Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices</b> .....	23

(standards.iteh.ai)

[SIST EN 1060-3:2000+A2:2010](https://standards.iteh.ai/catalog/standards/sist/014f421d-bf1e-4c35-b63f-ec44a11ab831/sist-en-1060-3-2000a2-2010)

<https://standards.iteh.ai/catalog/standards/sist/014f421d-bf1e-4c35-b63f-ec44a11ab831/sist-en-1060-3-2000a2-2010>

## Foreword

This document (EN 1060-3:1997+A2:2009) has been prepared by Technical Committee CEN/TC 205 "Non-active medical devices", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2010, and conflicting national standards shall be withdrawn at the latest by May 2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document includes Amendment 1, approved by CEN on 2005-11-24 and Amendment 2, approved by CEN on 2009-10-17.

This document supersedes EN 1060-3:1997.

The start and finish of text introduced or altered by amendment is indicated in the text by tags **A1** **A1** and **A2** **A2**.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

This European Standard 'Non-invasive sphygmomanometers' consists of the following parts:

Part 1: General requirements

Part 2: Supplementary requirements for mechanical sphygmomanometers

Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive 93/42/EEC.

**A2** Annexes A and ZA are given for information and do not form normative parts of this European Standard. **A2**

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



**3.1****auscultatory method**

technique whereby sounds (known as Korotkov sounds) are heard over an occluded artery as the occluding pressure is slowly released, the appearance of sounds coinciding with systolic and the disappearance of sounds with diastolic blood pressure

**3.2****electro-mechanical blood pressure measuring system**

system that consists of:

- a) at least one cuff, which is connected to the pneumatic system;
- b) at least one electro-mechanical transducer to measure cuff pressure;
- c) at least one measured value display;
- d) if needed, signal inputs and outputs.

**3.3****electro-mechanical pressure transducer**

component that transforms pressure signals into electrical signals

**3.4****oscillometric method**

method wherein a cuff is placed on the limb and the pressure in the cuff is increased until the blood flow in the artery is interrupted and then the pressure in the cuff is slowly reduced

NOTE During the inflation and deflation of the cuff small pressure changes (oscillations) occur in the cuff as a result of the arterial blood pressure pulses.

These oscillations, which first increase and then decrease, are detected and stored together with the corresponding cuff pressure values in the measurement system. With these stored values the systolic, diastolic and mean arterial blood pressure values can be mathematically derived using an appropriate algorithm. It is possible to carry out the measurement during the inflation phase

**3.5****zero setting**

procedure that corrects a deviation of the pressure reading to 0 mmHg at atmospheric pressure (gauge pressure: 0 mmHg)

**3.6****patient simulator**

device for simulating the oscillometric cuff pulses and/or auscultatory sounds during inflation and deflation

NOTE This device is not used for testing accuracy but is required in assessing stability of performance.

**A1**

**3.7****short term automatic mode**

mode in which as many automatic measurements as possible are made within a specified time period **A1**

**4 Cuff**

Clause 4 of EN 1060-1:1995 shall apply.

**EN 1060-3:1997+A2:2009 (E)****5 Display**

Clause 5 of EN 1060-1:1995 shall apply.

**6 Units**

Clause 6 of EN 1060-1:1995 shall apply.

**7 Requirements****7.1 General**

Equipment or parts thereof, using materials or having forms of construction different from those detailed in this Part of EN 1060, shall be accepted if it can be demonstrated that an equivalent degree of safety and performance is obtained.

**7.2 Limits of the error of the cuff pressure indication**

7.1.1 of EN 1060-1:1995 shall apply.

**7.3 Effect of voltage variations of the power source variations****7.3.1 Internal electrical power source**

- iTeh STANDARD PREVIEW**  
(standards.iteh.ai)
- a) **A1** Automated, cyclical measuring non-invasive blood pressure systems in which the cuff pressure is generated by an electrical pump shall comply with 56.7 c) 1) and 2) of EN 60601-2-30:2000; **A1**
- b) **A1** Non-cyclical measuring non-invasive blood pressure systems in which the cuff pressure is generated by an electrical pump shall comply with 56.7 c) 2) of EN 60601-2-30:2000; **A1**
- c) Outside of the working range, no cuff pressure reading and no result of the blood pressure measurement shall be displayed.

Testing shall be carried out in accordance with **8.2.1** and **8.3.1**.

**7.3.2 External electrical power source**

- a) **A1** Blood pressure measuring systems in which the cuff pressure is generated by an electrical pump shall comply with 49.3 of EN 60601-2-30:2000; **A1**
- b) **A1** For devices operated in manual mode 49.3 a) only applies. **A1**

Testing shall be carried out in accordance with **8.2.2** and **8.3.2** (alternating current) or **8.2.3** and **8.3.3** (direct current).

- c) Incorrect values resulting from voltage variations outside the limits given in **7.3.2** b) shall not be displayed.

Testing shall be carried out in accordance with **8.2.4** (alternating current) or **8.2.5** (direct current).



## 7.4 Pneumatic system

### 7.4.1 Air leakage

Air leakage shall not exceed a pressure drop of 6 mmHg/min (0,8 kPa/min).

For those devices in which the blood pressure is determined manually with the aid of a stethoscope the air leakage shall not exceed a pressure drop of 4 mmHg/min (0,5 kPa/min). Testing shall be carried out in accordance with 8.4.

### 7.4.2 Pressure reducing system for devices using the auscultatory method

The pressure reducing system for manually operated and automated deflation valves shall be capable of maintaining a deflation rate of 2 mmHg/s to 3 mmHg/s (0,3 kPa/s to 0,4 kPa/s) within the target range of systolic and diastolic blood pressure. For devices which control the pressure reduction as a function of the pulse rate, a deflation rate of between 2 mmHg/pulse and 3 mmHg/pulse (0,3 kPa/pulse and 0,4 kPa/pulse) shall be maintained.

NOTE Manually operated deflation valves should be easily adjustable to these values.

Testing shall be carried out in accordance with 8.5.

### 7.4.3 Rapid exhaust

During the rapid exhaust of the pneumatic system with fully opened valve, the time for the pressure reduction from 260 mmHg to 15 mmHg (34,7 kPa to 2,0 kPa) shall not exceed 10 s.

For blood pressure measuring systems having the capability to measure in a neonatal/infant mode, the time for the pressure reduction from 150 mmHg to 5 mmHg (20,0 kPa to 0,7 kPa) during the rapid exhaust of the pneumatic system with fully opened valve shall not exceed 5 s.

Testing shall be carried out in accordance with 8.6.

### 7.4.4 Zero setting

Blood pressure measuring systems shall be capable of automatic zero setting. The zero setting shall be carried out at appropriate intervals, at least starting after switching on the device. At the moment of the zero setting a gauge pressure of 0 mmHg (0 kPa) shall exist and be displayed thereafter.

Devices performing zero setting only immediately after switching on, shall switch off automatically before the drift of the pressure transducer and the analog signal processing exceeds 1 mmHg (0,1 kPa).

Testing shall be carried out in accordance with 8.7 and 8.8.

## 7.5 Environmental performance

### 7.5.1 Storage

Blood pressure measuring systems shall maintain the requirements specified in this Part of EN 1060 after storage for 24 h at a temperature of - 5 °C and for 24 h at a temperature of + 50 °C and a Relative Humidity of 85 % (non-condensing).

Testing shall be carried out at environmental conditions (see subclause 7.1.1 of EN 1060-1:1995) in accordance with 8.1 of EN 1060-1:1995 after the test sample has been placed for 24 h at a temperature of - 5 °C and immediately afterwards for 24 h at a temperature of + 50 °C in a climatic chamber.

**EN 1060-3:1997+A2:2009 (E)**

NOTE Integrated multiparameter monitors may contain components which can be damaged during storage. The general temperature range in EN 1060-1:1995 has therefore been reduced.

**7.5.2 Temperature, relative humidity**

7.1.2.2 of EN 1060-1:1995 shall apply.

The signal processing for the determination of the blood pressure values shall not be influenced within the range of temperature and Relative Humidity specified in 7.1.2.2 of EN 1060-1:1995.

Testing shall be carried out in accordance with **8.9**.

**7.5.3 Electromagnetic compatibility**

Either:

- a) Electrical and/or electromagnetic interferences shall not lead to degradations in the cuff pressure indication or in the result of the blood pressure measurement, or;
- b) If electrical and/or electromagnetic interferences lead to an abnormality, the abnormality shall be clearly indicated and it shall be possible to restore normal operation within 30 s after cessation of the electromagnetic disturbance.

Testing shall be carried out in accordance with  $\text{A}_2$  EN 60601-1-2  $\text{A}_2$ .

**7.6 Stability of the cuff pressure indication**

The change of the cuff pressure indication shall not be more than 3 mmHg (0,4 kPa) throughout the pressure range after 10 000 stimulated measurement cycles.

Testing shall be carried out in accordance with **8.10**.

**7.7 Pressure indicating device****7.7.1 Nominal range and measuring range**

The nominal range for the cuff pressure measurement shall be specified by the manufacturer. The measuring and indication ranges of the cuff pressure shall be equal to the nominal range. Values of blood pressure measurement results outside the nominal range of cuff pressure shall be clearly indicated as out of range.

Testing shall be carried out by visual inspection.

**7.7.2 Digital indication**

The numerical step shall be 1 mmHg (0,1 kPa).

$\text{A}_2$  Numbers shall be clearly legible in accordance with 7.1.2 of EN 60601-1:2006.  $\text{A}_2$

If the measured value of a parameter is to be indicated on more than one display, all the displays shall indicate the same numerical value.

Measured, numerical values on the display(s), and the symbols defining the units of measurement shall be arranged in such a way so as to avoid misinterpretation.

Testing shall be carried out by visual inspection.

## 7.8 Signal input and output parts

The construction of the signal input and output parts (excluding internal interfaces, e.g. microphone signal input) relevant to the non-invasive blood pressure measurement shall ensure that incorrectly fitted or defective accessories shall not result in erroneous indication of cuff pressure or erroneous indication of blood pressure.

Testing shall be carried out in accordance with 8.11.

## 7.9 Overall system accuracy

<sup>A1</sup> Except for short term automatic mode and devices in which blood pressure is determined manually with the aid of a stethoscope, the following overall system accuracy values shall apply:

- a) maximum mean error of measurement:  $\pm 5$  mmHg ( $\pm 0,7$  kPa);
- b) maximum experimental standard deviation: 8 mmHg (1,1 kPa).

Testing shall be performed in accordance with EN 1060-4.

Upon request the manufacturer shall provide evidence to the Notified Body that these requirements are met. <sup>A1</sup>

## 7.10 Alarms

If alarms are used, they shall be of at least medium priority and shall comply with <sup>A2</sup> EN 60601-1-8 <sup>A2</sup>.

## 7.11 Safety

### 7.11.1 Cuff pressure

It shall be possible to abort any blood pressure measurement at any time by single key operation and this shall lead to a rapid exhaust (see 7.4.3).

Testing shall be carried out in accordance with 8.12.

### 7.11.2 Unauthorised access

All controls that affect accuracy shall be sealed against unauthorised access.

Testing shall be carried out by visual inspection.

### 7.11.3 Tubing connectors

Luer lock connectors shall not be used.

NOTE In order to avoid possible misconnection with intravascular systems Luer slip connectors should not be used.

## 8 Test methods

### 8.1 General

An uncertainty of 1 mmHg (0,1 kPa) for digital indications shall be allowed in any displayed value because the display system cannot indicate a change of less than one unit.