



SLOVENSKI STANDARD

SIST EN 15149-2:2011

01-januar-2011

Nadomešča:

SIST-TS CEN/TS 15149-2:2006

Trdna biogoriva - Določevanje porazdelitve velikosti delcev - 2. del: Vibracijska zaslonska metoda z uporabo sita z odprtinami 3,15 mm in manj

Solid biofuels - Determination of particle size distribution - Part 2: Vibrating screen method using sieve apertures of 3,15 mm and below

Feste Biobrennstoffe - Bestimmung der Partikelgrößenverteilung - Teil 2: Rüttelsiebverfahren mit Sieb-Lochgrößen von 3,15 mm und darunter

Biocombustibles solides - Détermination de la distribution granulométrique - Partie 2: Méthode au tamis vibrant d'ouverture de maille inférieure ou égale à 3,15 mm

Ta slovenski standard je istoveten z: EN 15149-2:2010

ICS:

75.160.10 Trda goriva Solid fuels

SIST EN 15149-2:2011 en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 15149-2:2011

<https://standards.iteh.ai/catalog/standards/sist/0c305a52-73fa-41a3-a4a9-d9d70f6128cf/sist-en-15149-2-2011>

EUROPEAN STANDARD

EN 15149-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2010

ICS 75.160.10

Supersedes CEN/TS 15149-2:2006

English Version

Solid biofuels - Determination of particle size distribution - Part 2: Vibrating screen method using sieve apertures of 3,15 mm and below

Biocombustibles solides - Détermination de la distribution granulométrique - Partie 2: Méthode au tamis vibrant d'ouverture de maille inférieure ou égale à 3,15 mm

Feste Biobrennstoffe - Bestimmung der Partikelgrößenverteilung - Teil 2: Rüttelsiebverfahren mit Sieb-Lochgrößen von 3,15 mm und darunter

This European Standard was approved by CEN on 26 September 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/0c305a52-751a-41a3-a4a9-d9d70f6128cf/sist-en-15149-2-2011>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Principle.....	5
5 Apparatus	5
6 Sample preparation	7
6.1 Sample size	7
6.2 Moisture conditions.....	7
7 Procedure	7
8 Calculation.....	8
9 Performance characteristics	9
10 Test report	9
Annex A (normative) Determination of the median value of a particle size distribution.....	10
A.1 Definition	10
A.2 Procedure (example)	11
A.3 Calculation.....	11
Annex B (informative) Guidance data on performance characteristics	12
Bibliography.....	13

Foreword

This document (EN 15149-2:2010) has been prepared by Technical Committee CEN/TC 335 “Solid biofuels”, the secretariat of which is held by SIS.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2011, and conflicting national standards shall be withdrawn at the latest by April 2011.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 15149-2:2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

EN 15149, *Solid biofuels — Determination of particle size distribution*, consists of the following parts:

- *Part 1: Oscillating screen method using sieve apertures of 1 mm and above*
- *Part 2: Vibrating screen method using sieve apertures of 3,15 mm and below*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

EN 15149-2:2010 (E)**Introduction**

Part 1 describes the reference method for size classification of samples with a nominal top size of 1 mm and above.

Part 2 describes the reference method for size classification of samples with a nominal top size below 3,15 mm.

Manual sieving is not included in this standard, as no data is available which support that manual sieving operations are comparable to the here described mechanical sieving operations.

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

SIST EN 15149-2:2011

<https://standards.iteh.ai/catalog/standards/sist/0c305a52-73fa-41a3-a4a9-d9d70f6128cf/sist-en-15149-2-2011>

1 Scope

This European Standard specifies a method for the determination of the size distribution of particulate biofuels by the vibrating screen method. The method described is meant for particulate biofuels only, namely materials that either have been reduced in size, such as most wood fuels, or are physically in a particulate form. This document applies to particulate uncompressed fuels with a nominal top size of 3,5 mm and below (e.g. sawdust).

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 14588:2010, *Solid biofuels — Terminology, definitions and descriptions*

EN 14774-1:2009, *Solid biofuels — Determination of moisture content — Oven dry method — Part 1: Total moisture — Reference method*

EN 14774-2:2009, *Solid biofuels — Determination of moisture content — Oven dry method — Part 2: Total moisture — Simplified method*

prEN 14778, *Solid biofuels — Sampling*

prEN 14780, *Solid biofuels — Sample preparation*

EN 14961-1, *Solid biofuels — Fuel specifications and classes — Part 1: General requirements*

ISO 3310-1, *Test sieves — Technical requirements and testing — Part 1: Test sieves of metal wire cloth*

ISO 3310-2, *Test sieves — Technical requirements and testing — Part 2: Test sieves of perforated metal plate*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in EN 14588:2010 apply.

4 Principle

A sample is subjected to sieving through vibrating sieves, sorting the particles in decreasing size classes by mechanical means.

NOTE A manual sieving is excluded due to the fact that small sieve holes can easily be clogged by particles.

5 Apparatus

5.1 Sieves.

For the test an appropriate number of either circular or rectangular sieves with a minimum effective sieve area of 250 cm² is required. The geometry of the apertures, the thickness of the sieves, the hole distances and the diameter of the holes shall be in accordance with the requirements of ISO 3310-1 and -2. The frame of the sieves shall have a height that enables the sieves to contain the samples and allows a free movement of the sample during the sieving process.

The number of sieves and the aperture sizes of the sieves shall be chosen according to the size specification of the actual sample material, see also EN 14961-1. For sawdust and similar fine grade materials it is recommended to use the following set of sieves:

EN 15149-2:2010 (E)

- 3,15 mm round holes;
- 2,8 mm mesh wire cloth;
- 2,0 mm mesh wire cloth;
- 1,4 mm mesh wire cloth;
- 1,0 mm mesh wire cloth;
- 0,5 mm mesh wire cloth;
- 0,25 mm mesh wire cloth.

5.2 Collecting pans.


For weighing of the size classes an adequate number of collecting pans is required.

5.3 Flat brush.

For cleaning the sieves a flat brush is required.

5.4 Mechanical sieving equipment.

The mechanical device shall apply a vibrating operation. For principle drawing of the sieving operation, see Figure 1.

**Key**

- x Increasing hole diameters
- y Material addition
- z Material flow direction

Figure 1 — Principle of the sieving operation

5.5 Balance, capable of measuring the mass of the sample to be sieved to the nearest 0,01 g.

6 Sample preparation

6.1 Sample size

The minimum size of the test sample for the determination of the size distribution shall be 50 g and shall be sampled according to prEN 14778. To prevent overloading of the screens the layer height on the upper sieve shall never exceed 2 cm. In such a case the test portion is divided into two or more subsamples which are to be processed subsequently, the results of the separate determinations are combined according to Clause 8.

NOTE The sample should include material for determination of size distribution and moisture content.

6.2 Moisture conditions

The sample shall be sieved at a moisture content below 20 w-% wet basis, thus preventing the particles from sticking together or losing moisture during the sieving process. If necessary the sample has to be pre-dried. Drying is done according to prEN 14780.

NOTE By pre-drying, as described in prEN 14780, the sample is brought into equilibrium with the humidity of the surrounding atmosphere.

Determine the moisture content of the material to be sieved on a separate sub-sample following the procedure given in EN 14774-1:2009, Clause 7 or EN 14774-2:2009, Clause 7. The moisture content shall be determined and reported concurrently with the particle size distribution determination.

7 Procedure

iTeh STANDARD PREVIEW

Assemble and operate the mechanical shaking device with the appropriate sieve sizes with decreasing aperture size ending with the collecting pan. Weigh the sample to the nearest 0,01 g. If the size of the test sample is significantly larger than the minimum 50 g stated in 6.1, the test portion shall be divided into two or more sub-portions which are to be processed subsequently.

Spread the sample in an even layer on the top sieve and start the sieving operation. The sieving operation shall be continued until the mass changes between two sequential sieves do not exceed a maximum of 0,3 % of the total sample mass per 1 min time of sieving operation. This pre-testing requirement can be avoided by applying the sieving operation for the fixed duration of 30 min.

NOTE 1 The required minimum sieving time should be determined for each equipment and type of fuel in separate pre tests.

Avoid losing any particles when determining individual weight differences during such pre tests.

NOTE 2 Be aware that an excessive sieving time, which is significantly longer than the minimum sieving time, can cause abrasion and a higher portion of the fine fraction.

NOTE 3 After approximately half of the sieving operation, check if the sample is even distributed on the sieves. If not, turn each of the sieves approximately 180° and complete the sieving operation.

Weigh the retained net material on each sieve and in the collecting pan to an accuracy of 0,01 g and record each net mass in a scheme equal to Table 1. In case that a particle sticks in a sieving hole, it shall be removed with the brush and added to the fraction, which remained on the sieve (as if it did not pass the hole).

NOTE 4 During the sieving operation particles can stick on the edge of the sieves due to static electricity. Since this problem mostly is connected to the intensity of the mechanical shaking operation it should be taken into consideration in separate pre test for each equipment and biofuel sample materials. Grounding the sieves using a copper wire can reduce the problem with static electricity.

NOTE 5 In size classification by sieving, some of the thin particles, which are longer than the hole diameter, will pass the sieve and mix with the particles in the smaller size fractions.