

SLOVENSKI STANDARD
kSIST-TS FprCEN/TS 17083:2017
01-marec-2017

Živila - Določevanje akrilamida v živilih in kavi s plinsko kromatografijo/masno spektrometrijo (GC-MS)

Foodstuffs - Determination of acrylamide in food and coffee by gas chromatography-mass spectrometry (GC-MS)

Lebensmittel - Bestimmung von Acrylamid in Lebensmitteln und Kaffee mit Gaschromatographie-Massenspektrometrie (GC-MS)

Produits alimentaires - Dosage par CG-SM des produits alimentaires et café avec GC-MS

Ta slovenski standard je istoveten z: FprCEN/TS 17083

[SIST-TS CEN/TS 17083:2017](https://standards.iteh.ai/catalog/standards/sist/80cee5c3-b1fd-423d-a3c9-3b757a4ebf7d/sist-ts-cen-ts-17083-2017)

<https://standards.iteh.ai/catalog/standards/sist/80cee5c3-b1fd-423d-a3c9-3b757a4ebf7d/sist-ts-cen-ts-17083-2017>

ICS:

| | | |
|--------|--|--|
| 67.050 | Splošne preskusne in analizne metode za živilske proizvode | General methods of tests and analysis for food products |
|--------|--|--|

kSIST-TS FprCEN/TS 17083:2017

en,fr,de

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

FINAL DRAFT
FprCEN/TS 17083

January 2017

ICS 67.050; 67.060; 67.140.20; 67.080.20

English Version

**Foodstuffs - Determination of acrylamide in food and
coffee by gas chromatography-mass spectrometry (GC-MS)**

Produits alimentaires - Dosage par CG-SM des produits
alimentaires et café avec GC-MS

Lebensmittel - Bestimmung von Acrylamid in
Lebensmitteln und Kaffee mit Gaschromatographie-
Massenspektrometrie (GC-MS)

This draft Technical Specification is submitted to CEN members for Vote. It has been drawn up by the Technical Committee CEN/TC 275.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a Technical Specification. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a Technical Specification.

Document Preview

SIST-TS CEN/TS 17083:2017

<https://standards.iteh.ai/catalog/standards/sist/80cee5c3-b1fd-423d-a3c9-3b757a4cbf7d/sist-ts-cen-ts-17083-2017>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

| | |
|--|----|
| European foreword..... | 3 |
| 1 Scope | 4 |
| 2 Normative references | 4 |
| 3 Principle | 4 |
| 4 Reagents | 4 |
| 5 Standard solution preparation | 6 |
| 6 Apparatus..... | 7 |
| 7 Procedure..... | 8 |
| 7.1 General..... | 8 |
| 7.2 Sample treatment..... | 9 |
| 7.3 Test sample preparation | 9 |
| 7.4 Bromination | 9 |
| 7.5 Partition cleanup..... | 9 |
| 8 GC-MS analysis..... | 10 |
| 8.1 General..... | 10 |
| 8.2 Injection conditions for split/splitless injector | 10 |
| 8.3 Oven conditions | 10 |
| 8.4 Column conditions | 10 |
| 8.5 Transfer line conditions | 10 |
| 8.6 Mass spectrometer conditions | 11 |
| 8.7 Mass spectrometer acquisition parameters and peak identification | 11 |
| 8.8 Calibration line | 11 |
| 8.9 Sample analysis..... | 12 |
| 9 Calculating and reporting | 12 |
| 9.1 Calculation | 12 |
| 9.2 Reporting | 13 |
| 10 Test report..... | 13 |
| 11 Precision data..... | 13 |
| 11.1 Single laboratory validation..... | 13 |
| 11.2 Interlaboratory trial | 13 |
| Annex A (informative) Typical chromatograms | 14 |
| Annex B (informative) GC-MS troubleshooting..... | 16 |
| B.1 General..... | 16 |
| B.2 Alternative GC columns..... | 17 |
| Annex C (informative) Single laboratory validation | 18 |
| Annex D (informative) Interlaboratory trial precision data | 20 |
| Bibliography..... | 22 |

European foreword

This document (FprCEN/TS 17083:2017) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This document is currently submitted to the Formal Vote.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

Annexes A, B C and D are informative.

WARNING 1— The use of this Technical Specification can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

WARNING 2— Some precaution is required when using polyacrylamide-based plastics because acrylamide may leach from these materials.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[SIST-TS CEN/TS 17083:2017](https://standards.iteh.ai/catalog/standards/sist/80cee5c3-b1fd-423d-a3c9-3b757a4ebf7d/sist-ts-cen-ts-17083-2017)

<https://standards.iteh.ai/catalog/standards/sist/80cee5c3-b1fd-423d-a3c9-3b757a4ebf7d/sist-ts-cen-ts-17083-2017>

1 Scope

This Technical Specification specifies a method for the determination of acrylamide in cereal-based products, potato-based products and coffee by gas-chromatography mass spectrometry (GC-MS).

The method has been single-laboratory validated via the analysis of spiked samples (French fries (uncooked), bread, water biscuit, infant cereal, biscuit, green coffee, roast coffee and instant coffee, ranging from 30 µg/kg to 1 500 µg/kg acrylamide).

The results from the single laboratory validation were obtained by a laboratory with significant experience in acrylamide analysis. In addition, this method has also been studied by inter laboratory trial via the analysis of samples containing incurred acrylamide, ranging from approximately 200 µg/kg to 2 000 µg/kg. Critical points of the method are identified in 7.5 and Clause 8.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696:1995, *Water for analytical laboratory use. Specification and test methods*

3 Principle

The test portion is extracted with hot water and isotopically labelled acrylamide is added as internal standard. High-fat samples are defatted with hexane, cleared with Carrez solution and centrifuged. Sample extracts are brominated and extracted with ethyl acetate. Following removal of the ethyl acetate by evaporation, triethylamine is added to partially debrominate, after which a portion of the sample extract is injected onto a GC-MS system for quantification. The chromatographic separation is obtained on a mid-polarity capillary GC column. The acrylamide derivative is ionized at 70 eV and recorded in selected ion monitoring (SIM) mode, and quantified by comparison with the stable isotope labelled analogue.

4 Reagents

SIST-TS CEN/TS 17083:2017

<https://standards.iteh.ai/catalog/standards/sist/80cee5c3-b1fd-423d-a3c9-3b757a4cbf7d/sist-ts-cen-ts-17083-2017>

Use only reagents of recognized analytical grade and water complying with grade 1 of EN ISO 3696:1995 (electrical conductivity below 0,1 µS/cm at 25 °C), unless specified otherwise. Standard solutions are preferably prepared gravimetrically. An analytical balance (6.1) is used for the preparation of both native and stable isotope labelled acrylamide.

WARNING — Acrylamide has been classified by the International Agency for Research on Cancer (IARC) as probably carcinogenic to humans (see [2]). Bromine is very toxic and corrosive and hydrobromic acid is corrosive and hazardous. The prepared brominating solution (see 4.17) shall be considered as very toxic and corrosive.

Protective equipment such as laboratory coat, disposable gloves and safety glasses shall be used. All handlings of acrylamide, bromine, hydrobromic acid, brominating solution and organic solvents shall be performed in a fume cupboard with adequate air flow.

IMPORTANT — Dispose chemical waste according to applicable environmental rules and regulations. Bromine is very toxic to aquatic organisms hence discharge into the environment shall be avoided. Waste shall be disposed of appropriately.

4.1 Acrylamide, purity ≥ 99 %.

4.2 [1,2,3-¹³C₃]-acrylamide, (*acrylamide-¹³C₃*) isotopic ¹³C purity 99 %, supplied as 1 000 µg/ml solution¹⁾

It is permissible to use deuterated acrylamide (acrylamide-d₃) as an alternative internal standard. In the following sections of the procedure, MS detection and calculation are prescribed for acrylamide-¹³C₃ only.

4.3 Bromine, purity ≥ 99 %.

4.4 Ethyl acetate, MS grade.

4.5 Helium purified compressed gas, (purity equivalent to 99,995 % or better).

4.6 n-Hexane, MS grade.

4.7 Hydrobromic acid, 48 % in water, assay ≥ 99,9 %.

4.8 Ice, crushed.

4.9 Magnesium sulfate, anhydrous powder.

4.10 Potassium bromide, purity ≥ 98,5 %.

4.11 Potassium hexacyanoferrate trihydrate (II), K₄Fe(CN)₆·3H₂O, purity ≥ 98,5 %.

4.12 Sodium chloride.

4.13 Sodium thiosulfate pentahydrate, Na₂S₂O₃·5 H₂O, purity ≥ 99,5 %.

4.14 Sodium thiosulfate solution, substance concentration c = 1 mol/l. Dissolve 24,82 g of sodium thiosulfate pentahydrate in 100 ml of water.

4.15 Triethylamine, N(CH₃)₃, purity ≥ 99 %.

4.16 Zinc acetate dihydrate, Zn(CH₃COO)₂ · 2 H₂O, purity ≥ 98,5 %.

4.17 Brominating solution.

¹⁾ This is an example of a suitable product available commercially, manufactured by Cambridge Isotope Laboratories Inc, Andover MA, USA. CLM-813-1.2 or similar. This information is given for the convenience of users of this Technical Specification and does not constitute an endorsement of these products by CEN. Equivalent products may be used if they provide similar results.

FprCEN/TS 17083:2017 (E)

Wear gloves and carry out the following steps in a fume cupboard. Prepare fresh reagent every three months.

For preparation of bromine water (saturated), place a 1 l glass bottle containing ca. 400 ml of water in a crushed ice bath and stir with a magnetic stirrer for ca. 10 min. Using a measuring cylinder, add $15 \text{ ml} \pm 1 \text{ ml}$ of bromine (4.3), stopper the flask and continue stirring for at least 60 min. Cap securely and store at 4°C to 6°C .

Weigh $400 \text{ g} \pm 2 \text{ g}$ of potassium bromide (4.10) into a 2 l screw-capped glass bottle and add $1\,000 \text{ ml} \pm 5 \text{ ml}$ of water. Shake or stir to dissolve then add $20 \text{ ml} \pm 1 \text{ ml}$ of hydrobromic acid (4.7) using a measuring cylinder.

Using measuring cylinders, add $320 \text{ ml} \pm 5 \text{ ml}$ of saturated bromine water followed by $660 \text{ ml} \pm 5 \text{ ml}$ of water to the glass bottle (6.3) and mix well. Stopper and store at 4°C to 6°C . Discard after 3 months. To dispose of excess saturated bromine water, add solid sodium thiosulfate pentahydrate (4.13) and swirl until the brown colour disappears.

4.18 Carrez solution I.

Dissolve 10,6 g of potassium hexacyanoferrate trihydrate (II) (4.11) in 100 ml of water. Discard after 6 months.

4.19 Carrez solution II.

Dissolve 21,9 g of zinc acetate dehydrate (4.16) in 100 ml of water. Discard after 6 months.

5 Standard solution preparation**5.1 General**

Prepare all standard solutions preferably gravimetrically. The tare masses of all recipients and the masses after each preparation step are recorded and used for calculation of the standard concentrations. Volumetric standard preparation may be also be applied provided that the volumetric glassware used complies with EN ISO 1042 (see [1]).

5.2 Acrylamide stock standard solution, mass concentration ρ approximately $1\,000 \mu\text{g/ml}$.

Weigh $0,1 \text{ g} \pm 0,001 \text{ g}$ of acrylamide into a 100 ml volumetric flask. Add 30 ml of water, swirl to dissolve and dilute to the mark with water and mix well. The stock solution is stable for 1 month when stored in a refrigerator (4°C to 6°C) and protected from light.

NOTE Alternatively, commercially-available certified standard solutions may be used if available.

5.3 Acrylamide working standard solution, ρ approximately $1 \mu\text{g/ml}$.

Pipette $100 \mu\text{l}$ of acrylamide stock standard solution (5.2) into a 100 ml volumetric flask, dilute to 100 ml with water and mix well. The solution is stable for 2 weeks when stored in a refrigerator (4°C to 6°C) and protected from light.

5.4 Acrylamide spiking solution, ρ approximately $100 \mu\text{g/ml}$.

Pipette 10 ml of acrylamide stock solution (5.2) into a 100 ml volumetric flask, dilute to 100 ml with water and mix well. The solution is stable for 2 weeks when stored in a refrigerator (4°C to 6°C) and protected from light.

5.5 $^{13}\text{C}_3$ -Acrylamide internal standard solution, ρ approximately $20 \mu\text{g/ml}$.

Pipette 200 µl of $^{13}\text{C}_3$ -acrylamide stock standard solution (4.2) into a 10 ml volumetric flask dilute to 10 ml with water. The solution is stable for 2 weeks when stored in a refrigerator (4 °C to 6 °C) and protected from light.

5.6 Calibration standards

Prepare calibration standards of approximately 0 µg/l, 1 µg/l, 3 µg/l, 5 µg/l, 10 µg/l, 20 µg/l and 100 µg/l according to the following scheme. Pipette 15 ml of water into each of seven 40 ml screw-cap vials and add acrylamide working standard solution (5.3) and $^{13}\text{C}_3$ -acrylamide internal standard solution (5.5) to each vial as detailed in Table 1. Carry out the bromination step (7.4) on each calibration standard.

Alternatively, the volume of water may be adjusted to give the same concentration of internal standard in each vial (i.e. by using a 15 ml volumetric flask which will give a nominal internal standard concentration of 20 µg/l). The concentration (i.e. linear working) range of the calibration standards is for guidance only. Other standard concentrations may be prepared if required but shall take into account expected acrylamide levels.

Table 1 — Calibration standards (nominal values)

| Water (ml) | Working standard solution (5.3) (µl) | Internal standard solution (5.5) (µl) | Concentration acrylamide (A) (µg/l) | Concentration internal standard (B) (µg/l) | Ratio of B:A |
|------------|--------------------------------------|---------------------------------------|-------------------------------------|--|--------------|
| 15 | 0 | 15 | 0 | 19,98 | 0 |
| 15 | 15 | 15 | 1,00 | 19,96 | 0,05 |
| 15 | 45 | 15 | 2,99 | 19,92 | 0,15 |
| 15 | 75 | 15 | 4,97 | 19,88 | 0,25 |
| 15 | 150 | 15 | 9,89 | 19,78 | 0,50 |
| 15 | 300 | 15 | 19,59 | 19,59 | 1,00 |
| 15 | 1 500 | 15 | 90,83 | 18,17 | 5,00 |

6 Apparatus

WARNING — All glassware shall be meticulously cleaned (except disposable glassware).

Usual laboratory glassware and equipment and, in particular, the following:

- 6.1 Analytical balance**, capable of weighing to an accuracy of $\pm 0,0001$ g.
- 6.2 Bottle, glass or plastic (polypropylene)**, ca. 175 ml to 250 ml capacity.
- 6.3 Bottle**, glass 1 l and 2 l capacity.
- 6.4 Centrifuge**, capable of $\geq 3\,000$ g, suitable for 50 ml centrifuge tubes.
- 6.5 Centrifuge tubes**, 50 ml capacity.
- 6.6 Gas chromatography mass spectrometry (GC-MS) apparatus**, comprising the following:
 - 6.6.1 Injection system**, split-splitless injector, suitable for temperatures up to 200 °C.

FprCEN/TS 17083:2017 (E)

A programmed temperature vaporizing (PTV) injector may be used as an alternative to a split/splitless injector.

6.6.2 GC oven, suitable for temperatures up to 300 °C and capable of temperature programming.

6.6.3 Sample carousel, suitable for use with vials and caps (6.13).

6.6.4 GC capillary column, DB 35²⁾ (35 %-Phenyl)-dimethylpolysiloxane), length of 30 m, internal diameter of 0,25 mm, $d_f = 0,25 \mu\text{m}$ ($\beta = 250$), or any column with comparable separation characteristics.

The use of a suitable guard column is highly recommended e.g. deactivated silica 5 m.

6.6.5 An interface to a mass spectrometer, with a temperature control device, suitable for temperatures up to 250 °C.

6.6.6 A mass spectrometer with the following characteristics:

- electron ionization source.
- ionization energy of 70 eV.
- mass resolution of at least 1.
- temperature control devices for the ion source (280 °C), the GC-MS interface (280 °C). Optionally a temperature control device for the quadrupole (150 °C).
- tuning stability of at least 48 h (allowing for the analysis of a sequence of samples and standards).
- response linearity range of at least two orders of magnitude.

6.7 Heater, dry-block or similar with nitrogen stream, thermostatically controlled at $40 \text{ °C} \pm 1 \text{ °C}$.

6.8 Pipettes, calibrated positive displacement 15 μl to 1 500 μl capacity.

6.9 Pipettes, glass or calibrated air displacement 4 ml to 10 ml capacity.

6.10 Shaker, orbital type.

6.11 Syringe filters, polytetrafluoroethylene (PTFE) 0,45 μm , 13 mm diameter.

6.12 Vials, glass screw cap 4 ml and 40 ml capacity.

6.13 Vials, glass for use with GC injection apparatus (6.6.3).

7 Procedure

7.1 General

Residues of acrylamide have occasionally been detected in laboratory wares e.g. filters. Acrylamide may also be formed as an artefact in certain analytical procedures e.g. during extraction or in the GC injector

²⁾ This is an example of a suitable product available commercially. This information is given for the convenience of the users of this Technical Specification and does not constitute an endorsement by CEN of the product named. Equivalent products may be used if they provide similar results.