
Analize živil - Določevanje benzo(a)pirena, benz(a)antracena, krizena in benzo(b)fluorantena v živilih s plinsko kromatografijo z masno spektrometrijo (GS-MS)

Food analysis - Determination of benzo[a]pyrene, benz[a]anthracene, chrysene and benzo[b]fluoranthene in foodstuffs by gas chromatography mass spectrometry (GC-MS)

Lebensmittelanalytik - Bestimmung von Benzo[a]pyren, Benz[a]anthracen, Chrysen und Benzo[b]fluoranthen in Lebensmitteln mit Gaschromatographie und Massenspektrometrie (GC-MS)

Analyse des produits alimentaires - Dosage du benzo(a)pyrène, benzo(a)anthracène, chrysène et benzo(b)fluoranthène dans les denrées alimentaires par chromatographie en phase gazeuse couplée à la spectrométrie de masse (CG-SM)

Ta slovenski standard je istoveten z: EN 16619:2015

ICS:

67.050	Splošne preskusne in analizne metode za živilske proizvode	General methods of tests and analysis for food products
--------	--	---

SIST EN 16619:2015**en,fr,de**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 16619:2015

<https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998-5cb3a045cd76/sist-en-16619-2015>

EUROPEAN STANDARD

EN 16619

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2015

ICS 67.050

English Version

Food analysis - Determination of benzo[a]pyrene,
benz[a]anthracene, chrysene and benzo[b]fluoranthene in
foodstuffs by gas chromatography mass spectrometry (GC-MS)

Analyse des produits alimentaires - Dosage du
benzo(a)pyrène, benzo(a)anthracène, chrysène et
benzo(b)fluoranthène dans les denrées alimentaires par
chromatographie en phase gazeuse couplée à la
spectrométrie de masse (CG-SM)

Lebensmittelanalytik - Bestimmung von Benzo[a]pyren,
Benz[a]anthracen, Chrysen und Benzo[b]fluoranthen in
Lebensmitteln mit Gaschromatographie und
Massenspektrometrie (GC-MS)

This European Standard was approved by CEN on 7 February 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents	5
5 Standard preparation	9
6 Apparatus	12
7 Procedure	15
8 GC-MS analysis.....	17
9 Calculation and reporting	22
10 Quality control.....	22
11 Precision data	24
12 Test report	27
Annex A (informative) Typical chromatograms	28
Annex B (informative) Precision data.....	30
Annex C (informative) Precision data from single laboratory validation.....	35
Bibliography	37

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 16619:2015

https://standards.iteh.ai/catalog/standards/sist/16619-2015/en-16619-2015

5cb3a045cd76/sist-en-16619-2015

Foreword

This document (EN 16619:2015) has been prepared by Technical Committee CEN/TC 275 "Food analysis - Horizontal methods", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2015 and conflicting national standards shall be withdrawn at the latest by October 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN 16619:2015](https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998-5cb3a045cd76/sist-en-16619-2015)

<https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998-5cb3a045cd76/sist-en-16619-2015>

EN 16619:2015 (E)

1 Scope

This European Standard specifies a method for the determination of 4 of the 16 EU priority polycyclic aromatic hydrocarbons (PAHs), identified as target PAHs. They are benz[a]anthracene (BaA), benzo[a]pyrene (BaP), benzo[b]fluoranthene (BbF) and chrysene (CHR). The method allows their quantification in the presence of the other 12 EU priority PAHs (benzo[j]fluoranthene (BjF), cyclopenta[cd]pyrene (CPP), benzo[k]fluoranthene (BkF), dibenz[a,h]anthracene (DhA), benzo[c]fluorene (BcL), dibenzo[a,e]pyrene (DeP), benzo[ghi]perylene (BgP), dibenzo[a,h]pyrene (DhP), dibenzo[a,i]pyrene (DiP), dibenzo[a,l]pyrene (DlP), indeno[1,2,3-cd]pyrene (IcP), 5-methylchrysene (5MC)) in extruded wheat flour, smoked fish, dry infant formula, sausage meat, freeze-dried mussels, edible oil and wheat flour, by gas-chromatography mass-spectrometry (GC-MS). The extraction of PAHs from solid samples is performed by pressurized liquid extraction (PLE). Soxhlet extraction was applied by some participants in the method validation study by collaborative trial as alternative to PLE. The sample cleanup is performed by applying the following techniques in the reported sequence: size exclusion chromatography (SEC), and solid phase extraction (SPE).

This method complies with the performance characteristics specified in Commission Regulation (EU) No 836/2011 (see [1]). In particular the specifications for the limit of detection (LOD) and of the limit of quantification (LOQ) (0,30 µg/kg and 0,90 µg/kg respectively) were met.

The method has been validated in an interlaboratory study via the analysis of both naturally contaminated and spiked samples, ranging from 0,5 µg/kg to 11,9 µg/kg. However, linearity of the instrument response was proven for the concentration range 0,5 µg/kg to 20 µg/kg.

For the determination of PAHs in edible fats and oils, two other standards are also available, EN ISO 22959 and EN ISO 15753, for more information see [2] and [3].

ITEH STANDARD PREVIEW
(standards.iteh.ai)

2 Normative references

SIST EN 16619:2015

<https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998->

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 1042:1999, *Laboratory glassware - One-mark volumetric flasks (ISO 1042:1998)*

EN ISO 3696:1995, *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*

3 Principle

The sample is homogenized. A test portion is mixed with desiccant, sand and the stable isotope labelled internal standard solution. It is then extracted with *n*-hexane or cyclohexane by pressurized liquid extraction, or alternatively by Soxhlet extraction. If applicable, co-extracted water is separated from the organic phase of the extract. The organic extract is evaporated to a small volume, filtered and purified by SEC, using a mixture of ethyl acetate and cyclohexane as eluent.

After SEC, 200 µl of toluene are added as a keeper to the collected SEC fraction. The SEC fraction is evaporated to about 200 µl, and cleaned up by SPE on silica, using cyclohexane as eluent. The cleaned up sample extract is evaporated again to about 200 µl. Finally, an injection standard solution is added to the sample prior to measurement by GC-MS.

The injection is performed with a PTV, or split/splitless injection port. The chromatographic separation is obtained on a mid-polar capillary column with high selectivity for PAHs. The analytes are ionised by electron ionization (EI) at 70 eV. The target PAHs are recorded in Single Ion Monitoring (SIM) mode, and quantified by comparison with the stable isotope labelled analogues.

4 Reagents

4.1 General

Use only reagents of recognized analytical grade and water complying with grade 1 of EN ISO 3696:1995, unless otherwise specified. All reagents and standard solutions shall be stored according to the specifications given by the supplier. The specifications given in this procedure for opened commercial solutions or for in-house prepared solutions aim to minimize solvent evaporation and to protect the analytes (PAHs) from degradation.

Standard solutions are preferably prepared gravimetrically. Depending on the handled amount of substance a micro-balance (6.4) and/or an analytical balance (6.5) are used for the preparation of solutions of both native and stable isotope labelled PAHs. All concentrations are expressed as mass per mass. If necessary, the concentrations expressed as mass per volume could be obtained applying the density equation (Formula (1)).

$$\rho = \frac{m}{v} \quad (1)$$

where

ρ density (in g/ml);

m measured mass of the substance (in g);

v volume of the solution (in ml).

iTeh STANDARD PREVIEW
(standards.iteh.ai)

The density of toluene at 20 °C is 0,8669 g/ml. Comprehensive information on the density of solvents at various temperatures is given in [4].

[SIST EN 16619:2015](https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998-3c65a045cd76/sist-en-16619-2015)

All solutions and substances are used at room temperature.

WARNING 1 — Some PAHs are considered carcinogenic. Persons using this document should be familiar with normal laboratory practices. It is the responsibility of the user of this document to apply practices which are in agreement with applicable occupational safety and health practices.

WARNING 2 — Dispose chemical waste according to applicable environmental rules and regulations.

WARNING 3 — PAHs are degraded by UV light. Protect PAHs solutions from light (keep in the dark, use aluminium foil or amber glassware).

WARNING 4 — Some precaution is needed when using plastics as polypropylene or PTFE because the analytes may be absorbed onto these materials.

4.2 Helium purified compressed gas (purity equivalent to 99,995 % or better).

4.3 Nitrogen purified compressed gas (purity equivalent to 99,995 % or better).

4.4 Disodium sulfate, (Na₂SO₄), anhydrous, granular.

4.5 Poly(acrylic acid), partial sodium salt-graft-poly(ethylene oxide) granular, 90 µm to 850 µm particle size.

4.6 Sand, 50 mesh to 70 mesh particle size.

4.7 *n*-Hexane.

EN 16619:2015 (E)**4.8 Acetone.****4.9 Cyclohexane.****4.10 Toluene.****4.11 Ethyl acetate.****4.12 SEC eluent**

Mix 1 part per volume of cyclohexane (4.9) with 1 part per volume of ethyl acetate (4.11).

4.13 SPE column

For the solid phase extraction cleanup, a silica SPE column is used. Commercial cartridges of 500 mg – 4 ml or self-filled cartridges of the same size containing 500 mg activated silica are used. The surface area of the silica should be around 500 m²/g.

NOTE Commercial SPE columns made of polypropylene were used in the method validation study by collaborative trial.

4.14 Reference material for quality control

A certified reference material, or any other suitable quality control material (e.g left over proficiency test material) may be applied for this purpose. The CITAC/Eurachem Guide to Quality in Analytical Chemistry may be consulted for guidance, see [5].

Analyse this material with every sample batch and use it to control the method performances along time (see 10.4).








[SIST EN 16619:2015](https://standards.iteh.ai/catalog/standards/sist/3493e801-4711-4e29-b998-5cb3a045cd76/sist-en-16619-2015)

4.15 Native reference substances - commercially available neat material or solutions of PAHs

The list of native substances analysed with this method is provided in Table 1. The target analytes are given in bold font. Commercially available, preferably certified, standard solutions are preferred due to the higher level of safety in handling.

Triphenylene, benzo[*j*]fluoranthene, and benzo[*k*]fluoranthene are potentially interfering with the target analytes and are therefore used for evaluation of selectivity.

Table 1 — Names and structures of the native PAHs

Name ^a	CAS number	Structure	Name ^a	CAS number	Structure
Benz[a]anthracene (BaA)	56–55–3	 Figure 1	Benzo[b]fluoranthene (BbF)	205–99–2	 Figure 2
Benzo[a]pyrene (BaP)	50–32–8	 Figure 3	Chrysene (CHR)	218–01–9	 Figure 4
Triphenylene (TRP)	217–59–4	 Figure 5	Benzo[j]fluoranthene (BjF)	205–82–3	 Figure 6
Benzo[k]fluoranthene (BkF)	207–08–9	 Figure 7			

^a The acronym is given in parenthesis, the target analytes are given in bold.

4.16 Stable isotope labelled reference standards (in the form of commercially available stable isotope labelled PAH solutions)

SIST EN 16619:2015

<https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998-3c65a0493d70/sist-en-16619-2015>

The stable isotope labelled analogues, applied for the quantification of the target PAHs are listed in Table 2. The commercial solutions used in the method validation study by collaborative trial contained the stable isotope labelled PAHs at a level of about 100 µg/kg in nonane.





Preference is given to ¹³C labelled analogues as their chemical properties best match those of the native analytes.

However, alternatively to ¹³C labelled substances, deuterated analogues of the target analytes may be applied. The concentration levels of these solutions should be similar to the levels specified for the ¹³C labelled PAH solutions.

NOTE 1 Highly deuterated PAHs are separated on the specified GC-column at least partially from their native analogues.

NOTE 2 Both forms of benz[a]anthracene-¹³C₆, which are displayed in Table 2, are equally suitable for the purpose of this standard. The ¹³C labelled reference material might be even supplied as a mixture, which was the case in the method validation study by collaborative trial.

Table 2 — Names and structures of ^{13}C labelled PAHs

Name	CAS number	Structure
Benz[a]anthracene- $^{13}\text{C}_6$	not available	 Figure 8
Benzo[a]pyrene- $^{13}\text{C}_4$	not available	 Figure 9
Benzo[b]fluoranthene- $^{13}\text{C}_6$	not available	 Figure 10
Chrysene- $^{13}\text{C}_6$	not available	 Figure 11

The recovery of the method is calculated based on the stable isotope labelled standards (see 10.2). Their physical-chemical properties are considered equivalent to the native PAHs. Table 3 indicates the correspondence between each native PAH and the stable isotope labelled analogue applied for its quantification.

Table 3 — Correspondence stable isotope labelled PAHs and native PAHs


Stable isotope labelled PAHs	Native PAH ^a
Benz[a]anthracene $^{13}\text{C}_6$	Benz[a]anthracene (BaA)
Benzo[a]pyrene $^{13}\text{C}_4$	Benzo[a]pyrene (BaP)
Benzo[b]fluoranthene $^{13}\text{C}_6$	Benzo[b]fluoranthene (BbF)
Chrysene $^{13}\text{C}_6$	Chrysene (CHR)

^a The acronym is given in parenthesis.

4.17 9-fluorobenzo[k]fluoranthene (FBkF), as injection standard, neat or in form of a commercially available solution

9-fluorobenzo[k]fluoranthene (FBkF) is used as injection standard and added to the sample extract prior to injection into the GC-MS. Both neat material and commercially available solutions may be used. The concentration of a commercial solution is preferably 100 µg/kg in toluene.

Table 4 — Name and structure of the injection standard

Name ^a	CAS number	Structure
9-fluorobenzo[k]fluoranthene (FBkF)	113600-15-0	 Figure 12

^a The acronym is given in parenthesis.

5 Standard preparation

5.1 General

All standard solutions are preferably prepared gravimetrically. The tare masses of all recipients and the masses after each preparation step are recorded and used for calculation of the standard concentrations.

Volumetric preparation of standard solutions may be applied as well, provided that the used volumetric glassware complies with EN ISO 1042:1999.

Standard solutions may be prepared from neat materials, from commercial single substance solutions, or from commercial mixes. However, compatibility of the solvents of commercial solutions with toluene shall be taken into account.

Preferably, a commercial mixed PAH standard solution should be used for the preparation of intermediate standard solution (5.7) and calibration solutions (5.8). The concentration of this solution shall be in the same order of magnitude as the mixed PAH stock solution (5.4).

5.2 Injection standard solution

Prepare a solution of FBkF (4.17) in toluene (4.10) with a concentration of approximately 400 ng/g. The conversion of mass-per-mass units (ng/g) to mass-per-volume units (ng/ml) is done via the density equation (Formula (1)). For toluene a density value of 0,866 9 g/ml is applied.

This solution will be used for the spiking of the sample extract before measurement by GC-MS (see 7.9) to assess the recovery of stable isotope labelled PAHs (see 10.2).

Store this solution in the dark and at a temperature below 10 °C. A solution stored in this way is stable for at least six months. If longer stability is proven, the solution can still be applied.

<https://standards.iteh.ai/catalog/standards/sist/3403e80a-4711-4e29-b998-5cb3a045cd76/sist-en-16619-2015>

5.3 PAH stock solutions

A solution in toluene (4.10) with a concentration in the range from 50 µg/g to 150 µg/g shall be prepared for each of the native PAHs listed in Table 1 in case neat reference materials are applied for the preparation of calibrants. These PAH stock solutions are prepared by weighing from 1 mg to 5 mg of each neat substance into glass weighing cylinders (6.1) using the microbalance (6.4). The weighing cylinder (6.1) is transferred with tweezers into a 100 ml amber glass volumetric flask (6.2). About 40 ml of toluene is added and weighed with an analytical balance (6.5). Other amber glass recipients may be applied, provided that solvent evaporation is minimized during standard preparation. Each flask shall be sonicated for a couple of minutes to achieve complete dissolution of the native PAHs in the solvent.

Once the solutions are homogeneous, they are transferred for storage into 40 ml amber glass vials (6.3). These solutions will be used for the preparation of mixed PAH stock solution (5.4) and, finally, of calibration solutions (see 5.8).

Store these solutions in the dark and at a temperature below 10 °C. A solution stored in this way is stable for at least 12 months. If longer stability is proven, the solution can still be applied.

5.4 Mixed PAH stock solutions

Prepare, from the PAH stock solutions (5.3), a solution in toluene (4.10) with a concentration of approximately 2 µg/g. For this purpose, both the PAH stock solutions (5.3) and the toluene (4.10) are weighed using an analytical balance (6.5).

This solution will be used for the preparation of the intermediate standard solutions (5.7) and, finally, of the calibration solutions (see 5.8).

EN 16619:2015 (E)

Store this solution in the dark, at a temperature below 10 °C. A solution stored in this way was proven to be stable for at least 12 months. If longer stability is proven, the solution can still be applied.

5.5 Labelled PAH stock solution

Prepare with the individual solutions of stable isotope labelled PAHs (4.16) listed in Table 2, a solution in toluene (4.10) with a concentration of approximately 700 ng/g. Use an analytical balance (6.5) for this purpose.

This solution will be used for the preparation of the process solution (5.6).

Store this solution in the dark and at a temperature below 10 °C. A solution stored in this way is stable for at least 12 months. If longer stability is proven, the solution can still be applied.

5.6 Process solution

Prepare, from the Labelled PAH stock solution (5.5), a solution in toluene (4.10) with a concentration of approximately 150 ng/g. This concentration is obtained by adding 4,5 ml of the labelled PAH stock solution (5.5) to 16 ml of toluene (4.10). The exact amounts of the two components are determined gravimetrically with an analytical balance (6.5).

This solution will be used for spiking of the test portion (see 7.2).

Store this solution in the dark and at a temperature below 10 °C. A solution stored in this way is stable for at least six months. If longer stability is proven, the solution can still be applied.

5.7 Intermediate solutions

Prepare the intermediate solutions which will be used for calibration from the mixed PAH stock solution (5.4) by dilution in toluene (4.10).

These solutions will be used for the preparation of the calibration solutions (see 5.8).

The concentrations of PAHs in these solution shall be approximately two times the concentrations of PAHs in the calibration solutions (see 5.8), hence in the range from 10 ng/g to 250 ng/g. The required amounts of mixed PAH stock solution (5.4) are listed in Table 5. They are pipetted into a 100 ml amber volumetric flask (6.2) and made up to volume with toluene (4.10). The standard concentrations are calculated from gravimetical data, which are recorded at each preparation step. Use an analytical balance (6.5) for weighing.