

---

---

**Cereals and pulses — Determination of  
the nitrogen content and calculation of  
the crude protein content — Kjeldahl  
method**

*Céréales et légumineuses — Détermination de la teneur en azote et  
calcul de la teneur en protéines brutes — Méthode de Kjeldahl*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 20483:2006](https://standards.iteh.ai/catalog/standards/sist/c0b813f6-7925-411d-8e65-628b92f146de/iso-20483-2006)

<https://standards.iteh.ai/catalog/standards/sist/c0b813f6-7925-411d-8e65-628b92f146de/iso-20483-2006>



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 20483:2006

<https://standards.iteh.ai/catalog/standards/sist/c0b813f6-7925-411d-8e65-628b92f146de/iso-20483-2006>

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

Page

Foreword.....	iv
<b>1 Scope .....</b>	<b>1</b>
<b>2 Normative references .....</b>	<b>1</b>
<b>3 Terms and definitions.....</b>	<b>1</b>
<b>4 Principle.....</b>	<b>2</b>
<b>5 Reagents .....</b>	<b>2</b>
<b>6 Apparatus .....</b>	<b>3</b>
<b>7 Sampling.....</b>	<b>3</b>
<b>8 Preparation of test sample.....</b>	<b>3</b>
<b>9 Determination of the moisture content.....</b>	<b>4</b>
<b>10 Procedure .....</b>	<b>4</b>
10.1 General.....	4
10.2 Test portion .....	4
10.3 Determination .....	4
10.4 Blank test.....	5
10.5 Test with reference material (Check test) .....	5
<b>11 Expression of results .....</b>	<b>5</b>
11.1 Nitrogen content .....	5
11.2 Crude protein content .....	6
<b>12 Precision.....</b>	<b>6</b>
12.1 Interlaboratory test .....	6
12.2 Repeatability.....	6
12.3 Reproducibility .....	6
12.4 Critical difference.....	7
<b>13 Test report .....</b>	<b>7</b>
<b>Annex A (informative) Results of interlaboratory tests.....</b>	<b>8</b>
<b>Annex B (informative) Critical difference and practical application of the repeatability and reproducibility limits to different protein contents .....</b>	<b>10</b>
<b>Annex C (informative) Factors for converting nitrogen content to protein content.....</b>	<b>12</b>
<b>Bibliography .....</b>	<b>13</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 20483 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 4, *Cereals and pulses*.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 20483:2006](https://standards.iteh.ai/catalog/standards/sist/c0b813f6-7925-411d-8e65-628b92f146de/iso-20483-2006)

<https://standards.iteh.ai/catalog/standards/sist/c0b813f6-7925-411d-8e65-628b92f146de/iso-20483-2006>

# Cereals and pulses — Determination of the nitrogen content and calculation of the crude protein content — Kjeldahl method

**WARNING** — The use of this standard may involve hazardous materials, operations and equipment. This standard does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety practices and determine the applicability of regulatory limitations prior to use.

## 1 Scope

This International Standard specifies a method for the determination of the nitrogen content of cereals, pulses and derived products, according to the Kjeldahl method, and a method for calculating the crude protein content.

The method does not distinguish between protein nitrogen and non-protein nitrogen. If it is important to determine the non-protein nitrogen content, an appropriate method can be applied.

NOTE In certain cases, full recovery of the nitrogen in nitrates and nitrites is not possible by this method.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 712, *Cereals and cereal products — Determination of moisture content — Routine reference method*

ISO 6540, *Maize — Determination of moisture content (on milled grains and on whole grains)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1

#### **nitrogen content**

quantity of nitrogen determined after application of the procedure described in this International Standard

NOTE It is expressed as mass fraction of dry product, in percent.

### 3.2

#### **crude protein content**

quantity of crude protein obtained from the nitrogen content as determined by applying the method described in this document, calculated by multiplying this content by an appropriate factor depending on the type of cereal or pulse

NOTE It is expressed as mass fraction of dry product, in percent.

## 4 Principle

A test portion is digested by sulfuric acid in the presence of a catalyst. The reaction products are made alkaline then distilled. The liberated ammonia is collected in a boric acid solution, which is titrated with a sulfuric acid solution, to determine the nitrogen content and calculate the crude protein content.

## 5 Reagents

Use only nitrogen-free reagents of recognized analytical grade, except for the reference materials, and distilled or demineralized water or water of equivalent purity.

**WARNING — The reagents described in 5.4, 5.8, 5.11 and 5.12 shall be handled with precautions.**

**5.1 Potassium sulfate** ( $K_2SO_4$ ).

**5.2 Copper(II) sulfate pentahydrate** ( $CuSO_4 \cdot 5H_2O$ ).

**5.3 Titanium oxide** ( $TiO_2$ ).

**5.4 Sulfuric acid**,  $c(H_2SO_4) = 18 \text{ mol/l}$ ,  $\rho_{20}(H_2SO_4) = 1,84 \text{ g/ml}$ .

**5.5 Paraffin oil.**

**5.6 Acetanilide** ( $C_8H_9NO$ ), having a melting point of  $114^\circ C$  and nitrogen content of  $10,36 \text{ g/100 g}$ .

**5.7 Tryptophan** ( $C_{11}H_{12}N_2O_2$ ), having a melting point of  $282^\circ C$  and nitrogen content of  $13,72 \text{ g/100 g}$ .

**5.8 Phosphorus pentoxide** ( $P_2O_5$ ).

**5.9 Boric acid**, aqueous solution,  $\rho_{20}(H_3BO_3) = 40 \text{ g/l}$ , or any other concentration recommended for the apparatus being used.

**5.10 Coloured indicator**

Add volumes of Solution A (5.10.1) and Solution B (5.10.2) as recommended for the apparatus being used (for example: 5 volumes of Solution A and 1 volume of Solution B).

NOTE 1 It is possible to use a ready-to-use solution of boric acid containing the coloured indicator (5.9 + 5.10).

NOTE 2 The ratio of Solutions A and B may be adjusted depending on the apparatus.

The titration may also be carried out potentiometrically by the use of pH-electrode, which has to be checked every day.

### 5.10.1 Solution A

Bromocresol green ( $C_{21}H_{14}Br_4O_5S$ ): 200 mg.

Ethanol ( $C_2H_5OH$ ), with a volume fraction of 95 %: quantity sufficient for 100 ml solution.

### 5.10.2 Solution B

Methyl red ( $C_{15}H_{15}N_3O_2$ ): 200 mg.

Ethanol ( $C_2H_5OH$ ), with a volume fraction of 95 %: quantity sufficient for 100 ml solution.

**5.11 Sodium hydroxide**, aqueous solution (NaOH), with a mass fraction of 33 %, or a mass fraction of 40 %, with nitrogen content less than or equal to 0,001 %.

Sodium hydroxide of technical grade may also be used when its nitrogen content is less than or equal to 0,001 %.

**5.12 Sulfuric acid**, standard volumetric solution,  $c(\text{H}_2\text{SO}_4) = 0,05 \text{ mol/l}$ .

The use of  $\text{H}_2\text{SO}_4$  instead of HCl is recommended because  $\text{H}_2\text{SO}_4$  does not have the tendency to produce bubbles in the connecting tubes.

**5.13 Ammonium sulfate**, standard volumetric solution,  $c(\text{NH}_4)_2\text{SO}_4 = 0,05 \text{ mol/l}$ .

Alternatively, a salt such as  $(\text{NH}_4)_2\text{Fe}(\text{SO}_4)_2 \cdot 6\text{H}_2\text{O}$  may be used.

**5.14 Pumice stone**, granulated, washed in hydrochloric acid and ignited.

**5.15 Sucrose** (optional), free from nitrogen.

## 6 Apparatus

**6.1 Mechanical grinder.**

**6.2 Sieve**, with aperture size 0,8 mm.

**6.3 Analytical balance**, capable of weighing to the nearest 0,001 g.

**6.4 Digestion, distillation and titration apparatus.**

The homogeneous temperature distribution of the digestion unit should be ascertained.

The assessment of temperature homogeneity should be done by carrying out a whole test with one of the two reference materials (5.6 or 5.7) and considering the recovery rates obtained.

The distillation apparatus should also be verified by conducting the distillation of a known quantity of ammonium salt [e.g. 10 ml of an ammonium sulfate solution (5.13)] and by checking that the recovery rate is greater than or equal to 99,8 %.

## 7 Sampling

A representative sample should have been sent to the laboratory. It should not have been damaged or changed during transport or storage.

Sampling is not part of the method specified in this International Standard. Recommended sampling methods are given in ISO 6644 and ISO 13690.

## 8 Preparation of test sample

If necessary, grind the sample so that it passes entirely through a sieve with 0,8 mm aperture size. For grains, a mass of at least 200 g should be ground. Mix the ground sample thoroughly.

## 9 Determination of the moisture content

Determine the moisture content ( $w_H$ ) of the test sample from an aliquot of the sample prepared according to Clause 8. Carry out the determination by following the method adapted to the product under test (i.e. ISO 712 for cereals and cereal products, ISO 6540 for maize, or by the method described in Reference [10] for certain pulses).

## 10 Procedure

### 10.1 General

If it is required to check that the requirements given concerning the repeatability limit (12.2) are fulfilled, carry out two separate determinations in accordance with 10.2 to 10.5.

### 10.2 Test portion

Weigh, to the nearest 0,001 g, a mass of test sample prepared according to Clause 8, chosen on the basis of the assumed nitrogen content, so that the test portion contains between 0,005 g and 0,2 g of nitrogen and preferably more than 0,02 g.

### 10.3 Determination

#### 10.3.1 Digestion

iTeh STANDARD PREVIEW

**WARNING** — The following operations should be conducted under a well-ventilated, sulfuric acid resistant hood.

Transfer the test portion (10.2) to the digestion flask, then add  
<https://standards.iteh.ai/catalog/standards/sist/c0b813f6-7925-411d-8e65-628b92f146de/iso-20483-2006>

- 10 g of potassium sulfate (5.1),
- 0,30 g of copper(II) sulfate pentahydrate (5.2),
- 0,30 g of titanium oxide (5.3) (a catalyst in pellet form corresponding to the described composition may be used), and
- 20 ml of sulfuric acid (5.4).

The quantity of acid may be adjusted depending on the apparatus, but only after having made certain that this measure indeed leads to a recovery rate of 99,5 % for acetanilide and 99,0 % for tryptophan.

Carefully mix so as to ensure a thorough wetting of the test portion.

Place the flasks in the digestion unit preheated to  $(420 \pm 10) ^\circ\text{C}$ .

After a minimum of 2 h of digestion counted from the time the unit temperature again reached  $(420 \pm 10) ^\circ\text{C}$ , leave to cool.

**NOTE** It is advisable to add pumice stone (5.14) as boiling regulator and an antifoaming agent such as paraffin oil (5.5).

The minimum digestion time shall be checked on that reference material with which it was most difficult to reach the recovery rate (see 10.5).

Follow the recommendations of the equipment manufacturer as far as evacuation of the vapours is concerned, because too strong a suction can result in a loss of nitrogen.



### 10.3.2 Distillation

Cautiously add to the cooled flask 50 ml of water and leave to cool. Transfer into the collecting flask 50 ml of boric acid (5.9) and, for visual colorimetry or using an optical probe, at least 10 drops of coloured indicator (5.10).

Add an **excess** of 5 ml of the sodium hydroxide solution (5.11) required for neutralizing the quantity of sulfuric acid used. Then carry out the distillation.

Depending on the equipment, the quantities of reagents used may vary.

### 10.3.3 Titration

Carry out the titration using the sulfuric acid solution (5.12), either continuously during the distillation or at the end of distillation on all of the distillate.

The end-point determination is judged by visual colorimetry, or using an optical probe, or by potentiometric analysis with a pH measurement system.

### 10.4 Blank test

Perform a blank test with the reagents used in 10.3.1 to 10.3.3 but without the test sample (10.2).

NOTE It is possible to replace the test sample with 1 g of sucrose (5.15).

### 10.5 Test with reference material (Check test)

Dry the used reference material(s) at a temperature between 60 °C and 80 °C, under vacuum, in the presence of phosphorus pentoxide (5.8).

Carry out a check test on a test portion of a minimum of 0,15 g by determining the nitrogen content of the tryptophan (5.7) and/or of the acetanilide (5.6).

NOTE It is possible to add 1 g of sucrose (5.15) to reference material.

The nitrogen recovery rate from acetanilide shall be at least 99,5 % and at least 99,0 % from tryptophan.

## 11 Expression of results

### 11.1 Nitrogen content

The nitrogen content,  $w_N$ , expressed as a mass fraction of dry product, in percent, is obtained using the following equation:

$$w_N = \frac{(V_1 - V_0) \times T \times 0,014 \times 100}{m} \times \frac{100}{100 - w_H} = \frac{140 T (V_1 - V_0)}{m(100 - w_H)}$$

where

$V_0$  is the volume, in millilitres, of the sulfuric acid solution (5.12) required for the blank test;

$V_1$  is the volume, in millilitres, of the sulfuric acid solution (5.12) required for the test portion;

0,014 is the value, in grams, of the quantity of nitrogen equivalent to the use of 1 ml of a 0,5 mol/l sulfuric acid solution;