



SLOVENSKI STANDARD
SIST ISO 9236-1:2011

01-julij-2011

Fotografija - Senzimetrija zaslonsko-filmskih sistemov za medicinsko radiografijo - 1. del: Določanje oblike senzimetrične krivulje, splošne občutljivosti in povprečnega gradienta

Photography - Sensitometry of screen/film systems for medical radiography - Part 1: Determination of sensitometric curve shape, speed and average gradient

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Photographie - Sensitométrie des ensembles film/écran pour la radiographie médicale - Partie 1: Détermination de la forme de la courbe sensitométrique, de la sensibilité et du contraste moyen

<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011>

Ta slovenski standard je istoveten z: ISO 9236-1:2004

ICS:

37.040.25 Radiografski filmi Radiographic films

SIST ISO 9236-1:2011 **en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 9236-1:2011

<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011>

INTERNATIONAL STANDARD

ISO
9236-1

Second edition
2004-02-01

Photography — Sensitometry of screen/film systems for medical radiography —

Part 1:

Determination of sensitometric curve shape, speed and average gradient

iTeh STANDARD PREVIEW
(standards.iteh.ai)

*Photographie — Sensitométrie des ensembles film/écran pour la
radiographie médicale —*

*Partie 1. Détermination de la forme de la courbe sensitométrique, de la
sensibilité et du contraste moyen*

SIST ISO 9236-1:2011
<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-ab11-43d5-9849-711814927043/sist-iso-9236-1-2011>



Reference number
ISO 9236-1:2004(E)

© ISO 2004

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 9236-1:2011](https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011)

<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction	v
1 Scope.....	1
2 Normative references	1
3 Terms and definitions.....	1
4 General requirements	2
4.1 Storage and handling conditions	2
4.2 Safelights	2
4.3 X-ray equipment	3
4.4 Air kerma meter.....	3
4.5 Processing.....	3
4.6 Densitometry	4
5 Determination of sensitometric curve shape.....	4
5.1 General	4
5.2 Beam qualities.....	5
5.3 Geometry for curve shape determination.....	5
5.4 Exposure	8
5.5 Evaluation	8
6 Determination of average gradient.....	8
7 Determination of speed	10
7.1 Definition.....	10
7.2 Beam qualities.....	10
7.3 Geometry	13
7.4 Exposure	13
7.5 Evaluation	13
8 Speed and average gradient determination without sensitometric curve.....	16
9 Uncertainty	16
10 Test report	17
Annex A (informative) Rationale	18
Bibliography	20

ISO 9236-1:2004(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9236-1 was prepared by Technical Committee ISO/TC 42, *Photography*.

This second edition cancels and replaces the first edition (ISO 9236-1:1996), which has been technically revised to incorporate the following technical and major editorial changes:

- a spherical ionization chamber, or an equivalent detector, is required for dosimetry;
- only high frequency or 12-pulse high-voltage generators are allowed, 6-pulse high-voltage generators are excluded;
- the allowed uncertainty for the density measurement has been increased in order to comply with the other parts of the ISO 9236 series;
- the exposure times for the determination of speed and sensitometric curve shape have been reduced to match the current state of the art;
- the phantom of Technique IV has been changed (leaving the beam quality unchanged) in order to reduce the air kerma rate;
- the distances between the focal spot of the x-ray tube and the screen-film combination when determining speed and average gradient may now be in the range from 1,5 m to 4,0 m;
- the use of a monitoring detector is no longer mandatory, because the precision of modern x-ray tubes and high-voltage generators is often superior to that of monitoring detectors;
- the total uncertainty which can be reached has been changed;
- an informative annex has been added in order to describe the background of speed and curve shape measurements, the choice of phantoms, and the energy dependence of speed values.

ISO 9236 consists of the following parts, under the general title *Photography — Sensitometry of screen/film systems for medical radiography*:

- *Part 1: Determination of sensitometric curve shape, speed and average gradient*
- *Part 3: Determination of sensitometric curve shape, speed and average gradient for mammography*

The following part is under preparation:

- *Part 2: Method for determining modulation transfer function (MTF)*

Introduction

This part of ISO 9236 provides methods for determining the sensitometric curve shape, the average gradient and the speed of radiographic screen/film/filmholder/processing systems used in medical radiography, except in mammography and dental radiography.

The sensitometric curve shape, which is also needed for the determination of other properties (as, for example, the modulation transfer function), is measured under low scatter conditions via intensity scale X-ray sensitometry, preferably using an inverse square sensitometer. For the determination of the sensitometric curve shape, as well as for a subsequent determination of the average gradient from the measured curve, but not for speed, the irradiation of the screen/film/filmholder combination need to be measured only in relative units.

Speed is measured in a separate way, under exposure conditions which simulate medical practice more closely, including realistic fractions of scattered radiation. Different types of medical exposures are simulated by using appropriate phantoms and X-ray tube voltages, and the screen/film/filmholder combination is exposed behind the respective phantom. The irradiation is measured in absolute units of air kerma (gray, Gy) in order to determine the speed.

Four different techniques are defined, differing in beam quality and fraction of scattered radiation, simulating the imaging of extremities, skull, lumbar spine and colon, and chest. Speed may be measured for each technique of interest. Owing to its dependence on X-ray energy and scatter, screen/film system speed varies widely in medical practice. The four measurement conditions described in this part of ISO 9236 provide values that are representative of those found under practical conditions.

[SIST ISO 9236-1:2011](https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011)

<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 9236-1:2011

<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011>

Photography — Sensitometry of screen/film systems for medical radiography —

Part 1: Determination of sensitometric curve shape, speed and average gradient

1 Scope

This part of ISO 9236 specifies methods for the determination of the sensitometric curve shape, average gradient and speed of a single sample of a screen/film/filmholder/processing system for medical radiography. It is not applicable to special radiographic applications such as mammography, dental radiography and direct-exposing medical radiographic systems (see for example ISO 5799 [3]).

The filmholder can be any means that ensures close screen/film contact and prevents the film from being exposed to ambient light. In particular, the filmholder can be a light-tight vacuum bag, as often used in the laboratory, or a radiographic cassette as used in medical radiography.

2 Normative references

SIST ISO 9236-1:2011

[https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-](https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011)

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5-2:2001, *Photography — Density measurements — Part 2: Geometric conditions for transmission density*

ISO 5-3:1995, *Photography — Density measurements — Part 3: Spectral conditions*

ISO 554:1976, *Standard atmospheres for conditioning and/or testing — Specifications*

IEC 60522:1999, *Determination of the permanent filtration of X-ray tube assemblies*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

screen/film system

radiographic imaging system consisting of screen(s), film, filmholder and film processing

NOTE Hereafter, screen/film/filmholder combinations will be referred to as “combinations” and will be referred to as “systems” when the processing is included.

ISO 9236-1:2004(E)**3.2****air kerma** K

sum of the initial kinetic energies of all charged particles (e.g., electrons) liberated by uncharged particles (e.g., X-ray photons) from air molecules, divided by the mass of air in that volume where the charged particles are liberated

NOTE The unit is the gray (Gy).

3.3**sensitometric curve**

plot of the density of a processed photographic film as a function of the logarithm to the base 10 of the exposure

3.4**speed** S

quantitative measure of the response of the screen/film system to radiant energy for the specified conditions of exposure, processing and density measurement

3.5**average gradient** \bar{G}

slope of the straight line joining two specified points on a sensitometric curve

3.6**net density** D

density of an exposed and processed film minus the density of an unexposed and processed sample of that film

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST ISO 9236-1:2011](https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011)

<https://standards.iteh.ai/catalog/standards/sist/5bc0958e-abf0-43d5-9849-711814927043/sist-iso-9236-1-2011>

3.7**coverage factor** k

numerical factor, used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty

NOTE The coverage factor is explained in the *Guide to the expression of uncertainty in measurement* [8]. Its value is typically in the range of 2 to 3. The coverage factor is chosen based on the level of confidence desired. A coverage factor (k) of 2 generally will result in a level of confidence of approximately 95 %, and a coverage factor of 3 generally will result in a level of confidence of approximately 99 %. This association of confidence level and coverage factor is based on an assumption regarding the probability distribution of measurement results.

4 General requirements**4.1 Storage and handling conditions**

The film and screens shall be stored according to the manufacturer's recommendations. Before and during exposures, the temperatures of the films and screens shall be maintained at $23\text{ °C} \pm 2\text{ °C}$ (see ISO 554) and the moisture content of the film shall be such that it will be in equilibrium at a relative humidity of $(50 \pm 20)\%$.

4.2 Safelights

To eliminate the possibility of safelight illumination affecting the sensitometric results, all films shall be kept in total darkness during handling, exposure and processing.

4.3 X-ray equipment

For all tests described in this part of ISO 9236, high frequency (multipulse) high voltage generators or at least 12-pulse high voltage generators shall be used.

For all tests described in this part of ISO 9236, X-ray tubes equipped with fixed anodes or rotating anodes may be used. In either case, the target material shall be tungsten or a tungsten-based alloy.

NOTE 1 The target is that part of the anode onto which the electron beam is directed to produce X-radiation. For technological reasons it is common practice to use alloys of tungsten with up to 10 % rhenium for the target, while other parts of the anode can consist of other materials (e.g. molybdenum).

The permanent filtration of the X-ray tube and its housing, as defined in IEC 60522, shall be equivalent to $2,5 \text{ mm} \pm 0,2 \text{ mm}$ of aluminium.

NOTE 2 The permanent filtration of the X-ray tube and its housing is effected by permanently fixed materials intercepting the X-ray beam, which are not intended to be removed for any application. As the permanent filtration is usually stated on the X-ray tube housing and in the accompanying documents, its measurement, as described in IEC 60522, is not necessary.

4.4 Air kerma meter

For the air kerma measurement, calibrated detectors shall be used. The uncertainty of air kerma measurement (level of confidence 95 %) shall be less than 3 % for collimated beams without scatter, and less than 5 % for radiation measurements behind the phantom when scattered radiation is included.

A spherical ionization chamber of 30 cm^3 to 100 cm^3 volume should be used for measurements where scattered radiation is involved. The chamber shall be calibrated for the beam qualities given in Table 2, including scattered radiation. The centre of the spherical chamber is to be considered the reference point; the stem of the spherical chamber should point in a direction opposite to the radiation source.

NOTE During calibration of the air kerma meter and during usage, scattered radiation originating not from the phantom but from, for example, the stem of the chamber, can be minimized in order to meet the specified uncertainty requirement.

4.5 Processing

Screen/film systems, including either manual or automatic processing, may be tested in accordance with this part of ISO 9236. Processing should be carried out in accordance with the film manufacturer's recommendations. Nothing shall be construed to require the disclosure of proprietary information.

No processing specifications are described in this part of ISO 9236 in recognition of the wide range of chemicals and equipment used. Speed and average gradient values provided by film manufacturers generally apply to the system when the film is processed in accordance with their recommendations so that the photographic characteristics specified for the process are produced. Processing information shall be provided by the film manufacturer or others who quote speed and average gradient values and shall specify the processing chemicals, times, temperatures, agitation, equipment and procedures used for each of the processing steps, and any additional information required to obtain the sensitometric results described. The values for speed and average gradient obtained using other processing procedures may differ significantly. The processing conditions selected by a person using this part of ISO 9236 are, in any case, part of the system being tested.

NOTE 1 Different speeds for a particular film can be achieved by varying the processes. However, these variations to the processes can cause other undesirable changes.

In order to minimize any effects due to latent-image instability or process variability, all film samples shall be processed together, neither less than 30 min nor more than 4 h after exposure. Between exposure and processing, the temperature of the film shall be maintained at $23 \text{ °C} \pm 2 \text{ °C}$ and its moisture content shall be such that the film will be in equilibrium at a relative humidity of $(50 \pm 20) \%$.