
**Dentistry — Dental handpieces —
Electrical-powered scalers and scaler tips**

*Art dentaire — Pièces à main dentaires — Instruments pour détartrage
électriques et parties actives des instruments pour détartrage*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 22374:2005](https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cf839d1/iso-22374-2005)

[https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-
ba48cf839d1/iso-22374-2005](https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cf839d1/iso-22374-2005)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 22374:2005](https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005)

<https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005>

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 Requirements	2
4.1 General design of handpiece connection	2
4.2 Scaler tip.....	2
4.3 Performance	2
4.4 Supply of cooling liquid	3
4.5 Noise level	3
4.6 Resistance to sterilization	3
4.7 Energy for light source (if applicable)	3
4.8 Electrical power supply.....	3
5 Sampling.....	3
6 Test methods.....	3
6.1 General.....	3
6.2 Visual inspection	3
6.3 Scaler tip.....	4
6.4 Frequency.....	4
6.5 Amplitude, unloaded scaler tip	5
6.6 Amplitude, loaded scaler tip.....	5
6.7 Supply of cooling liquid	6
6.8 Noise level	6
6.9 Resistance to sterilization	7
6.10 Energy for light supply (if applicable)	7
7 Instructions for use, maintenance and service	8
8 Marking	9
9 Labelling	9
10 Packaging	9
Annex A (informative) Performance and general design	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 22374 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 22374:2005](https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cf839d1/iso-22374-2005)

<https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cf839d1/iso-22374-2005>

Dentistry — Dental handpieces — Electrical-powered scalers and scaler tips

1 Scope

This International Standard specifies requirements and test methods for electrical-powered scalers and scaler tips, including piezo, ferrostrictive and magnetostrictive type ultrasonic scalers, operated as stand-alone items or connected to dental units, for use on patients. It also contains specifications on manufacturers' instructions, marking and packaging.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 7494-1, *Dentistry — Dental units — Part 1: General requirements and test methods*

ISO 9687, *Dental equipment — Graphical symbols*

ISO 15223, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied*

ISO 17664, *Sterilization of medical devices — Information to be provided by the manufacturer for the processing of resterilizable medical devices*

IEC 60601-1:1988, *Medical electrical equipment — Part 1: General requirements for safety*

IEC 61012, *Filters for the measurement of audible sound in the presence of ultrasound*

IEC 61672-1, *Electroacoustics — Sound level meters — Part 1: Specifications*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

3.1

electrical powered scaler

instrument with a oscillating scaler tip used in dental treatment consisting of a combination of a dental unit, a handpiece, the tubing connecting the handpiece to the dental unit, and an interchangeable scaler tip

NOTE The system also includes a connection to a supply of cooling liquid.

**3.2
scaler insert system**

interchangeable instrument used in an electrical-powered scaler and consisting of a working part and a transducer for dental scaling

**3.3
scaler tip**

fixed or interchangeable instrument used in an electrical-powered scaler and consisting of a shaft and a working part for dental procedures

**3.4
operating area of the scaler tip**

area for use, as described by the manufacturer for different scaler tips

4 Requirements

4.1 General design of handpiece connection

The configuration, dimensions and tolerances of hose connections, tested in accordance with 6.2, shall be in accordance with the manufacturer's instructions for use.

4.2 Scaler tip

4.2.1 Extraction force

iTeh STANDARD PREVIEW
(standards.iteh.ai)

When installed in accordance with the manufacturer's instructions for use, the scaler tip, tested in accordance with 6.3.1, shall withstand, without displacement, a minimum axial tension force of 20 N.

[ISO 22374:2005](https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005)

4.2.2 Torque (for screw-in-tips only)

<https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005>

The following requirement is only applicable for screw-in-tips.

When installed in accordance with the manufacturer's instructions, the scaler tip, tested in accordance with 6.3.2, shall withstand, without rotation displacement, a minimum torque of 200 N·mm.

4.2.3 Insertion of the scaler tip

The force required to insert and secure the scaler tip in the handpiece shall not exceed 30 N when tested in accordance with 6.3.3.

4.3 Performance

4.3.1 Frequency

When operated at the settings recommended by the manufacturer, the frequency of the scaler tip, tested in accordance with 6.4, shall be between 18 000 Hz and 60 000 Hz.

4.3.2 Amplitude, unloaded scaler tip

When operated at the maximum power recommended by the manufacturer, the maximum unloaded peak to peak excursion of the scaler tip, tested in accordance with 6.5, shall not exceed 200 µm.

4.3.3 Amplitude, loaded scaler tip

When operated with the maximum power recommended by the manufacturer, in a direction perpendicular to the plane of vibration (or to the vibration direction) and under a load of 1 N, the maximum amplitude of the working tip, tested in accordance with 6.6, shall not exceed 200 μm .

4.4 Supply of cooling liquid

When operated at the maximum power recommended by the manufacturer, the liquid cooling capability to the operating area of the scaler tip, tested in accordance with 6.7, shall be less than 50 ml/min.

4.5 Noise level

When operated at the maximum power recommended by the manufacturer, the A-weighted sound pressure value generated, tested in accordance with 6.8, shall not exceed 70 dB.

4.6 Resistance to sterilization

Electrically powered scaler handpieces and scaler tips shall be capable of withstanding a minimum of 250 sterilization cycles as defined in the manufacturer's instructions for use without deterioration in appearance or performance.

Single use handpieces or the disposable (non-reusable) parts of other handpieces, tested in accordance with 6.9, shall be supplied sterile or be capable of withstanding one sterilization cycle, as defined in the manufacturer's instructions, without deterioration in appearance or performance.

4.7 Energy for light source (if applicable)

The voltage of the light source in the handpiece, tested in accordance with 6.10, shall not exceed a nominal value of 25 V a.c. or 60 V d.c. on the transformer or converter, between conductors in an earth-free circuit which is isolated from the supply main by a safety transformer or by a device with an equivalent separation.

4.8 Electrical power supply

This shall be as specified by the manufacturer and complying with ISO 7494-1.

5 Sampling

At least one handpiece or insert (scaler tip type) for each model series shall be tested for compliance with this International Standard.

6 Test methods

6.1 General

All tests described in this International Standard are type tests.

6.2 Visual inspection

Visual inspection shall be carried out at normal visual acuity without magnification.

6.3 Scaler tip

6.3.1 Extraction force

6.3.1.1 Apparatus

6.3.1.1.1 **Force gauge**, with an accuracy of $\pm 0,5$ N, to measure the extraction force.

6.3.1.2 Procedure

Install the scaler tip in the handpiece in accordance with the manufacturer's instructions. Operate the handpiece at the recommended liquid flow rate and maximum frequency for at least 1 min and then switch off. Adjust the force gauge to register the maximum force exerted. Apply the device and record the required force to extract the moving scaler tip.

6.3.2 Torque (for screw-in-tips only)

6.3.2.1 Apparatus

6.3.2.1.1 **Torque watch or dynamometer**, capable of measuring the torque in (N·mm) to an accuracy of ± 10 %.

6.3.2.2 Procedure

Install the scaler tip in the handpiece in accordance with the manufacturer's instructions. Operate the handpiece at the recommended maximum liquid flow rate and maximum frequency for at least 1 min and then switch off. Adjust the measuring device to register the maximum torque exerted. Apply the device and record the required torque to unlock the scaler tip from the electrical-powered scaler.

<https://standards.itech.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005>

6.3.3 Insertion of the scaler tip

6.3.3.1 Apparatus

6.3.3.1.1 **Force gauge**, with an accuracy of $\pm 0,5$ N, to measure the insertion force.

6.3.3.2 Procedure

Install the scaler tip in the handpiece under application of the device in accordance with the manufacturer's instructions. Record the required force to lock the scaler tip in the handpiece.

6.4 Frequency

6.4.1 Apparatus

6.4.1.1 **A non-contacting frequency measurement device**, with an electronic frequency counter or a calibrated time base and operating with an accuracy of ± 10 % of the measured value.

6.4.2 Procedure

Install the scaler tip in the handpiece in accordance with the manufacturer's instructions. Operate the scaler tip at the recommended maximum liquid flow rate and maximum power for at least 1 min without any applied load. Measure the frequency of the scaler tip at the operating area.

6.5 Amplitude, unloaded scaler tip

6.5.1 Apparatus

6.5.1.1 A non-contacting optical or electrical length measurement device, with an accuracy of $\pm 10\%$ of the measured value.

6.5.2 Procedure

Install the scaler tip in the handpiece in accordance with the manufacturer's instructions. Operate the handpiece at the maximum power recommended by the manufacturer with or without cooling liquid and without any applied load for 1 min. Measure the peak to peak distance of the scaler tip in the time range between 5 s and 10 s after using the scaler tip in all directions. Record the measured amplitude of the moving scaler tip.

6.6 Amplitude, loaded scaler tip

6.6.1 Apparatus

6.6.1.1 A non-contacting optical or electrical length measurement device, with an accuracy of $\pm 10\%$ of the measured value.

6.6.1.2 Flat, smooth glass surface, 50 mm \times 50 mm, 2 mm thick with the top surface coloured.

NOTE The colouring may be achieved with the use of a permanent marker pen.

6.6.1.3 Microscope, with a magnification of at least $\times 100$ and a calibrated eyepiece reticule or micrometer.

[ISO 22374:2005](https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005)

6.6.2 Procedure <https://standards.iteh.ai/catalog/standards/sist/afe30661-8cb0-4704-80d6-ba48cfa839d1/iso-22374-2005>

Press the handpiece laterally, i.e. vertically to the plane of vibration or vibration direction, with a load of 1 N on the coloured glass surface (registering plane). Only the end of the working tip may touch the glass top. See Figure 1.

NOTE A deviation of the direction of maximum 10° (from 0° up to 10°) to the registering plane (glass top) is permitted in order to simplify the measurement.

Move the tip on the coloured glass surface or move the glass top under the tip in a direction parallel to the registering level and perpendicular to the vibration direction so that the track of the tip is recorded.

Measure the amplitude of the track with and without power supply to the working tip.