

---

---

**Milk — Determination of urea content —  
Enzymatic method using difference in pH  
(Reference method)**

*Lait — Détermination de la teneur en urée — Méthode enzymatique  
utilisant les fluctuations du pH (Méthode de référence)*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 14637:2004](https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004)

<https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004>



Reference numbers  
ISO 14637:2004(E)  
IDF 195:2004(E)

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. Neither the ISO Central Secretariat nor the IDF accepts any liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies and IDF national committees. In the unlikely event that a problem relating to it is found, please inform the ISO Central Secretariat at the address given below.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[ISO 14637:2004](https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004)

<https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004>

© ISO and IDF 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO or IDF at the respective address below.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

International Dairy Federation  
Diamant Building • Boulevard Auguste Reyers 80 • B-1030 Brussels  
Tel. + 32 2 733 98 88  
Fax + 32 2 733 04 13  
E-mail [info@fil-idf.org](mailto:info@fil-idf.org)  
Web [www.fil-idf.org](http://www.fil-idf.org)

Published in Switzerland

## Contents

Page

Foreword.....	iv
<b>1 Scope.....</b>	<b>1</b>
<b>2 Terms and definitions.....</b>	<b>1</b>
<b>3 Principle.....</b>	<b>1</b>
<b>4 Reagents.....</b>	<b>1</b>
<b>5 Apparatus.....</b>	<b>2</b>
<b>6 Sampling.....</b>	<b>3</b>
<b>7 Preparation of test sample.....</b>	<b>3</b>
<b>8 Procedure.....</b>	<b>3</b>
8.1 General.....	3
8.2 Blank determination.....	3
8.3 Calibration.....	4
8.4 Checking the calibration.....	5
8.5 Determination.....	5
8.6 Checking the stability.....	5
8.7 Checking the contamination of the electrodes.....	5
8.8 Cleaning procedure.....	5
<b>9 Maintenance of the electrodes.....</b>	<b>5</b>
9.1 Regeneration.....	5
9.2 Strong regeneration.....	6
<b>10 Calculation and expression of results.....</b>	<b>6</b>
10.1 Calculation.....	6
10.2 Expression of results.....	6
<b>11 Precision.....</b>	<b>6</b>
11.1 Interlaboratory tests.....	6
11.2 Repeatability.....	6
11.3 Reproducibility.....	7
<b>12 Test report.....</b>	<b>7</b>
<b>Annex A (informative) Differential pH apparatus.....</b>	<b>8</b>
<b>Annex B (informative) Interlaboratory tests.....</b>	<b>9</b>
<b>Bibliography.....</b>	<b>11</b>

## Foreword

**ISO (the International Organization for Standardization)** is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committee are circulated to the member bodies for voting. Publication as an International standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 14637|IDF 195 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

(standards.iteh.ai)

[ISO 14637:2004](https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1ed5b2bd5ba/iso-14637-2004)

<https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1ed5b2bd5ba/iso-14637-2004>

## Foreword

**IDF (the International Dairy Federation)** is a worldwide federation of the dairy sector with a National Committee in every member country. Every National Committee has the right to be represented on the IDF Standing Committees carrying out the technical work. IDF collaborates with ISO and AOAC International in the development of standard methods of analysis and sampling for milk and milk products.

Draft International Standards adopted by the Action Teams and Standing Committees are circulated to the National Committees for voting. Publication as an International Standard requires approval by at least 50% of IDF National Committees casting a vote.

International Standard ISO 14637|IDF 195 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 5, *Milk and milk products*, and the International Dairy Federation (IDF), in collaboration with AOAC International. It is being published jointly by ISO and IDF and separately by AOAC International.

All work was carried out by the Joint ISO/IDF/AOAC Action Team, *Nitrogen compounds*, of the Standing Committee on *Main components of milk*, under the aegis of its project leader, Mr Ph. Trossat (FR).

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 14637:2004](https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004)

<https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 14637:2004

<https://standards.iteh.ai/catalog/standards/sist/355edd7c-f16d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004>

# Milk — Determination of urea content — Enzymatic method using difference in pH (Reference method)

## 1 Scope

This International Standard specifies an enzymatic method for the determination of the urea content of milk by measurement of the difference in pH.

## 2 Terms and definitions

For the purpose of this document, the following terms and definitions apply.

### 2.1

#### urea content

mass fraction of substances determined by the procedure specified in this International Standard

NOTE The urea content is expressed in milligrams per litre.

## 3 Principle

[ISO 14637:2004  
https://standards.itech.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004](https://standards.itech.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004)

Urease is added to the test sample to split urea into ammonia and carbon dioxide. At pH 6,7, ammonia immediately hydrolyses thereby releasing hydroxyl ions, and carbon dioxide liberates protons that partly neutralize these hydroxyl ions. The balance between the ammonia and carbon dioxide hydrolysis and the resulting neutralization induces a change in pH. The pH change varies as a function of the urea content of the sample and is measured by using a differential pH analyser.

## 4 Reagents

Use only reagents of recognized analytical grade, unless otherwise specified, and distilled or demineralized water or water of equivalent purity.

### 4.1 Reagents for urea determination.

#### 4.1.1 Buffer solution, pH 6,7.

Dissolve 1,777 g of potassium monohydrogenphosphate ( $K_2HPO_4$ ), 1,388 g of potassium dihydrogenphosphate ( $KH_2PO_4$ ), 7,600 g of potassium chloride (KCl), 1,00 g of sodium azide ( $NaN_3$ ), 0,010 g of acetazolamide (5-acetamido-1,3,4-thiadiazole-2-sulfonamide), 1,040 g of magnesium chloride hexahydrate ( $MgCl_2 \cdot 6H_2O$ ), 2 g of Triton X100, 1 g of Brij 35 and 20 ml of LM<sup>1)</sup> in a 1 000 ml volumetric flask (5.5). Dilute to the mark with water and mix.

1) This detergent is available from Valetudo S.r.l., BG, Italy, and is an example of a suitable product available commercially. This information is given for the convenience of users of this International Standard and does not constitute an endorsement by ISO or IDF of this product.

The buffer solution may be kept for 6 months if stored at 4 °C.

#### 4.1.2 Urease enzyme solution.

Dissolve 360 mg of lyophilized urease (EC 3.5.1.5) in 1 ml of a 50 % (volume fraction) aqueous solution of glycerol. The activity of the obtained urease enzyme solution shall be 2 100 units/ml  $\pm$  300 units/ml<sup>2</sup>).

The urease enzyme solution may be kept for 6 months if stored at 4 °C.

#### 4.1.3 Urea standard solution.

Dissolve 1,000 g of dry urea (N<sub>2</sub>H<sub>4</sub>CO) (dried under vacuum in an oven at 90 °C  $\pm$  1 °C for 1 day), 7,45 g of potassium chloride (KCl) and 1,0 g of sodium azide (NaN<sub>3</sub>) in a 1 000 ml volumetric flask (5.5). Dilute to the mark with water and mix.

The urea standard solution may be kept for 6 months if stored at 4 °C.

#### 4.1.4 Zero milk.

Add 20  $\mu$ l of urease solution (4.1.2) to 1 ml of raw milk. Mix and incubate the thus-prepared raw milk for 10 min in the water bath (5.3) set at 40 °C.

### 4.2 Reagents for cleaning and maintenance of electrodes.

#### 4.2.1 Cleaning solution.

Dissolve 1,742 g of potassium monohydrogenphosphate (K<sub>2</sub>HPO<sub>4</sub>), 1,361 g of potassium dihydrogenphosphate (KH<sub>2</sub>PO<sub>4</sub>), 7,455 g of potassium chloride (KCl), 1,00 g of sodium azide (NaN<sub>3</sub>), 2 g of Triton X100, 2 g of Brij 35 and 3 g of LM1<sup>1</sup>) in a 1 000 ml volumetric flask (5.5). Dilute to the mark with water and mix.

The cleaning solution may be kept for 1 year if stored at room temperature.

#### 4.2.2 Regenerating solution.

Use hydrochloric acid of concentration,  $c(\text{HCl}) = 0,1 \text{ mol/l}$ .

The regenerating solution may be kept for 1 year if stored at room temperature.

#### 4.2.3 Strong regenerating solution.

Dissolve 30 g of nitric acid (HNO<sub>3</sub>) with a mass fraction of approximately 69 %, 30 g of hydrochloric acid (HCl) with a mass fraction of approximately 37 %, 30 g of sodium fluoride (NaF) and 1 g of Triton X100 in a 1 000 ml volumetric flask (5.5). Dilute to the mark with water and mix.

The strong regenerating solution may be kept for 1 year if stored at room temperature.

## 5 Apparatus

Usual laboratory equipment and, in particular, the following.

### 5.1 Analytical balance, capable of weighing to the nearest 1 mg.

---

2) This unit (often called the International Unit or Standard Unit) is defined as the amount of enzyme which will catalyse the transformation of one micromole of substrate per minute under standard conditions.

**5.2 Micropipettes** (positive displacement), of capacities 15 µl and 20 µl.

**5.3 Water bath**, capable of being maintained at 38 °C ± 1 °C and at 40 °C ± 1 °C.

**5.4 Differential pH apparatus**, generally operating according to the scheme shown in Annex A.

The arrangement and components used may be different.

The differential pH apparatus consists of peristaltic pumps to circulate liquids, a mixing chamber, two glass capillary flow-through electrodes (EL1 and EL2) and an electronic system for measurement.

**5.5 One-mark volumetric flasks**, of capacity 1 000 ml.

## 6 Sampling

It is important that the laboratory receive a sample which is truly representative and has not been damaged or changed during transport or storage.

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 707.

## 7 Preparation of test sample

Heat the test sample in the water bath (5.3), set at 38 °C, to that temperature while mixing. Cool to 20 °C just before the preparation of the test portion.

## 8 Procedure

ISO 14637:2004  
<https://standards.iteh.ai/catalog/standards/sist/355edd7c-fl6d-4cd3-8116-1edf5b2bd5ba/iso-14637-2004>

### 8.1 General

Since various types of differential pH apparatus (5.4) available differ in design and handling, the operator shall carefully follow the instrument manufacturer's instructions for setting up, calibration and operation of the instrument. Switch the instrument on and allow its operating conditions to stabilize.

If the time between two consecutive measurements is 5 min or more, renew the buffer solution (4.1.1) in the mixing chamber of the apparatus.

### 8.2 Blank determination

Fill the flow-through electrodes, EL1 and EL2, of the pH apparatus (5.4) with buffer solution (4.1.1). Measure the offset differential pH ( $D_1$ ) between the electrodes. The difference between the two electrodes shall be between the limits ± 150 mpH (millipH) units.

Using a micropipette (5.2), add 15 µl of urease enzyme solution (4.1.2) to the mixing chamber of the apparatus and mix. Only fill the flow-through electrode EL2 with the buffer/enzyme mixture, Measure again the offset differential pH ( $D_2$ ) between the two electrodes.

Calculate the difference in pH for the blank,  $\Delta H_0$ , by using the following equation

$$\Delta H_0 = D_2 - D_1$$

where

$\Delta H_0$  is the difference in pH units between the two offset differential pH measurements,  $D_1$  and  $D_2$ , for the blank determination;