



INTERNATIONAL STANDARD ISO 11143:1999 TECHNICAL CORRIGENDUM 1

Published 2002-06-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Dental equipment — Amalgam separators

TECHNICAL CORRIGENDUM 1

Matériel dentaire — Séparateurs d'amalgame

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 11143:1999 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 6, *Dental equipment*. The corrections concern in particular minimum and maximum parameters.

STANDARD PREVIEW
(standards.iteh.ai)
<https://standards.iteh.ai/catalog/standards/sist/5c885f8e-1e27-47c2-a5ef-f145dfbeb1bb/iso-11143-1999-cor-1-2002>

Page 3, subclause 5.1

Add the following sentence at the end of the existing text:

“If the efficiency depends on different operations or supply parameters, the test shall also be performed taking into consideration the minimum and maximum values (parameters) as given by the manufacturers.”

Page 10, subclause 9.3.2.3

Add the following sentence at the end of the existing text:

“If the resulting water flowrate is lower than 0,5 l/min, then set the water flowrate to 0,5 l/min.

For additional testing see 5.1, second sentence.”

Page 10, subclause 9.3.2.5

Add the following sentence at the end of the existing text:

“For additional testing see 5.1, second sentence.”