



SLOVENSKI STANDARD
oSIST prEN 13205-1:2012

01-november-2012

Zrak na delovnem mestu - Ocenjevanje lastnosti merilnikov za merjenje koncentracij lebdečih delcev - 1. del: Splošne zahteve

Workplace exposure - Assessment of sampler performance for measurement of airborne particle concentrations - Part 1: General requirements

Exposition am Arbeitsplatz - Bewertung der Leistungsfähigkeit von Sammlern für die Messung der Konzentration luftgetragener Partikel - Teil 1: Allgemeine Anforderungen

[SIST EN 13205-1:2014](https://standards.iteh.ai/catalog/standards/sist/d9b0e5a2-2d18-4385-ae1f-2a1e50cc3849/sist-en-13205-1-2014)

[https://standards.iteh.ai/catalog/standards/sist/d9b0e5a2-2d18-4385-ae1f-](https://standards.iteh.ai/catalog/standards/sist/d9b0e5a2-2d18-4385-ae1f-2a1e50cc3849/sist-en-13205-1-2014)

[2a1e50cc3849/sist-en-13205-1-2014](https://standards.iteh.ai/catalog/standards/sist/d9b0e5a2-2d18-4385-ae1f-2a1e50cc3849/sist-en-13205-1-2014)

Ta slovenski standard je istoveten z: prEN 13205-1

ICS:

13.040.30 Kakovost zraka na delovnem mestu Workplace atmospheres
mestu

oSIST prEN 13205-1:2012

en,fr,de

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

DRAFT
prEN 13205-1

October 2012

ICS 13.040.30

Will supersede EN 13205:2001

English Version

Workplace exposure - Assessment of sampler performance for measurement of airborne particle concentrations - Part 1: General requirements

Exposition am Arbeitsplatz - Beurteilung der
Leistungsfähigkeit von Sammlern für die Messung der
Konzentration luftgetragener Partikel - Teil 1: Allgemeine
Anforderungen

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 137.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

2a1e50cc3849/sist-en-13205-1-2014

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	6
4 Symbols and abbreviations	10
5 Requirements	12
5.1 Summary of requirements	12
5.2 Expanded uncertainty for an aerosol sampler	13
5.3 Expanded uncertainty for a measuring procedure	14
6 Test methods.....	14
6.1 General.....	14
6.2 Critical review in order to delimit the performance test	17
6.3 Overview of test methods	18
7 Types of evaluation	19
7.1 Sampler evaluation	19
7.2 Evaluation of a measuring procedure	19
8 Instructions for use	19
9 Marking, quality control	20
9.1 Marking	20
9.2 Quality control.....	20
Annex A (normative) Calculation of expanded uncertainty for a measuring procedure.....	21
Bibliography	29

Foreword

This document (prEN 13205-1:2012) has been prepared by Technical Committee CEN/TC 137 "Assessment of workplace exposure to chemical and biological agents", the secretariat of which is held by DIN.

This document is currently submitted to the CEN Enquiry.

This document together with prEN 13205-2, FprCEN/TR 13205-3, prEN 13205-4, prEN 13205-5 and prEN 13205-6 supersedes prEN 13205:2010 and will supersede EN 13205:2001.

This part 1 of EN 13205 *Workplace exposure – Assessment of sampler performance for measurement of airborne particle concentrations* belongs to a series of documents, the other parts of which are the following:

- *Part 2: Laboratory performance test based on determination of sampling efficiency;*
- *Part 3: Analysis of sampling efficiency data;*
- *Part 4: Laboratory performance test based on comparison of concentrations;*
- *Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces;*
- *Part 6: Transport and handling tests.*

iteh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 13205-1:2014](https://standards.iteh.ai/catalog/standards/sist/d9b0e5a2-2d18-4385-aef1-2a1e50cc3849/sist-en-13205-1-2014)

<https://standards.iteh.ai/catalog/standards/sist/d9b0e5a2-2d18-4385-aef1-2a1e50cc3849/sist-en-13205-1-2014>

Introduction

EN 481 defines sampling conventions for the particle size fractions to be collected from workplace atmospheres in order to assess their impact on human health. Conventions are defined for the inhalable, thoracic and respirable aerosol fractions. These conventions represent target specifications for aerosol samplers, giving the ideal sampling efficiency as a function of particle aerodynamic diameter. In general, the sampling efficiency of real aerosol samplers will deviate from the target specification, and the aerosol mass collected will therefore differ from that which an ideal sampler would collect. In addition, the behaviour of real samplers is influenced by many factors such as external wind speed. In many cases there is an interaction between the influence factors and fraction of the airborne size distribution of the environment in which the sampler is used.

EN 482 contains general performance requirements for methods used for determining the concentrations of chemical agents in workplace atmospheres. These performance requirements include maximum values of expanded uncertainty (a combination of random and non-random measurement uncertainty) achievable under prescribed laboratory conditions for the methods to be used. The requirements of EN 482 apply to a complete measuring procedure, a combination of the stages consisting of sampling, sample transport/storage and sample preparation/analysis.

This part of EN 13205 gives performance requirements for samplers for the inhalable, thoracic or respirable aerosol fractions. Requirements for the aerosol sampler and transport of loaded collection samplers are stated. Furthermore, the method for calculating the expanded uncertainty for a measuring procedure based on aerosol sampling is described.

Different test procedures and types of evaluation are described in the other parts of EN 13205 in order to enable application of EN 13205 to a wide variety of instruments. In detail, three different performance tests for sampled concentration and a transport test of loaded collection substrates are described. The three tests differ in the amount of information obtained by the test and its corresponding cost. The first test method determines the sampling efficiency curve of a candidate sampler, the second compares concentrations sampled from three laboratory test atmospheres by a candidate sampler and a (previously) validated sampler, and the third method compares concentrations sampled from a specific workplace by a candidate sampler and a (previously) validated sampler. Additionally a method for determining equivalence between aerosol samplers at specific workplaces and an alternative handling test are presented.

EN 13205 enables manufacturers and users of aerosol sampling instruments to adopt a consistent approach to sampler validation, and provide a framework for the assessment of sampler performance with respect to EN 481 and EN 482.

It is the responsibility of the manufacturer of aerosol samplers to inform the user of the sampler performance under the laboratory conditions¹⁾ specified in other parts of this European Standard. It is the responsibility of the user to ensure that the actual conditions of intended use are within what the manufacturer specifies as acceptable conditions according to the performance test.

¹⁾ The inhalable convention is undefined for particle sizes in excess of 100 µm or for wind speeds greater than 4 m/s. The tests required to assess performance are therefore limited to these conditions. Should such large particle sizes or wind speeds actually exist at the time of sampling, it is possible that different samplers meeting this part of EN 13205 give different results.

1 Scope

This European Standard specifies performance requirements that are specific to aerosol samplers, primarily inhalable, thoracic and respirable aerosol samplers. These performance requirements, which include conformity with the EN 481 sampling conventions, are applicable only to the process of sampling the airborne particles from the air, not to the process of analysing particles collected by the process of sampling. Although analysis of samples collected in the course of testing is usually necessary in order to evaluate the sampler performance, the specified test methods ensure that analytical errors are kept very low during testing and do not contribute significantly to the end result.

This part of EN 13205 specifies how the performance of aerosol measuring procedures is assessed with respect to the general requirements of EN 482, through the combination of errors arising in the sampling, sample transportation/storage and sample preparation/analysis stages.

This part of EN 13205 is applicable to all samplers used for the health-related sampling of particles in workplace air.

This part of EN 13205 is not applicable to the determination of analytical errors and factors related to them (for example the bias, precision and limit of detection of the analytical method). Where the aerosol sampler requires the use of an external (rather than integral) pump, the pump is not subject to the requirements of this part of EN 13205.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 481, *Workplace atmospheres — Size fraction definitions for measurement of airborne particles*

EN 482:2012, *Workplace exposure — General requirements for the performance of procedures for the measurement of chemical agents*

EN 1232, *Workplace atmospheres — Requirements and test methods for pumps used for personal sampling of chemical agents in the workplace*

EN 1540:2011, *Workplace exposure — Terminology*

EN 12919, *Workplace atmospheres — Pumps for the sampling of chemical agents with a volume flow rate of over 5 l/min — Requirements and test methods*

prEN 13205-2:2012, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations – Part 2: Laboratory performance test based on determination of sampling efficiency*

prEN 13205-4:2012, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations – Part 4: Laboratory performance test based on comparison of concentrations*

prEN 13205-5:2012, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations – Part 5: Aerosol sampler performance test and sampler comparison carried out at workplaces*

prEN 13205-6:2012, *Workplace exposure — Assessment of sampler performance for measurement of airborne particle concentrations – Part 6: Transport and handling tests*

prEN 13205-1:2012 (E)

EN 13890, *Workplace exposure — Procedures for measuring metals and metalloids in airborne particles — Requirements and test methods*

EN 14530, *Workplace atmospheres — Determination of diesel particulate matter — General requirements*

ISO 15767, *Workplace atmospheres — Controlling and characterising uncertainty in weighing collected aerosols*

ISO 21438-1, *Workplace atmospheres — Determination of inorganic acids by ion chromatography — Part 1: Non-volatile acids (sulfuric acid and phosphoric acid)*

ISO 21438-3, *Workplace atmospheres — Determination of inorganic acids by ion chromatography — Part 3: Hydrofluoric acid and particulate fluorides*

ISO 24095, *Workplace air — Guidance for the measurement of respirable crystalline silica*

3 Terms and definitions

For the purpose of this document, the term and definitions given in EN 1540 and the following apply.

3.1 Terms related to sampling and transportation

Note 1 to entry: In addition to the terms and definitions given by entry numbers 3.1.1 to 3.1.22, in particular, the following general terms, terms related to the physical and chemical process of air sampling and terms related to the analytical method of EN 1540 are used in this document as well: respirable fraction, sampling efficiency, thoracic fraction, measuring procedure, analysis, analytical method, measurand and occupational exposure limit value.

3.1.1**airborne particles**

fine matter, in solid or liquid form, dispersed in air

Note 1 to entry: Smoke, fume, mist and fog consist of airborne particles.

Note 2 to entry: For the purpose of this document, the term comprises all particles surrounded by air with a terminal settling speed less than a critical value, whereas the critical value will depend on the application. The idea is to exclude particles that are too large for a specific application.

Note 3 to entry: For the purpose of this document, particles surrounded by air but with a terminal velocity exceeding a critical value are not to be considered part of the sample.

[SOURCE: EN 1540:2011, 2.2.3, modified – Notes 2 and 3 to entry have been added.]

3.1.2**aerosol**

airborne particles and the gas (and vapour) mixture in which they are suspended

Note 1 to entry: The airborne particles can be in or out of equilibrium with their own vapours.

Note 2 to entry: In occupational hygiene, the carrier gas is air, possibly contaminated by other gases and vapours.

[SOURCE: EN 1540:2011, 2.2.4, modified – Note 2 to entry has been added.]

3.1.3**aerosol sampler**

(airborne) particle sampler

(airborne) particulate sampler

sampler that is used to transport airborne particles to a collection substrate

Note 1 to entry: The term aerosol sampler is commonly used although it is not in line with the definition of aerosol given in EN 1540:2011, 2.2.4.

Note 2 to entry: The transport can be either active or passive.

Note 3 to entry: For the purpose of this document, a sampler is not a pump or an air mover, but can include either of them in specific cases.

[SOURCE: EN 1540:2011, 3.2.1.5, modified – Note 3 to entry has been added.]

3.1.4

candidate sampler

any aerosol sampler that can be used to collect airborne particles in order to determine their concentration and whose performance is subjected to performance tests

Note 1 to entry: A candidate sampler that meets the performance criteria will be termed a validated sampler.

3.1.5

collected sample

product of the process of air sampling that consists of the collected chemical and/or biological agents only

Note 1 to entry: For the purpose of this document the collected sample comprises of airborne particles collected and retained on the sampling substrate for subsequent analysis.

[SOURCE: EN 1540:2011, 3.1.2, modified – Note 1 to entry has been added.]

3.1.6

collection substrate

sampling substrate

collection medium

sampling medium

medium on which airborne chemical and/or biological agents are collected for subsequent analysis

Note 1 to entry: Filters, polyurethane foams and sampling cassettes are examples of collection substrates for airborne particles.

Note 2 to entry: Activated carbon, silica gel and reagent impregnated filters are examples of collection substrates for gases and vapours.

Note 3 to entry: Agar media are examples of collection substrates for bioaerosols.

Note 4 to entry: As an example of the converse, the 25-mm or 37-mm plastic filter cassette often used for “total dust” sampling (with gravimetric analysis) in either its closed-face or open-face version is not part of the substrate in the definition above, since it is not weighed.

[SOURCE: EN 1540:2011, 3.3.6, modified – Note 4 to entry has been added.]

3.1.7

collection efficiency

efficiency of collection and retention of sampled particles by the collection substrate

Note 1 to entry: The collection efficiency can, for example be influenced by the amount of particles deposited in the collection substrate.

Note 2 to entry: The collection efficiency (of a collection substrate) should not be confused with the sampling efficiency (of a sampler). For the definition of sampling efficiency see EN 1540:2011, 3.3.10.

3.1.8

inhalable fraction

mass fraction of total airborne particles which is inhaled through the nose and mouth

prEN 13205-1:2012 (E)

Note 1 to entry: The inhalable fraction is specified in EN 481.

Note 2 to entry: The inhalable fraction depends on the speed and direction of the air movement, on breathing rate and other factors.

[SOURCE: EN 1540:2011, 2.3.1.1, modified – Note 2 to entry has been added.]

3.1.9**inhalable sampler**

aerosol sampler that is used to collect the inhalable fraction

Note 1 to entry: An inhalable sampler collects the inhalable fraction or airborne particles, as defined in EN 481, with a performance as stipulated in this document.

[SOURCE: EN 1540:2011, 3.2.1.5.1, modified – Note 1 to entry has been added.]

3.1.10**nominal flow rate**

design flow rate recommended by the sampler manufacturer or measuring procedure

3.1.11**particle aerodynamic diameter**

diameter of a sphere of 1 g cm^{-3} density with the same terminal settling velocity in calm air as the particle, under the prevailing conditions of temperature, pressure and relative humidity

Note 1 to entry: The particle aerodynamic diameter depends on the size, density and shape of the particle.

Note 2 to entry: For particles of aerodynamic diameter less than $0,5 \mu\text{m}$, the particle thermodynamic diameter should be used instead of the particle aerodynamic diameter.

[SOURCE: EN 1540:2011, 2.3.2, modified – Note 2 to entry has been added.]

3.1.12**sampler inlet efficiency**

for each particle aerodynamic diameter the ratio of aerosol concentration passing through the sampler inlet system to the corresponding total airborne particle concentration

Note 1 to entry: The sampler inlet efficiency is the product of the aspiration efficiency, which characterises the aerodynamic behaviour of the sampler orifice, and the size-dependent effects of particle bounce and losses both inside and outside the inlet. The inlet losses can, for some samplers, also depend on external factors such as wind speed and aerosol size distribution.

3.1.13**penetration**

internal penetration

for each particle aerodynamic diameter the ratio of the sampling efficiency to the sampler inlet efficiency

Note 1 to entry: The penetration describes the efficiency with which particles pass through the stage of inertial separation, as for example foams, cyclones or impactors.

3.1.14**personal sampler**

sampler, attached to a person, that collects gases, vapours or airborne particles in the breathing zone to determine exposure to chemical and/or biological agents

Note 1 to entry: For the purpose of this document “agent” means airborne particles.

[SOURCE: EN 1540:2011, 3.2.2, modified – Note 1 to entry has been added.]

3.1.15**respirable sampler**

aerosol sampler that is used to collect the respirable fraction

Note 1 to entry: A respirable sampler collects the respirable fraction or airborne particles, as defined in EN 481, with a performance as stipulated in this document.

[SOURCE: EN 1540:2011, 3.2.1.5.3 modified – Note 1 to entry has been added.]

3.1.16**sampler specimen**

sampler individual

single individual of a given type of aerosol sampler

3.1.17**sampling cassette**

cassette mounted inside a sampler, designed in such a way that its collection substrate consists of all its interior surfaces (bounding the air-stream with sampled particles), and usually containing a filter or another suitable collection substrate

3.1.18**sampling method**

part of the measuring procedure that describe the overall process of sampling, including sampler preparation and sample transport

3.1.19**sampling process**

physical mechanisms by which particles are selectively aspirated into a sampler inlet, graded by means of inertial or other forces, transported to the collection substrate or to other internal surfaces, or lost from the collection substrate

3.1.20**thoracic sampler**

aerosol sampler that is used to collect the thoracic fraction

Note 1 to entry: A thoracic sampler collects the thoracic fraction or airborne particles, as defined in EN 481, with a performance as stipulated in this document.

[SOURCE: EN 1540:2011, 3.2.1.5.2 modified – Note 1 to entry has been added.]

3.1.21**validated sampler (type A)**

sampler that has previously been tested using the methods described in prEN 13205-2 resulting in a performance as required by prEN13205-1

Note 1 to entry: In a performance test according to prEN 13205-2, the sampling efficiency curve of the candidate sampler will be determined as a function of particle size (and possibly other influencing factors).

3.1.22**validated sampler (type B)**

sampler that has previously been tested using the methods described in prEN 13205-4 resulting in a performance as required by prEN13205-1

Note 1 to entry: In a performance test according prEN 13205-4, the concentration sampled by the candidate sampler will be compared with the concentration sampled by a validated reference sampler, for at least three test aerosols.

prEN 13205-1:2012 (E)**3.1.23****validated sampler (type C)**

sampler that has previously been tested using the methods described in prEN 13205-5 resulting in a performance as required in prEN 13205-1

Note 1 to entry: In a performance test according prEN 13205-5, the concentration sampled by the candidate sampler at a workplace will be compared with the concentration sampled by a validated reference sampler.

3.2 Terms related to performance

Note 1 to entry: In addition to the terms and definitions listed below, in particular, the following terms of EN 1540 related to (method) performance are used in this document as well: random uncertainty, non-random uncertainty, standard uncertainty, combined standard uncertainty, expanded uncertainty, uncertainty (of measurement), coverage factor, measuring range and precision.

3.2.1**bias**

difference between the expectation of a test result or measurement result and a true value

Note 1 to entry: Bias is the total non-random error as contrasted to random error. There can be one or more non-random error components contributing to the bias. A larger systematic difference from the true value is reflected by a larger bias value.

Note 2 to entry: In practice, the accepted reference value is substituted for the true value. The accepted reference value (for definition see ISO 3534-2) can be, for example, the certified value of a reference material, the concentration of a standard test atmosphere or the target value of an interlaboratory comparison.

Note 3 to entry: In prEN 13205-2, FprCEN/TR 13205-3, prEN 13205-4 and prEN 13205-5 the true value of the concentration of a chemical agent in air will be the concentration calculated to be sampled by an ideal sampler with a sampling efficiency identical to the sampling convention or sampled by a validated sampler.

Note 4 to entry: The definition has originally been taken from ISO 3534-2:2006, 3.3.2.

[SOURCE: EN 1540:2011, 5.3.1 modified – Notes 3 and 4 to entry have been added.]

3.2.2**sampler bias**

bias of the sampling method

4 Symbols and abbreviations

NOTE The letter for each Annex in which the symbol is used is shown after each symbol,

4.1 Symbols**4.1.1 Latin**

C_{OEL}	relevant occupational exposure limit, [mg/m^3]
$C_{0,1}$	concentration corresponding to 10% of the relevant C_{OEL} , [mg/m^3]
$C_{0,5}$	concentration corresponding to 50% of the relevant C_{OEL} , [mg/m^3]
C_2	concentration corresponding to 200% of the relevant C_{OEL} , [mg/m^3]
m_{Analysed}	mass of the chemical compound analysed in the collected sample, [mg]