
**Health informatics — Interoperability of
telehealth systems and networks —**

**Part 2:
Real-time systems**

*Informatique de santé — Interopérabilité des systèmes et des réseaux
de télésanté*

iTeh STANDARD PREVIEW
Partie 2: Systèmes en temps réel
(standards.iteh.ai)

[ISO/TR 16056-2:2004](https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004)

<https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/TR 16056-2:2004](https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004)

<https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004>

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction.....	vi
1 Scope	1
2 Normative References.....	1
3 Terms and definitions	2
4 Abbreviations	12
5 Multimedia Conferencing Standards	13
5.1 General	13
5.2 H.320 recommendations	16
5.3 H.321 and H.310 Recommendations	17
5.4 H.323 Recommendations	19
5.4.1 H.323 Terminals	19
5.4.2 Gateways	21
5.4.3 Gatekeeper	21
5.4.4 Multipoint control unit	21
5.5 H.324 Recommendations	22
5.6 T.120 Recommendations	23
5.6.1 T.120 Overview	23
5.6.2 T.120 Layer composition	23
5.6.3 T.120 Interoperability	24
6 Telehealth applications	24
6.1 General	24
6.2 Telelearning	25
6.3 Teleconsultation	26
7 Interoperability issues	28
7.1 General	28
7.2 Standard-related sources of interoperability issues.....	28
7.2.1 Loose adaptation of previous protocols	29
7.2.2 Loosely defined encoding and decoding methods	29
7.2.3 Inconsistent definition of mandatory/optional requirements	29
7.2.4 Gaps in the recommendations.....	29
7.2.5 No specifications for data encoding in the H.320/H.323 Recommendations	30
7.2.6 Evolving T.120 Recommendations	30
7.2.7 Loosely defined specifications for an MCU	30
7.2.8 Evolving/undefined specifications for Gateways and Gatekeepers	30
7.2.9 Loosely defined bandwidth management.....	31
7.2.10 Poor standardization of media conferencing	31
7.3 Product-related Sources of Interoperability Issues	31
7.3.1 Time delay between the standards (ever-evolving) and products	31
7.3.2 Implementation of a different subset of the standards	32
7.3.3 Proprietary solutions	32
7.4 Implementation-related sources of interoperability issues.....	32
7.4.1 Undefined specifications for Gatekeeper Communication.....	32
7.4.2 Gateway to Gatekeeper Communication.....	32
7.4.3 A decentralized multipoint conference	33
8 Interoperability requirements	33
8.1 General	33
8.2 Call Setup	34
8.3 Acquisition, processing, and transmission of multimedia data	35
8.4 Controlling near- and far-end devices	36

8.5	Call termination.....	36
9	Interoperability in heterogeneous telehealth networks	36
9.1	General	36
9.2	Interoperability of H.320 systems over frame relay network.....	36
9.3	Interoperability of h.3xx-compliant systems over heterogeneous networks.....	37
9.3.1	Gatekeeper functions.....	38
9.3.2	MCU functions.....	39
9.3.3	Design requirements – Telehealth solutions	40
10	Framework for interoperable architectures	41
10.1	General	41
10.2	Telehealth component boundaries.....	41
10.2.1	User interface	41
10.2.2	Medical Devices	42
10.2.3	Data manager	42
10.2.4	Processing Manager	42
10.2.5	Communications Manager	43
10.2.6	Resource Manager	43
10.2.7	Integration Manager.....	43
	Bibliography.....	45

List of Figures

Figure 1 - The H.320 Terminal	16
Figure 2 - A multipoint configuration	17
Figure 3 - The H.321 Terminal	18
Figure 4 - The H.323 Terminal	20
Figure 5 - The H.324 Terminal	22
Figure 6 - The T.120 Recommendations	23
Figure 7 - Parties contributing to telehealth interoperability.....	28
Figure 8 - Interoperability dimensions for real-time telehealth services	34
Figure 9 - H.320 video services incorporated into existing data and voice networks over frame relay	37
Figure 10 - Connectivity in a heterogeneous system and network environment.....	38

List of Tables

Table 1 - Summary of the multimedia conferencing standards.....	15
Table 2 - Summary of the H.320 Recommendations.....	17
Table 3 - The H.323 and T.120 protocol stack	20
Table 4 - Summary of the H.323 Recommendations.....	21
Table 5 - Summary of functionality of a telelearning system	25
Table 6 - Summary of functionality of a teleconsultation system	26

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

In exceptional circumstances, when a technical committee has collected data of a different kind from that which is normally published as an International Standard ("state of the art", for example), it may decide by a simple majority vote of its participating members to publish a Technical Report. A Technical Report is entirely informative in nature and does not have to be reviewed until the data it provides are considered to be no longer valid or useful.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO/TR 16056-2 was prepared by Technical Committee ISO/TC 215, *Health informatics*.

ISO/TR 16056 consists of the following parts, under the general title *Health informatics — Interoperability of telehealth systems and networks*.

— *Part 1: Introduction and definitions*

— *Part 2: Real-time systems*

Introduction

Delivery of health care services by means of telehealth is advancing rapidly. Telehealth enables providing these services with the use of information and telecommunications technologies. This includes a broad spectrum of capabilities including acquisition, storage, presentation, and management of patient information (represented in different digital forms such as video, audio, or data), and communication of this information between care facilities with the use of communications links.

Telehealth interactions may be carried out in three ways: **real-time**, **store-and-forward** or with the use **media streaming** methods. While real-time interactions imply that all parties directly participate in the telehealth session, store-and-forward interactions involve sending, reviewing, and returning an opinion over a period of time. Streaming is a method of delivery real-time or stored data such as audio, video, documents, still images, or other data type across networks with a reasonable amount of Quality of Services (QoS). With streaming, a receiving system can start displaying (or playing) the data before the entire content arrives.

Real-time telehealth sessions usually involve **synchronous** data transmission while store-and-forward can usually be regarded as **asynchronous**. Streaming uses time-synchronized streams of continuous media during transmission. However, data presentation uses buffering, if the receiving system receives data more quickly than required. If the data is not received quickly enough, the presentation of the data is interrupted.

Interoperability of telehealth systems and networks is critical in ensuring the telehealth technology serves well the care recipients and providers and meets their expectations. While this requirement is essential to the long-term sustainability of telehealth, interoperability is difficult to achieve. There are many reasons that make telehealth interoperability difficult, however, the following three need urgent addressing: (1) too broad definition of telehealth, (2) lack of standards specifically designed for telehealth, and (3) collaboration between the information technology and telecommunications industries.

There are multiple definitions of telehealth. The services provided by telehealth cover a broad spectrum of activities ranging from videoconferencing through exchange of health information to providing care services in emergency and complex clinical cases. From a technology perspective, the scope of these services is too broad and this makes it difficult to develop telehealth standards and products.

There is no 'official' telehealth standard. The telehealth industry uses high-level health care guidelines and technical standards developed for various technology sectors including multimedia conferencing, information technology, data communications, and security. These guidelines and standards focus on functional and operational requirements and do not address interoperability. To further complicate the problem, all of these standards as well as the telehealth needs and practices are rapidly changing.

Telehealth, more than any other recent development, bridges the boundaries between telecommunications and information technologies. The business goals and attitudes of these two industries are different. Telecommunications industry has a history of regulation, standardization, and control of the customer premises equipment. Interoperability and reliability have been the key factors to growth. The information technology industry (the desktop computing industry in particular) has achieved success through encouraging innovation, diversity, and tremendous cost-efficiency not always paying attention to interoperability aspects of the technology. The marriage of these two cultures and the integration of their respective technologies proved to be challenging.

To address the needs for interoperable telehealth systems and networks, telehealth services must be clearly defined in terms of their scope and interrelationships with other health-related services, a set of telehealth-specific standards must be developed, and subsequently implemented by the respective industries.

This two-part ISO Technical Report addresses interoperability issues in telehealth systems and networks. This document has been structured as follows:

Part 1: Introduction and definitions. Covers an introduction to telehealth and includes the definitions of telehealth, interoperability, and related terms.

Part 2: Real-time systems. Defines the scope of the technical standards related to real-time applications, (including video, audio, and data conferencing), identifies gaps and overlaps in the standards, defines requirements for interoperable telehealth systems and networks, and identifies building blocks for interoperable telehealth solutions.

This Technical Report is to be complemented by two other documents that will cover interoperability of store-and-forward and media streaming telehealth applications.

The target users of these documents are care providers and health care organizations, telehealth equipment vendors and implementers of telehealth solutions, professional organizations, and governments.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO/TR 16056-2:2004](https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004)

<https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO/TR 16056-2:2004](https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004)

<https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004>

Health informatics — Interoperability of telehealth systems and networks — Part 2: Real-time systems

1 SCOPE

This Technical Report entitled *Interoperability of telehealth systems and networks – Part 2: Real-time systems* builds on the introduction to telehealth described in *ISO/IEC TR 16056-1 Health informatics - Interoperability of telehealth systems and networks - Part 1: Introduction and definitions* and focuses on the technical standards related to real-time applications, (including video, audio, and data conferencing) and interoperability aspects of telehealth systems and networks.

Specifically, this document addresses four main areas:

- i) **Standards for real-time telehealth systems.** This Technical Report describes the technical standards related to real-time telehealth applications, including audio, video, and data conferencing capabilities. It also identifies gaps, overlaps and inconsistencies in the standards, and provides some guidance about how they need to evolve.
- ii) **Interoperability issues in telehealth applications.** This Technical Report examines interoperability aspects of real-time multimedia conferencing standards and telehealth products, and identifies areas of concerns from the interoperability perspective that need to be resolved.
- iii) **Requirements for interoperable telehealth systems and networks.** This Technical Report defines interoperability requirements at different levels of interaction between telehealth systems and provides some guidelines on how interoperability can be achieved.
- iv) **Framework for interoperable architectures.** This Technical Report identifies interoperable building blocks for telehealth solutions and interactions between these building blocks, and explores the possibility of standardization of these building blocks.

The scope of the Technical Report does not include conformity and interoperability tests or functional specifications for telehealth systems and networks.

2 NORMATIVE REFERENCES

This Technical Report incorporates by dated or undated reference, provisions from other publications. These normative references are cited in the appropriate places in the text, and the publications are listed hereafter.

For dated references, subsequent amendments and revisions of any of these publications apply to this ISO Technical Report only when incorporated in it by amendment and revision. For undated references, the latest edition of the referenced document (including any amendments) applies.

CEN/TC 251/N99-097 (1999)	<i>Health Informatics - Interoperability of Healthcare Multimedia Report Systems. Final draft CEN Report</i>
ISO/IEC 17000:2004	<i>Conformity assessment – Vocabulary and general principles</i>
ITU-T Recommendation G.711 (1988)	<i>Pulse code modulation (PCM) of voice frequencies.</i>
ITU-T Recommendation G.722 (1993)	<i>7 KHz audio - coding within 64 kbit/s.</i>
ITU-T Recommendation G.728 (1992)	<i>Coding of speech at 16 kbit/s using low-delay code excited linear prediction.</i>
ITU-T Recommendation H.221 (1993)	<i>Frame structure for a 64 to 1920 kbit/s channel in audiovisual teleservices.</i>

ISO TR 16056-2:2004(E)

ITU-T Recommendation H.230 (1997)	<i>Frame-synchronous control and indication signals for audiovisual systems.</i>
ITU-T Recommendation H.242 (1996)	<i>System for establishing communication between audiovisual terminals using digital channels up to 2 Mbit/s.</i>
ITU-T Recommendation H.243 (1997)	<i>Procedures for establishing communication between three or more audiovisual terminals using digital channels up to 1920 kbit/s.</i>
ITU-T Recommendation H.224 (1994)	<i>A real time control protocol for simplex applications using the H.221 LSD/HSD/HLP channels.</i>
ITU-T Recommendation H.281 (1994)	<i>A far end camera control protocol for videoconferences using H.224.</i>
ITU-T Recommendation H.233 (1996)	<i>Confidentiality System for Audiovisual Services.</i>
ITU-T Recommendation H.234 (1996)	<i>Encryption key management and authentication system for audiovisual services.</i>
ITU-T Recommendation H.320 (1996)	<i>Narrow-band visual telephone systems and terminal equipment.</i>
ITU-T Recommendation T.120 (1996)	<i>Data protocols for multimedia conferencing.</i>
ITU-T Recommendation T.121 (1996)	<i>Generic application template.</i>
ITU-T Recommendation T.122 (1993)	<i>Multipoint communication service for audiographics and audiovisual conferencing service definition.</i>
ITU-T Recommendation T.123 (1994)	<i>Protocol stacks for audiographic and audiovisual teleconference applications.</i>
ITU-T Recommendation T.124 (1995)	<i>Generic conference control.</i>
ITU-T Recommendation T.125 (1994)	<i>Multipoint communication service protocol specification.</i>
ITU-T Recommendation T.126 (1995)	<i>Multipoint still image and annotation protocol.</i>
ITU-T Recommendation T.127 (1995)	<i>Multipoint binary file transfer protocol.</i>

3 TERMS AND DEFINITIONS

For the purposes of this Technical Report, the following definitions apply.

3.1

accreditation

third party attestation related to a conformity assessment body conveying formal demonstration of its competence to carry out specific conformity assessment tasks

3.2**A-law**

variant of the G.711 audio encoding used primarily in North America and Japan

NOTE Related terms include μ -law and G.711

3.3**asynchronous transmission**

transmission of individual bytes without time-dependency between the bytes

3.4**audiographics terminal**

terminal that has audio and graphics capabilities, but no video capability

3.5**audiovisual terminal**

terminal that has audio, video, and graphics capabilities

3.6**basic rate interface****BRI**

ISDN service comprising two B (bearer) channels operating at 64 Kbps each and one D (data) channel operating at 16 Kbps

3.7**call**

point-to-point multimedia communication between two H.32x endpoints

3.8**call setup**

process of establishing a group of communication users and includes the initialization of any shared application and other resources which the user may require to be available

3.9**call signalling channel**

reliable channel used to convey call setup messages following Q.931

3.10**call teardown**

process of ending a call and freeing any resources reserved for that call

3.11**centralized multipoint conference**

conference call in which all participating terminals communicate in a point-to-point fashion with an MCU

3.12**certification**

third-party attestation related to products, processes, systems or persons

NOTE 1 Certification of a management system is sometimes also called registration.

NOTE 2 Certification is applicable to all objects of conformity assessment except for conformity assessment bodies themselves, to which accreditation is applicable.

3.13

channel service unit

CSU

interface used to connect a terminal or computer to a digital medium in the same way that a modem is used for connection to an analogue medium

3.14

charge coupled device

CCD

device used in cameras as an optical scanning mechanism.

NOTE It consists of a shift register that stores samples of analog signals. An analog charge is sequentially passed along the device by the action of stepping voltages and stored in potential wells formed under electrodes. The charge is moved from one well to another by the stepping voltages.

3.15

COder/DECoder

COmpression/DECompression

CODEC

hardware and/or software used with interactive video systems that converts an analog signal to digital, then compresses it so that lower bandwidth telecommunications lines can be used

NOTE The signal is decompressed and converted back to analog output by a compatible CODEC at the receiving end. The compression method (algorithm) may be proprietary or standards-based.

3.16

common intermediate format

CIF

ITU-T standard video picture scanning format where information is stored in luminance (brightness) and two color difference (chrominance) components

iTeh STANDARD PREVIEW
(standards.iteh.ai)

NOTE CIF represents 352 pixels/line by 288 lines/image for luminance and 176 pixels/line by 144 lines/image for chrominance. See also QCIF.

<https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004>

3.17

composite video

type of video signal in which all information -- the red, blue, and green signals, and sometimes audio signals as well, are mixed together

NOTE Composite video is used by NTSC-compliant devices (see NTSC Standard).

3.18

conformity assessment

demonstration that specified requirements relating to a product, process, system, person or body are fulfilled

NOTE Conformity to a set of specifications is a prerequisite to interoperability. However, conformity to the specifications alone does not guarantee interoperability of systems.

3.19

data service unit

DSU

device used in digital transmission for connecting a CSU to data terminal equipment (a terminal or computer), in the same way that a modem is used for connection to an analogue medium

Note See also CSU.

3.20

decentralized multipoint conference

conference in which the participating terminals multicast to all other participating terminals without an MCU

3.21**endpoint**

terminal, gateway, or MCU

3.22**G.711**

ITU-T recommendation for the digital representation of speech up to 3.4 KHz of frequency producing a 64 Kbps data stream

NOTE Commonly used in telephone networks. It comes in two variants: A-law and μ -law.

3.23**G.722**

ITU-T recommendation for the digital representation of audio up to 7 KHz of frequency producing a 64 Kbps data stream with a much higher quality than G.711

3.24**G.728**

ITU-T recommendation for the digital representation of audio producing a 16 Kbps data stream producing near-telephone quality audio.

3.25**gatekeeper**

H.323 entity that provides address translation, control access, and sometimes bandwidth management to the LAN for H.323 terminals, gateways, and MCUs

3.26**gateway**

H.323 entity, which provides real-time, two-way communications between H.323 terminals on the LAN and other ITU terminals on a WAN, or to another H.323 gateway

3.27**generic conference call****GCC**

set of conference services described in the ITU-T T.124 Recommendation

3.28**H.221**

ITU-T recommendation defining how to multiplex video and audio into frames using 64-1920 Kbps channels for switched and leased network services, excluding packetized networks

3.29**H.225D**

ITU-T recommendation that specifies messages for call control including signaling, registration and admissions, and packetization/synchronization of media systems

3.30**H.230**

ITU-T recommendation that specifies the frame-synchronous control and indication signals for audiovisual systems

3.31**H.231**

ITU-T recommendation that specifies the multipoint control unit

3.32**H.235**

ITU-T recommendation that defines the security framework used to provide authentication, encryption, and integrity for H.323 systems

3.33

H.242

ITU-T recommendation that specifies how to establish the communication between audiovisual terminals using digital channels with speeds up to 2 Mbps

3.34

H.243

ITU-T recommendation that specifies the establishment of communication between three or more audiovisual terminals using digital channels with speeds up to 2 Mbps

3.35

H.245

ITU-T recommendation that specifies messages for opening and closing channels for media streams, and other commands, requests and indications between two H.323 endpoints

3.36

H.261

ITU-T recommendation that specifies the video encoding and compression algorithm for two video resolutions: 352 x 288 CIF and 176 x 144 QCIF

NOTE H.261 is used in both H.320 and T.120.

3.37

H.263

ITU-T recommendation that specifies a new video codec for video over packet-switched networks or POTS

NOTE H.263 optimizes H.261 for very low bit rate of video coding below 64 Kbps. H.263 provides better motion compensation, more accurate motion vectors, optimized quantization for very low bit rates, and arithmetic coding.

3.38

H.310

family of ITU-T standards ratified in 1995 that describes the technical specifications for adapting narrow-band ISDN visual telephone terminals, as defined in H.320, to broadband ISDN (BISDN) and ATM environments

NOTE H.310 adds the MPEG-2 video-compression algorithm that provides MPEG-2 video quality.

3.39

H.320

family of ITU-T standards, ratified in 1990, that specifies how voice and video conferencing systems communicate over ISDN or leased networks, using a bandwidth from 64 Kbps to 1920 Kbps

3.40

H.323

family of ITU-T standards, ratified in 1996, that extends H.320 to computer networks, including LANs and the Internet

NOTE H.323 supports both point-to-point and multipoint operations. In addition, H.323 shares many components of the H.32x specification, such as the H.261 video codec, the G.711 audio codec, the H.263 video codec, G.722, G.723 and G.728. As a new feature, H.323 specifies a gatekeeper component that allows LAN administrators to manage video traffic for QoS. The H.323 specification also defines a LAN/H.320 gateway that permits a H.323 node to interoperate with H.320/H.324 terminals.

3.41

H.324

family of ITU-T standards, ratified in 1996, that allows video conferencing over standard analog phone lines with features similar to those in H.320

NOTE The H.324 standard uses H.263, which contains a better codec for POTS than H.261. H.263 is an improved version of H.261 that adds a 128 x 96 sub-QCIF (SQCIF) format. By using a 28.8 or 36.6 Kbps modem, H.263 may produce frame rates approaching those achieved by H.320 systems over ISDN.

3.42**interoperability**

the ability of two or more systems (computers, communication devices, networks, software, and other information technology components) to interact with one another and exchange information according to a prescribed method in order to achieve predictable results

3.43**interoperability testing**

an assessment of the ability of two or more systems to interact with one another and exchange usable electronic data

NOTE As conformity to the specifications alone does not guarantee interoperability of systems, interoperability testing is required to assess the ability of two or more systems to interact with one another and exchange usable electronic data. Interoperability testing does not include assessment of performance, robustness or reliability nor does it measure the conformity of an implementation. Two systems can be interoperable but still not compliant to the standard or specification.

3.44**μ-law**

variant of the G.711 audio encoding used primarily in North America and Japan

NOTE See also G.711 and A-law.

3.45**multipoint control unit****MCU**

endpoint on the LAN which enables three or more terminals and gateways to participate in a multipoint conference

NOTE The MCU includes a mandatory MC and optional MPs.

3.46**multipoint controller
MC**

An entity that provides for the control of three or more terminals in a multipoint conference

[ISO/TR 16056-2:2004](https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004)

<https://standards.iteh.ai/catalog/standards/sist/ccd856a0-53c2-4d53-91fd-584c96d4a064/iso-tr-16056-2-2004>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

3.47**multipoint processor****MP**

entity that provides for the processing of audio, video, and/or data streams in a multipoint conference

NOTE The MP provides for the mixing, switching, or other processing of media streams under the control of the MC.

3.48**multipoint conference**

conference between three or more terminals, which may be on the LAN or on the circuit switched network

3.49**NTSC Standard**

standard for television broadcasting established by the National Television Standards Committee (NTSC)

NOTE Used in North America, Japan and some other countries. NTSC format: Lines / frame: 525; Frames per second (fps): 30; Interlace ratio: 2:1; Aspect ratio: 4:3; Color matrix equation: $Y = 0.3 * R + 0.59 * G + 0.11 * B$; $I = 0.6 * R - 0.28 * G - 0.32 * B$; $Q = 0.21 * R - 0.52 * G + 0.31 * B$; where R = red, G = green, and B = blue.

3.50**point-to-point protocol**

protocol defined in RFC 1661, the Internet standard for transmitting network layer datagrams (e.g. IP packets) over serial point-to-point links