

SLOVENSKI STANDARD SIST EN ISO 6508-3:2015

01-maj-2015

Nadomešča:

SIST EN ISO 6508-3:2006

Kovinski materiali - Preskus trdote po Rockwellu - 3. del: Umerjanje primerjalnih ploščic (ISO 6508-3:2015)

Metallic materials - Rockwell hardness test - Part 3: Calibration of reference blocks (ISO 6508-3:2015)

Metallische Werkstoffe - Härteprüfung nach Rockwell - Teil 3: Kalibrierung von Härtevergleichsplatten (ISO 6508-3:2015)

Matériaux métalliques - Essai de dureté Rockwell-3: Partie 3: Étalonnage des blocs de référence (ISO 6508#3:2045) rds. iteh. ai/catalog/standards/sist/995ac2ad-7d97-47a6-aaef-6f832d317874/sist-en-iso-6508-3-2015

Ta slovenski standard je istoveten z: EN ISO 6508-3:2015

ICS:

77.040.10 Mehansko preskušanje kovin Mechanical testing of metals

SIST EN ISO 6508-3:2015 en

SIST EN ISO 6508-3:2015

iTeh STANDARD PREVIEW (standards.iteh.ai)

EUROPEAN STANDARD NORME EUROPÉENNE **EN ISO 6508-3**

EUROPÄISCHE NORM

March 2015

ICS 77.040.10

Supersedes EN ISO 6508-3:2005

English Version

Metallic materials - Rockwell hardness test - Part 3: Calibration of reference blocks (ISO 6508-3:2015)

Matériaux métalliques - Essai de dureté Rockwell - Partie 3: Étalonnage des blocs de référence (ISO 6508-3:2015) Metallische Werkstoffe - Härteprüfung nach Rockwell - Teil 3: Kalibrierung von Härtevergleichsplatten (ISO 6508-3:2015)

This European Standard was approved by CEN on 10 January 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

SIST EN ISO 6508-3:2015

https://standards.iteh.ai/catalog/standards/sist/995ac2ad-7d97-47a6-aaef-6f832d317874/sist-en-iso-6508-3-2015



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

EN ISO 6508-3:2015 (E)

Contents

	Page
Foreword	3

iTeh STANDARD PREVIEW (standards.iteh.ai)

EN ISO 6508-3:2015 (E)

Foreword

This document (EN ISO 6508-3:2015) has been prepared by Technical Committee ISO/TC 164 "Mechanical testing of metals" in collaboration with Technical Committee ECISS/TC 101 "Test methods for steel (other than chemical analysis)" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2015, and conflicting national standards shall be withdrawn at the latest by September 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 6508-3:2005.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANEndorsement notice VIEW

The text of ISO 6508-3:2015 has been approved by CEN as EN ISO 6508-3:2015 without any modification.

SIST EN ISO 6508-3:2015

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 6508-3:2015

INTERNATIONAL STANDARD

ISO 6508-3

Third edition 2015-03-01

Metallic materials — Rockwell hardness test —

Part 3: Calibration of reference blocks

Matériaux métalliques — Essai de dureté Rockwell —

iTeh STPartie 3: Étalonnage des blocs de référence (standards.iteh.ai)

<u>SIST EN ISO 6508-3:2015</u> https://standards.iteh.ai/catalog/standards/sist/995ac2ad-7d97-47a6-aaef-6f832d317874/sist-en-iso-6508-3-2015



Reference number ISO 6508-3:2015(E)

ISO 6508-3:2015(E)

iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN ISO 6508-3:2015</u> https://standards.iteh.ai/catalog/standards/sist/995ac2ad-7d97-47a6-aaef-6f832d317874/sist-en-iso-6508-3-2015



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

ISO 6508-3:2015(E)

Co	ontents	Page
Fore	reword	iv
1	Scope Normative references	1
2	Normative references	1
3	Manufacture of reference blocks	1
4	Calibration machine and calibration indenter 4.1 General 4.2 Calibration machine 4.3 Calibration diamond indenter 4.4 Calibration ball indenter	
5	Calibration procedure	4
6	Number of indentations	5
7	Uniformity of hardness	5
8	Marking	6
9	Calibration certificate	7
10	Validity	7
Ann	nex A (normative) Uniformity of reference blocks	8
	nex B (informative) Uncertainty of the mean hardness value of hardness-reference	
Ann	nex C (normative) Requirements for reference diamond indenters	16
	oliography	

SIST EN ISO 6508-3:2015

https://standards.iteh.ai/catalog/standards/sist/995ac2ad-7d97-47a6-aaef-6f832d317874/sist-en-iso-6508-3-2015

ISO 6508-3:2015(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 164, *Mechanical testing of metals*, Subcommittee SC 3, *Hardness testing*.

SIST EN ISO 6508-3:2015

This third edition cancels and areplaces athealse conduced tions (ISO-765087362005), which has been technically revised. 6f832d317874/sist-en-iso-6508-3-2015

ISO 6508 consists of the following parts, under the general title *Metallic materials* — *Rockwell hardness test*:

- Part 1: Test method
- Part 2: Verification and calibration of testing machines and indenters
- Part 3: Calibration of reference blocks

Metallic materials — Rockwell hardness test —

Part 3:

Calibration of reference blocks

1 Scope

This part of ISO 6508 specifies a method for the calibration of reference blocks to be used for the indirect and daily verification of Rockwell hardness testing machines, as specified in ISO 6508-2:2015.

Attention is drawn to the fact that the use of hard metal for ball indenters is considered to be the standard type of Rockwell indenter ball. Steel indenter balls can be used only when complying with ISO 6508-1:2015, Annex A.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 376, Metallic materials — Calibration of force-proving instruments used for the verification of uniaxial testing machines

ISO 6508-1:2015, Metallic materials — Rockwell hardness test — Part 1: Test method

https://standards.iteh.ai/catalog/standards/sist/995ac2ad-7d97-47a6-aaef-

ISO 6508-2:2015, Metallic materials Rockwell hardness test — Part 2: Verification and calibration of testing machines and indenters

3 Manufacture of reference blocks

3.1 The block shall be specially manufactured for use as a hardness-reference block.

NOTE Attention is drawn to the need to use a manufacturing process, which will give the necessary homogeneity, stability of structure, and uniformity of surface hardness.

- **3.2** Each hardness reference block shall be of a thickness not less than 6 mm. To minimize the effect of hardness change with increasing number of indents, thicker blocks should be used.
- **3.3** The reference blocks shall be free of magnetism. It is recommended that the manufacturer ensure that the blocks, if made of steel, have been demagnetized at the end of the manufacturing process (before calibration).
- **3.4** The deviation from surface flatness of the top and bottom surfaces shall be ≤ 0.01 mm. The bottom of the blocks shall not be convex. The deviation from parallelism of the top and bottom surfaces shall be ≤ 0.02 mm per 50 mm.
- **3.5** The test surface and lower surface shall be free from damage, such as notches, scratches, oxide layers, etc., which can interfere with the measurement of the indentations. The surface roughness, Ra, shall not exceed 0,000 3 mm for the test surface and 0,000 8 mm for the bottom surface. Sampling length is l = 0.8 mm (see ISO 4287:1997, 3.1.9).