



SLOVENSKI STANDARD
SIST ISO 15337:2013

01-april-2013

Zunanji zrak - Titracija plinske faze - Umerjanje analizatorjev ozona

Ambient air - Gas phase titration - Calibration of analysers for ozone

Air ambiant - Titrage en phase gazeuse - Étalonnage des analyseurs d'ozone

Ta slovenski standard je istoveten z: ISO 15337:2009

[SIST ISO 15337:2013](https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013)

<https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013>

ICS:

13.040.20 Kakovost okoljskega zraka Ambient atmospheres

SIST ISO 15337:2013

en,fr

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 15337:2013

<https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013>

**Ambient air — Gas phase titration —
Calibration of analysers for ozone**

*Air ambiant — Titrage en phase gazeuse — Étalonnage des analyseurs
d'ozone*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST ISO 15337:2013

[https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-
bc4ceb15c3b6/sist-iso-15337-2013](https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15337:2013](https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013)

<https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15337 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 3, *Ambient atmospheres*.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15337:2013](https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013)

<https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013>

Introduction

For ambient ozone (O₃) analysers, the primary standard measurement principle for calibration is UV photometry. This International Standard provides an alternative secondary measurement principle and method based on gas phase titration of an O₃ gas mixture with excess nitric oxide (NO). When using this method, the generated O₃ calibration gases are traceable to a certified primary NO measurement standard.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST ISO 15337:2013](https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013)

<https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013>

Ambient air — Gas phase titration — Calibration of analysers for ozone

1 Scope

This International Standard specifies the gas phase titration (GPT) method for the calibration of ambient air ozone (O₃) analysers. The method is applicable to the calibration of O₃ concentrations in the range 10 µg m⁻³ (5 nmol mol⁻¹ mole fraction) to 2 000 µg m⁻³ (1 000 nmol mol⁻¹ mole fraction). This International Standard uses the reference conditions of 25 °C and 101,325 kPa; however, reference temperatures of 0 °C and 20 °C are also acceptable.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7996, *Ambient air — Determination of the mass concentration of nitrogen oxides — Chemiluminescence method*

SIST ISO 15337:2013

<https://standards.iteh.ai/catalog/standards/sist/da450900-0da5-46d1-8872-bc4ceb15c3b6/sist-iso-15337-2013>

3 Principle

Gas phase titration (GPT) is based on the simple gas phase bimolecular reaction:



with a bimolecular reaction rate constant of $1,8 \times 10^{-14}$ cm³ molecule⁻¹ s⁻¹ (or 0,44 µmol mol⁻¹ s⁻¹ when concentrations are expressed as mole fractions) at 298 K.

This reaction is fast, and equilibrium lies far to the right hand side of Expression (1) if the kinetic conditions specified in Annex A are satisfied. This International Standard is based on mixing O₃ with excess nitric oxide (NO) in a dynamic flow system where, first, the NO and O₃ are mixed at relatively high concentrations (to effect essentially complete reaction of the O₃) and, second, the reaction products and any excess NO are then diluted with zero reference gas (e.g. synthetic air) to final calibration concentrations. Since the reaction is stoichiometric, the molar decrease in measured NO is equal to the added O₃; this is also equal to the nitrogen dioxide (NO₂) reaction product.

The NO is obtained from a certified NO measurement standard, a calibrated gas mixture in a compressed gas cylinder. Other calibrated sources could be used (e.g. see VDI 2453-3 [11]), and adapted to the GPT calibration apparatus. The change in NO concentration at the GPT output manifold is measured by a chemiluminescence NO analyser. A stable O₃ generator is used to produce variable concentrations of the gas to cover the calibration range of interest.

4 Reagents and materials

4.1 Sample line and connectors, made of material that is inert to O₃ and NO, such as glass or fluorocarbon polymer [e.g. perfluoro(alkoxy alkane) (PFA), polytetrafluoroethylene (PTFE) or perfluoro (ethylene-propylene) plastic (FEP) are acceptable]; these shall be as short as possible to keep the residence time to a minimum.

NOTE Whenever a sampling line is cleaned or replaced, it can take several hours to equilibrate with ambient conditions.

4.2 Zero reference gas, for calibration of the GPT procedure. If synthetic air is used, the oxygen (O₂) content shall be at the normal atmospheric concentration of 20,9 % ± 2 % volume fraction. No O₃, nitrogen oxides or any other interfering substance that can cause an undesired measurable positive or negative response in the analysis shall be detectable in the zero air.

NOTE Details on a system for making zero air from ambient air can be found in ASTM D5011 [9].

4.3 NO measurement standard, stored in a compressed gas cylinder and containing a known concentration of NO in nitrogen, in the range 10 μmol mol⁻¹ to 100 μmol mol⁻¹, for use in the calibration procedure. This NO cylinder shall be traceable to a primary measurement standard (e.g. a certified reference material), and the NO₂ impurity shall be less than 0,5 % mass fraction of the NO concentration.

5 Apparatus

Usual laboratory equipment and, in particular, the following.

5.1 Ozone generator, capable of producing steady O₃ concentrations in the required range throughout the period of the calibration. Conventional UV low-pressure mercury vapour lamps are adequate for this purpose; however, both voltage and temperature regulation shall be provided for a stable O₃ output.

CAUTION — Ozone is a toxic gas and good laboratory practice should limit indoor ozone concentrations to less than 200 μg m⁻³ (100 nmol mol⁻¹). Consult a reference text for more details on hazards of ozone and appropriate safety precautions. Any excess should be vented into an activated charcoal scrubber (with negligible back-pressure) or outdoors well away from any sampling intake. Comply with any local regulations currently in force for handling, use and disposal of ozone.

5.2 Gas flow controllers and meters: there are two options for controlling and measuring the gas flows; see 5.2.1 and 5.2.2. Electronic mass flow controllers (5.2.1) are recommended because of their inherent low measurement uncertainty and greater precision.

5.2.1 Electronic mass flow controllers, calibrated and capable of maintaining constant gas flow rates within ± 0,5 % throughout the calibration period. Components in contact with NO shall be of a non-reactive material.

5.2.2 Manual gas flow control and meters, capable of maintaining constant gas flow rates within ± 2 % throughout the calibration period. The gas flow meters shall be capable of measuring the required gas flows within ± 2 %.

5.3 Reaction chamber, to provide a suitable environment for the quantitative reaction between NO and O₃ at high concentration. This chamber shall be made of materials inert to O₃ and nitrogen oxides, such as borosilicate glass, PFA, FEP or PTFE. Its volume shall be limited so that the residence time of the gas mixture in this volume is less than 60 s (see Annex A for predetermining the volume for given flow conditions).

5.4 Dilution chamber, to provide a suitable environment for the mixing of reaction products and dilution air. It shall be made of materials inert to O₃ and nitrogen oxides, such as borosilicate glass, PFA, FEP or PTFE. Its volume should be sufficiently large to allow complete mixing of the gas components, but small enough to limit the residence time to less than 60 s.

5.5 Output manifold, to serve as a multi-port interface to allow sampling of the output from the GPT calibration system. It shall be made of materials inert to O₃, such as borosilicate glass, PFA, FEP or PTFE. It shall be of sufficient diameter and be vented to ensure an insignificant pressure drop from inside to outside the manifold. The vent outlet shall be located downstream of the other outlet ports so as to prevent intrusion of ambient air.

5.6 Temperature sensor, to measure the temperature of the detection cell of the O₃ analyser, readable to within $\pm 0,5$ °C.

5.7 Pressure meter, to indicate the pressure in the detection cell of the O₃ analyser, readable to within ± 2 hPa.

5.8 Chemiluminescence nitrogen oxides analyser, whose NO channel meets the requirements of ISO 7996. The purpose of this instrument is to determine quantitatively the decrease in NO response equivalent to the O₃ added in the GPT system.

NOTE Additional details on calibrating such an analyser can be found in VDI 2453-2 ^[10].

5.9 Pressure regulator for source gas cylinder, whose internal components are inert to NO.

6 Calibration procedure

6.1 Overview

A schematic diagram of a calibration system is shown in Figure 1. The following procedure is written for the option where

- a) both NO and O₃ analysers are simultaneously sampling the output manifold of the GPT system,
- b) concentrations are expressed as mole fractions, in nanomoles per mole.

A suitable and accurately known NO concentration in air is generated and measured with the chemiluminescence NO analyser (5.8). Then, O₃ is generated to titrate some of the NO. The decrease in NO is equal to the added O₃. Finally, with the O₃ generator (5.1) still on, the NO flow is turned off so that the ambient O₃ analyser can measure the known O₃ concentration via the common manifold. By varying the O₃ generator output, other known O₃ concentrations can be generated in a similar manner. A linear least squares analysis of the O₃ analyser responses and corresponding calculated O₃ concentrations will produce a linear calibration function for the O₃ analyser.

The linearity of the chemiluminescence NO analyser shall have been recently verified by means of a linear least squares analysis on its calibration data, and the calculated correlation coefficient shall be better than 0,99 for the NO calibration range of interest. (It should be noted that the calibration function for the NO analyser need not be used in the following calibration calculations.)

6.2 Calibration of the ambient ozone analyser

6.2.1 Install the instruments in a suitable location and provide temperature control of the measurement room to minimize any temperature dependence of the instruments. Follow the manufacturer's operating instructions for the analysers to set the various operating parameters correctly, including the sample flow rate and, if applicable, activation of the electronic temperature and pressure compensation on the O₃ analyser. Also, follow the diagnostic procedure as outlined in the manufacturer's operations manual to verify that instrument functions are within their performance specifications. The measured concentrations should be recorded by means of a suitable recording device (e.g. chart recorder or electronic data acquisition system).