



# SLOVENSKI STANDARD

## SIST EN 14793:2017

01-julij-2017

Nadomešča:

SIST-TS CEN/TS 14793:2005

---

### Emisije nepremičnih virov - Dokazovanje enakovrednosti alternativne metode z referenčno metodo

Stationary source emissions - Demonstration of equivalence of an alternative method with a reference method

Emissionen aus stationären Quellen - Nachweis der Gleichwertigkeit eines Alternativverfahrens mit einem Referenzverfahren

Émissions de sources fixes - Démonstration de l'équivalence d'une méthode 'alternative' avec une méthode de référence

Ta slovenski standard je istoveten z: EN 14793:2017

---

#### ICS:

13.040.40 Emisije nepremičnih virov Stationary source emissions

SIST EN 14793:2017

en,fr,de

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 14793:2017

<https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017>

EUROPEAN STANDARD

EN 14793

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2017

ICS 13.040.40

Supersedes CEN/TS 14793:2005

English Version

## Stationary source emissions - Demonstration of equivalence of an alternative method with a reference method

Émissions de sources fixes - Démonstration de l'équivalence d'une méthode alternative avec une méthode de référence

Emissionen aus stationären Quellen - Nachweis der Gleichwertigkeit eines Alternativverfahrens mit einem Referenzverfahren

This European Standard was approved by CEN on 26 September 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

**iTeh STANDARD PREVIEW**

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

<b>Contents</b>	<b>Page</b>
European foreword.....	3
Introduction .....	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions .....	6
4 Symbols.....	9
5 Contents of the demonstration of equivalence.....	11
5.1 General.....	11
5.2 Description of the alternative method .....	11
5.3 Determination of performance characteristics .....	12
5.3.1 General.....	12
5.3.2 Manual method .....	12
5.3.3 Automatic method .....	13
5.4 Calculation of the expanded uncertainty of the AM.....	13
5.5 Demonstration of equivalence in the field.....	14
5.5.1 Coverage of in field demonstration .....	14
5.5.2 Evaluation of repeatability and trueness in relation to the RM .....	15
6 Report of the demonstration of equivalence.....	18
Annex A (informative) Example of comparison of repeatability and trueness of Thorin Method and Ion Chromatography Method for SO <sub>2</sub> measurement in stack.....	21
Annex B (informative) Critical values for Grubbs test.....	32
Annex C (normative) Minimum requirements for a test bench .....	33
Bibliography.....	34

## European foreword

This document (EN 14793:2017) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This document supersedes CEN/TS 14793:2005.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2017, and conflicting national standards shall be withdrawn at the latest by July 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 14793:2017](https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017)

<https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017>

## Introduction

Much has been published in the literature concerning method validation by collaborative study. CEN/TC 264 working groups try to follow these method validations when a new standard is prepared and the collaborative study is probably the preferred way of carrying out the validation. However, it is not always a suitable option for accredited laboratories. The application for which the method is required can be esoteric to the extent that no other laboratories would be interested in collaboration. Those that might be interested can be competitors.

This European Standard provides one of possible methods of testing the equivalence of an alternative method (AM) with the standard reference method (SRM) or with a reference method (RM) if the legislator has not defined a standard reference method.

NOTE The term “reference method” is used in this standard to cover reference methods as well as standard reference methods.

In the framework of certification of automated measuring systems used for the measurement of stationary source emissions this European Standard can be used in conjunction with EN 15267-4:2017 to demonstrate the equivalence of portable automated measuring systems (P-AMS) based on an AM with the standard reference method (SRM).

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 14793:2017](https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017)

<https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017>

## 1 Scope

This European Standard specifies a procedure to demonstrate the equivalence of an alternative method (AM) with the reference method (RM) or the standard reference method (SRM), both implemented to determine the same measurand.

In particular, this European Standard provides the statistical tools and different criteria to evaluate the alternative method. This does not release the body performing the equivalence testing from bearing technical and analytical judgement on the evaluation of the different criteria.

Three steps are required for demonstration of equivalence:

- description of the alternative method and setting of the field of application (measurement range and type of gas matrix);
- determination of the performance characteristics of the alternative method and calculation of the expanded uncertainty where appropriate and check of compliance with the maximum expanded uncertainty allowed for the reference method;
- check of repeatability and lack of systematic deviation of the alternative method in the field or on a recognized test bench in comparison with the reference method for the type of matrix defined in the field of equivalence.

This European Standard requires that a reference method has been defined and validated.

This European Standard only considers the case of linear quantitative methods.

This European Standard is applicable to manual and automated methods.

This European Standard has been drawn up for laboratories working in air quality measurements and consequently an example taken from this sector are presented in Annex A.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

ISO 5725-1:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*

ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

**EN 14793:2017 (E)****3 Terms and definitions**

For the purposes of this document, the following terms and definitions apply.

**3.1  
standard reference method**
**SRM**

reference method prescribed by European or national legislation

[SOURCE: EN 15259:2007]

Note 1 to entry: Standard reference methods are used e.g. to calibrate and validate automated measuring systems permanently installed at stacks and for periodic measurements to check compliance with limit values.

**3.2  
reference method**
**RM**

measurement method taken as a reference by convention, which gives the accepted reference value of the measurand

Note 1 to entry: A reference method is fully described.

Note 2 to entry: A reference method can be a manual or an automated method.

Note 3 to entry: Alternative methods can be used if equivalence to the reference method has been demonstrated.

**STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SOURCE: EN 15259:2007]

[SIST EN 14793:2017](https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017)

<https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017>

**3.3  
alternative method**
**AM**

measurement method which complies with the criteria given by this European Standard with respect to the reference method

Note 1 to entry: An alternative method can consist of a simplification of the reference method.

**3.4  
measurement method**

method described in a written procedure containing all the means and procedures required to sample and analyse, namely field of application, principle and/or reactions, definitions, equipment, procedures, presentation of results, other requirements and measurement report

**3.5  
calibration**

set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring method or measuring system, and the corresponding values given by the applicable reference

**3.6  
accepted reference value**

value which serves as a reference value (or conventionally true value) of the sample, provided by the reference method or standard reference method



**3.7****demonstration of equivalence**

act of subjecting a measurement method to a study, which is based on a standardised and/or recognised protocol and which provides proof that, for its field of application, the measurement method satisfies pre-established performance criteria

Note 1 to entry: In the framework of this European Standard, the demonstration of equivalence of a method is mainly based on an “in field” study that includes comparison to a reference method.

**3.8****field of application**

combination of the different types of matrix and the range of concentrations of the measured component covered, to which the measurement method is applied

Note 1 to entry: As well as being an indication of all the satisfactory performance conditions for each factor, the field of application of the measurement method can also include warnings concerning known interferences caused by other components, or the inapplicability of certain matrices or conditions.

**3.9****matrix**

all the components of the sample other than the measured component

Note 1 to entry: Some components of the matrix can influence the result of measurement. These components are called interferents.

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

**3.10****measurand**

particular quantity subject to measurement

[SIST EN 14793:2017](#)

[SOURCE: EN 15259:2007] <https://standards.iteh.ai/catalog/standards/sist/83458ced-9804-4bf2-a0b2-073a21ef5292/sist-en-14793-2017>

Note 1 to entry: The measurand is a quantifiable property of the stack gas under test, for example mass concentration of a measured component, temperature, velocity, mass flow, oxygen content and water vapour content.

**3.11****measured component**

constituent of the waste gas for which a defined measurand is to be determined by measurement

[SOURCE: EN 15259:2007]

Note 1 to entry: Measured component is also called determinand.

**3.12****reference material**

substance or mixture of substances, with a known concentration within specified limits, or a device of known characteristics

**EN 14793:2017 (E)****3.13****linearity**

capacity of a measurement method, within certain limits, to provide an instrument response or results proportional to the quantity of the measurand to be determined in the sample

Note 1 to entry: This proportionality is expressed through a defined a priori mathematical expression.

Note 2 to entry: The linearity limits are the concentration limits in the experiment between which a linear calibration model can be applied with a known level of confidence.

**3.14****lack of fit**

systematic deviation, within the measurement range, between the accepted value of a reference material applied to the measuring system and the corresponding result of measurement produced by the calibrated measuring system

Note 1 to entry: In common language lack of fit is often called “linearity” or “deviation from linearity”. Lack of fit test is often called “linearity test”.

**3.15****detection limit**

$L_D$

smallest quantity of the measurand which can be detected, but not quantified, in the experiment conditions described for the measurement method

**3.16****quantification limit**

$L_Q$

smallest quantity of the measurand which can be quantified, in the experiment conditions described for the measurement method

**3.17****repeatability**

closeness of agreement between independent test results obtained under stipulated conditions

Note 1 to entry: Repeatability depends exclusively upon the distribution of random errors and has no relation with the true or specified value.

Note 2 to entry: The measure of repeatability is calculated from the standard deviation of test results. A lower level of repeatability is reflected by a greater standard deviation.

Note 3 to entry: The term “independent test results” signifies results obtained in such a way as not to be influenced by a previous result on the same or similar testing equipment. Quantitative measurements of repeatability depend critically upon the stipulated conditions. Repeatability and reproducibility conditions are specific groups of extreme conditions.

**3.18****trueness**

closeness of agreement between the average value obtained from a large series of test results and an accepted reference value

[SOURCE: ISO 5725-1:1994]

Note 1 to entry: The measure of trueness is generally expressed in terms of a bias or a systematic deviation.

## 4 Symbols

For the purpose of this document, the symbols listed in Table 1 apply.

**Table 1 — Symbols and formulae**

Symbol	Description	Formula
$C_0$	intercept of the orthogonal regression line between AM and RM values	$C_0 = \bar{\bar{x}} - \frac{s(\bar{\bar{x}})}{s(\bar{\bar{z}})} \bar{\bar{z}}$
$C_1$	slope of the orthogonal regression line between AM and RM values	$C_1 = \frac{s(\bar{\bar{x}})}{s(\bar{\bar{z}})}$
$dx_i$	difference between $x_{i1}$ and $x_{i2}$ for each value of $i$	$dx_i = (x_{i1} - x_{i2})$
$e_i$	ratio between $dx_i$ and $\bar{x}_i$ for each value of $i$	$e_i = \frac{dx_i}{\bar{x}_i}$
$\bar{e}$	average value of the values $e_i$	$\bar{e} = \frac{\sum_{i=1}^n e_i}{n}$
$G_i$	ratio between $(e_i - \bar{e})$ and $s(e_i)$	$G_i = \frac{ e_i - \bar{e} }{s(e_i)}$
$L_D$	detection limit	
$L_Q$	quantification limit	
$p$	number of trials	
$n_i$	number of parallel measurement for the AM and for the RM for a trial $i$	
$N$	total number of measurements for the AM and for the RM	$N = \sum_{i=1}^p n_i$
$r$	correlation coefficient	$r = \frac{SPD(\bar{\bar{x}}, \bar{\bar{z}})}{\sqrt{SSD(\bar{\bar{x}}) \times SSD(\bar{\bar{z}})}}$
$s(e_i)$	standard deviation of the population of the $e_i$	$s(e_i) = \sqrt{\frac{1}{n-1} \sum_{i=1}^n (e_i - \bar{e})^2}$
$s^2(\bar{\bar{x}})$	variance of the AM	$s^2(\bar{\bar{x}}) = \frac{SSD(\bar{\bar{x}})}{p-1}$
$s_r(\bar{\bar{x}})$	repeatability standard deviation of the AM	

## EN 14793:2017 (E)

$s_r^2(\bar{x})$	repeatability variance of the AM	$s_r^2(\bar{x}) = \frac{\sum_{i=1}^p \sum_{j=1}^{n_i} (x_{ij} - \bar{x}_i)^2}{N - p}$
$s^2(\bar{z})$	variance of the RM	$s^2(\bar{z}) = \frac{\text{SSD}(\bar{z})}{p - 1}$
$s_r(\bar{z})$	repeatability standard deviation of the RM	
$s_r^2(\bar{z})$	repeatability variance of the RM	$s_r^2(\bar{z}) = \frac{\sum_{i=1}^p \sum_{j=1}^{n_i} (z_{ij} - \bar{z}_i)^2}{N - p}$
$s_{r,\text{limit}}(\bar{z})$	maximum allowable repeatability standard deviation for the RM	
$s_R(\bar{z})$	reproducibility standard deviation given in the RM standard	
$\text{SPD}(\bar{x}, \bar{z})$	sum of the products of the deviations for two variables $\bar{x}$ and $\bar{z}$	$\text{SPD}(\bar{x}, \bar{z}) = \sum_{i=1}^p (\bar{x}_i - \bar{x}) \times (\bar{z}_i - \bar{z})$
$\text{SSD}(\bar{x})$	sum of the squares of the mean deviations for the AM	$\text{SSD}(\bar{x}) = \sum_{i=1}^p (\bar{x}_i - \bar{x})^2$
$\text{SSD}(\bar{z})$	sum of the squares of the mean deviations for the RM	$\text{SSD}(\bar{z}) = \sum_{i=1}^p (\bar{z}_i - \bar{z})^2$
$U_{\text{RM}}$	maximum permissible expanded uncertainty given in the RM standard	
$x_{ij}$	concentration obtained by the AM for a trial $i$ and repetition $j$	
$\bar{x}_i$	arithmetic mean of $x_{ij}$ for which $n_i$ measurements have been taken	$\bar{x}_i = \frac{\sum_{j=1}^{n_i} x_{ij}}{n_i}$
$\bar{x}$	grand average of $x_{ij}$ for which $N$ measurements have been taken	$\bar{x} = \frac{\sum_{i=1}^p \sum_{j=1}^{n_i} x_{ij}}{N}$
$x_p$	outlier	
$z_{ij}$	concentration obtained by the RM for a trial $i$ and repetition $j$	
$\bar{z}_i$	arithmetic mean of $z_{ij}$ for which $n_i$ measurements have been taken	$\bar{z}_i = \frac{\sum_{j=1}^{n_i} z_{ij}}{n_i}$