



# SLOVENSKI STANDARD

## SIST EN 13795-1:2019

01-julij-2019

Nadomešča:

SIST EN 13795:2011+A1:2013

---

### Operacijska oblačila in pokrivala - Zahteve in preskusne metode - 1. del: Operacijska pokrivala in plašči

Surgical clothing and drapes - Requirements and test methods - Part 1: Surgical drapes and gowns

Operationskleidung und -abdecktücher - Anforderungen und Prüfverfahren - Teil 1:  
Operationsabdecktücher und -mäntel

(standards.iteh.ai)

Vêtements et champs chirurgicaux - Exigences et méthodes d'essai - Partie 1 : Champs et casaques chirurgicaux

<https://standards.iteh.ai/catalog/standards/sist/90d45b1c-1b18-4772-9e89-81b9eff98dd/sist-en-13795-1-2019>

**Ta slovenski standard je istoveten z: EN 13795-1:2019**

---

#### **ICS:**

11.140

Oprema bolnišnic

Hospital equipment

**SIST EN 13795-1:2019**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN 13795-1:2019

<https://standards.iteh.ai/catalog/standards/sist/90d45b1c-1b18-4772-9e89-81bf9eff98dd/sist-en-13795-1-2019>

EUROPEAN STANDARD

EN 13795-1

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2019

ICS 11.140

Supersedes EN 13795:2011+A1:2013

English Version

## Surgical clothing and drapes - Requirements and test methods - Part 1: Surgical drapes and gowns

Vêtements et champs chirurgicaux - Exigences et méthodes d'essai - Partie 1 : Champs et casaques chirurgicaux

Operationskleidung und -abdecktücher - Anforderungen und Prüfverfahren - Teil 1: Operationsabdecktücher und -mäntel

This European Standard was approved by CEN on 24 October 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

<b>Contents</b>	<b>Page</b>
European foreword.....	4
Introduction .....	6
1 Scope.....	7
2 Normative references.....	7
3 Terms and definitions .....	8
4 Performance requirements.....	11
5 Manufacturing and processing requirements and documentation .....	13
6 Information to be supplied with the product .....	13
6.1 Information to be supplied to the user .....	13
6.2 Information to be supplied to the processor .....	14
Annex A (normative) Testing.....	15
A.1 General.....	15
A.2 Test methods and conformance.....	15
A.2.1 Test method for evaluation of cleanliness microbial/bioburden .....	15
A.2.2 Test method for evaluation of particle release .....	16
A.2.3 Test method for evaluation of liquid penetration.....	16
A.2.4 Test method for evaluation of bursting strength in dry and wet state.....	16
A.2.5 Test method for evaluation of tensile strength in dry and wet state .....	17
A.2.6 Test method for evaluation of dry microbial penetration .....	17
A.2.7 Test method for evaluation of wet microbial penetration.....	17
A.2.8 Test method for evaluation of biocompatibility .....	17
A.3 Treatment of results .....	18
Annex B (informative) Rationales .....	19
B.1 General.....	19
B.2 Cleanliness - microbial .....	19
B.3 Particle release .....	19
B.4 Resistance to liquid penetration .....	20
B.5 Bursting strength - dry and wet .....	21
B.6 Tensile strength - dry and wet.....	21
B.7 Resistance to microbial penetration - dry.....	21
B.8 Resistance to microbial penetration - wet .....	23
B.9 Labelling.....	23
B.10 Treatment of results .....	24

<b>Annex C (informative) Information on further characteristics .....</b>	<b>25</b>
<b>C.1 Comfort.....</b>	<b>25</b>
<b>C.2 Adhesion for fixation for the purpose of wound isolation.....</b>	<b>25</b>
<b>C.3 Liquid control.....</b>	<b>25</b>
<b>C.4 Flammability .....</b>	<b>26</b>
<b>C.5 Electrostatic discharge .....</b>	<b>26</b>
<b>Annex D (informative) Environmental aspects.....</b>	<b>27</b>
<b>Annex E (informative) Guidance to users for selecting products.....</b>	<b>28</b>
<b>E.1 Performance levels .....</b>	<b>28</b>
<b>E.2 Functional design.....</b>	<b>28</b>
<b>E.2.1 General .....</b>	<b>28</b>
<b>E.2.2 Critical and less critical areas .....</b>	<b>28</b>
<b>E.2.3 Size .....</b>	<b>29</b>
<b>E.2.4 Accessories.....</b>	<b>29</b>
<b>E.2.5 Comfort.....</b>	<b>29</b>
<b>E.2.5.1 General .....</b>	<b>29</b>
<b>E.2.5.2 Surgical gowns .....</b>	<b>29</b>
<b>E.2.5.3 Surgical drapes .....</b>	<b>30</b>
<b>E.3 Practical trials.....</b>	<b>30</b>
<b>Annex ZA (informative) Relationship between this European standard and the essential requirements of Directive 93/42/EEC [1993 OJ L 169] aimed to be covered .....</b>	<b>31</b>
<b>Bibliography .....</b>	<b>32</b>

iTech STANDARD PREVIEW  
(standards.itech.ai)

SIST EN 13795-1:2019

<https://standards.itech.ai/catalog/standards/sist/90d45b1c-1b18-4772-9a89-81b5e98dd/sist-en-13795-1-2019>

**EN 13795-1:2019 (E)****European foreword**

This document (EN 13795-1:2019) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2019, and conflicting national standards shall be withdrawn at the latest by October 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Together with EN 13795-2:2019, this document supersedes EN 13795:2011+A1:2013.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

EN 13795 consists of the following parts, under the general title *Surgical clothing and drapes — Requirements and test methods*:

- *Part 1: Surgical drapes and gowns* ([standards.iteh.ai](https://standards.iteh.ai))
- *Part 2: Clean air suits*

The following changes have been introduced:

- a) The product ‘clean-air suit’ has been moved to Part 2 of the EN 13795 standard series because of distinctive requirements and test methods;
- b) Alignment of the document title and the Scope;
- c) Revision of the Normative references and the Bibliography;
- d) Alignment of the Clause ‘Terms and definitions’;
- e) Review of the performance requirements in Table 1 and Table 2 especially with regard to ‘Cleanliness - Particulate matter’ and ‘Linting’, which have been combined as ‘Particle release’;
- f) Movement of former Clause 5 ‘Testing’ to A.1 and editorial alignment;
- g) Revision of Clause ‘Manufacturing and processing requirements’ by adding of documentary requirements and a section for the introduction of a QM system;
- h) Enhancement and improved structuring of Clause ‘Information to be supplied by the manufacturer or processor’;
- i) Deletion of the former Annex A ‘Details of significant changes between this document and the previous edition’ which consisted of 3 parts;
- j) Complete revision and extension of Annex A ‘Testing’ (formerly Annex B ‘Test methods’);

- k) Inclusion of a new Annex B 'Rationales' which provides precise reasons for the essential requirements of this document and which is intended for users aware of the subject of this document, but who have not participated in its development;
- l) Deletion of the former Annex C 'Prevention of infection in the operating room';
- m) Revision and extension of Annex C (formerly Annex D) 'Information on further characteristics'; e.g. inclusion of a Clause on 'Flammability' and 'Electrostatic discharge';
- n) Inclusion of a new Annex D 'Environmental aspects';
- o) Inclusion of a new Annex E 'Guidance to users for selecting products';
- p) Revision of Annex ZA on the relationship to the Medical Device Directive (93/42/EEC);
- q) Complete editorial revision.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## **iTeh STANDARD PREVIEW** **(standards.iteh.ai)**

SIST EN 13795-1:2019

<https://standards.iteh.ai/catalog/standards/sist/90d45b1c-1b18-4772-9e89-81bf9eff98dd/sist-en-13795-1-2019>

## Introduction

The transmission of infective agents during invasive surgical procedures can occur in several ways (see informative Annex B).

Surgical drapes, including the intended use as a sterile field, and surgical gowns are used to minimize the spread of infective agents to and from patients' operating wounds, thereby helping to prevent post-operative wound infections (see Annex B).

The performance required of coverings for patients, clinical staff and equipment varies with, for example, the type and duration of the procedure, the degree of wetness of the operation field, the degree of mechanical stress on the materials and the susceptibility of the patient to infection.

The use of surgical gowns with resistance to the penetration of liquids can also diminish the risk to the operating staff from infective agents carried in blood or body fluids.

This document is intended to assist the communication between manufacturers and third parties with regard to material or product characteristics and performance requirements.

Therefore, Annex B provides comprehensive information on characteristics, measurement of performance and performance requirements. Annex C clarifies that this document does not include environmental provisions. Annex D provides information on characteristics regarded relevant in context with surgical gowns and drapes, however but not covered normatively (i.e. without applicable performance requirements). Annex E explains the concept of performance levels and provides guidance to users for selecting products.

This document focuses on Essential Requirements arising from the Medical Device Directive 93/42/EEC, which are applicable to surgical drapes and gowns. The requirements and guidance in this document are expected to be of help to manufacturers and users when designing, processing, assessing and selecting products. It is the intention of this document to ensure the same level of safety from single-use and reusable surgical clothing and drapes throughout their useful life.

Surgical gowns are used to minimize the transmission of infective agents between patients and clinical staff during surgical and other invasive procedures. Hereby, surgical gowns contribute to the clinical condition and the safety of patients as well as to the safety and health of users following up essential requirement 1 of Directive 93/42/EEC on Medical Devices. This document addresses the same level of protection for patients and users (i.e. the surgical team) by not differentiating the performance requirements for surgical gowns respectively. However, this document does not formally address any basic health and safety requirements of the Directive 89/686/EEC or Regulation (EU) 2016/425 on Personal Protective Equipment and does not provide specific guidance for surgical gowns intended by the manufacturer for dual use as medical device and personal protective equipment.



## 1 Scope

This document specifies information to be supplied to users and third party verifiers in addition to the usual labelling of medical devices (see EN 1041 and EN ISO 15223-1), concerning manufacturing and processing requirements. This document gives information on the characteristics of single-use and reusable surgical gowns and surgical drapes used as medical devices for patients, clinical staff and equipment, intended to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures. This document specifies test methods for evaluating the identified characteristics of surgical drapes and gowns and sets performance requirements for these products.

This document does not cover requirements for resistance to penetration by laser radiation of products. Suitable test methods for resistance to penetration by laser radiation, together with an appropriate classification system, are given in EN ISO 11810.

This document does not cover requirements for incision drapes or films.

This document does not cover requirements for antimicrobial treatments for surgical gowns and drapes. Antimicrobial treatment can cause environmental risks such as resistance and pollution. However, antimicrobial treated surgical gowns and drapes fall under the scope of this document with respect to their use as surgical gowns and drapes.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 811:2018, *Textiles - Determination of resistance to water penetration - Hydrostatic pressure test (ISO 811:2018)* <https://standards.iteh.ai/catalog/standards/sist/90d45b1c-1b18-4772-9e89-81b9eff98dd/sist-en-13795-1-2019>

EN 29073-3:1992, *Textiles - Test methods for nonwovens - Part 3: Determination of tensile strength and elongation*

EN ISO 139:2005,<sup>1</sup> *Textiles — Standard atmospheres for conditioning and testing (ISO 139:2005 + Amd. 1:2011)*

EN ISO 9073-10:2004, *Textiles - Test methods for nonwovens - Part 10: Lint and other particles generation in the dry state (ISO 9073-10:2003)*

EN ISO 10993-1:2009, *Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009)*

EN ISO 11737-1:2018, *Sterilization of medical devices — Microbiological methods — Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018)*

EN ISO 13938-1:1999, *Textiles - Bursting properties of fabrics - Part 1: Hydraulic method for determination of bursting strength and bursting distension (ISO 13938-1:1999)*

---

<sup>1</sup> Impacted by EN ISO 139:2005+A1:2011

**EN 13795-1:2019 (E)**

EN ISO 22610:2006, *Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment - Test method to determine the resistance to wet bacterial penetration (ISO 22610:2006)*

EN ISO 22612:2005, *Clothing for protection against infectious agents - Test method for resistance to dry microbial penetration (ISO 22612:2005)*

**3 Terms and definitions**

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

**3.1 colony forming unit****CFU**

unit by which the culturable number of microorganisms is expressed

Note 1 to entry: The culturable number is the number of microorganisms, single cells or aggregates, able to form colonies on a solid nutrient medium.

**3.2 cleanliness**

freedom from unwanted foreign matter **(standards.iteh.ai)**

**iTeh STANDARD PREVIEW**

Note 1 to entry: Such matter can be microorganisms, organic residues or particulate matter.

<https://standards.iteh.ai/catalog/standards/sist/90d45b1c-1b18-4772-9e89-81bf9eff98dd/sist-en-13795-1-2019>

**3.2.1 cleanliness — microbial**

freedom from population of viable micro-organisms on a product and/or a package

Note 1 to entry: In practical use, microbial cleanliness is often referred to as 'bioburden'.

**3.3 critical product area**

product area with a greater probability to be involved in the transfer of infective agents to or from the wound, e.g. front and sleeves of surgical gowns

**3.4 infective agent**

micro-organism that has been shown to cause wound infections or that might cause infection in a member of the surgical team or the patient

**3.5 less critical product area**

product area less likely to be involved in the transfer of infective agents to or from the wound

**3.6****manufacturer**

natural or legal person with responsibility for the design, manufacture, packaging and labelling of a device before it is placed on the market under his own name, regardless of whether these operations are carried out by that person himself or on his behalf by a third party

Note 1 to entry: For more details, refer to the Medical Device Directive 93/42/EEC.

**3.7****particle release**

release of fibre fragments and other particles during mechanical stress simulating handling and use

**3.8****performance level**

discrete standard defined to classify products according to the performance requirements of this document

Note 1 to entry: With the introduction of two performance levels, this document acknowledges the fact that products are challenged to differing extents during surgical procedures, dependent upon the duration, mechanical stress and liquid challenge throughout the surgical procedure.

**3.8.1****standard performance**

classification addressing minimum performance requirements for various characteristics of products used as medical devices in invasive surgical procedures

**3.8.2****high performance**

classification addressing elevated performance requirements for various characteristics of products used as medical devices in invasive surgical procedures

Note 1 to entry: Examples of surgical procedures where elevated performance level should be considered are those where extensive exposure to liquid, mechanical stresses or longer surgical procedures can be expected.

**3.9****processor**

natural or legal person who processes products so that their performance complies with the requirements of this document

Note 1 to entry: A processor who places a product on the market is a manufacturer in the sense of this document.

Note 2 to entry: A processor of reusable products is often referred to as a 'reprocessor' and processing reusable products is often referred to as 'reprocessing' (as e.g. in Medical Device Directive 93/42/EEC). References in EN 13795-2 and this document to 'processors' include 'reprocessors' and to 'processing' include 'reprocessing'.

**3.10****product**

surgical gown, surgical drape including equipment covering

Note 1 to entry: In cases of surgical packs, each gown or drape is regarded as a product.

**EN 13795-1:2019 (E)****3.11****resistance to liquid penetration**

ability of material to withstand the penetration of liquid(s) from one side of the material through to the other

**3.12****resistance to microbial penetration**

ability of material(s) to withstand penetration of micro-organisms from one side of the material through to the other

**3.12.1****dry penetration**

effect of a combination of air movement and mechanical action by vibration on microbial penetration in dry condition

**3.12.2****wet penetration**

effect of combination of wetness, pressure and rubbing on microbial penetration

**3.13****reusable product**

product intended by the manufacturer to be reprocessed and reused

**3.14****single-use product**

product intended to be used once only for a single patient

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

**3.15****sterile field**

area created by sterile surgical drape material where aseptic technique is practised

SIST EN 13795-1:2019

<https://standards.iteh.ai/catalog/standards/sist/90d45b1c-1b18-4772-9e89-8118f892d4/sist-en-13795-1-2019>

Note 1 to entry:

A sterile field can be practised e.g. on a back table.

**3.16****surgical drape**

drape covering the patient or equipment to prevent transfer of infective agents

**3.17****surgical gown**

gown worn by a member of a surgical team to prevent transfer of infective agents

**3.18****surgical procedure**

surgical intervention performed by a surgical team

**3.18.1****invasive surgical procedure**

surgical procedure penetrating skin or mucosa