



SLOVENSKI STANDARD

SIST EN 16195:2013

01-maj-2013

Nadomešča:

SIST-TS CEN/TS 16195:2011

Gnojila - Določevanje kloridov v odsotnosti organske snovi

Fertilizers - Determination of chlorides in the absence of organic material

Düngemittel - Bestimmung von Chlorid bei Abwesenheit organischer Stoffe

Engrais - Dosage des chlorures en l'absence de matières organiques
iTeh STANDARD PREVIEW
(standards.iteh.ai)

Ta slovenski standard je istoveten z: ~~SIST EN 16195:2012~~ **EN 16195:2012**

<https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013>

ICS:

65.080

Gnojila

Fertilizers

SIST EN 16195:2013

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 16195:2013

<https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013>

EUROPEAN STANDARD

EN 16195

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2012

ICS 65.080

Supersedes CEN/TS 16195:2011

English Version

Fertilizers - Determination of chlorides in the absence of organic material

Engrais - Dosage des chlorures en l'absence de matières organiques

Düngemittel - Bestimmung von Chlorid bei Abwesenheit organischer Stoffe

This European Standard was approved by CEN on 6 October 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST EN 16195:2013](https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013)

<https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013>



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
Foreword.....	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Principle.....	4
5 Sampling.....	4
6 Reagents.....	4
7 Apparatus	5
8 Procedure	5
8.1 Preparation of the test portion and the solution	5
8.2 Determination.....	5
8.3 Blank test.....	6
8.4 Control test.....	6
9 Calculation and expression of the results	6
10 Precision.....	6
10.1 Inter-laboratory test.....	6
10.2 Repeatability.....	6
10.3 Reproducibility.....	6
11 Test report	7
Annex A (informative) Statistical results of the inter-laboratory test	8
Bibliography.....	9

iTech STANDARD PREVIEW
(standards.itech.ai)

SIST EN 16195:2013

<https://standards.itech.ai/standards/16195-2013>
8496-2572050618e2/sist-en-16195-2013

Foreword

This document (EN 16195:2012) has been prepared by Technical Committee CEN/TC 260 “Fertilizers and liming materials”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2013, and conflicting national standards shall be withdrawn at the latest by May 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16195:2011.

The following changes have been made to the former edition:

- a) the CEN Technical Specification has been adopted as a European Standard;
- b) the document has been editorially revised.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

EN 16195:2012 (E)**1 Scope**

This European Standard specifies a method for the determination of chlorides in the absence of organic material. The method is applicable to all fertilizers which are free from organic material.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN 12944-1:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 1: General terms*

EN 12944-2:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 2: Terms relating to fertilizers*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944 1:1999 and EN 12944-2:1999 apply.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

4 Principle

The chlorides, dissolved in water, are precipitated in an acid medium by an excess of standard solution of silver nitrate. The excess is titrated with a solution of ammonium thiocyanate in the presence of ferric ammonium sulfate (Volhard's method).

5 Sampling

Sampling is not part of the method specified in this document. A recommended sampling method is given in EN 1482-1.

Sample preparation shall be carried out in accordance with EN 1482-2. Grinding of the laboratory sample is recommended for homogeneity reasons.

6 Reagents

Use only reagents of recognised analytical grade.

6.1 Water, distilled or demineralised and free from chlorides.

6.2 Nitrobenzene or diethyl ether.

6.3 Nitric acid, $c = 10 \text{ mol/l}$.

6.4 Indicator solution.

Dissolve 40 g of ferric ammonium sulfate $\text{Fe}_2(\text{SO}_4)_3 \cdot (\text{NH}_4)_2\text{SO}_4 \cdot 24\text{H}_2\text{O}$, in water and make up to 1 l.

6.5 Silver nitrate standard solution, $c = 0,1$ mol/l.

Preparation: since this salt is hygroscopic and cannot be dried without risk of decomposition, it is advisable to weigh out an amount of approximately 18 g, dissolve in water and make up the volume to 1 l. Adjust to $c = 0,1$ mol/l substance concentration by titration of ammonium thiocyanate 0,1 mol/l.

Silver nitrate standard solution is also readily available commercially and may be used instead.

6.6 Ammonium thiocyanate, standard solution, $c = 0,1$ mol/l.

Preparation: since ammonium thiocyanate is hygroscopic and cannot be dried without risk of decomposition, it is advisable to weigh out an amount of approximately 9 g, dissolve in water and make up the volume to 1 l. Adjust to $c = 0,1$ mol/l substance concentration by titration of silver nitrate 0,1 mol/l.

Ammonium thiocyanate standard solution is also readily available commercially, and may be used instead.

7 Apparatus

7.1 Standard laboratory equipment.

7.2 Rotary shaker, 35 to 40 min^{-1} .

7.3 Burettes.

7.4 Graduated flask, capacity 500 ml and 1 000 ml.

7.5 Conical (Erlenmeyer) flask, capacity 250 ml.

8 Procedure

[SIST EN 16195:2013
https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013](https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013)

8.1 Preparation of the test portion and the solution

Place 5 g of the laboratory sample, weighed to the nearest 0,001 g, in a graduated flask (7.4) and add 450 ml of water (6.1). Mix for 0,5 h on the shaker (7.2), make up to 500 ml with water (6.1), mix and filter into a beaker.

8.2 Determination

Take an aliquot part of the test solution containing not more than 0,150 g of chloride: for example 25 ml (0,25 g), 50 ml (0,5 g) or 100 ml (1 g). If the amount of the test solution taken is smaller than 50 ml it is necessary to make up the volume to 50 ml with water (6.1).

Add 5 ml of nitric acid (6.3), 20 ml of the indicator solution (6.4), and two drops of ammonium thiocyanate standard solution (6.6) (this latter reagent is added with a burette adjusted to zero for this purpose).

With a burette (7.3) then add silver nitrate standard solution (6.5) until there is an excess of 2 ml to 5 ml. Add 5 ml of nitrobenzene or 5 ml of diethyl ether (6.2) and shake well to agglomerate the precipitate. Titrate the excess silver nitrate with ammonium thiocyanate (6.6) until a red-brown colour appears which remains after the flask has been shaken slightly.

NOTE Nitrobenzene or diethyl ether (but above all nitrobenzene) prevents the silver chloride from reacting with thiocyanate ions. Thus a clear colour change is obtained.

EN 16195:2012 (E)**8.3 Blank test**

Carry out a blank test (omitting the test portion) under the same conditions and allow for it when calculating the final result.

8.4 Control test

Before carrying out the estimations check the accuracy of the method by using an aliquot part of a freshly prepared solution of potassium chloride, such that this part contains a known quantity in the order of 100 mg of chloride.

9 Calculation and expression of the results

Express the result of the analysis as a percentage of chloride contained in the sample as it has been received for analysis.

Calculate the mass fraction of chlorides (Cl), w_{Cl} , in percent according to Formula (1).

$$w_{Cl} = 0,003\,546 \times \frac{(V_Z - V_{CZ}) - (V_a - V_{ca})}{M} \times 100 \quad (1)$$

where

V_Z is the amount of silver nitrate 0,1 mol/l, in millilitres;

V_{CZ} is the amount of silver nitrate 0,1 mol/l, used in the blank test, in millilitres;

V_a is the amount of ammonium thiocyanate 0,1 mol/l, in millilitres;

V_{ca} is the amount of ammonium thiocyanate 0,1 mol/l, used in the blank test, in millilitres;

M is the mass, in grams, of the test portion taken (8.1).

10 Precision**10.1 Inter-laboratory test**

An inter-laboratory test has been carried out in 2009 with 14/13 participating laboratories and two different samples. The repeatability and reproducibility were calculated according to ISO 5725-2.

The values derived from these inter-laboratory tests may not be applicable to concentration ranges and matrices other than those given in Annex A.

10.2 Repeatability

The absolute difference between two independent single test results, obtained with the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of the cases exceed the values of r given in Table 1.

10.3 Reproducibility

The absolute difference between two single test results, obtained with the same method on identical test material in different laboratories by different operators using different equipment, will not exceed values of R given in Table 1 in more than 5 % of the cases.

Table 1 — Mean values, repeatability and reproducibility limits

Sample	\bar{x} %	r %	R %
Patent kali	2,781	0,066	0,284
NPK2 (12-11-18+4+8)	0,478	0,076	0,312

11 Test report

The test report shall contain at least the following information:

- all information necessary for the complete identification of the sample;
- the test method used with reference to this document;
- the test results obtained;
- date of sampling and sampling procedure (if known);
- date when the analysis was finished;
- whether the requirement of the repeatability limit has been fulfilled;
- all operating details not specified in this document, or regarded as optional, together with details of any incidents occurred when performing the method, which might have influenced the test result(s).

[SIST EN 16195:2013](https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013)

<https://standards.iteh.ai/catalog/standards/sist/c5889540-4d01-4292-8496-2572050618e2/sist-en-16195-2013>