



**SLOVENSKI STANDARD**  
**SIST EN ISO 11073-30400:2013**  
**01-januar-2013**

---

**Zdravstvena informatika - Komunikacija medicinskih naprav na mestu oskrbe - 30400. del: Profil vmesnika - Ožičeni ethernet (ISO 11073-30400:2012)**

Health informatics - Point-of-care medical device communication - Part 30400: Interface profile - Cabled Ethernet (ISO 11073-30400:2012)

**iTeh STANDARD PREVIEW**

Informatique de santé - Communication entre dispositifs médicaux sur le site des soins - Partie 30400: Profil d'interface - Ethernet câblé (ISO 11073-30400:2012)

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/9f123-52a2-4232-96b4-2c85bd1c5261/sist-en-iso-11073-30400-2013)

**Ta slovenski standard je istoveten z: EN ISO 11073-30400:2012**

---

**ICS:**

35.240.80	Uporabniške rešitve IT v zdravstveni tehniki	IT applications in health care technology
-----------	--	---

**SIST EN ISO 11073-30400:2013**                      **en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013>

EUROPEAN STANDARD

EN ISO 11073-30400

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2012

ICS 35.240.80

English Version

## Health informatics - Point-of-care medical device communication - Part 30400: Interface profile - Cabled Ethernet (ISO 11073- 30400:2012)

Informatique de santé - Communication entre dispositifs  
médicaux sur le site des soins - Partie 30400: Profil  
d'interface - Ethernet câblé (ISO 11073-30400:2012)

This European Standard was approved by CEN on 20 October 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

**Contents**

Page

Foreword.....3

**iTeh STANDARD PREVIEW  
(standards.iteh.ai)**

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdf3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdf3261/sist-en-iso-11073-30400-2013>

## Foreword

This document (EN ISO 11073-30400:2012) has been prepared by Technical Committee ISO/TC 215 "Health informatics" in collaboration with Technical Committee CEN/TC 251 "Health informatics" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2013, and conflicting national standards shall be withdrawn at the latest by May 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## iTeh STANDARD PREVIEW Endorsement notice (standards.iteh.ai)

The text of ISO 11073-30400:2012 has been approved by CEN as a EN ISO 11073-30400:2012 without any modification.

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdf3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdf3261/sist-en-iso-11073-30400-2013>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013>

INTERNATIONAL STANDARD ISO/IEEE  
11073-30400

First edition  
2012-11-01

---

---

**Health informatics — Point-of-care  
medical device communication —  
Part 30400:  
Interface profile — Cabled Ethernet**

*Informatique de santé — Communication entre dispositifs médicaux sur  
le site des soins*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**  
*Partie 30400: Profil d'interface — Ethernet câblé*

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdf3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdf3261/sist-en-iso-11073-30400-2013>



Reference number  
ISO/IEEE 11073-30400:2012(E)

© ISO 2012  
© IEEE 2012

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013>



### **COPYRIGHT PROTECTED DOCUMENT**

© ISO 2012

© IEEE 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO or IEEE at the respective address below.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Institute of Electrical and Electronics Engineers, Inc.  
3 Park Avenue, New York • NY 10016-5997, USA  
E-mail [stds.ipr@ieee.org](mailto:stds.ipr@ieee.org)  
Web [www.ieee.org](http://www.ieee.org)

Published in Switzerland



## Contents

1. Overview .....	1
1.1 Scope .....	1
1.2 Purpose .....	1
2. Normative references.....	2
3. Definitions, acronyms, and abbreviations .....	2
3.1 Definitions .....	2
3.2 Acronyms and abbreviations .....	4
4. Clinical connectivity context.....	5
4.1 Clinical Point-of-Care deployment diagram.....	5
4.2 Use of normative references .....	6
4.3 High-level intent .....	6
4.4 Mapping “PoC reference points” to “high-level intent” .....	7
4.5 Compliance with other standards.....	8
5. Marking and cabling.....	8
5.1 Port marking.....	8
5.2 Cable and connector marking.....	9
5.3 Cabling requirements.....	9
6. IEEE 11073-30400 feature group definition .....	10
6.1 Section-level summary .....	10
6.2 Clause/annex level mapping.....	10
Annex A (informative) Bibliography .....	30
Annex B (informative) IEEE list of participants .....	31

## ISO/IEEE 11073-30400:2012(E)

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

IEEE Standards documents are developed within the IEEE Societies and the Standards Coordinating Committees of the IEEE Standards Association (IEEE-SA) Standards Board. The IEEE develops its standards through a consensus development process, approved by the American National Standards Institute, which brings together volunteers representing varied viewpoints and interests to achieve the final product. Volunteers are not necessarily members of the Institute and serve without compensation. While the IEEE administers the process and establishes rules to promote fairness in the consensus development process, the IEEE does not independently evaluate, test, or verify the accuracy of any of the information contained in its standards.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is called to the possibility that implementation of this standard may require the use of subject matter covered by patent rights. By publication of this standard, no position is taken with respect to the existence or validity of any patent rights in connection therewith. ISO/IEEE is not responsible for identifying essential patents or patent claims for which a license may be required, for conducting inquiries into the legal validity or scope of patents or patent claims or determining whether any licensing terms or conditions provided in connection with submission of a Letter of Assurance or a Patent Statement and Licensing Declaration Form, if any, or in any licensing agreements are reasonable or non-discriminatory. Users of this standard are expressly advised that determination of the validity of any patent rights, and the risk of infringement of such rights, is entirely their own responsibility. Further information may be obtained from ISO or the IEEE Standards Association.

ISO/IEEE 11073-30400 was prepared by the IEEE 11073 Standards Committee of the IEEE Engineering in Medicine and Biology Society (as IEEE Std 11073-30400-2010). It was adopted by Technical Committee ISO/TC 215, *Health informatics*, in parallel with its approval by the ISO member bodies, under the “fast-track procedure” defined in the Partner Standards Development Organization cooperation agreement between ISO and IEEE. IEEE is responsible for the maintenance of this document with participation and input from ISO member bodies.

ISO/IEEE 11073 consists of the following parts, under the general title *Health informatics — Personal health device communication* (text in parentheses gives a variant of subtitle):

- *Part 10101: (Point-of-care medical device communication) Nomenclature*
- *Part 10201: (Point-of-care medical device communication) Domain information model*
- *Part 10404: Device specialization — Pulse oximeter*
- *Part 10407: Device specialization — Blood pressure monitor*
- *Part 10408: Device specialization — Thermometer*
- *Part 10415: Device specialization — Weighing scale*

- Part 10417: Device specialization — Glucose meter
- Part 10420: Device specialization — Body composition analyzer
- Part 10421: Device specialization — Peak expiratory flow monitor (peak flow)
- Part 10471: Device specialization — Independent living activity hub
- Part 10472: Device specialization — Medication monitor
- Part 20101: (Point-of-care medical device communication) Application profiles — Base standard
- Part 20601: Application profile — Optimized exchange protocol
- Part 30200: (Point-of-care medical device communication) Transport profile — Cable connected
- Part 30300: (Point-of-care medical device communication) Transport profile — Infrared wireless
- Part 30400: (Point-of-care medical device communication) Interface profile — Cabled Ethernet
- Part 90101: (Point-of-care medical device communication) Analytical instruments — Point-of-care test
- Part 91064: (Standard communication protocol) Computer-assisted electrocardiography
- Part 92001: (Medical waveform format) — Encoding rules

ITeH STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013)

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2e83bdfc3261/sist-en-iso-11073-30400-2013>

## Introduction

This introduction is not part of IEEE Std 11073-30400-2010, Health informatics—Point-of-Care medical device communication—Part 30400: Interface profile—Cabled Ethernet.

ISO/IEEE 11073 standards enable communication between medical devices and external computer systems. They provide automatic and detailed electronic data capture of patient vital signs information and device operational data. The primary goals are as follows:

- To provide real-time plug-and-play interoperability for patient-connected medical devices
- To facilitate the efficient exchange of vital signs and medical device data, acquired at the PoC, in all health care environments

“Real time” means that data from multiple devices can be retrieved, time correlated, and displayed or processed in fractions of a second. “Plug and play” means that all the clinician has to do is make the connection between devices. The devices automatically detect, configure, and initiate communication without any other human interaction.

“Efficient exchange of medical device data” means that information that is captured at the Point of Care (e.g., patient vital signs data) can be archived, retrieved, and processed by many different types of applications without extensive software and equipment support, and without needless loss of information. The standards are especially targeted at acute and continuing care devices, such as patient monitors, ventilators, infusion pumps, electrocardiogram (ECG) devices, and so on. They comprise a family of standards that can be layered together to provide connectivity optimized for the specific devices being interfaced.

[SIST EN ISO 11073-30400:2013](https://standards.iteh.ai/catalog/standards/sist/fde9423-52a2-4232-96b4-2c650dc5261/sist-en-iso-11073-30400-2013)

This standard defines a communications interface profile. This profile is for a cable-connected, Ethernet-based local area network (LAN) for the interconnection of medical devices.

Specifically, this standard calls out layers 1 and 2 of the Open Systems Interconnection (OSI) reference model (physical and data link layers) communications services and protocols, as implemented in IEEE Std 802.3-2008,<sup>a</sup> that are appropriate for the medical communications environment. This standard is one part of the family of ISO/IEEE 11073 series of standards. It is compatible with the upper layer ISO/IEEE 11073 standards. It is expected that this standard will be combined, as appropriate, with other standards from the ISO/IEEE 11073 series.

The primary users of this standard are technical personnel who are creating or interfacing with a medical communications system. Familiarity with the ISO/IEEE 11073 family of standards is recommended. Familiarity with communications and networking technologies is also recommended.

<sup>a</sup> Information on references can be found in Clause 2.

# Health informatics — Point-of-care medical device communication —

## Part 30400: Interface profile — Cabled Ethernet

**IMPORTANT NOTICE:** This standard is not intended to ensure safety, security, health, or environmental protection. Implementers of the standard are responsible for determining appropriate safety, security, environmental, and health practices or regulatory requirements.

This IEEE document is made available for use subject to important notices and legal disclaimers. These notices and disclaimers appear in all publications containing this document and may be found under the heading “Important Notice” or “Important Notices and Disclaimers Concerning IEEE Documents.” They can also be obtained on request from IEEE or viewed at <http://standards.ieee.org/IPR/disclaimers.html>.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

### 1 Overview

#### 1.1 Scope

SIST EN ISO 11073-30400:2013

<https://standards.iteh.ai/catalog/standards/sist/fde9f423-52a2-4232-96b4-2106e5261881/iso-11073-30400-2013>

This document focuses on the application of the Ethernet family (IEEE Std 802.3™-2008<sup>1</sup>) of protocols for use in medical device communication. The scope is limited to referencing the appropriate Ethernet family specifications and to calling out any specific special needs or requirements of the ISO/IEEE 11073 environment, with a particular focus on easing interoperability and controlling costs.

#### 1.2 Purpose

This standard defines a comprehensive set of protocols consistent with the ISO/IEEE 11073 and Ethernet family of protocols for common use by medical devices in networked operating contexts. By providing this standard, the ISO/IEEE 11073 design goal to provide real-time plug-and-play interoperability will be extended to a broad set of network interfaces.

---

<sup>1</sup> Information on references can be found in Clause 2.