

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical devices – Application of usability engineering to medical devices

Dispositifs médicaux – Application de l'ingénierie de l'aptitude à l'utilisation aux dispositifs médicaux

[IEC 62366:2007](https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007)

<https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2007 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

▪ Catalogue of IEC publications: www.iec.ch/searchpub
The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

▪ IEC Just Published: www.iec.ch/online_news/justpub
Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

▪ Electropedia: www.electropedia.org
The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

▪ Customer Service Centre: www.iec.ch/webstore/custserv
If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch
Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

▪ Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm
Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

▪ Just Published CEI: www.iec.ch/online_news/justpub
Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

▪ Electropedia: www.electropedia.org
Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

▪ Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm
Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch
Tél.: +41 22 919 02 11
Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical devices – Application of usability engineering to medical devices

Dispositifs médicaux – Application de l'ingénierie de l'aptitude à l'utilisation aux dispositifs médicaux

[IEC 62366:2007](https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007)

<https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

XD

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 * Scope	7
2 Normative references	7
3 Terms and definitions	7
4 * Principles	11
4.1 General requirements.....	11
4.1.1 * USABILITY ENGINEERING PROCESS	11
4.1.2 RESIDUAL RISK.....	11
4.1.3 Information for SAFETY	12
4.2 * USABILITY ENGINEERING FILE	12
4.3 Scaling of the USABILITY ENGINEERING effort.....	12
5 * USABILITY ENGINEERING PROCESS.....	12
5.1 * Application specification.....	12
5.2 * Frequently used functions	13
5.3 Identification of HAZARDS and HAZARDOUS SITUATIONS related to USABILITY	13
5.3.1 Identification of characteristics related to SAFETY.....	13
5.3.2 * Identification of known or foreseeable HAZARDS and HAZARDOUS SITUATIONS.....	14
5.4 PRIMARY OPERATING FUNCTIONS	14
5.5 * USABILITY SPECIFICATION	15
5.6 USABILITY VALIDATION plan.....	15
5.7 * USER INTERFACE design and implementation.....	16
5.8 * USABILITY VERIFICATION.....	16
5.9 * USABILITY VALIDATION.....	17
6 * ACCOMPANYING DOCUMENT	17
7 * Training and materials for training.....	18
Annex A (informative) General guidance and rationale.....	19
Annex B (informative) Categories of USER action.....	31
Annex C (informative) Examples of USE ERRORS, ABNORMAL USE and possible causes.....	33
Annex D (informative) Guidance on the USABILITY ENGINEERING PROCESS.....	36
ANNEX E (informative) Questions that can be used to identify MEDICAL DEVICE characteristics associated with USABILITY that could impact on SAFETY.....	60
ANNEX F (informative) Examples of possible USABILITY related HAZARDOUS SITUATIONS.....	64
Annex G (informative) USABILITY goals: Illustrative example for a home parenteral infusion pump	67
ANNEX H (informative) Sample USABILITY SPECIFICATION and its inputs	77
Annex I (informative) Recommended reading list	87
Annex J (informative) Reference to the essential principles	95
Bibliography.....	96
Index of defined terms	98

Figure A.1 – A comparison of the RISK MANAGEMENT PROCESS (ISO 14971:2007) and the USABILITY ENGINEERING PROCESS (IEC 62366)	24
Figure B.1 – Categories of foreseeable USER action	32
Figure D.1 – A USER INTERFACE design cycle	39
Figure D.2 – Bubble diagram of the conceptual model of a physiological monitor	52
Figure F.1 – Pictorial representation of the relationship of HAZARD, sequence of events, HAZARDOUS SITUATION and HARM	65
Table D.1 – Sample of design flaws and associated USE ERRORS	37
Table D.2 – Mapping of Figure D.1 to the subclauses of this International Standard	39
Table D.3 – Examples of USER INTERFACE requirements	42
Table D.4 – Typical deliverables	47
Table D.5 – Examples of objective USABILITY goals	50
Table D.6 – Examples of subjective USABILITY goals	50
Table D.7 – Examples of USER INTERFACE modelling techniques	53
Table D.8 – Characteristics of a typical USABILITY testing effort	53
Table F.1 – Glossary of relevant RISK MANAGEMENT terms	64
Table F.2 – Examples of HARM due to USABILITY related HAZARDS	65
Table G.1 – Power on/off	70
Table G.2 – Program pump	70
Table G.3 – Start/stop infusion	71
Table G.4 – Monitor infusion status	72
Table G.5 – Install and change set	72
Table G.6 – Priming	73
Table G.7 – Respond to and inactivate ALARM SIGNALS ^a	73
Table G.8 – Lockouts	74
Table G.9 – Power management	74
Table G.10 – Preventative and routine maintenance	75
Table G.11 – Basic operation	76
Table G.12 – Advanced functions	76
Table J.1 – Correspondence between this document and the essential principles	95

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL DEVICES –
APPLICATION OF USABILITY ENGINEERING
TO MEDICAL DEVICES**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any medical device declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62366 has been prepared by a joint working group of subcommittee 62A: Common aspects of electrical medical equipment used in medical practice, of IEC technical committee 62: Electrical medical equipment in medical practice and technical committee ISO/TC 210: Quality management and corresponding general aspects for medical devices.

It is published as double logo standard.

The text of this standard is based on the following documents:

FDIS	Report of voting
62A/574/FDIS	62A/579/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 20 P-members out of 20 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this International Standard, the following print types are used:

- Requirements and definitions: roman type.
- *Means to assess compliance: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type
- TERMS DEFINED IN CLAUSE 3 OR AS NOTED: SMALL CAPITALS.

The requirements are followed by means to assess compliance.

Clause and subclauses for which a rationale is provided in informative Annex A are marked with an asterisk (*).

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[IEC 62366:2007](https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007)

<https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007>

INTRODUCTION

Medical practice is increasingly using MEDICAL DEVICES for observation and treatment of PATIENTS. USE ERRORS caused by inadequate MEDICAL DEVICE USABILITY have become an increasing cause for concern. Many of the MEDICAL DEVICES developed without applying a USABILITY ENGINEERING PROCESS are non-intuitive, difficult to learn and to use. As healthcare evolves, less skilled USERS including PATIENTS themselves are now using MEDICAL DEVICES and MEDICAL DEVICES are becoming more complicated. In simpler times, the USER of a MEDICAL DEVICE might be able to cope with an ambiguous, difficult-to-use USER INTERFACE. The design of a usable MEDICAL DEVICE is a challenging endeavour, yet many organizations treat it as if it were just “common sense”. The design of the USER INTERFACE to achieve adequate (safe) USABILITY requires a very different skill set than that of the technical implementation of that interface.

The USABILITY ENGINEERING PROCESS is intended to achieve reasonable USABILITY, which in turn is intended to minimise USE ERRORS and to minimise use-associated RISKS. Some, but not all, forms of incorrect use are amenable to control by the MANUFACTURER. The USABILITY ENGINEERING PROCESS is related to the RISK MANAGEMENT PROCESS as indicated in Figure A.1.

This International Standard describes a USABILITY ENGINEERING PROCESS, and provides guidance on how to implement and execute the PROCESS to provide SAFETY in MEDICAL DEVICES. It is intended to be useful not only for MANUFACTURERS of MEDICAL DEVICES, but also for technical committees responsible for the preparation of particular MEDICAL DEVICE standards.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

IEC 62366:2007

<https://standards.iteh.ai/catalog/standards/sist/60c07687-362a-49b6-80d0-f4c0898d78b4/iec-62366-2007>

MEDICAL DEVICES – APPLICATION OF USABILITY ENGINEERING TO MEDICAL DEVICES

1 * Scope

This International Standard specifies a PROCESS for a MANUFACTURER to analyse, specify, design, VERIFY and VALIDATE USABILITY, as it relates to SAFETY of a MEDICAL DEVICE. This USABILITY ENGINEERING PROCESS assesses and mitigates RISKS caused by USABILITY problems associated with CORRECT USE and USE ERRORS, i.e. NORMAL USE. It can be used to identify but does not assess or mitigate RISKS associated with ABNORMAL USE.

NOTE For the purposes of this standard, USABILITY (see 3.17) is limited to characteristics of the USER INTERFACE.

If the USABILITY ENGINEERING PROCESS detailed in this International Standard has been complied with and the acceptance criteria documented in the USABILITY VALIDATION plan have been met (see 5.9), then the RESIDUAL RISKS, as defined in ISO 14971, associated with USABILITY of a MEDICAL DEVICE are presumed to be acceptable, unless there is OBJECTIVE EVIDENCE to the contrary (see 4.1.2).

This International Standard does not apply to clinical decision-making relating to the use of a MEDICAL DEVICE.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE Informative references are listed in the bibliography beginning on page 96.

ISO 14971:2007, *Medical devices – Application of risk management to medical devices*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in ISO 14971:2007 and the following apply.

NOTE An index of defined terms is found beginning on page 98.

3.1

ABNORMAL USE

intentional act or intentional omission of an act by the RESPONSIBLE ORGANIZATION or USER of a MEDICAL DEVICE as a result of conduct that is beyond any further reasonable means of RISK CONTROL by the MANUFACTURER

NOTE 1 See also 4.1.3 and Annex B. Examples are given in Annex C.

NOTE 2 It is possible for the PATIENT to be the USER, e.g. when the MEDICAL DEVICE is used in the PATIENT'S home.

3.2

ACCOMPANYING DOCUMENT

document accompanying a MEDICAL DEVICE and containing information for those accountable for the installation, use and maintenance of the MEDICAL DEVICE or the USER, particularly regarding SAFETY

[ISO 14971:2007, definition 2.1, modified]

3.3

ALARM LIMIT

threshold used by an ALARM SYSTEM to determine an ALARM CONDITION

[IEC 60601-1-8:2006, definition 3.3]

NOTE This term is only used in notes and informative annexes.

3.4

ALARM OFF

state of indefinite duration in which an ALARM SYSTEM or part of an ALARM SYSTEM does not generate ALARM SIGNALS

[IEC 60601-1-8:2006, definition 3.4]

NOTE This term is only used in notes and informative annexes.

3.5

ALARM SIGNAL

type of signal generated by the ALARM SYSTEM to indicate the presence (or occurrence) of an ALARM CONDITION

[IEC 60601-1-8:2006, definition 3.9]

NOTE This term is only used in notes and informative annexes.

3.6

ALARM SYSTEM

parts of the MEDICAL DEVICE that detect ALARM CONDITIONS and, as appropriate, generate ALARM SIGNALS

[IEC 60601-1-8:2006, definition 3.11, modified]

NOTE This term is only used in notes and informative annexes.

3.7

CORRECT USE

NORMAL USE without USE ERROR

3.8

EFFECTIVENESS

measure of accuracy and completeness with which USERS achieve specified goals

[ISO 9241-11:1998, definition 3.2, modified]

NOTE This is a different concept than the 'clinical effectiveness'.

3.9

EFFICIENCY

EFFECTIVENESS in relation to the resources expended

3.10

INFORMATION SIGNAL

any signal that is not an ALARM SIGNAL or a REMINDER SIGNAL

EXAMPLE 1 ECG waveform

EXAMPLE 2 SpO₂ tone

EXAMPLE 3 Fluoroscopy beam-on indication

[IEC 60601-1-8:2006, definition 3.23]

NOTE This term is only used in notes and informative annexes.

3.11

MEDICAL DEVICE

any instrument, apparatus, implement, machine, appliance, implant, in vitro reagent or calibrator, software, material or other similar or related article, intended by the MANUFACTURER to be used, alone or in combination, for human beings for one or more of the specific purpose(s) of

- diagnosis, prevention, monitoring, treatment or alleviation of disease,
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury,
- investigation, replacement, modification, or support of the anatomy or of a physiological PROCESS,
- supporting or sustaining life,
- control of conception,
- disinfection of MEDICAL DEVICES,
- providing information for medical purposes by means of in vitro examination of specimens derived from the human body,

and which does not achieve its primary intended action in or on the human body by pharmacological, immunological or metabolic means, but which can be assisted in its function by such means.

[ISO 13485:2003, definition 3.7]

3.12

* NORMAL USE

operation, including routine inspection and adjustments by any USER, and stand-by, according to the instructions for use or in accordance with generally accepted practice for those MEDICAL DEVICES provided without instructions for use

[IEC 60601-1:2005, definition 3.71, modified]

NOTE 1 USE ERROR can occur in NORMAL USE.

NOTE 2 MEDICAL DEVICES that can be used safely without instructions for use are exempted from having instructions for use by some authorities with jurisdiction.

3.13

* PATIENT

living being (person) undergoing a medical, surgical or dental procedure

[IEC 60601-1:2005, definition 3.76, modified]

3.14

* PRIMARY OPERATING FUNCTION

function that involves USER interaction that is either frequently used or related to the SAFETY of the MEDICAL DEVICE

3.15

REMINDER SIGNAL

periodic signal that reminds the USER that the ALARM SYSTEM is in an ALARM SIGNAL-inactivation state

[IEC 60601-1-8:2006, definition 3.34, modified]

NOTE This term is only used in notes and informative annexes.

3.16

RESPONSIBLE ORGANIZATION

entity accountable for the use and maintenance of a MEDICAL DEVICE or combination of MEDICAL DEVICES

NOTE 1 The accountable entity can be, for example, a hospital, an individual clinician or a lay person. In home use applications, the PATIENT, USER and RESPONSIBLE ORGANIZATION can be one and the same person.

NOTE 2 Education and training is included in "use."

[IEC 60601-1:2005, definition 3.101, modified]

3.17

*** USABILITY**

characteristic of the USER INTERFACE that establishes EFFECTIVENESS, EFFICIENCY, ease of USER learning and USER satisfaction

3.18

USABILITY ENGINEERING

application of knowledge about human behaviour, abilities, limitations, and other characteristics related to the design of tools, devices, systems, tasks, jobs, and environments to achieve adequate USABILITY

3.19

*** USABILITY ENGINEERING FILE**

set of RECORDS and other documents that are produced by the USABILITY ENGINEERING PROCESS

3.20

USABILITY SPECIFICATION

documentation defining the USER INTERFACE requirements related to USABILITY

3.21

USE ERROR

act or omission of an act that results in a different MEDICAL DEVICE response than intended by the MANUFACTURER or expected by the USER

NOTE 1 USE ERROR includes slips, lapses, and mistakes.

NOTE 2 See also Annex B and D.1.3.

NOTE 3 An unexpected physiological response of the PATIENT is not in itself considered USE ERROR.

3.22

USE SCENARIO

specified sequence of events and tasks as performed by a specified USER in a specified environment

3.23

*** USER**

person using, i.e. operating or handling, the MEDICAL DEVICE

NOTE 1 This includes, but is not limited to, cleaners, maintainers and installers.

NOTE 2 PATIENTS or other laypersons can be USERS.

3.24

*** USER INTERFACE**

means by which the USER and the MEDICAL DEVICE interact

[ANSI/AAMI/HE 74:2001, definition 3.24, modified]

NOTE The ACCOMPANYING DOCUMENT is considered part of the MEDICAL DEVICE and its USER INTERFACE.

3.25

USER PROFILE

summary of the mental, physical and demographic traits of an intended USER population, as well as any special characteristics that can have a bearing on design decisions, such as occupational skills and job requirements

3.26

VALIDATION

confirmation, through the provision of OBJECTIVE EVIDENCE, that the requirements for a specific INTENDED USE or application have been fulfilled

NOTE 1 The term “VALIDATED” is used to designate the corresponding status.

NOTE 2 The use conditions for VALIDATION can be real or simulated.

[ISO 9000:2000, definition 3.8.5]

4 * Principles

4.1 General requirements

4.1.1 * USABILITY ENGINEERING PROCESS

The MANUFACTURER shall establish, document and maintain a USABILITY ENGINEERING PROCESS to provide SAFETY for the PATIENT, USER and others related to USABILITY. The PROCESS shall address USER interactions with the MEDICAL DEVICE according to the ACCOMPANYING DOCUMENT, including, but not limited to:

- * transport;
- * storage;
- installation;
- operation;
- maintenance and repair; and
- disposal.

NOTE See also D.3.1.

Consider compliance with the requirements of this clause to exist when the criteria of the relevant inspections and tests in this International Standard are achieved.

4.1.2 RESIDUAL RISK

If the USABILITY ENGINEERING PROCESS detailed in this International Standard has been complied with and the acceptance criteria documented in the USABILITY VALIDATION plan have been met (see 5.9), then, for the purposes of ISO 14971, the RESIDUAL RISKS associated with USABILITY of the MEDICAL DEVICE shall be presumed to be acceptable, unless there is OBJECTIVE EVIDENCE to the contrary.

NOTE 1 Such OBJECTIVE EVIDENCE can subsequently originate from post-market surveillance.

NOTE 2 ISO 14971:2007, Subclause 6.6 requires that design changes resulting from the USABILITY ENGINEERING PROCESS be reviewed to determine if other HAZARDS or HAZARDOUS SITUATIONS have been generated.

NOTE 3 ISO 14971:2007, Clause 7 requires that all RESIDUAL RISK be considered when evaluating the overall RESIDUAL RISK of the MEDICAL DEVICE, including the RESIDUAL RISK associated with USABILITY of the MEDICAL DEVICE.

Compliance is checked by inspection of the USABILITY ENGINEERING FILE.

4.1.3 Information for SAFETY

If information for SAFETY is used as a RISK CONTROL measure, the MANUFACTURER shall subject this information to the USABILITY ENGINEERING PROCESS.

EXAMPLE 1 Warnings or limitation of use in the ACCOMPANYING DOCUMENT

EXAMPLE 2 Markings

Disregarding such information for SAFETY shall be considered beyond any further reasonable means of RISK CONTROL. See also Annex B.

NOTE 1 Information for SAFETY is one element in a hierarchal approach to RISK CONTROL in which the MANUFACTURER uses one or more of the following in the priority listed (see ISO 14971:2007, 6.2):

- a) inherent SAFETY by design;
- b) protective measures in the MEDICAL DEVICE itself or in the manufacturing PROCESS, e.g. ALARM SYSTEMS;
- c) information for SAFETY, e.g. warnings in the instructions for use, display of a monitored variable, training and materials for training, maintenance details.

NOTE 2 The need to include information for SAFETY can be an input to the USABILITY ENGINEERING PROCESS (e.g. imposed by some other standard) or it can be discovered during the USABILITY ENGINEERING PROCESS.

Compliance is checked by inspection of the ACCOMPANYING DOCUMENT and the USABILITY ENGINEERING FILE.

4.2 * USABILITY ENGINEERING FILE

The results of the USABILITY ENGINEERING PROCESS shall be recorded in the USABILITY ENGINEERING FILE. The RECORDS and other documents that form the USABILITY ENGINEERING FILE may form part of other documents and files.

EXAMPLE 1 MANUFACTURER'S product design file

EXAMPLE 2 RISK MANAGEMENT FILE

Compliance is checked by inspection of the USABILITY ENGINEERING FILE.

4.3 Scaling of the USABILITY ENGINEERING effort

The USABILITY ENGINEERING PROCESS may vary in form and extent based on the nature of the MEDICAL DEVICE, its intended USER and its INTENDED USE (see D.3.2). In the case of the modification of a MEDICAL DEVICE design, the USABILITY ENGINEERING PROCESS may be scaled-up or scaled-down based on the significance of the modification as determined by the results of the RISK ANALYSIS (see D.3.2.2).

NOTE 1 The MANUFACTURER should conduct iterative design and development. USABILITY ENGINEERING, including USABILITY VALIDATION, should begin early and continue through the MEDICAL DEVICE design and development lifecycle.

NOTE 2 Due to the iterative nature of the USABILITY ENGINEERING PROCESS, the activities described in Clause 5 can be carried out in any convenient order (see Clause D.2).

Compliance is checked by inspection of the USABILITY ENGINEERING FILE.

5 * USABILITY ENGINEERING PROCESS

5.1 * Application specification

The MANUFACTURER shall specify the application of the MEDICAL DEVICE in the USABILITY ENGINEERING FILE.

This specification shall include:

- * intended medical indication;
 - EXAMPLE 1 Conditions(s) or disease(s) to be screened, monitored, treated, diagnosed, or prevented
- intended PATIENT population;
 - EXAMPLE 2 age
 - EXAMPLE 3 weight
 - EXAMPLE 4 health
 - EXAMPLE 5 condition
- intended part of the body or type of tissue applied to or interacted with;
- * intended USER PROFILE;
- * intended conditions of use; and
 - EXAMPLE 6 environment including hygienic requirements
 - EXAMPLE 7 frequency of use
 - EXAMPLE 8 location
 - EXAMPLE 9 mobility
- * operating principle.

NOTE Requirements for disclosing a summary of the MEDICAL DEVICE APPLICATION SPECIFICATION in the ACCOMPANYING DOCUMENT are found in Clause 6.

Compliance is checked by inspection of the USABILITY ENGINEERING FILE.

5.2 * Frequently used functions IEC 62366:2007

The MANUFACTURER shall determine the frequently used functions that involve USER interaction with the MEDICAL DEVICE and record them in the USABILITY ENGINEERING FILE.

Compliance is checked by inspection of the USABILITY ENGINEERING FILE.

5.3 Identification of HAZARDS and HAZARDOUS SITUATIONS related to USABILITY

5.3.1 Identification of characteristics related to SAFETY

An identification of characteristics related to SAFETY (part of a RISK ANALYSIS) that focuses on USABILITY shall be performed according to ISO 14971:2007, 4.2.

During the identification characteristics related to SAFETY, the following shall be considered:

- application specification, including USER PROFILE(S) (see 5.1); and
- frequently used functions (see 5.2).

The results of this identification characteristics related to SAFETY shall be recorded in the USABILITY ENGINEERING FILE.

NOTE 1 See also D.4.3, D.4.6.4, D.5.2, and D.5.4 for a discussion of methods that might be useful in identifying characteristics related to SAFETY.

NOTE 2 See 4.3 for scaling of the USABILITY ENGINEERING effort.

Compliance is checked by inspection of the USABILITY ENGINEERING FILE.