

SLOVENSKI STANDARD

SIST EN 16187:2015

01-september-2015

Nadomešča:

SIST-TS CEN/TS 16187:2011

Živila - Določevanje fumonizina B1 in fumonizina B2 v predelani hrani na osnovi koruze za dojenčke in majhne otroke - Metoda HPLC s čiščenjem z imunoafinitetno kolono in fluorescenčno detekcijo po predkolonski derivatizaciji

Foodstuffs - Determination of fumonisin B1 and fumonisin B2 in processed maize containing foods for infants and young children - HPLC method with immunoaffinity column cleanup and fluorescence detection after pre-column derivatisation

Lebensmittel - Bestimmung von Fumonisin B1 und Fumonisin B2 in Säuglings- und Kleinkindernahrung auf Maisbasis - HPLC-Verfahren mit Reinigung an einer Immunoaffinitätssäule und Fluoreszenzdetektion nach Vorsäulenderivatierung

Denrées alimentaires - Dosage de la fumonisine B1 et de la fumonisine B2 dans les aliments pour nourrissons et jeunes enfants contenant du maïs transformé - Méthode par CLHP avec purification sur colonne d'immunoaffinité et détection de fluorescence après dérivation précolonne

Ta slovenski standard je istoveten z: EN 16187:2015

ICS:

67.060	Žita, stročnice in proizvodi iz njih	Cereals, pulses and derived products
67.230	Predpakirana in pripravljena hrana	Prepackaged and prepared foods

SIST EN 16187:2015

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 16187:2015](#)

<https://standards.iteh.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015>

EUROPEAN STANDARD

EN 16187

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2015

ICS 67.230

Supersedes CEN/TS 16187:2011

English Version

Foodstuffs - Determination of fumonisin B1 and fumonisin B2 in processed maize containing foods for infants and young children - HPLC method with immunoaffinity column cleanup and fluorescence detection after pre-column derivatisation

Denrées alimentaires - Dosage de la fumonisine B1 et de la fumonisine B2 dans les aliments pour nourrissons et jeunes enfants contenant du maïs transformé - Méthode par CLHP avec purification sur colonne d'immunoaffinité et détection de fluorescence après dérivation précolonne

Lebensmittel - Bestimmung von Fumonisin B1 und Fumonisin B2 in Säuglings- und Kleinkindernahrung auf Maisbasis - HPLC-Verfahren mit Reinigung an einer Immunoaffinitätssäule und Fluoreszenzdetektion nach Vorsäulenderivatisierung

This European Standard was approved by CEN on 7 May 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

<https://standards.iteh.ai/catalog/standards/sist/629561e7-29ef-479d-ad16->

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword.....	3
1 Scope	4
2 Normative references	4
3 Principle	4
4 Reagents	4
5 Apparatus	7
6 Procedure	8
7 Calculation.....	10
8 Precision.....	11
9 Test report	13
Annex A (informative) Typical chromatograms	14
Annex B (informative) Precision data.....	15
Annex C (informative) Comparison between the method in this document and EN 14352:2004 and EN 13585:2001 on fumonisins in maize.....	18
Bibliography.....	19

[SIST EN 16187:2015](https://standards.itech.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015)

<https://standards.itech.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015>

Foreword

This document (EN 16187:2015) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2015, and conflicting national standards shall be withdrawn at the latest by December 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16187:2011.

No technical change has been introduced during the conversion of CEN/TS 16187:2011 into this final draft European Standard.

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN 16187:2015](https://standards.iteh.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015)

<https://standards.iteh.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015>

EN 16187:2015 (E)**1 Scope**

This European Standard specifies a method for the determination of fumonisin B₁ (FB₁) and fumonisin B₂ (FB₂) in processed maize-containing foods for infants and young children by high performance liquid chromatography (HPLC) with immunoaffinity cleanup and fluorescence detection (FLD). This method has been validated in an interlaboratory study via the analysis of both naturally contaminated and spiked samples ranging from 112 µg/kg to 458 µg/kg for FB₁+FB₂, 89 µg/kg to 384 µg/kg for FB₁ and 22 µg/kg to 74 µg/kg for FB₂.

For further information on the validation, see Clause 8 and Annex B.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696:1995, *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*

3 Principle

Fumonisin is extracted from the sample with a mixture of citrate-phosphate buffer with methanol and acetonitrile. The filtered extract is diluted with water and applied to an immunoaffinity column containing antibodies specific to fumonisin. Fumonisin is eluted from the column with methanol and water and quantified by HPLC-FLD with pre-column derivatisation with *o*-phthalaldehyde (OPA) reagent.

[SIST EN 16187:2015](https://standards.iteh.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015)

<https://standards.iteh.ai/catalog/standards/sist/629561e7-29ef-479d-ad16-0954777940af/sist-en-16187-2015>

4 Reagents

Use only reagents of recognized analytical grade and water complying with grade 1 of EN ISO 3696:1995, unless otherwise specified. Solvents shall be of quality for HPLC analysis, unless otherwise specified. Commercially available solutions with equivalent properties to those listed may be used.

WARNING — Dispose of waste solvents according to applicable environmental rules and regulations. Decontamination procedures for laboratory wastes have been reported by the International Agency for Research on Cancer (IARC), see [1].

4.1 Acetonitrile

WARNING — Acetonitrile is hazardous and samples shall be shaken using an orbital shaker which is housed within a fume cupboard. After shaking, samples shall be filtered inside a fume cupboard.

4.2 Methanol**4.3 O-phthalaldehyde (OPA)****4.4 Citric acid solution**, substance concentration $c(\text{C}_6\text{H}_8\text{O}_7 \cdot \text{H}_2\text{O}) = 0,1 \text{ mol/l}$

Dissolve 21,0 g of $\text{C}_6\text{H}_8\text{O}_7 \cdot \text{H}_2\text{O}$ in water and dilute to 1 l.

4.5 Disodium hydrogen phosphate solution, $c(\text{Na}_2\text{HPO}_4) = 0,2 \text{ mol/l}$

Dissolve 28,4 g of Na_2HPO_4 in water and dilute to 1 l.

4.6 2-mercaptoethanol

4.7 Citrate-phosphate buffer solution

Mix 1 part per volume of citric acid solution (4.4) with 1 part per volume of disodium hydrogen phosphate solution (4.5).

4.8 Extraction solvent

Mix 2 parts per volume of citrate-phosphate buffer solution (4.7) with 1 part per volume of methanol (4.2) and 1 part per volume of acetonitrile (4.1).

4.9 Glacial acetic acid

4.10 Phosphate buffered saline (PBS) solution, $c(\text{NaCl}) = 137 \text{ mmol/l}$, $c(\text{KCl}) = 2,7 \text{ mmol/l}$, $c(\text{phosphate buffer}) = 10 \text{ mmol/l}$, $\text{pH} = 7,4$ at $T = 25 \text{ }^\circ\text{C}$

Dissolve one tablet of commercially available PBS material in 200 ml of water.

4.11 Mixture of acetonitrile and water A

Mix 1 part per volume of acetonitrile (4.1) with 1 part per volume of water. Use this solvent to prepare spiking solutions.

4.12 Mixture of acetonitrile and water B

Mix 3 parts per volume of acetonitrile (4.1) with 7 parts per volume of water. Use this solvent to prepare calibration solutions and to redissolve dried extracts from immunoaffinity cleanup.

4.13 Sodium tetraborate solution, $c(\text{Na}_2\text{B}_4\text{O}_7 \cdot 10\text{H}_2\text{O}) = 0,1 \text{ mol/l}$

Dissolve 3,8 g of $\text{Na}_2\text{B}_4\text{O}_7 \cdot 10\text{H}_2\text{O}$ in 100 ml of water.

4.14 OPA reagent solution

Dissolve 40 mg of OPA (4.3) in 1 ml of methanol (4.2) and dilute with 5 ml of sodium tetraborate solution (4.13). Add 50 μl of 2-mercaptoethanol (4.6) and mix for 1 min. This reagent solution is stable for up to one week at room temperature in the dark in a capped amber vial.

4.15 HPLC mobile phase

4.15.1 HPLC mobile phase A

Mix 30 parts per volume of acetonitrile (4.1) with 69 parts per volume of water and 1 part per volume of glacial acetic acid (4.9).

4.15.2 HPLC mobile phase B

Mix 60 parts per volume of acetonitrile (4.1) with 39 parts per volume of water and 1 part per volume of glacial acetic acid (4.9).

4.16 Immunoaffinity column

The immunoaffinity column shall contain antibodies raised against FB_1 and FB_2 . The column shall have a capacity of not less than 5 μg of fumonisins and shall give a recovery of not less than 80 % for the sum of FB_1 and FB_2 when applied as a standard solution in PBS containing 5 μg of fumonisins. The columns shall be warmed up to room temperature before use. With these columns the working range of the method reaches

EN 16187:2015 (E)

5 000 µg/kg of total fumonisins. The use of columns with a capacity less than 5 µg of fumonisins will reduce the working range of the method.

4.17 Certified standard solution of fumonisin B₁ (FB₁), mass concentration $\rho(\text{FB}_1) = 50 \text{ µg/ml}$ in a mixture of 1 part per volume of acetonitrile and 1 part per volume of water (e.g. Biopure RK 002003 ¹⁾ or equivalent)

4.18 Certified standard solution of fumonisin B₂ (FB₂), $\rho(\text{FB}_2) = 50 \text{ µg/ml}$ in a mixture of 1 part per volume of acetonitrile and 1 part per volume of water (e.g. Biopure RK 002004 ¹⁾ or equivalent)

WARNING — Fumonisins are nephrotoxic, hepatotoxic and carcinogenic to rats and mice and classified as possible human carcinogen by IARC. These compounds should be treated with extreme caution. Gloves and safety glasses shall be worn at all times and all standard and sample preparation stages shall be carried out in a fume cupboard.

4.19 Mixed FB₁ and FB₂ stock solution

Prepare a mixed FB₁ and FB₂ stock solution by pipetting 2 000 µl of the FB₁ certified standard solution (4.17) and 500 µl of the FB₂ certified standard solution (4.18) into a vial. Cap the vial and shake well to obtain a stock solution containing 40,0 µg/ml of FB₁ and 10,0 µg/ml of FB₂.

4.20 Diluted mixed FB₁ and FB₂ stock solution

Pipette 500 µl of the mixed stock solution (4.19) into a 10 ml calibrated volumetric flask. Fill up to the mark with the mixture of acetonitrile water B (4.12) and shake well to obtain a diluted mixed stock solution containing 2,0 µg/ml of FB₁ and 0,5 µg/ml of FB₂. Store the solution at less than 4 °C. This solution is stable for at least 6 months at these conditions.

4.21 Mixed FBs calibration solutions for HPLC

Prepare five HPLC mixed calibration solutions in 5 ml calibrated volumetric flasks by further diluting the diluted mixed FBs stock solution (4.20) according to Table 1. Make up each calibration solution to volume (5 ml) with the mixture of acetonitrile and water B (4.12) and mix well. Store the solution at less than 4 °C. This solution is stable for at least 6 months at these conditions.

Table 1 — Preparation of mixed FBs calibration solutions for HPLC

HPLC calibration solution	Diluted mixed FBs stock solution (4.20) µl	Final concentration of mixed FBs calibration solutions (4.21)		Sample equivalent levels of FB ₁ and FB ₂ ^a	
		FB ₁ µg/ml	FB ₂ µg/ml	FB ₁ µg/kg	FB ₂ µg/kg
1	100	0,04	0,01	20,0	5,0
2	200	0,08	0,02	40,0	10,0
3	400	0,16	0,04	80,0	20,0
4	1 000	0,40	0,10	200,0	50,0
5	2 400	0,96	0,24	480,0	120,0

^a For a sample extract reconstituted in 500 µl of the mixture of acetonitrile and water B (4.12).

¹⁾ Biopure RK 002003 and RK 002004 are examples of suitable products available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of these products. Equivalent products may be used if they can be shown to lead to the same results.

5 Apparatus

Usual laboratory glassware and equipment and, in particular, the following:

- 5.1 **Analytical balance**, capable of weighing to 0,000 1 g.
- 5.2 **Laboratory balance**, capable of weighing to 0,1 g.
- 5.3 **Thermostated water bath**.
- 5.4 **Conical flasks**, of 250 ml capacity with screw caps.
- 5.5 **Orbital shaker**.
- 5.6 **Centrifuge**, capable of a centrifugal force up to 3 000 *g*.
- 5.7 **Centrifuge bottles**, of 250 ml capacity with screw caps.
- 5.8 **Calibrated microliter pipettes or microliter syringes**, of 100 μ l, 200 μ l or 1 000 μ l capacity.
- 5.9 **Displacement pipettes**, of 5 ml, 10 ml or 25 ml capacity.
- 5.10 **Vacuum manifold**, to accommodate immunoaffinity columns (4.16).
- 5.11 **Reservoirs** (of 25 ml capacity and attachments to fit to columns).
- 5.12 **Vacuum pump**.
- 5.13 **Filter paper**, e.g. qualitative, strong, fast flow, 24 cm diameter, 30 μ m pore size, prefolded or equivalent.
- 5.14 **Glass microfibre filter paper**, e.g. 1,6 μ m pore size or equivalent.
- 5.15 **Heating block with nitrogen or air gas supply**.
- 5.16 **Vials**, of 4 ml to 12 ml capacity with screw caps.
- 5.17 **HPLC autosampler vials**, of 1,8 ml capacity with caps.
- 5.18 **Glass flat bottom vial insert**, of 250 μ l volume capacity.
- 5.19 **Vortex mixer**, or equivalent.
- 5.20 **HPLC apparatus**, comprising the following:
 - 5.20.1 **Injection system**.
 - 5.20.2 **Mobile phase pump**, (binary, ternary or quaternary pump), capable of generating a binary gradient at 1 ml/min.
 - 5.20.3 **Autosampler**, capable of performing automated pre-column derivatisation with OPA reagent according to 6.4.1.
 - 5.20.4 **Analytical reverse-phase HPLC separating column**, e.g. C18 reverse-phase column, 150 mm x 4,6 mm, 5 μ m preceded by a suitable pre-column or guard filter, which provides acceptable retention and resolution for FB₁ and FB₂.

EN 16187:2015 (E)

Waters C18 SymmetryShield™²⁾, Agilent Zorbax SB-C18²⁾ or similar have been found to be suitable.

5.20.5 Column oven, capable to operate at 20 °C.

NOTE C18 reverse-phase columns can also operate at ambient temperatures.

5.20.6 Fluorescence detector, fitted with a flow cell and suitable for measurements with excitation wavelength of 335 nm and emission wavelength of 440 nm.

5.20.7 Recorder, integrator or computer based data processing system.

5.21 Linear gradient settings

The gradient conditions are given in Table 2.

Table 2 — Gradient conditions

Time min	Flow rate ml/min	Mobile phase A %	Mobile phase B %
0,00	1,00	60	40
5,00	1,00	60	40
26,00	1,00	12	88
29,00	1,00	12	88
29,10	1,00	100	100
30,00	1,00	-	100
30,10	1,00	60	40
38,00	1,00	60	40

The gradient programme above has proven to give acceptable resolution for FB₁ and FB₂ by using a Waters C18 SymmetryShield™²⁾ column. The use of a different column may necessitate adjustment of these conditions until acceptable resolution is achieved.

6 Procedure**6.1 Extraction**

The sample should be finely ground to pass through a 0,5 mm sieve and thoroughly mixed to ensure complete homogenization.

Weigh, to the nearest 0,1 g, a 20 g test portion of the ground sample into a conical flask (5.4). Add 100 ml of extraction solvent (4.8) and close the flask with screw cap. Heat at 55 °C (static condition) in the thermostated water bath (5.3) for 60 min. Put the flask on the orbital shaker (5.5) and shake for 30 min at room temperature. Transfer the sample into a centrifuge bottle (5.7) and centrifuge (5.6) at 3 000 g for 15 min. Filter through a filter paper (5.13) and collect 10 ml filtrate. Dilute the 10 ml filtrate with 40 ml of water and mix. Filter through a

²⁾ Waters C18 SymmetryShield™ and Agilent Zorbax SB-C18 are examples of suitable products available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of these products. Equivalent products may be used if they can be shown to lead to the same results.

microfibre filter (5.14), collect 25 ml filtrate (equivalent to 1 g of test sample) and proceed rapidly to immunoaffinity cleanup (i.e. within the next hour).

6.2 Immunoaffinity column cleanup

Accommodate the immunoaffinity column (4.16) to the vacuum manifold (5.10) and attach the reservoir. Do not empty storage solution from column. Add the 25 ml of the diluted filtered extract to the reservoir. Elute the extract at a flow rate of about one drop per second. Wash the column with 10 ml of PBS solution (4.10) followed by 2 ml of water at a flow rate of about one drop per second. Pass air or nitrogen for 5 s through the column to expel water. Place a vial or tube under the column.

Elute fumonisins with two times 1 ml of methanol (4.2) and 2 ml water in the following way.

Add 1 ml of methanol on the column. Press the liquid through the column with slight over pressure of air (e.g. with Pasteur balloon or syringe) until the first drop comes out. Close the stopcock immediately. Wait for 1 min. Open the stopcock. Allow elution to take place by gravity force, until the liquid reaches the top layer of the column. Close the stopcock immediately and wait for 1 min. Add the second 1 ml of methanol. Open the stopcock and let half of the liquid (approx. 500 µl) flow through the column by gravity force. Close the stopcock and wait 1 min. Open the stopcock and elute the remaining liquid by gravity force. Pass air or nitrogen through the column to remove the last methanol from the column.

Add 2 ml of water and elute the liquid through the column by gravity force. Pass air or nitrogen through the column to remove the last liquid from the column. Evaporate the eluates to dryness under a stream of air or nitrogen at ca 55 °C (5.15). Redissolve the residue in 500 µl of the mixture of acetonitrile and water B (4.12). Cap the vial and shake on a vortex mixer (5.19) for 30 s, making sure the lower part of the vial is thoroughly rinsed by the solvent and store at 4 °C until HPLC analysis that should be performed within one week.

6.3 Spiking procedure

To determine recovery, spike a fumonisin-free representative ground baby food material with a spiking solution prepared by appropriately diluting the mixed FB1 and FB2 stock solution (4.19) with the mixture of acetonitrile and water A (4.11). No more than 1 ml of spiking solution should be added to each test portion. The fumonisin mass concentration of the spiking solution should be prepared according to the spiking level chosen. Leave the spiked sample to stand overnight to ensure evaporation of the solvent.

6.4 Derivatisation and HPLC determination

6.4.1 Automated pre-column derivatisation programme

Transfer 110 µl of extract to an HPLC autosampler vial containing a glass flat bottom vial insert (5.18). Put the vial containing the sample extract in the autosampler.

The following conditions have proven to produce satisfying results.

Aspirate an air segment of 5 µl to separate the wash solvent in the buffer tubing from the OPA reagent (4.14). Aspirate 50 µl of OPA reagent to flush the tubing and needle. Unload the syringe to the syringe-waste position. Aspirate 110 µl of OPA reagent and dispense it to the vial insert containing 110 µl of sample extract or calibration solution. Rinse the buffer tubing and needle with the mixture of acetonitrile and water A (4.11). Aspirate 55 µl of derivatised solution and dispense it back into the vial insert; repeat this four times. Rinse the buffer tubing and needle with wash solvent (4.11). Wait for 20 s. Inject 50 µl in full loop mode. Rinse the buffer tubing and needle with wash solvent (4.11) five times. This automated derivatisation procedure takes 2,5 min.

NOTE Different pre-column derivatisation conditions (volumes, derivatisation programme, etc.) could be used according to your autosampler, bearing in mind that the signals (peak area or height) of FB₁ or FB₂ in the sample extract should fall within the calibration range.