
**Water quality — Determination of
ammonium nitrogen — Method by flow
analysis (CFA and FIA) and spectrometric
detection**

*Qualité de l'eau — Dosage de l'azote ammoniacal — Méthode par
analyse en flux (CFA et FIA) et détection spectrométrique*

iteh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 11732:2005

<https://standards.iteh.ai/catalog/standards/iso/8deb819b-6ba2-4db1-b9fc-99f7f6c3f65f/iso-11732-2005>



Reference number
ISO 11732:2005(E)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 11732:2005

<https://standards.iteh.ai/catalog/standards/iso/8deb819b-6ba2-4db1-b9fc-99f7f6c3f65f/iso-11732-2005>

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Determination of ammonium nitrogen by flow injection analysis (FIA) and spectrometric detection	1
4 Determination of ammonium nitrogen by continuous flow analysis (CFA) and spectrometric detection	7
5 Calculation	10
6 Precision	10
7 Test report	11
Annex A (informative) Examples of flow analysis systems	12
Annex B (informative) Precision data	16
Bibliography	18

IteH Standards
 (https://standards.iteh.ai)
 Document Preview

ISO 11732:2005

<https://standards.iteh.ai/catalog/standards/iso/8deb819b-6ba2-4db1-b9fc-99f7f6c3f65f/iso-11732-2005>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11732 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

This second edition cancels and replaces the first edition (ISO 11732:1997), which has been technically revised.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

ISO 11732:2005

<https://standards.iteh.ai/catalog/standards/iso/8deb819b-6ba2-4db1-b9fc-99f7f6c3f65f/iso-11732-2005>

Introduction

Methods using flow analysis are automating wet chemical procedures and are therefore particularly suitable for the processing of large sample series at a high analysis frequency (up to 100 samples per hour).

It is differentiated between flow injection analysis (FIA)^{[1],[2]} and continuous flow analysis (CFA)^[3]. Both methods consist of the automatic dosage of the sample introduced into a flow system (manifold) in which the sample analytes react with the reagent solutions on their way through the manifold. The sample preparation may be integrated into the manifold. The reaction product is measured in a flow detector.

The user should be aware that particular problems could require the specification of additional marginal conditions.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[ISO 11732:2005](https://standards.iteh.ai/catalog/standards/iso/8deb819b-6ba2-4db1-b9fc-99f7f6c3f65f/iso-11732-2005)

<https://standards.iteh.ai/catalog/standards/iso/8deb819b-6ba2-4db1-b9fc-99f7f6c3f65f/iso-11732-2005>

Water quality — Determination of ammonium nitrogen — Method by flow analysis (CFA and FIA) and spectrometric detection

WARNING — Persons using this International Standard should be familiar with normal laboratory practice. This International Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this standard be carried out by suitably trained staff.

1 Scope

This International Standard specifies methods suitable for the determination of ammonium nitrogen in various types of waters (such as ground, drinking, surface, and waste waters) in mass concentrations ranging from 0,1 mg/l to 10 mg/l (in the undiluted sample), applying either FIA (Clause 3) or CFA (Clause 4). In particular cases, the range of application may be adapted by varying the operating conditions.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes*

ISO 5667-2, *Water quality — Sampling — Part 2: Guidance on sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of water samples*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

3 Determination of ammonium nitrogen by flow injection analysis (FIA) and spectrometric detection

3.1 Principle

The sample containing ammonium is injected into a continuous carrier stream by means of an injection valve and is mixed with a continuously streaming flow of an alkaline solution. The ammonia formed is separated in a diffusion cell from the solution over a hydrophobic semipermeable membrane and taken up by a streaming recipient flow containing a pH indicator. Due to the resulting pH shift, the indicator solution will change its colour which is measured continuously in the flow photometer. Additional information about this analytical technique is given in references [4], [5], [6], [7] and [8].

NOTE Equipment following this principle using CFA instead of FIA is commercially available, however it has not been validated.

3.2 Interferences

Volatile amines, if present, diffuse through the membrane and lead to a pH shift. If the concentrations of the volatile amines (e.g. methylamine or ethylamine) are equal to those of the ammonium, higher results may then be expected [12]. Significant concentrations of volatile amines can be reduced by distilling the sample adjusted to pH 5,8 prior to the analysis.

In rare cases, it may be possible that the sample does not reach a pH of 12 after the addition of the alkaline reagent solution, thus leading to a loss of ammonium because it will not be converted quantitatively into ammonia. This may particularly occur with strong acidic or buffered samples. In these cases, the pH of the sample shall be adjusted to the range of 3 to 5 by the addition of sodium hydroxide solution (3.3.2 or 3.3.3).

A high concentration of metal ions that can precipitate as hydroxides will give poorly reproducible results. The addition of a suitable complexing agent, such as ethylenedinitrilotetraacetic acid (EDTA), disodium salt, to the alkaline reaction solution (3.3.17) in sufficiently large concentration will prevent the interference by Cu, Zn, Fe, Ca, Mg, and Al. A concentration of 30 g/l of ethylenedinitrilotetraacetic acid, disodium salt (3.3.4), in solution R1 (3.3.17) is adequate for metal concentrations up to 0,2 mg/l each.

In the case of samples containing particulate matter, see 3.5 (last paragraph).

Samples with a total salt concentration of > 10 g/l shall be diluted prior to the measurement.

3.3 Reagents

Apart from the reagents dealt with in the 3.3.6 and 3.3.7, use only reagents of "analytical grade quality for the determination of nitrogen", or, if not available, those of recognized "analytical grade quality". The ammonium content of the blank shall regularly be checked (3.6.3).

3.3.1 Water, of grade 1 as specified in ISO 3696:1987.

3.3.2 Sodium hydroxide solution I, $c(\text{NaOH}) = 5 \text{ mol/l}$.

3.3.3 Sodium hydroxide solution II, $c(\text{NaOH}) = 0,01 \text{ mol/l}$.

3.3.4 Ethylenedinitrilotetraacetic acid (EDTA)¹⁾, disodium salt, monohydrate, $\text{C}_{10}\text{H}_{14}\text{N}_2\text{Na}_2\text{O}_8 \cdot \text{H}_2\text{O}$.

3.3.5 Bromocresol purple, $\text{C}_{21}\text{H}_{16}\text{Br}_2\text{O}_5\text{S}$.

3.3.6 Bromothymol blue, $\text{C}_{27}\text{H}_{28}\text{Br}_2\text{O}_5\text{S}$.

3.3.7 Cresol red, $\text{C}_{21}\text{H}_{18}\text{O}_5\text{S}$.

3.3.8 Ammonium chloride, NH_4Cl , dried at $105^\circ\text{C} \pm 2^\circ\text{C}$ to constant mass.

3.3.9 Potassium chloride, KCl .

3.3.10 Boric acid, H_3BO_3 .

3.3.11 Hydrochloric acid solution I, $c(\text{HCl}) = 0,01 \text{ mol/l}$.

3.3.12 Hydrochloric acid solution II, $c(\text{HCl}) = 0,1 \text{ mol/l}$.

3.3.13 Hydrochloric acid solution III, $c(\text{HCl}) = 1,0 \text{ mol/l}$.

3.3.14 Sulfuric acid, $\rho(\text{H}_2\text{SO}_4) = 1,84 \text{ g/ml}$.

1) Commonly known as ethylenediaminetetraacetic acid.

3.3.15 Mixed indicator.

In a mortar prepare a dry mixture consisting of 10 g of bromocresol purple (3.3.5), 5 g of bromothymol blue (3.3.6), 2,5 g of cresol red (3.3.7), and 45 g of potassium chloride (3.3.9).

The given quantities can be reduced (e.g. by one tenth).

3.3.16 Carrier solution, C (see Figure A.1).

Use water according to 3.3.1, degassed e.g. by reduced pressure.

3.3.17 Alkaline reaction solution, R1 (see Figure A.1).

Dissolve in a graduated flask, nominal capacity 1 000 ml, 30 g of ethylenedinitrilotetraacetic acid, disodium salt (3.3.4) in approximately 800 ml of water (3.3.1), and add 12,4 g of boric acid (3.3.10).

Add 100 ml of sodium hydroxide solution I (3.3.2) dropwise to the suspension and make up to volume with water (3.3.1).

Degas the solution by filtering it through the membrane filter assembly (3.4.2).

The pH of the solution will be approximately 13. Being stored in a plastics bottle (polyethene) at room temperature, it will be stable for one month.

3.3.18 Indicator solution.

Dissolve in a 200 ml graduated flask 1 g of the mixed indicator (3.3.15) in a mixture of 50 ml of sodium hydroxide solution II (3.3.3). Make up to volume with water (3.3.1).

The solution should have a dark reddish colour.

Filter off any undissolved particles.

This solution may be stored at room temperature for three months in an amber glass bottle.

3.3.19 Ammonia recipient solution, R2 (see Figure A.1).

Dilute 10 ml of the indicator solution (3.3.18) with approximately 480 ml of water (3.3.1).

The absorbance of the solution should be 0,3 to 0,5. Otherwise, add dropwise sodium hydroxide solution II (3.3.3) or hydrochloric acid solution III (3.3.13) until an absorbance value of 0,3 to 0,5 (path length 10 mm, wavelength 590 nm) is obtained. Make up to 500 ml with water (3.3.1).

Degas and purify the solution by the membrane filter assembly (3.4.2), fill it into the reagent reservoir and let it stand for at least 2 h.

Immediately before starting the measurement (3.6), check the absorbance again and adjust, if need be, to the absorbance range specified above by adding sodium hydroxide solution II (3.3.3) or hydrochloric acid I, II or III (3.3.11 to 3.3.12) respectively.

This solution may be stored at room temperature for two weeks in a glass bottle.

3.3.20 Ammonium stock solution, $\rho(N) = 1\,000\text{ mg/l}$.

Dissolve in a 1 000 ml graduated flask 3,819 g of ammonium chloride (3.3.8) in approximately 900 ml of water (3.3.1), acidify to pH 2 by dropwise addition of sulfuric acid (3.3.14), and make up to volume with water (3.3.1).

This solution may be stored in a refrigerator for at most three months.

3.3.21 Ammonium standard solution I, $\rho(\text{N}) = 100 \text{ mg/l}$.

Pipette 10 ml of the ammonium stock solution (3.3.20) into a 100 ml graduated flask, add approximately 80 ml of water (3.3.1), acidify by dropwise addition of sulfuric acid (3.3.14), and make up to volume with water (3.3.1).

This solution may be stored in a refrigerator for at most one week.

3.3.22 Ammonium standard solution II, $\rho(\text{N}) = 10 \text{ mg/l}$.

Pipette 1 ml of the ammonium stock solution (3.3.20) or 10 ml of the ammonium standard solution I (3.3.21) into a 100 ml graduated flask, add approximately 80 ml of water, acidify to pH 2 by dropwise addition of sulfuric acid (3.3.14), and make up to volume with water (3.3.1).

This solution may be stored in a refrigerator for at most one week.

3.3.23 Calibration solutions

Prepare calibration solutions by diluting the ammonium standard solutions I or II (3.3.21 or 3.3.22). At least five calibration standards per working range are recommended. As an example, if six standards are applied, proceed for the working ranges I or II respectively, as follows:

a) Working range I (1 mg/l to 10 mg/l):

- Pipette into a series of 100 ml graduated flasks, 1 ml, 2 ml, 4 ml, 6 ml, 8 ml and 10 ml respectively, of the ammonium standard solution I (3.3.21), and make up to volume with water (3.3.1).
- The mass concentrations of ammonium, expressed as nitrogen, in these calibration solutions are 1 mg/l, 2 mg/l, 4 mg/l, 6 mg/l, 8 mg/l and 10 mg/l.

b) Working range II (0,1 mg/l to 1,0 mg/l):

- Pipette into a series of 100 ml graduated flasks, 1 ml, 2 ml, 4 ml, 6 ml, 8 ml and 10 ml respectively, of the ammonium standard solution II (3.3.22), and make up to volume with water (3.3.1).
- The mass concentrations of ammonium, expressed as nitrogen, in these calibration solutions are 0,1 mg/l, 0,2 mg/l, 0,4 mg/l, 0,6 mg/l, 0,8 mg/l and 1,0 mg/l.

Prepare all calibration solutions freshly before use.

3.4 Apparatus

3.4.1 Flow injection system.

In general, the flow injection system consists of the following components (see Figure A.1).

3.4.1.1 Reagent reservoirs.

3.4.1.2 Low pulsation pump.

3.4.1.3 Suitable pump tubes, if required.

3.4.1.4 Injection valve with a suitable injection volume.

For working range I, an injection volume of 40 μl , for working range II, a volume of e.g. 360 μl or 400 μl is recommended.