
**Ophthalmic instruments — Fundamental
requirements and test methods —**

Part 1:
**General requirements applicable
to all ophthalmic instruments**

iTeh STANDARD PREVIEW

*Instruments ophtalmiques — Exigences fondamentales et méthodes
d'essai —*
(standards.iteh.ai)

*Partie 1: Exigences générales applicables à tous les instruments
ophtalmiques* 4-1:2006

[https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-
fl d6e91cf7e0/iso-15004-1-2006](https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-fl d6e91cf7e0/iso-15004-1-2006)



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[ISO 15004-1:2006](https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-fl d6e91cf7e0/iso-15004-1-2006)

<https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-fl d6e91cf7e0/iso-15004-1-2006>

© ISO 2006

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	2
4 Fundamental requirements (for non-active and active ophthalmic instruments).....	2
4.1 General.....	2
4.2 Design	2
4.3 Performance	2
4.4 Combination of different devices	2
4.5 Materials	3
4.6 Protection against contaminants	3
4.7 Scales and displays.....	3
4.8 Thermal hazards	3
4.9 Mechanical hazards	3
5 Environmental conditions (for non-active and active ophthalmic instruments).....	3
5.1 Environmental conditions of use.....	3
5.2 Storage conditions	4
5.3 Transport conditions	4
6 Particular requirements for active ophthalmic instruments	5
6.1 Electrical safety.....	5
6.2 Inapplicable clauses of IEC 60601-1:2005.....	5
6.3 Optical radiation hazard	5
7 Test methods.....	5
7.1 Ignitability	5
7.2 Surface temperatures	5
7.3 Environmental conditions.....	5
7.4 Checking electrical safety.....	7
8 Information supplied by the manufacturer	7
8.1 Accompanying documents	7
8.2 Marking	7
Annex A (informative) Product-related International Standards for ophthalmic instruments	8

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 15004-1 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This first edition together with ISO 15004-2 cancels and replaces ISO 15004:1997, which has been technically revised as follows:

- iTeh STANDARD PREVIEW**
(standards.iteh.ai)
- a) all reference to light hazard (definitions 3.4 to 3.9, 6.3, 7.5, Annexes A, C and D of ISO 15004:1997) has essentially been moved to ISO 15004-2; <https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-d75e91cf7c9/iso-15004-1-2006>
 - b) ignitability requirement/testing changed (4.5.2 and 7.1 of ISO 15004:1997);
 - c) environmental requirements/testing partly changed [Table 1; 5.2.2 and 8.1 f) of ISO 15004:1997];
 - d) normative Annex B (now informative Annex A) entirely updated;
 - e) normative (dated) reference updated to use IEC 60601-1:2005 edition.

ISO 15004 consists of the following parts, under the general title *Ophthalmic instruments — Fundamental requirements and test methods*:

- *Part 1: General requirements applicable to all ophthalmic instruments*
- *Part 2: Light hazard protection*

Ophthalmic instruments — Fundamental requirements and test methods —

Part 1: General requirements applicable to all ophthalmic instruments

1 Scope

This part of ISO 15004 specifies fundamental requirements for non-invasive, active and non-active ophthalmic instruments. This part of ISO 15004 is also applicable to low-vision aids and tonometers, but not to other ophthalmic instruments which are used in contact with the globe of the eye.

This part of ISO 15004 is not applicable to operation microscopes, endoscopes and devices intended for laser investigation or laser treatment of the eye.

iTeh STANDARD PREVIEW

2 Normative references (standards.iteh.ai)

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 9022-2:2002, *Optics and optical instruments — Environmental test methods — Part 2: Cold, heat and humidity*

ISO 9022-3:1998, *Optics and optical instruments — Environmental test methods — Part 3: Mechanical stress*

ISO 15004-2:—¹⁾, *Ophthalmic instruments — Fundamental requirements and test methods — Part 2: Light hazard protection*

IEC 60601-1:2005, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*

IEC 60601-1-1:1992, *Medical electrical equipment — Part 1-1: General requirements for safety — Collateral standard: Safety requirements for medical electrical systems*

IEC 60695-2-10:2000, *Fire hazard testing — Part 2-10: Glowing/hot-wire based test methods — Glow-wire apparatus and common test procedure*

IEC 60695-2-11:2000, *Fire hazard testing — Part 2-11: Glowing/hot-wire based test methods — Glow-wire flammability test method for end-products*

1) To be published. (Revision of ISO 15004:1997)

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

ophthalmic instrument

device designed to have an application to the eye

3.2

non-invasive ophthalmic instrument

ophthalmic instrument which does not in whole or in part penetrate inside the body, either through a body orifice or through the surface of the body

3.3

active ophthalmic instrument

any ophthalmic instrument that depends on a source of electrical energy or any source of power other than that directly generated by the human body or gravity and that acts by converting this energy

NOTE Ophthalmic devices intended to transmit energy, substances or other elements between an active ophthalmic instrument and the patient, without any significant change, are not considered to be an active ophthalmic instrument.

3.4

manufacturer

(ophthalmic instrument) natural or legal person who places the ophthalmic instrument on the market

iTeh STANDARD PREVIEW
(standards.iteh.ai)

4 Fundamental requirements (for non-active and active ophthalmic instruments)

4.1 General

ISO 15004-1:2006

This part of ISO 15004 takes precedence over the corresponding requirements of IEC 60601-1:2005 and IEC 60601-1-1:1992, if differences exist.

The general requirements specified in this part of ISO 15004 for ophthalmic instruments shall be applied in conjunction with those of the relevant product-related International Standard, if it exists. Annex A provides for information the list of relevant product-related International Standards.

4.2 Design

Ophthalmic instruments shall be so designed that, when used for the performance of the intended function(s) in accordance with instructions provided by the manufacturer, the risks associated with such use are reduced to a level compatible with the generally acknowledged state of the art.

4.3 Performance

The ophthalmic instrument shall achieve the performance stipulated by the manufacturer for the intended function(s) under the intended conditions of use.

4.4 Combination of different devices

If another device is intended for use in combination with an ophthalmic instrument, the connecting system shall not impair the specified performance of either instrument.

For coupling with active ophthalmic instruments, the provisions of IEC 60601-1-1 shall apply.

4.5 Materials

4.5.1 Components of the ophthalmic instrument which are designed to come into direct contact with the skin of the patient or operator shall be made of materials which are neither toxic nor known to create significant allergic reactions, when used as intended by the manufacturer.

4.5.2 Materials used shall not ignite. When tested as described in 7.1, combustion shall not continue after withdrawal of the glow-wire.

4.6 Protection against contaminants

Parts of the ophthalmic instrument which are designed to come into contact with the patient or the operator shall either be capable of easy disinfection or be protected by a disposable cover.

4.7 Scales and displays

Scales and displays of ophthalmic instruments shall be designed and placed in accordance with ergonomic principles, taking into account the intended purpose of the instrument.

4.8 Thermal hazards

The temperature of parts of the ophthalmic instrument held by the operator or accessible to the patient shall not exceed the allowable maximum temperatures given in Tables 22, 23 and 24 of IEC 60601-1:2005, 11.1.

4.9 Mechanical hazards

The ophthalmic instrument shall be designed so that, when used to perform the intended function(s) in conformance with the user's instructions, the risk of physical injury when using this instrument is reduced as much as is practicable.

iTech STANDARD PREVIEW
(standards.iteh.ai)
<https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-f1d6e91cf7e0/iso-15004-1-2006>

5 Environmental conditions (for non-active and active ophthalmic instruments)

NOTE The requirements specified in 5.1, 5.2 and 5.3 are verified as described in 7.3.

5.1 Environmental conditions of use

The ophthalmic instrument shall conform to all safety, optical, mechanical and accuracy requirements under the environmental conditions given in Table 1.

Table 1 — Environmental conditions of use

Criterion	Environmental conditions
Temperature	+ 10 °C to + 35 °C
Relative humidity	30 % to 90 %
Atmospheric pressure	800 hPa to 1 060 hPa
Shock (without packing) ^a	10 g, duration 6 ms
a Applicable to hand-held instruments only.	

5.2 Storage conditions

5.2.1 After being stored under the conditions given in Table 2, the ophthalmic instrument shall conform to all safety, optical, mechanical and accuracy requirements under the environmental conditions of use given in Table 1 after being fully adapted to these conditions.

Table 2 —Storage conditions

Criterion	Environmental conditions
Temperature	– 10 °C to + 55 °C
Relative humidity	10 % to 95 %
Atmospheric pressure	700 hPa to 1 060 hPa

5.2.2 Certain ophthalmic instruments may have components, critical for them to perform their design function, that are not able to meet the temperature requirement of 5.2.1 for storage because the temperature extreme would degrade their operating characteristics. When this is the case, the instrument may still fulfil the requirements of this standard if the manufacturer states in bold extra large writing on the shipping container and again in the instructions accompanying the instrument, as required by 8.1 the following:

THIS INSTRUMENT DOES NOT MEET THE TEMPERATURE REQUIREMENTS OF ISO 15004-1 FOR STORAGE.
DO NOT STORE THIS INSTRUMENT IN CONDITIONS WHERE THE TEMPERATURE MAY RISE ABOVE _____ °C OR FALL BELOW _____ °C.

ITeH STANDARD PREVIEW
(standards.iteh.ai)

5.3 Transport conditions

It is recommended that the instrument, in its original packaging, be tested for ability to withstand transport conditions.

<https://standards.iteh.ai/catalog/standards/sist/df5009bc-f83b-435c-ad09-f16c91477031/iso-15004-1-2006>

If ability to withstand exposure to the transport conditions listed in Table 3 of this part of ISO 15004 is claimed [see 8.1 c)], the following shall apply:

After exposure of the ophthalmic instrument in its original packaging to the range of transport conditions given in Table 3, the ophthalmic instrument shall conform to all safety, optical, mechanical and accuracy requirements under the environmental conditions of use given in Table 1 after being fully adapted to these conditions.

Table 3 — Transport conditions

Criterion	Environmental conditions
Temperature	– 40 °C to + 70 °C
Relative humidity	10 % to 95 %
Atmospheric pressure	500 hPa to 1 060 hPa
Vibration, sinusoidal	10 Hz to 500 Hz: 0,5 g
Shock	30 g, duration 6 ms
Bump	10 g, duration 6 ms

6 Particular requirements for active ophthalmic instruments

6.1 Electrical safety

With respect to electrical safety, IEC 60601-1 shall apply.

Compliance with the requirements shall be verified as described in 7.4.

6.2 Inapplicable clauses of IEC 60601-1:2005

The requirements on mechanical strength as specified in 15.3 of IEC 60601-1:2005 shall not apply.

6.3 Optical radiation hazard

NOTE This clause replaces 10.4, 10.5, 10.6 and 10.7 of IEC 60601-1:2005.

The possibility of an optical radiation hazard will be present only for those types of ophthalmic instruments with very high levels of radiation output that are capable of causing high irradiance on the retina and other ocular tissue. ISO 15004-2 specifies the requirements for optical radiation safety for ophthalmic instruments, and the limit values specified therein are considered acceptable with respect to the risks when weighted against the performances intended.

7 Test methods

All tests described in this part of ISO 15004 are type tests.

7.1 Ignitability

ISO 15004-1:2006

Ignitability testing shall be carried out in accordance with IEC 60695-2-11:2000, utilizing the test temperature $650\text{ °C} \pm 10\text{ °C}$ and using the test equipment specified in IEC 60695-2-10:2000.

7.2 Surface temperatures

The requirements given in 4.8 shall be verified at the highest ambient temperature specified in Table 1.

7.3 Environmental conditions

The requirement specified in Clause 5 shall be verified by the tests according to the appropriate part of ISO 9022 given in Table 4.