



**INTERNATIONAL STANDARD ISO 11979-8:1999**  
**TECHNICAL CORRIGENDUM 1**

Published 2003-11-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

**Ophthalmic implants — Intraocular lenses —**  
**Part 8:**  
**Fundamental requirements**

TECHNICAL CORRIGENDUM 1

*Implants ophtalmiques — Lentilles intraoculaires —*

*Partie 8: Exigences fondamentales*

RECTIFICATIF TECHNIQUE 1

**ITeH STANDARD PREVIEW**  
**(standards.iteh.ai)**

<https://standards.iteh.ai/catalog/standards/sist/8a99d8ed-4b4d-4f1d-bb8-892eb16896/iso-11979-8-1999-cor-1-2003>

Technical Corrigendum 1 to ISO 11979-8:1999 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

---

*Page 2, Clause 8*

At the end of this clause add the following text:

IOLs sterilized by ethylene oxide shall have residual ethylene oxide levels and release rates in accordance with the limits given in ISO 10993-7.