

Designation: E1079 - 05

# Standard Practice for Calibration of Transmission Densitometers<sup>1</sup>

This standard is issued under the fixed designation E1079; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\varepsilon$ ) indicates an editorial change since the last revision or reapproval.

# 1. Scope

1.1 This practice<sup>2</sup> covers the calibration of transmission densitometers used to perform radiographic film density measurements (see Note 1).

1.2 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

NOTE 1—For further information on the design and use of densitometers, the following literature is suggested as additional background information: ISO 5–2:2001, ISO 5–1:2001, and ISO 14807:2001.

#### 2. Referenced Documents

2.1 ASTM Standards:<sup>3</sup>

- E1316 Terminology for Nondestructive Examinations
- 2.2 ISO Standards:<sup>4</sup>
- ISO 5–2:2001 Photography-Density Measurements—Part 2: Geometric Conditions for Transmission Density
- ISO 5–1:2001 Photography-Density Measurements—Part 1: Terms, Symbols, and Notations

ISO 14807:2001 Photography-Transmission and Reflection Densitometers—Method for Determining Performance

https://standards.iteh.ai/catalog/standards/sist/5bbd00 3. Terminology

3.1 *Definitions*—For definitions of terms used in this practice, see Terminology E1316.

### 4. Significance and Use

4.1 This practice provides a means for calibrating transmission densitometers used for the measurement of radiographic film density. A transmission densitometer calibrated in accordance with this practice provides the assurance that accurate density values of radiographs are obtained.

## 5. Apparatus

5.1 Apparatus should consist of the following:

5.1.1 A calibrated step tablet shall be used. The step tablet may be a NIST X-ray Step Tablet (Standard Reference Material SRM 1001)<sup>5</sup> or alternately a step tablet from another supplier which is traceable to the NIST SRM 1001 X-ray Step Tablet. The step tablet shall have at least five step densities ranging from 0.9 through 4.1. The step tablet may have additional step densities less than 0.9 and greater than 4.1. A calibration certificate shall be provided with the step tablet indicating the tablet ID and recorded values for each step density. For suppliers of step tablets other than NIST, the certificate shall indicate conformance of traceability to NIST instrumentation used in the calibration process, applicable ANSI standards used, verification of measurement on a NIST SRM 1001 step tablet, the ID number of the SRM 1001 step tablet, and calibration date of the step tablet. Precautions should be taken in the storage, handling, and use of the step tablet. In the event it becomes scratched, blemished, or exhibits other signs of deleterious wear, it should be replaced immediately. The SRM 1001 (or alternate, if used) step tablet shall be replaced four years from the date of first use.<sup>6</sup>

5.1.2 *Transmission Densitometers*, with either direct-scale readout or digital readout displays specifically manufactured for the purpose of measuring the range of film densities described in 5.1.1 may be used.

5.1.3 Manufacturer's Operating Instructions for Appropriate Transmission Densitometer.

Copyright © ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States.

<sup>&</sup>lt;sup>1</sup> This practice is under the jurisdiction of ASTM Committee E07 on Nondestructive Testing and is the direct responsibility of Subcommittee E07.01 on Radioglogy (X and Gamma) Method.

Current edition approved January 1, 2005. Published January 2005. Originally approved in 1985. Last previous edition approved in 2000 as E1079 - 00. DOI: 10.1520/E1079-05.

 $<sup>^2\,{\</sup>rm For}$  ASME Boiler and Pressure Vessel Code applications see related Practice SE-1079 in Section II of that Code.

<sup>&</sup>lt;sup>3</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>&</sup>lt;sup>4</sup> Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036.

<sup>&</sup>lt;sup>5</sup> Standard Reference Material 1001 is available from the National Institute of Standards and Technology, Gaithersburg, MD.

<sup>&</sup>lt;sup>6</sup> Expiration interval of the SRM 1001 or alternate step tablet may be different than the requirements of this practice. Unless otherwise specified, requirements of this practice shall apply.