



SLOVENSKI STANDARD

SIST EN 15058:2017

01-marec-2017

Nadomešča:
SIST EN 15058:2006

Emisije nepremičnih virov - Določevanje masne koncentracije ogljikovega monoksida - Standardna referenčna metoda: nedisperzna infrardeča spektrometrija

Stationary source emissions - Determination of the mass concentration of carbon monoxide - Standard reference method: non-dispersive infrared spectrometry

iTeh STANDARD PREVIEW

Emissionen aus stationären Quellen - Bestimmung der Massenkonzentration von Kohlenmonoxid - Standardreferenzverfahren: Nicht-dispersive Infrarotspektrometrie

<http://standards.iTeh.si>

Emissions de sources fixes - Détermination de la concentration massique de monoxyde de carbone - Méthode de référence normalisée : spectrométrie infra-rouge non dispersive

Ta slovenski standard je istoveten z: EN 15058:2017

ICS:

13.040.40 Emisije nepremičnih virov Stationary source emissions

SIST EN 15058:2017

en,fr,de

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 15058:2017

<https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017>

EUROPEAN STANDARD

EN 15058

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2017

ICS 13.040.40

Supersedes EN 15058:2006

English Version

Stationary source emissions - Determination of the mass concentration of carbon monoxide - Standard reference method: non-dispersive infrared spectrometry

Émissions de sources fixes - Détermination de la concentration massique de monoxyde de carbone - Méthode de référence normalisée : spectrométrie infrarouge non dispersive

Emissionen aus stationären Quellen - Bestimmung der Massenkonzentration von Kohlenmonoxid - Standardreferenzverfahren: Nicht-dispersive Infrarotspektrometrie

This European Standard was approved by CEN on 26 September 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions	6
4 Symbols and abbreviations	13
4.1 Symbols.....	13
4.2 Abbreviated terms	14
5 Principle	14
5.1 General.....	14
5.2 Measuring principle	14
6 Description of the measuring system.....	15
6.1 General.....	15
6.2 Sampling and sample gas conditioning system.....	16
6.2.1 Sampling probe	16
6.2.2 Filter.....	16
6.2.3 Sample gas line.....	16
6.2.4 Sample gas conditioning system.....	16
6.2.5 Sample gas pump.....	17
6.2.6 Secondary filter.....	17
6.2.7 Flow controller and flow meter	17
6.3 Analyser equipment.....	17
6.3.1 General.....	17
6.3.2 Pressure and temperature effects	18
6.3.3 Sampling pump for the analyser.....	18
6.3.4 Interferences due to infrared absorbing gases	18
7 Performance characteristics of the SRM.....	18
8 Suitability of the measuring system for the measurement task	20
9 Field operation.....	20
9.1 Measurement planning.....	20
9.2 Sampling strategy.....	21
9.2.1 General.....	21
9.2.2 Measurement section and measurement plane.....	21
9.2.3 Minimum number and location of measurement points.....	21
9.2.4 Measurement ports and working platform	21
9.3 Choice of the measuring system	21
9.4 Setting of the measuring system on site	22
9.4.1 General.....	22
9.4.2 Preliminary zero and span check, and adjustments	22
9.4.3 Zero and span checks after measurement	23

10	Ongoing quality control.....	23
10.1	Introduction.....	23
10.2	Frequency of checks	23
11	Expression of results	24
12	Equivalence of an alternative method	25
13	Measurement report	25
Annex A (informative) Validation of the method in the field.....		26
A.1	General	26
A.2	Characteristics of installations	26
A.3	Repeatability and reproducibility in the field.....	27
A.3.1	General	27
A.3.2	Repeatability	28
A.3.3	Reproducibility.....	29
Annex B (informative) Schematics of non-dispersive infrared spectrometer		31
Annex C (informative) Calculation of the uncertainty associated with a concentration expressed on dry gas and at an oxygen reference concentration.....		33
C.1	Uncertainty associated with a concentration expressed on dry gas.....	33
C.2	Uncertainty associated with a concentration expressed at a oxygen reference concentration.....	35
Annex D (informative) Example of assessment of compliance of non-dispersive infrared method for CO with requirements on emission measurements.....		37
D.1	General	37
D.2	Elements required for the uncertainty determinations.....	37
D.2.1	Model equation.....	37
D.2.2	Combined uncertainty	38
D.2.3	Expanded uncertainty.....	38
D.2.4	Determination of uncertainty contributions in case of rectangular distributions.....	39
D.2.5	Determination of uncertainty contributions by use of sensitivity coefficients	39
D.3	Example of an uncertainty calculation.....	40
D.3.1	Site specific conditions	40
D.3.2	Performance characteristics	41
D.3.3	Determination of the uncertainty contributions	42
D.3.4	Result of uncertainty calculation	45
D.3.4.1	Standard uncertainties.....	45
D.3.4.2	Combined uncertainty	46
D.3.4.3	Expanded uncertainty.....	46
D.3.4.4	Evaluation of the compliance with the required measurement quality.....	46
Annex E (informative) Example of correction of data from drift effect		47

EN 15058:2017 (E)

Annex F (informative) Significant technical changes.....	49
Bibliography	50

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN 15058:2017](https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017)

<https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017>

European foreword

This document (EN 15058:2017) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This document supersedes EN 15058:2006.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2017, and conflicting national standards shall be withdrawn at the latest by July 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

Annex F provides details of significant technical changes between this document and the previous edition.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom. (standards.iteh.ai)

[SIST EN 15058:2017](https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017)

<https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017>

EN 15058:2017 (E)**1 Scope**

This European Standard specifies the standard reference method (SRM) based on the infrared (IR) absorption principle. It includes the sampling and the gas conditioning system, and allows the determination of the carbon monoxide CO in flue gases emitted to the atmosphere from ducts and stacks.

This European Standard specifies the characteristics to be determined and the performance criteria to be fulfilled by portable automated measuring systems (P-AMS) using the IR measurement method. It applies for periodic monitoring and for the calibration or control of automated measuring systems (AMS) permanently installed on a stack, for regulatory or other purposes.

This European Standard specifies criteria for demonstration of equivalence of an alternative method (AM) to the SRM by application of EN 14793:2017.

This European Standard has been validated during field tests on waste incineration, co-incineration and large combustion plants and on a recognized test bench. It has been validated for CO concentrations with sampling periods of 30 min in the range of 0 mg/m³ to 400 mg/m³ for large combustion plants and 0 mg/m³ to 740 mg/m³ for waste and co-incineration. Directive 2010/75/EU lays down emission values which are expressed in mg/m³, on dry basis at a specified value of oxygen and at standard conditions (273 K and 101,3 kPa).

NOTE The characteristics of installations, the conditions during field tests and the values of repeatability and reproducibility in the field are given in Annex A.

2 Normative references

ITeh STANDARD PREVIEW
(standards.iteh.ai)

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 14793:2017, *Stationary source emissions - Demonstration of equivalence of an alternative method with a reference method*

EN 15259:2007, *Air quality - Measurement of stationary source emissions - Requirements for measurement sections and sites and for the measurement objective, plan and report*

EN 15267-4:2017, *Air quality — Certification of automated measuring systems — Part 4: Performance criteria and test procedures for automated measuring systems for periodic measurements of emissions from stationary sources*

EN ISO 14956:2002, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956:2002)*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1**standard reference method****SRM**

reference method prescribed by European or national legislation

[SOURCE: EN 15259:2007]

3.2 reference method RM

measurement method taken as a reference by convention, which gives the accepted reference value of the measurand

Note 1 to entry: A reference method is fully described.

Note 2 to entry: A reference method can be a manual or an automated method.

Note 3 to entry: Alternative methods can be used if equivalence to the reference method has been demonstrated.

[SOURCE: EN 15259:2007]

3.3 measurement method

method described in a written procedure containing all the means and procedures required to sample and analyse, namely field of application, principle and/or reactions, definitions, equipment, procedures, presentation of results, other requirements and measurement report

[SOURCE: EN 14793:2017]

3.4 alternative method AM

measurement method which complies with the criteria given by this European Standard with respect to the reference method

Note 1 to entry: An alternative method can consist of a simplification of the reference method.

[SOURCE: EN 14793:2016]

3.5 measuring system

set of one or more measuring instruments and often other devices, including any reagent and supply, assembled and adapted to give information used to generate measured quantity values within specified intervals for quantities of specified kinds

[SOURCE: JCGM 200:2012]

3.6 automated measuring system AMS

entirety of all measuring instruments and additional devices for obtaining a result of measurement

Note 1 to entry: Apart from the actual measuring device (the analyser), an AMS includes facilities for taking samples (e.g. probe, sample gas lines, flow meters and regulator, delivery pump) and for sample conditioning (e.g. dust filter, pre-separator for interferences, cooler, converter). This definition also includes testing and adjusting devices that are required for functional checks and, if applicable, for commissioning.

Note 2 to entry: The term “automated measuring system” (AMS) is typically used in Europe. The term “continuous emission monitoring system” (CEMS) is also typically used in the UK and USA.

[SOURCE: EN 15267-4:2017]

EN 15058:2017 (E)**3.7
portable automated measuring system****P-AMS**

automated measuring system which is in a condition or application to be moved from one to another measurement site to obtain measurement results for a short measurement period

Note 1 to entry: The measurement period is typically 8 h for a day.

Note 2 to entry: The P-AMS can be configured at the measurement site for the special application but can be also set-up in a van or mobile container. The probe and the sample gas lines are installed often just before the measurement task is started.

[SOURCE: EN 15267-4:2017]

**3.8
calibration**

set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring method or measuring system, and the corresponding values given by the applicable reference

Note 1 to entry: In case of automated measuring system (AMS) permanently installed on a stack the applicable reference is the standard reference method (SRM) used to establish the calibration function of the AMS.

Note 2 to entry: Calibration should not be confused with adjustment of a measuring system.

**3.9
adjustment**

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured

Note 1 to entry: The adjustment can be made directly on the instrument or using a suitable calculation procedure.

**3.10
span gas**

test gas used to adjust and check a specific point on the response line of the measuring system

**3.11
measurand**

particular quantity subject to measurement

[SOURCE: EN 15259:2007]

Note 1 to entry: The measurand is a quantifiable property of the stack gas under test, for example mass concentration of a measured component, temperature, velocity, mass flow, oxygen content and water vapour content.

**3.12
interference**

negative or positive effect upon the response of the measuring system, due to a component of the sample that is not the measurand

3.13**influence quantity**

quantity that is not the measurand but that affects the result of the measurement

Note 1 to entry: Influence quantities are e.g. presence of interfering gases; ambient temperature, pressure of the gas sample.

3.14**ambient temperature**

temperature of the air around the measuring system

3.15**emission limit value****ELV**

limit value given in regulations such as EU Directives, ordinances, administrative regulations, permits, licences, authorisations or consents

Note 1 to entry: ELV can be stated as concentration limits expressed as half-hourly, hourly and daily averaged values, or mass flow limits expressed as hourly, daily, weekly, monthly or annually aggregated values.

3.16**measurement site**

place on the waste gas duct in the area of the measurement plane(s) consisting of structures and technical equipment, for example working platforms, measurement ports, energy supply

Note 1 to entry: Measurement site is also known as sampling site.

[SOURCE: EN 15259:2007]

[SIST EN 15058:2017](https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017)

3.17**measurement plane**

plane normal to the centreline of the duct at the sampling position

Note 1 to entry: Measurement plane is also known as sampling plane.

[SOURCE: EN 15259:2007]

3.18**measurement port**

opening in the waste gas duct along the measurement line, through which access to the waste gas is gained

Note 1 to entry: Measurement port is also known as sampling port or access port.

[SOURCE: EN 15259:2007]

3.19**measurement line**

line in the measurement plane along which the measurement points are located, bounded by the inner duct wall

Note 1 to entry: Measurement line is also known as sampling line.

[SOURCE: EN 15259:2007]

EN 15058:2017 (E)**3.20****measurement point**

position in the measurement plane at which the sample stream is extracted or the measurement data are obtained directly

Note 1 to entry: Measurement point is also known as sampling point.

[SOURCE: EN 15259:2007]

3.21**performance characteristic**

one of the quantities (described by values, tolerances, range) assigned to equipment in order to define its performance

3.22**response time**

duration between the instant when an input quantity value of a measuring instrument or measuring system is subjected to an abrupt change between two specified constant quantity values and the instant when a corresponding indication settles within specified limits around its final steady value

Note 1 to entry: By convention time taken for the output signal to pass from 0 % to 90 % of the final variation of indication.

3.23**short-term zero drift**

difference between two zero readings at the beginning and at the end of the measurement period

3.24**short-term span drift**

difference between two span readings at the beginning and at the end of the measurement period

3.25**lack of fit**

systematic deviation, within the measurement range, between the measurement result obtained by applying the calibration function to the observed response of the measuring system measuring test gases and the corresponding accepted value of such test gases

Note 1 to entry: Lack of fit can be a function of the measurement result.

Note 2 to entry: The expression "lack of fit" is often replaced in everyday language by "linearity" or "deviation from linearity".

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 15058:2017

<https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-16dce972578c/en-15058-2017>

3.26**repeatability in the laboratory**

closeness of the agreement between the results of successive measurements of the same measurand carried out under the same conditions of measurement

Note 1 to entry: Repeatability conditions include:

- same measurement method;
- same laboratory;
- same measuring system, used under the same conditions;
- same location;
- repetition over a short period of time.

Note 2 to entry: Repeatability can be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 3 to entry: In this European Standard the repeatability is expressed as a value with a level of confidence of 95 %.

3.27**repeatability in the field**

closeness of the agreement between the results of simultaneous measurements of the same measurand carried out with two sets of equipment under the same conditions of measurement

Note 1 to entry: These conditions include:

- same measurement method; <https://standards.iteh.ai/catalog/standards/sist/6ad7f00f-596a-4487-90f0-18dfcd897f23/sist-en-15058-2017>
- two sets of equipment, the performances of which are fulfilling the requirements of the measurement method, used under the same conditions;
- same location;
- implemented by the same laboratory;
- typically calculated on short periods of time in order to avoid the effect of changes of influence parameters (e.g. 30 min).

Note 2 to entry: Repeatability can be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 3 to entry: In this European Standard, the repeatability under field conditions is expressed as a value with a level of confidence of 95 %.

EN 15058:2017 (E)**3.28****reproducibility in the field**

closeness of the agreement between the results of simultaneous measurements of the same measurand carried out with several sets of equipment under the same conditions of measurement

Note 1 to entry: These conditions are called field reproducibility conditions and include:

- same measurement method;
- several sets of equipment, the performances of which are fulfilling the requirements of the measurement method, used under the same conditions;
- same location;
- implemented by several laboratories.

Note 2 to entry: Reproducibility can be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 3 to entry: In this European Standard, the reproducibility under field conditions is expressed as a value with a level of confidence of 95 %.

3.29**residence time in the measuring system**

time period for the sample gas to be transported from the inlet of the probe to the inlet of the measurement cell

3.30**uncertainty**

parameter associated with the result of a measurement, that characterises the dispersion of the values that could reasonably be attributed to the measurand

3.31**standard uncertainty** **u**

uncertainty of the result of a measurement expressed as a standard deviation

3.32**combined uncertainty** **u_c**

standard uncertainty attached to the measurement result calculated by combination of several standard uncertainties according to the principles laid down in ISO/IEC Guide 98-3 (GUM)

3.33**expanded uncertainty** **U**

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

$$U = k \times u_c$$

Note 1 to entry: In this European Standard, the expanded uncertainty is calculated with a coverage factor of $k = 2$, and with a level of confidence of 95 %.

Note 2 to entry: The expression overall uncertainty is sometimes used to express the expanded uncertainty.

3.34**uncertainty budget**

calculation table combining all the sources of uncertainty according to EN ISO 14956 or ISO/IEC Guide 98-3 in order to calculate the combined uncertainty of the method at a specified value

4 Symbols and abbreviations**4.1 Symbols**

For the purposes of this document, the following symbols apply.

$A(t_0)$	(result given by the analyser after adjustment at t_0 at span point – result given by the analyser after adjustment at t_0 at zero point) / (calibration gas concentration at span point – calibration gas concentration at zero point)
$B(t_0)$	result given by the analyser after adjustment at t_0 at zero point
C	measured concentration
C_{corr}	measured concentration corrected for drift
Drift(A)	{[(result given by the analyser during the drift check at t_{end} at span point – result given by the analyser during the drift check at t_{end} at zero point) / (calibration gas concentration at span point – calibration gas concentration at zero point)] – $A(t_0)$ } / ($t_{\text{end}} - t_0$)
Drift(B)	(result given by the analyser during the drift check at t_{end} at zero point – result given by the analyser after adjustment at t_0 at zero point) / ($t_{\text{end}} - t_0$)
f	volume fraction
k	coverage factor
M_{mol}	molar mass
S_R	reproducibility standard deviation
$S_{r,\text{limit}}$	maximum allowable repeatability standard deviation
t	time
t_0	time of adjustment
t_{end}	time of check for drift at the end of the measurement period
u	standard uncertainty
u_c	combined uncertainty
U	expanded uncertainty
V_{mol}	molar volume