
**Ophthalmic optics — Contact lenses
and contact lens care products —
Guidance for clinical investigations**

*Optique ophtalmique — Lentilles de contact et produits d'entretien
pour lentilles de contact — Directives pour les investigations cliniques*

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 11980:2009

<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 11980:2009

<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2009

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Clinical investigational requirements.....	1
4.1 General	1
4.2 Additional requirements	2
4.2.1 Study design	2
4.2.2 Variables.....	4
4.3 Other considerations	4
Annex A (informative) Elements of a clinical investigation.....	5
Annex B (informative) Procedures for the evaluation of safety, physiological performance and effect on ocular tissues	18
Annex C (informative) The evaluation of visual, refractive and lens performance and subject acceptance	24
Bibliography.....	28

ISO 11980:2009

<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11980 was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This second edition cancels and replaces the first edition (ISO 11980:1997), of which it constitutes a technical revision. It also incorporates the Technical Corrigendum ISO 11980:1997/Cor.1:1998.

ISO 11980:2009
<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>

Introduction

Currently, contact lenses and contact lens care products are regulated in different ways in different countries. This International Standard has been developed to encourage global harmonization. Widespread adoption of this International Standard should represent yet another step toward mutual recognition. This International Standard can also be used as a basis to fulfil design elements of ISO 9001^[1].

iTeh STANDARD PREVIEW (standards.iteh.ai)

[ISO 11980:2009](https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009)

<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

ISO 11980:2009

<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>

Ophthalmic optics — Contact lenses and contact lens care products — Guidance for clinical investigations

1 Scope

This International Standard gives guidelines for the clinical investigation (CI) of the safety and performance of contact lenses and contact lens care products.

NOTE This International Standard attempts to harmonize the recognized regulatory requirements for the conduct of a CI to meet the marketing and labelling requirements for contact lenses and contact lens care products around the world. However, national requirements vary greatly. Wherever national practice or regulations dictate some legal requirement, this requirement takes precedence over this International Standard.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14155-1, *Clinical investigation of medical devices for human subjects — Part 1: General requirements*
ISO 11980:2009

ISO 14155-2, *Clinical investigation of medical devices for human subjects — Part 2: Clinical investigation plans*
<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>

ISO 14534, *Ophthalmic optics — Contact lenses and contact lens care products — Fundamental requirements*

ISO 18369-1, *Ophthalmic optics — Contact lenses — Part 1: Vocabulary, classification system and recommendations for labelling specifications*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 14155-1, ISO 14155-2, ISO 14534 and ISO 18369-1 apply.

4 Clinical investigational requirements

4.1 General

The general requirements for a CI and for a clinical investigation plan (CIP) given in ISO 14155-2 shall apply, with additional requirements given below.

4.2 Additional requirements

4.2.1 Study design

4.2.1.1 General

- a) The inclusion criteria for subject selection shall relate to the study objectives and should include:
- 1) subjects with normal eyes who are not using any ocular medications, aged 18 years or over [except when contact lens investigations have a special indication for use in “children” (for the purposes of this International Standard, persons less than 18 years of age) such as orthokeratology and paediatric aphakic lenses];
 - 2) lens powers within the range available for the test lenses;
 - 3) the manifest cylinder less than or equal to 0,75 D (for a study with only spherical power correcting lenses);
 - 4) best spectacle corrected visual acuity greater than or equal to 20/25 (less than or equal to LogMAR 0,1).
- b) The exclusion criteria for subject selection shall relate to the study objectives and should include, but not be limited to:
- 1) anterior segment infection, inflammation or abnormality;
 - 2) any active anterior segment ocular disease that would contraindicate contact lens wear;
 - 3) the use of systemic or ocular medications that would contraindicate contact lens wear;
 - 4) history of herpetic keratitis;
 - 5) history of refractive surgery or irregular cornea (except when the contact lenses under investigation are indicated for irregular cornea, keratoconus or refractive surgery);
 - 6) slit lamp findings that are equal to or more serious than trace findings (greater than or equal to grade 1);
 - 7) corneal vascularization greater than 1 mm of penetration;
 - 8) a pathologically dry eye;
 - 9) participation of the subject in a contact lens or contact lens care product clinical trial within the previous 30 days.
- c) The CIP shall provide a description of the monitoring procedure to ensure consistent quality of data collection and recording.
- d) The CIP shall include a statistical analysis plan. Sample size shall be justified, calculated by a validated statistical software package.

4.2.1.2 Contact lenses

4.2.1.2.1 General. A CI of contact lenses, including daily wear and extended wear hydrogel, silicone hydrogel, and rigid gas-permeable contact lenses, shall be designed as one of 4.2.1.2.2 or 4.2.1.2.3.

For CIPs to demonstrate safety and performance, as well as special claims (e.g. comfort), labelling or additional indications, the following is required: a pre-determined statistical analysis plan (including sample

size calculations) shall be specified in the clinical protocol. Where feasible, the CIP shall define objective endpoints to help support such claims.

NOTE 1 Inter-subject controls are preferred to intra-subject controls due to the potential dependence between the two eyes and concerns regarding subject compliance.

NOTE 2 Annex A provides guidance for the design of a CI.

4.2.1.2.2 As a prospective, concurrently controlled study. For investigations evaluating hydrogel, silicone hydrogel or rigid gas-permeable contact lenses, a prospective, concurrent control study design shall be followed. Either a bilateral crossover design or a contra-lateral eye (i.e. intra-subject) design or inter-subject controls shall be utilized. If inter-subject controls are utilized, the ratio of test subjects to control subjects may be either 2:1 or 1:1. The control lens shall be a currently marketed contact lens in use for the same modality. Randomization and masking (subject, investigator and evaluator) shall be employed where possible to minimize the potential for bias. Subjects shall be divided evenly between study investigators.

4.2.1.2.3 As an uncontrolled study. Here, results are compared to a historical control. Alternative investigational study designs, such as historical controls, shall be utilized when a sponsor has a clinical database on a marketed contact lens to use as a comparator. If any historical control is used, the control group shall be defined and adequately characterized for comparison to the test group. Compatibility of test and control groups shall be demonstrated by comparison of the selection criteria, demographics, refractive characteristics, contact lens wearing history and CIPs used.

4.2.1.3 Contact lens care products

For investigations evaluating contact lens care products, a prospective concurrent control study design shall be followed. It is recommended that the ratio of test to control subjects be either 2:1 or 1:1. The control care product shall be a currently marketed contact lens care product. Randomization and masking (subject, investigator and evaluator) shall be employed where possible to minimize the potential for bias. Subjects shall be divided evenly between study investigators. Alternative investigational study designs, such as use of historical controls, may be utilized when a manufacturer has a clinical database on a marketed care product to use for comparison. If any historical control is used, the control group should be defined and adequately characterized for comparison to the test group. Compatibility of test and control groups should be demonstrated by comparison of the selection criteria and CIPs used.

For CIPs to demonstrate safety and performance, as well as special claims (e.g. comfort), labelling or additional indications, the following is required for the care products: a pre-determined statistical analysis plan (including sample size calculations) shall be specified in the clinical protocol. Where feasible, the protocol should define objective endpoints to help support such claims.

NOTE 1 Inter-subject controls are preferred to intra-subject controls due to the potential dependence between the two eyes and concerns regarding subject compliance.

In a contact lens care product investigation, a daily wear schedule shall be followed for most products in order to maximize the subject's exposure to those products. However, a study of a lens or a periodic cleaner, used at weekly intervals, may provide more valuable clinical data concerning efficacy when extended wear subjects are enrolled than a similar investigation with daily wear subjects.

When a daily wear schedule is used and safety is a primary objective, one post-dispensing visit should be done 1 h to 2 h after lens insertion in order to permit observation of corneal and conjunctival staining caused by an immediate toxicity reaction.

A contact lens care product with a cleaning indication shall have an objective measure of lens cleanliness on at least one lens collected from each subject at the end of the clinical study.

If the manufacturer of a contact lens care product wishes to recommend its use with a specific type of lens in the labelling, the compatibility with the lens type should be confirmed pre-clinically and during the clinical trial.

If the CI has not collected any data on use with a particular type of lens material (such as silicone hydrogel lenses), the product label should clearly state this fact.

NOTE 2 Annex A provides guidance for the design of a CI.

4.2.2 Variables

4.2.2.1 Contact lenses

The following variables should be considered during the CI for contact lenses, in addition to the variables listed in 4.2.2.2:

- a) visual performance;
- b) refractive performance;
- c) keratometric measurements;
- d) lens centration;
- e) lens movement;
- f) lens surface wettability;
- g) lens surface deposits;
- h) subject acceptance of comfort;
- i) subject acceptance of vision;
- j) subject acceptance of handling.

Additional variables can be studied in the CI to support specific claims.

NOTE Annex C provides guidance on classifications for each of these variables.

4.2.2.2 Contact lens care products (standards.iteh.ai)

The following variables should be assessed during the CI for contact lens care products:

- a) corneal oedema;
- b) corneal infiltrates;
- c) endothelial irregularity;
- d) corneal vascularization;
- e) corneal staining;
- f) conjunctival observations;
- g) palpebral conjunctival observations;
- h) corneal ulcers;
- i) corneal opacification;
- j) hyphema;
- k) hypopyon;
- l) iritis;
- m) corneal scarring.

Additional variables can be studied in the CI to support specific claims.

NOTE Annex B provides guidance on classifications for some of these variables.

4.3 Other considerations

Serious ophthalmic adverse events and all adverse device effects shall be reported using a special case report form and forwarded to the sponsor as required. All other ophthalmic adverse events shall be reported using the standard visit case report forms and shall be collected during monitoring.

Annex A (informative)

Elements of a clinical investigation

A.1 General

The following are elements of a CIP which can assist in collecting data for the purpose of determining the safety and performance of contact lenses and contact lens care products.

A.2 Study size and duration

A.2.1 Contact lens investigations

**Table A.1 — Guide to the subject numbers (completed subjects)
suggested for contact lens clinical investigations (informative)**

Wearing modality	Subject number completed per group at end of trial	Duration	Material and design
Daily wear	50	3 months	Containing new or new ratio material components; significant design changes
	30	30 days	All materials and designs
Daily wear orthokeratology	50	3 months or longer if necessary to reach defined stability	All materials and designs
Extended wear, up to 7 days	160	12 months	All materials and designs
Extended wear, up to 30 days	570	12 months	All materials and designs
Overnight wear (may include orthokeratology)	300	12 months	All materials and designs

A.2.2 Contact lens care product investigations

A.2.2.1 Contact lens care products, including saline solutions, daily cleaners, periodic cleaners, disinfecting solutions, neutralizers, “in-eye” solutions, conditioning solutions, and multipurpose solutions that have any new active ingredient, or any active ingredient outside the concentration range used in a comparable marketed product, should undergo a 3 month clinical study.

A.2.2.2 Products for use with soft (hydrophilic) lenses: sample size (completed) should be 30 subjects in the test solution and 15 subjects in the control solution (a currently marketed solution for the same indication) for each appropriate representative category such as:

- Group I;
- Group IV;
- A separate group for each silicone hydrogel lens. If more than one lens is made by a given manufacturer, and they all have the same general chemistry, it is sufficient to use only the lens of highest water content.

A.2.2.3 Products for use with rigid lenses: sample size (completed) should be 15 or 30 subjects using the test solution and 15 subjects using the control solution (a currently marketed solution for the same indication) for each appropriate material group.

A.2.2.4 For a contact lens solution that does not contain any new active ingredients (as described in A.2.2.1), but contains any active ingredient lower than the concentration range used in a comparable marketed product, a 1 month clinical study should be conducted. In this case, the sample size should be about half of that recommended in A.2.2.2 and A.2.2.3, using the same general distribution of subjects.

A.2.3 Statistical considerations for extended wear evaluations

A.2.3.1 General

Primary safety analysis: the key safety endpoint should be the frequency of serious and significant adverse events.

The null hypothesis, H_0 , is that the test rate of endpoint adverse events, p_t , minus the control rate of endpoint adverse events, p_c , is greater than or equal to the clinically insignificant difference, δ , between the two rates.

The alternative hypothesis, H_a , is that the test rate of endpoint adverse events, p_t , minus the control rate of endpoint adverse events, p_c , is less than a clinically insignificant difference, δ , between the two rates.

$$H_0: p_t - p_c \geq \delta$$

$$H_a: p_t - p_c < \delta$$

iTeh STANDARD PREVIEW
(standards.iteh.ai)

where

p_t is the proportion in the test population;

p_c is the proportion in the control population.

[ISO 11980:2009](https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009)

<https://standards.iteh.ai/catalog/standards/sist/da767159-622b-4ba8-909a-e4438a3c4c84/iso-11980-2009>

When using a 1:1 ratio of patient allocation between treatment and control, the minimum number, n , of completed patients necessary for each treatment group is determined by:

$$n = \frac{(Z_{1-\beta} + Z_{1-\alpha})^2 \times [p_t(1 - p_t) + p_c(1 - p_c)]}{\delta^2}$$

where

α is the significance level (also known as the type 1 error rate);

$1 - \beta$ is the power of the test;

Z is the standard normal quantile.

The following is an example of the calculation that makes assumptions found to be reasonable for clinical studies of 7 day extended wear hydrogel or silicone hydrogel contact lenses. With a control rate, p_c , and a test rate, p_t , of 0,033 (under H_a), a clinically insignificant difference, δ , of 0,05, a power, $1 - \beta$, of 0,80, and a significance level, α , of 0,05, the minimum number of completed patients per treatment group is:

$$n = \frac{(0,84 + 1,64)^2 \times [0,033 \times (1 - 0,033) + 0,033 \times (1 - 0,033)]}{0,05^2} \approx 158$$