



Designation: D 2330 – 88 (Reapproved 1995)^{ε1}

AMERICAN SOCIETY FOR TESTING AND MATERIALS
100 Barr Harbor Dr., West Conshohocken, PA 19428
Reprinted from the Annual Book of ASTM Standards. Copyright ASTM

Standard Test Method for Methylene Blue Active Substances¹

This standard is issued under the fixed designation D 2330; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense. Consult the DoD Index of Specifications and Standards for the specific year of issue which has been adopted by the Department of Defense.

^{ε1} NOTE—Editorial changes were made throughout in June 1995.

1. Scope

1.1 This test method covers the determination of compounds that react with methylene blue under the conditions specified in the test procedure. They are referred to as methylene blue active substances (MBAS), and are calculated and reported in terms of the reference material, linear alkyl benzene sulfonate, LAS.

1.2 This test method is applicable for determining MBAS in water and wastewater. It is the user's responsibility to ensure the validity of this test method for waters of untested matrices.

1.3 This test method is a simple, rapid, control procedure suitable for monitoring the effectiveness of a biodegradation or other linear alkyl benzene sulfonate (LAS) removal process. For greater specificity and interference removal, the pretreatment procedure in Annex A1 should be used. Data derived without the pretreatment procedure should be interpreted with care. This test method is applicable in the range from 0.03 to 1.5 mg/L for a 100-mL sample.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For a specific hazard statement, see Note 2.

2. Referenced Documents

2.1 ASTM Standards:

- D 459 Terminology Relating to Soap and Other Detergents²
- D 1129 Terminology Relating to Water³
- D 1193 Specification for Reagent Water³
- D 2777 Practice for Determination of Precision and Bias of Applicable Methods of Committee D-19 on Water³
- D 3370 Practices for Sampling Water from Closed Conduits³

¹ This test method is under the jurisdiction of ASTM Committee D-19 on Water and is the direct responsibility of Subcommittee D19.06 on Methods of Analysis for Organic Substances in Water.

Current edition approved Jan. 29, 1988. Published May 1988. Originally published as D2330 – 65 T. Last previous edition D2330 – 82.

² Annual Book of ASTM Standards, Vol 15.04.

³ Annual Book of ASTM Standards, Vol 11.01.

D 3856 Guide for Good Laboratory Practices in Laboratories Engaged in Sampling and Analysis of Water³

D 4210 Practice for Intralaboratory Quality Control Procedures and a Discussion on Reporting Low-Level Data³

E 60 Practice for Photometric and Spectrophotometric Methods for Chemical Analysis of Metals⁴

E 131 Terminology Relating to Molecular Spectroscopy⁵

E 275 Practice for Describing and Measuring Performance of Ultraviolet, Visible, and Near Infrared Spectrophotometers⁵

3. Terminology

3.1 *Definitions*—For definitions of terms used in this test method, refer to Terminology D 1129 and E 131.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *alkyl benzene sulfonate (ABS)*⁶—the generic name applied to the neutralized product resulting from the sulfonation of a branched-chain alkylated benzene. See also Terminology D 459.

3.2.2 *linear alkyl benzene sulfonate (LAS)*⁶—a form of alkyl benzene sulfonate (ABS) in which the alkyl group is linear rather than a branched chain. See also Terminology D 459.

4. Summary of Test Method⁷

4.1 This test method is based upon the formation of a blue-colored chloroform extractable ion pair by the reaction of cationic methylene blue and an anionic surfactant (including LAS, other sulfonates, and sulfate esters).

4.2 The sample is mixed with an acidified, aqueous solution of methylene blue. Any resulting hydrophobic ion pair which may be formed is extracted successfully with chloroform. The combined chloroform extracts are washed with an acid solution to remove the less hydrophobic ion pairs (having low partition coefficients) that can be formed by potentially interfering

⁴ Annual Book of ASTM Standards, Vol 03.05.

⁵ Annual Book of ASTM Standards, Vol 03.06.

⁶ For a more complete discussion of terms relating to synthetic detergents and their significance, refer to "Syndets and Waste Disposal" by McKinney, R. E., *Sewage and Industrial Wastes*, Vol 29, Part 6, June 1957, pp. 654–666.

⁷ Adopted from "Surfactants (Anionic) Methylene Blue Methods," *Standard Methods for the Examination of Water and Waste Water*, Twelfth Ed., 1965.

substances. The chloroform layer retains the highly hydrophobic methylene blue-LAS ion pairs.

4.3 The intensity of the blue color remaining in the chloroform extract is measured photometrically at the wavelength of maximum absorption near 650 nm. This intensity is related to the concentration of LAS by means of a calibration curve or chart.

5. Significance and Use

5.1 The widespread use and discharge of detergents into surface waters can result in a lowering of its aesthetic quality by foam formation and by causing toxicity to aquatic wildlife. This test method is capable of detecting small concentrations of detergents as MBAS so that they can be controlled to prevent such problems.

5.2 Biodegradable linear alkyl benzene sulfonates (LAS) have replaced the branched-chain alkyl benzene sulfonates (ABS) in detergent formulations, which were more resistant to biodegradation. Differentiation between linear and branched-chain alkyl benzene sulfonates, as well as differentiation of the various positional isomers of either type, is not possible by this test method. While the methylene blue method may be employed to monitor studies designed to measure biodegradability, it cannot be used to predict this quality.

6. Interferences

6.1 Any organic or inorganic compound that will form a chloroform extractible ion pair will interfere by producing high results, unless the ion pair formed is eliminated by the treatment described in 4.1. These positive interferences include organic sulfonates, carboxylates, phosphates, and phenols, as well as inorganic cyanates, chlorides, nitrates, and thiocyanates.

6.2 Any compound effectively competing with methylene blue to form a LAS ion pair will give negative results. This negative interference is demonstrated by some amines and has analytical significance in the case of quaternary ammonium compounds.

6.3 An evaluation of the effect of various potential interferences is summarized in Table 1. The listed compounds, in the concentrations indicated, were added to solutions containing 1 mg/L LAS.

6.4 When interferences are present, the pretreatment procedure described in Annex A1 should be used. Table 2 shows the interferences that can be present even though the pretreatment was used.

6.5 When a concentrated acid chromate cleaning solution is used to clean glassware, including separatory funnels, between samples, care must be taken to completely flush all of the acid chromate cleaning solution from all surfaces and, in particular, from the space between the barrel and plug of the separatory funnel stopcock. Failure to remove the acid can result in an error in results.

6.5.1 Never use a detergent to clean any glassware used in this test method as a detergent is difficult to remove from surfaces. Any residual detergent could cause a high result.

7. Apparatus

7.1 *Filter Photometer or Spectrophotometer*, suitable for

TABLE 1 Evaluation of Potential Interferences in the Methylene Blue Method

Added to 1.0 mg/L LAS Solution	Concentration, mg/L	Indicated LAS, mg/L
Acetic acid	100	1.0
Ammonium diethylphosphorodithioate	20	1.1
Benzene sulfonic acid	100	1.3
Cholesterol	100	1.0
2,4-dichlorophenol	100	1.0
Diethanolamine	1000	1.0
Disodium phenylphosphate	10	1.0
Isopropylamine	14	1.0
Leucine	10	1.0
M-1-(naphthylethylenediamine) hydrochloride	100	0.9
Nonyl phenol + 9 EtO	100	1.0
Phenol	100	1.0
Picric acid	5	4.6
Potassium chloride	100	1.0
Potassium cyanate	100	1.0
Potassium nitrate	100	1.0
Potassium thiocyanate	2	1.0
Potassium thiocyanate	100	4.1
Proteins (Knox gelatine)	100	0.9
Sodium dodecyl sulfate	10	14.6
Sodium dodecane sulfonate	5	5.0
Sodium naphthalene sulfonate	5	5.1
Sodium stearate	100	1.0

TABLE 2 Evaluation of Potential Interferences in the Methylene Blue Method with Pretreatment Described in Annex A1

Added to 1.0 mg/L LAS Solution	Concentration, mg/L	Indicated LAS, mg/L
Sodium dodecane sulfonate	5	3.7
Sodium benzene sulfonate	100	1.2
Sodium dodecyl sulfate	10	0.9
Potassium thiocyanate	100	1.0
Picric acid	10	1.0

measurement at a wavelength in the region near 650 nm and equipped with 50-mm and 10-mm light path absorption cells.

NOTE 1—Photometers and photometric practices prescribed in this test method shall conform to Practice E 60. Spectrophotometers shall conform to Practice E 275.

7.2 *Separatory Funnels*, 250-mL size, Squibb-type, glass-stoppered, preferably with TFE-fluorocarbon stopcocks.

8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society.⁸ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification D 1193, Type II.

8.3 *Chloroform* (CHCl₃).

⁸ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopoeia and National Formulary*, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

NOTE 2—Warning: Chloroform (CHCl_3) is toxic and is suspected of being a possible carcinogen: avoid ingestion, inhalation, or absorption through the skin. Use a well-ventilated fume hood to carry off chloroform vapors during analysis.

8.4 *Linear Alkyl Benzene Sulfonate Solution, Stock* (1.0 mL = 1.0 mg LAS)—Weigh the amount of reference material⁹ necessary to provide the equivalent of 1.000 g of LAS on a 100 % active basis. Dissolve in water and dilute to 1 L, mixing gently to prevent foam formation. Record the molecular weight of the LAS reference material as supplied. The stock solution may be stored at 4°C in the dark for 12 months in a well-stoppered flask without deterioration.¹⁰

8.5 *Linear Alkyl Benzene Sulfonate Standard Solution*, (1.0 mL = 0.01 mg LAS)—Dilute 10.0 mL of the foam-free stock solution (8.4) to 1 L with water that has been previously adjusted to pH 2 with sulfuric acid and mix. The standard solution may be stored at 4°C in the dark for at least 12 months in a well-stoppered flask without deterioration.¹⁰

8.6 *Methylene Blue Solution* (30 mg/L)—Dissolve 0.1 g of methylene blue chloride in 100 mL of water. Transfer 30 mL of this solution to a 1-L volumetric flask and add 500 mL of water. Add carefully 50 mL of 14 % sulfuric acid stock solution (8.10) and 50 g of sodium dihydrogen phosphate monohydrate ($\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$). Shake until solution is complete and then dilute to 1 L with water and mix.

8.7 *Phenolphthalein Indicator Solution* (5.0 g/L)—Dissolve 0.5 g of phenolphthalein in 50 mL of 95 % ethyl alcohol and dilute to 100 mL with water and mix.

NOTE 3—Specially denatured ethyl alcohol conforming to Formula No. 3A or 30 of the U. S. Bureau of Alcohol, Tobacco, and Firearms may be substituted for 95 % ethyl alcohol.

8.8 *Phosphate Wash Solution*—Dissolve 50 g of sodium dihydrogen phosphate monohydrate ($\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$) in 500 mL of water in a 1-L volumetric flask. Add carefully 50 mL of 14 % sulfuric acid stock solution (8.10) and dilute to volume with water and mix. The solution has a pH of approximately 1.8.

8.9 *Sodium Hydroxide Solution* (10 g/L)—Dissolve 10 g of sodium hydroxide (NaOH) in water, dilute to 1 L and mix.

8.10 *Sulfuric Acid Stock Solution* (14 % volume per volume)—Add carefully 140 mL of concentrated sulfuric acid (H_2SO_4 , sp gr 1.84) to 700 mL of cold (0 to 5°C) water with good stirring, dilute to 1 L with water and mix.

8.11 *Sulfuric Acid Solution, Dilute* (0.7 % volume per volume)—Dilute carefully 50 mL of 14 % sulfuric acid stock solution (8.10) to 1 L with water and mix.

9. Sampling

9.1 Collect the sample in accordance with Practices D 3370.

9.2 Samples may be preserved against biological oxidation by adding concentrated sulfuric acid (H_2SO_4) to adjust the sample to pH 2 or less and storing at 4°C. Analyze the

preserved sample as soon as possible, or within 1 week after collection. Data on decomposition are not available.

9.3 Rinse the sample container and cap well to free them of detergent if they have been used previously and cleaned prior to recycling.

10. Preparation of Apparatus

10.1 Glassware Conditioning:

10.1.1 All glassware used for the determination of LAS should be free of scratches and etch marks because of the tendency of surface-active materials to adsorb on this type of surface. All volumetric flasks and photometer cells, projected for use in LAS determinations, should, as instructed herein, be preconditioned as follows: Obtain the chloroform extract from 12.0 mL of the standard LAS solution as described in 11.4. Transfer sequentially to each of the volumetric flasks and photometer cells and permit a minimum contact time, in each case, of 5 min. Rinse thoroughly with chloroform and drain (**Warning**—see Note 2).

11. Calibration

11.1 Prepare a series of standards by adding the standard solution (8.5) from a 25-mL buret to a series of 250-mL separatory funnels (see 6.5) and dilute the standards to 100 mL volume with water, yielding solutions as follows:

Standard, mL (1.0 mL = 0.01 mg LAS)	LAS, mg (per 100 mL of extract)
0.00	0.00
1.00	0.01
3.00	0.03
5.00	0.05
7.00	0.07
9.00	0.09
12.00	0.12

NOTE 4—If desired, additional standards in the range from 0.00 to 0.12 mg of LAS may be prepared for the calibration series.

11.2 Add 3 drops of phenolphthalein solution (8.7) and just enough sodium hydroxide solution (8.9) to produce a pink color. Add dilute sulfuric acid solution (8.11), in small increments until the pink color is barely discharged.

11.3 Add 25 mL of methylene blue solution (8.6) and mix. Add 25 mL of chloroform (**Warning**—see Note 2) and mix thoroughly for 30 s with shaking. Vent carefully, permit the phases to separate and then drain the chloroform layer into a second 250-mL separatory funnel (see 6.5). Leave any emulsion layer in the first separatory funnel. Repeat the extraction, serially, with two additional 25-mL portions of chloroform.

NOTE 5—Vent the separatory funnel through the stopcock with the funnel tip directed away from the face to avoid contact with any sample spray (**Warning**—see Note 2).

11.4 Add 50 mL of phosphate wash solution (8.8) to the combined chloroform extracts in the second separatory funnel and shake vigorously for 30 s (see Note 5). Hold the separatory funnel in a vertical position and swirl the contents. Permit settling for 1 min. Filter the chloroform layer through a glass wool plug into a conditioned (see 10.1) 100-mL volumetric flask. Add 20 mL of chloroform to the second separatory funnel and repeat the shaking, swirling, and settling steps (see Note 5). Combine the chloroform layer through the glass wool into

⁹ Linear alkyl benzene sulfonate reference material may be obtained from the United States Environmental Protection Agency, Environmental Monitoring and Support Laboratory, Cincinnati, OH 45268.

¹⁰ Data supporting the precision and bias statements and the stability of the LAS Stock and Standard Solutions are available from ASTM Headquarters. Request RR:D19-169.