# INTERNATIONAL STANDARD

ISO 127

Third edition 1995-12-01

## Rubber, natural latex concentrate — Determination of KOH number

## iTeh STANDARD PREVIEW

Latex concentré de caoutchous naturel — Détermination de l'indice de potasse da rus. Iten. al )



#### **Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 127 was prepared by Technical Committee ISO/TC 45, Rubber and rubber products, Subcommittee SC 3, Raw materials (including latex) for use in the rubber industry.

(standards.iteh.ai)

This third edition cancels and replaces the second edition (ISO 127:1984), which has been technically revised.

ISO 127:1995

Annexes A and B of this International Standard are for information only 0e25-151a-404c-9a1f-694825cebea7/iso-127-1995

© ISO 1995

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

### Rubber, natural latex concentrate — Determination of KOH number

WARNING — Persons using this standard should be familiar with normal laboratory practice. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

#### 1 Scope

ISO 125:1990, Natural rubber latex concentrate — DeiTeh STANDARD termination of alkalinity.

This International Standard specifies a method for the determination of the KOH number of natural rubber. (SO 976:1), Rubber latices — Determination of pH. latex concentrate which is preserved wholly or in part with ammonia. The method is applicable to latices27:1995 SO 1802:1985, Natural rubber latex concentrate containing boric acid. The methodisdnot applicable tolards/sis Petermination of boric acid content. latices preserved with potassium hydroxide 4tt is not 7/iso-127-1995 necessarily suitable for latices from natural sources other than Hevea brasiliensis, or for latices of synthetic rubber, compounded latex, vulcanized latex or artificial dispersions of rubber.

#### 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 123:1985, Rubber latex — Sampling.

ISO 124:1992. Rubber latices — Determination of total solids content.

For the purposes of this International Standard, the following definition applies.

**3.1 KOH number (of latex):** The number of grams of potassium hydroxide equivalent to the acid radicals combined with ammonia in latex concentrate containing 100 g of total solids.

(Definition taken from ISO 1382:1982, Rubber — Vocabulary/Add.2:1982.)

#### 4 Reagents

During the analysis, use only reagents of recognized analytical grade and only distilled water free of dissolved carbon dioxide, or water of equivalent purity.

<sup>3</sup> Definition

<sup>1)</sup> To be published.

ISO 127:1995(E) © ISO

- **4.1 Potassium hydroxide**, standard volumetric solution,  $c(KOH) = 0.1 \text{ mol/dm}^3$ , carbonate-free.
- **4.2 Potassium hydroxide**, standard volumetric solution,  $c(KOH) = 0.5 \text{ mol/dm}^3$ , carbonate-free.
- **4.3** Formaldehyde, 45 g to 50 g in 1 dm<sup>3</sup> of solution  $[c(HCHO) = 1,5 \text{ mol/dm}^3 \text{ to } 1,67 \text{ mol/dm}^3]$ , acid-free, prepared by diluting concentrated formaldehyde with water and neutralizing with 0,1 mol/dm<sup>3</sup> potassium hydroxide solution (4.1), using as indicator the faint pink colour of phenolphthalein.

Determine the concentration of the formaldehyde solution as described in annex A.

#### 5 Apparatus

Standard laboratory glassware, plus the following:

**5.1 pH-meter**, conforming to ISO 976 but capable of being read to 0,01 units.

- **5.2 Glass electrode**, of a type suitable for use in solutions of pH up to 12,0.

  Solutions of pH up to 12,0.

  180 127:1995

  190 191 19 10,3 coa higher, omit the initial ad-
- **5.3 Mechanical stirrer**, with earthed motor and glass paddle, or **magnetic stirrer**.

NOTE 1 An automatic titrator may be used provided it has been checked as giving the same result as the standard method.

#### 6 Sampling

Carry out the sampling in accordance with one of the methods specified in ISO 123.

#### 7 Procedure

Calibrate the pH-meter by the method specified in ISO 976. If the total solids ( $w_{\rm TS}$ ) and alkalinity (A) of the latex are not known, determine them in accordance with ISO 124 and ISO 125 respectively. If the latex contains boric acid and the content is not known, determine it in accordance with ISO 1802.

Carry out the determination in duplicate.

Weigh, to the nearest 0,1 g, into a 400 cm<sup>3</sup> beaker a test portion (mass m) of the test sample containing approximately 50 g of total solids. If necessary, adjust the alkalinity to  $(0,5 \pm 0,1)$  % ammonia calculated with respect to the water phase by adding, while stirring, the required quantity of formaldehyde solution (4.3).

Calculate the volume of formaldehyde solution to be added from the formula

$$\frac{m(100 - w_{TS})(A - 0.5)}{1134 c(HCHO)}$$

where c(HCHO) is the actual concentration, expressed in moles per cubic decimetre, of the formaldehyde solution (4.3).

Dilute the latex with water to about 30 % total solids.

Insert the electrodes of the pH-meter (5.1) into the diluted latex concentrate and record the pH.

lf the initial pH is less than 10,3, slowly add 5 cm<sup>3</sup> of 0,5 mol/dm<sup>3</sup> potassium hydroxide solution (4.2) while stirring slowly with the glass paddle or magnetic stirrer (5.3). Record the resultant equilibrium pH reading. With continued stirring, add 0,5 mol/dm<sup>3</sup> potassium hydroxide solution (4.2) in 1 cm<sup>3</sup> increments at regular (e.g. 15 s) intervals, recording the resultant equilibrium pH after each addition. Continue until the end-point has been passed.

694825cebea7dition 7of 95 cm<sup>3</sup> at one time and proceed directly to "add 0,5 mol/dm<sup>3</sup> potassium hydroxide solution (4.2) in 1 cm<sup>3</sup> increments" as described above.

The end-point of the titration is the point of inflexion of the titration curve of the pH-value against the volume, in cubic centimetres, of potassium hydroxide solution. At this point, the slope of the curve, i.e. the first differential, reaches a maximum and the second differential changes from a positive to a negative value. The end-point shall be calculated from the second differential on the assumption that the change from a positive to a negative value bears a linear relation to the addition of potassium hydroxide during the 1 cm<sup>3</sup> interval involved.

An example of a typical titration and the calculation of the end-point is given in annex B.

The results of duplicate determinations shall agree to within 5 % (m/m).

#### 8 Expression of results

Calculate the KOH number, expressed as a percentage by mass, of the latex concentrate from the formula

$$\frac{561c \cdot V}{w_{\text{TS}} \cdot m}$$

where

 is the actual concentration, expressed in moles of KOH per cubic decimetre, of the potassium hydroxide solution (4.2);

V is the volume, in cubic centimetres, of the nominally 0,5 mol/dm<sup>3</sup> potassium hydroxide solution (4.2) required to reach the end-point;

 $w_{\rm TS}$  is the total solids content, expressed as a percentage by mass, of the latex concentrate;

m is the mass, in grams, of the test portion.

If the latex concentrate contains boric acid, subtract the KOH number equivalent to the boric acid from the KOH number obtained above. Calculate the KOH number equivalent to the boric acid present from the formula

$$91 \times \frac{w_{\rm BA}}{w_{\rm TS}}$$

where  $w_{\rm BA}$  is the boric acid content, expressed as a percentage by mass.

#### 9 Test report

The test report shall include the following information:

- a) a reference to this International Standard:
- b) all details necessary for complete identification of the sample;
- all details necessary for complete identification of the pH-meter used;
- d) the result obtained;
- e) the correction applied for boric acid, if present;
- details of any operation not included in this International Standard or regarded as optional;
- g) the date of the test.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

#### Annex A

(informative)

#### **Determination of formaldehyde**

The method given in previous editions of this International Standard appears not to have been widely used due to the fact that standard solutions of ammonia are considered to be unsatisfactory. Having regard to the consistent quality of analytical-grade concentrated formaldehyde solution, the majority of users prepare a standard solution of formaldehyde directly.

Where it is necessary to determine the concentration of the diluted formaldehyde, a variety of methods exist and users are referred to the Encyclopaedia of Industrial Chemical Analysis, Vol. 13, published by Interscience (1971). The method given below is for information only.

Transfer 100 cm<sup>3</sup> of the solution to a 500 cm<sup>3</sup> conical flask. Accurately weigh into the flask 6,0 g to 8,0 g of the nominally 50 g/dm³ formaldehyde solution (4.3) <sup>2)</sup> and swirl to mix thoroughly. Allow to stand for 5 min, then titrate with 0,25 mol/dm³ sulfuric acid (A.1.2) to the first colourless end-point using thymolphthalein (A.1.3) as indicator. Run a blank determination with the soldium sulfite solution.

#### A.3 Expression of results

Calculate the formaldehyde content, expressed as a iTeh STANDA percentage by mass, of the formaldehyde solution from the formula

#### Reagents

(standards.iteh.ai  $V_2$ ) × 2c(H<sub>2</sub>SO<sub>4</sub>)

 $V_1$ 

**A.1.1 Sodium sulfite**, anhydrous, analytical grade.

 $10m_{1}$ 

https://standards.iteh.ai/catalog/standarwhereb3bf0e25-151a-404c-9a1f-A.1.2 Sulfuric acid, standard volumetric solution\_cebea7/iso-127-1995

 $c(H_2SO_4) = 0.25 \text{ mol/dm}^3.$ 

is the volume, in cubic centimetres, of sulfuric acid (A.1.2) used for the titration of the test portion of formaldehyde solution;

A.1.3 Thymolphthalein, indicator solution.

Dissolve 80 mg of thymolphthalein in 100 cm<sup>3</sup> of ethyl alcohol and dilute with 100 cm<sup>3</sup> of distilled water.

 $V_2$ is the volume, in cubic centimetres, of sulfuric acid (A.1.2) required for the blank;

A.2 Procedure

 $c(H_2SO_4)$  is the actual concentration, in moles of H<sub>2</sub>SO<sub>4</sub> per cubic decimetre, of the sulfuric acid;

Prepare a solution of 125 g of anhydrous sodium sulfite (A.1.1) in 500 cm<sup>3</sup> of water and dilute to 1 dm<sup>3</sup>.

is the mass, in grams, of the test por $m_1$ tion of formaldehyde solution.

<sup>2)</sup> When analysing concentrated formaldehyde solution, 1,8 g to 2,0 g of solution is a more convenient amount to take.

#### **Annex B**

(informative)

### Example of a typical titration and calculation of end-point

Table B.1 — Example of a typical titration to show variation in pH

Volume of KOH solution added cm <sup>3</sup>	pH reading	First difference ΔpH/cm <sup>3</sup>	Second difference $\Delta^2$ pH/cm $^3$
Initial	10,09		
5	10,46		
6	10,55	0,09	0,01
7	10,65	0,10	0,01
8	10,76	0,11	0,03
9	Teh STONDAR	D PREVIEW	0,04
10	(stændards	iteh.ai)	0,06
11	11.32	0,24	0,07
12 http	ISO 127:1 s://standards.itehla <b>63</b> :talog/standard	s/sist/b3bf0e25-151a-404c-9a1f-	-0,01
13	694825cebea7/iso 11,93		-0,09
14	12,14	0,21	

In this example, the first difference reaches a maximum of 0,31 between 11 cm $^3$  and 12 cm $^3$  of potassium hydroxide solution. The precise point of inflexion is calculated from the ratio of the adjacent second-difference values, i.e. 0,07/(0,07 + 0,01) = 0,875 of the difference between 11 cm $^3$  and 12 cm $^3$ , i.e. 11,875 cm $^3$ .

Figure B.1 presents this data graphically to show the point of inflexion.




Figure B.1 — Illustrative curve showing change of pH during titration

# iTeh STANDARD PREVIEW This page intentionally left blank (standards.iteh.ai)