
INTERNATIONAL STANDARD



INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Acoustics — Measurement of sound insulation in buildings and of building elements — Part I : Requirements for laboratories

*Acoustique — Mesurage de l'isolation acoustique des immeubles et des éléments de construction —
Partie I : Spécifications relatives aux laboratoires*

standards.iteh.ai

First edition — 1978-07-15

[ISO 140-1:1978](https://standards.iteh.ai/catalog/standards/sist/8f489bea-9c9f-48df-88fa-c98fcb092000/iso-140-1-1978)

<https://standards.iteh.ai/catalog/standards/sist/8f489bea-9c9f-48df-88fa-c98fcb092000/iso-140-1-1978>

UDC 534.833.522.4.08

Ref. No. ISO 140/1-1978 (E)

Descriptors: acoustics, acoustic measurement, acoustic insulation, buildings, structural members, tests, testing conditions, laboratories.

Price based on 2 pages

FOREWORD

ISO (the International Organization for Standardization) is a worldwide federation of national standards institutes (ISO member bodies). The work of developing International Standards is carried out through ISO technical committees. Every member body interested in a subject for which a technical committee has been set up has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council.

International Standard ISO 140/1 was developed by Technical Committee ISO/TC 43, *Acoustics*, and was circulated to the member bodies in May 1976.

(standards.iteh.ai)

It has been approved by the member bodies of the following countries :

Australia
Austria
Belgium
Canada
Czechoslovakia
Denmark
Finland
France
Germany
Hungary

India
Israel
Italy
Japan
Korea, Rep. of
Mexico
Netherlands
New Zealand
Norway
Poland

ISO 140-1:1978

Romania
South Africa, Rep. of
Sweden
Switzerland
Turkey
United Kingdom
U.S.A.
U.S.S.R.

<https://standards.iteh.ai/catalog/standards/sist/8f489bea-9c9f-48df-88fa-c98fcb022010/iso-140-1-1978>

The member body of the following country expressed disapproval of the document on technical grounds :

Spain

This International Standard, together with International Standards ISO 140/III, IV, VI and VII, cancel and replace ISO Recommendation R 140-1960, of which they constitute a technical revision.

Acoustics — Measurement of sound insulation in buildings and of building elements —

Part I : Requirements for laboratories

0 INTRODUCTION

The purpose of this International Standard is to provide a co-ordinated statement of requirements for laboratories used for measurements of sound insulation of building elements.

Laboratories for the determination of airborne and impact sound reduction of structural elements should be constructed in such a way that the measurement results can be directly or indirectly applied to the conditions in actual buildings.

In the case of laboratories with suppressed radiation from flanking elements, the behaviour of the element in the actual building can be concluded only indirectly and only in some cases from the measurement results in the laboratory. The test rooms described in this International Standard belong to this group of laboratories. This group includes laboratories where the specimen is structurally isolated from both test rooms, and laboratories where the test specimen is connected to one or both of the test rooms, the radiation from flanking elements being reduced either by use of heavy elements or by use of appropriate linings.

A direct application of the results of laboratory measurements is possible if the flanking transmission is included. For this purpose, the test rooms and coupling of test specimen to the flanking construction must resemble the situation in usual buildings ("laboratories with flanking transmission", "mockups"). The requirements for such laboratories are under consideration.

1 SCOPE AND FIELD OF APPLICATION

This International Standard lays down specifications concerning laboratories for sound insulation measurements of building elements. It applies to laboratories with suppressed radiation from flanking elements.

2 REFERENCES

ISO 140/III, *Acoustics — Measurement of sound insulation in buildings and of building elements — Part III : Laboratory measurements of airborne sound insulation of building elements.*

ISO 140/VI, *Acoustics — Measurement of sound insulation in buildings and of building elements — Part VI : Laboratory measurements of impact sound insulation of floors.*

ISO 140/VIII, *Acoustics — Measurement of sound insulation in buildings and of building elements — Part VIII : Laboratory measurements of the reduction of transmitted impact noise by floor coverings on a standard floor.*

3 LABORATORIES FOR AIRBORNE SOUND INSULATION MEASUREMENTS UNDER DIFFUSE CONDITIONS

The laboratory test facility consists of two adjacent reverberant rooms with a test opening between them in which the test specimen is inserted.

3.1 Rooms

Volumes and shapes of the two test rooms should not be exactly the same. A difference in room volumes of at least 10% is recommended. The volumes of the test rooms should be at least 50 m³.

The ratios of the room dimensions should be so chosen that the natural frequencies in the low-frequency region are spaced as uniformly as possible.

If necessary, diffusing elements should be installed in the rooms to obtain a diffuse sound field.

NOTES

1 The volume of the rooms and the size of the test opening as well as the position of the test specimen within this opening are under consideration. Theoretical calculation as well as some experiments have indicated that it may be advisable that the specimen should cover a total side wall or ceiling of the test room, i.e. the test opening should extend from wall to wall and/or from ceiling to floor. In such a case, a volume of 50 m³ is appropriate in view of the recommended size of the test opening.

2 The reverberation time in the rooms should not be excessively long. Where the reverberation time at low frequencies exceeds 2 s, a check should be made to determine whether the measured sound reduction index depends on the reverberation time. When such a dependence is found — even with diffusors in the rooms — the room should be modified to reduce the reverberation time to not more than 2 s at low test frequencies.

The background level in the receiving room must be sufficiently low to permit a measurement of the sound transmitted from the source room, considering the power output in the source room and the sound insulation of the specimens for which the laboratory is intended.

In laboratory test facilities for measuring the sound reduction index, the sound transmitted by any indirect path should be negligible compared with the sound trans-

mitted through the test specimen. One way to achieve this in such facilities is to provide sufficient structural isolation between source and receiving room. Another method might be to cover all surfaces of both rooms with linings that reduce the radiation sufficiently.

NOTE — A suitable measure to use as a reference value is R'_{\max} defined in ISO 140/III.

3.2 Test opening

It is recommended that the size of the test opening should be approximately 10 m² for walls, and between 10 m² and 20 m² for floors, with the shorter edge length not less than 2,3 m for both walls and floors.

A smaller size may be used, if the wavelength of free flexural waves at the lowest frequency considered is smaller than half the minimum dimension of the specimen. The smaller the specimen, however, the more sensitive the results will be to edge constraint conditions and to local variations in sound fields.

A smaller size may also be appropriate for tests of windows, doors and similar components. Test openings for doors should be so arranged that the lower edge is situated as near to the level of the floor of the test rooms as will reproduce conditions in the actual building.

NOTE — Test openings should be so arranged that the test specimen can be installed in a manner as similar as possible to the actual construction with a careful simulation of normal connections and sealing conditions at the perimeter and at joints within the partition.

4 LABORATORIES FOR IMPACT SOUND INSULATION OF FLOORS AND FLOOR COVERINGS

4.1 Receiving room

The volume of the receiving room should be at least 50 m³.

The ratios of the receiving room dimensions should be so chosen that the natural frequencies in the low-frequency region are spaced as uniformly as possible.

If necessary, diffusing elements should be installed in the receiving room to obtain a diffuse sound field.

NOTES

1 The volume of the receiving room and the size of the test opening as well as the position of the test specimen within this opening are under consideration. Theoretical calculation as well as some experiments have indicated that it may be advisable that the specimen should cover the total ceiling of the receiving room, i.e. the test opening should extend from wall to wall. In such a case, a volume of 50 m³ is appropriate in view of the recommended size of the test opening.

2 The reverberation time in the receiving room should not be excessively long. Where the reverberation time at low frequencies exceeds 2 s, a check should be made to determine whether the measured impact sound insulation depends on the reverberation time. When such a dependence is found — even with diffusors in the room — the room should be modified to reduce the reverberation time to not more than 2 s at low test frequencies.

The background level in the receiving room must be sufficiently low to permit a measurement of the transmitted impact sound, considering the properties of the tapping machine and the sound insulation of the specimens for which the laboratory is intended.

The airborne sound insulation between the receiving room and the source room must be sufficiently high that the sound field measured in the receiving room is only that generated by the impact excitation of the test floor.

4.2 Test opening

It is recommended that the size of the test opening for floors should be between 10 m² and 20 m², with the shorter edge length not less than 2,3 m.

NOTE — When measuring reduction in impact sound pressure level by floor coverings according to ISO 140/VIII, special precautions with respect to the test opening may not be necessary.