

# **SLOVENSKI STANDARD**

## **SIST EN ISO 8536-4:2013**

**01-maj-2013**

**Nadomešča:**

**SIST EN ISO 8536-4:2010**

---

**Infuzijska oprema za uporabo v medicini - 4. del: Infuzijski seti za enkratno uporabo, delujoči na osnovi gravitacije (ISO 8536-4:2010)**

Infusion equipment for medical use - Part 4: Infusion sets for single use, gravity feed (ISO 8536-4:2010)

Infusionsgeräte zur medizinischen Verwendung - Teil 4: Infusionsgeräte für Schwerkraftinfusionen zur einmaligen Verwendung (ISO 8536-4:2010)

Matériel de perfusion à usage médical - Partie 4: Appareils de perfusion non réutilisables, à alimentation par gravité (ISO 8536-4:2010)

**Ta slovenski standard je istoveten z: EN ISO 8536-4:2013**

---

**ICS:**

11.040.20	Transfuzijska, infuzijska in injekcijska oprema	Transfusion, infusion and injection equipment
-----------	---	---

**SIST EN ISO 8536-4:2013**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-4:2013](https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)

<https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN ISO 8536-4**

February 2013

ICS 11.040.20

Supersedes EN ISO 8536-4:2010

English Version

**Infusion equipment for medical use - Part 4: Infusion sets for  
single use, gravity feed (ISO 8536-4:2010)**

Matériel de perfusion à usage médical - Partie 4: Appareils  
de perfusion non réutilisables, à alimentation par gravité  
(ISO 8536-4:2010)

Infusionsgeräte zur medizinischen Verwendung - Teil 4:  
Infusionsgeräte für Schwerkraftinfusionen zur einmaligen  
Verwendung (ISO 8536-4:2010)

This European Standard was approved by CEN on 8 January 2013.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.

[SIST EN ISO 8536-4:2013](https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)

<https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

	Page
Foreword.....	3
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices.....	4

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8536-4:2013](https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)  
<https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013>

## Foreword

The text of ISO 8536-4:2010 has been prepared by Technical Committee ISO/TC 76 “Transfusion, infusion and injection, and blood processing equipment for medical and pharmaceutical use” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 8536-4:2013 by Technical Committee CEN/TC 205 “Non-active medical devices” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2013, and conflicting national standards shall be withdrawn at the latest by August 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8536-4:2010.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

[SIST EN ISO 8536-4:2013](https://standards.iteh.ai/catalog/standards/sist/9d40c1e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)

<https://standards.iteh.ai/catalog/standards/sist/9d40c1e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013>

### Endorsement notice

The text of ISO 8536-4:2010 has been approved by CEN as EN ISO 8536-4:2013 without any modification.

## Annex ZA (informative)

### Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on Medical Devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

**Table ZA.1 — Correspondence between this European Standard and Directive 93/42/EEC on Medical Devices**

Clause(s)/subclause(s) of this European Standard	Essential Requirements (ERs) of Directive 93/42/EEC	Qualifying remarks/Notes
6.1, 6.7, 8.3, 8.4, 8.5	7.2	
8.1	7.5	Presumption of conformity with the Essential Requirements relating to the biological evaluation can only be provided if the manufacturer chooses to apply the EN ISO 10993 series of standards.
6.2, 6.4, 6.5	7.6	
6.11, 6.13	8	
3.2	8.1	
10	8.3	
8.2	8.4	
6.3, 6.12	9.1	
6.9, 6.10	10	
6.3	12.7.1	
6.6, 6.8, 6.9, 6.10	12.8	
9	13	The part of ER 13.3 a) relating to the authorized representative is not addressed.  ERs 13.3 f) and 13.6 h) relating to single-use are not fully addressed.  ER 13.6 q) is not addressed.
4	13.3	

**WARNING — Other requirements and other EU Directives may be applicable to the product(s) falling within the scope of this standard.**

# INTERNATIONAL STANDARD

**ISO**  
**8536-4**

Fifth edition  
2010-10-01

---

---

## Infusion equipment for medical use — Part 4: Infusion sets for single use, gravity feed

*Matériel de perfusion à usage médical —*

*Partie 4: Appareils de perfusion non réutilisables, à alimentation par  
gravité*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 8536-4:2013](https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)

[https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-  
47b1ad6adde6/sist-en-iso-8536-4-2013](https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)



Reference number  
ISO 8536-4:2010(E)

© ISO 2010

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 8536-4:2013](https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013)

<https://standards.iteh.ai/catalog/standards/sist/8d49c4e8-70ad-4671-8067-47b1ad6adde6/sist-en-iso-8536-4-2013>

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2010

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland



# Contents

Page

Foreword .....	iv
1 Scope .....	1
2 Normative references .....	1
3 General requirements .....	1
4 Designation .....	4
4.1 Infusion set .....	4
4.2 Air-inlet device .....	4
5 Materials .....	4
6 Physical requirements .....	5
6.1 Particulate contamination .....	5
6.2 Leakage .....	5
6.3 Tensile strength .....	5
6.4 Closure-piercing device .....	5
6.5 Air-inlet device .....	5
6.6 Tubing .....	6
6.7 Fluid filter .....	6
6.8 Drip chamber and drip tube .....	6
6.9 Flow regulator .....	6
6.10 Flow rate of infusion fluid .....	6
6.11 Injection site .....	6
6.12 Male conical fitting .....	6
6.13 Protective caps .....	6
7 Chemical requirements .....	7
7.1 Reducing (oxidizable) matter .....	7
7.2 Metal ions .....	7
7.3 Titration acidity or alkalinity .....	7
7.4 Residue on evaporation .....	7
7.5 UV absorption of extract solution .....	7
8 Biological requirements .....	7
8.1 General .....	7
8.2 Sterility .....	7
8.3 Pyrogenicity .....	7
8.4 Haemolysis .....	7
8.5 Toxicity .....	8
9 Labelling .....	8
9.1 Unit container .....	8
9.2 Shelf or multi-unit container .....	8
10 Packaging .....	9
Annex A (normative) Physical tests .....	10
Annex B (normative) Chemical tests .....	14
Annex C (normative) Biological tests .....	16
Bibliography .....	17

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 8536-4 was prepared by Technical Committee ISO/TC 76, *Transfusion, infusion and injection equipment for medical and pharmaceutical use*.

This fifth edition cancels and replaces the fourth edition (ISO 8536-4:2007), of which it constitutes a minor revision. In detail, 7.1 was more clarified in alignment with B.2, and A.2.2 was changed in order to go back with the leakage test pressure to 20 kPa and to restrict the leakage test for  $(40 \pm 1) ^\circ\text{C}$ .

ISO 8536 consists of the following parts, under the general title *Infusion equipment for medical use*:

- *Part 1: Infusion glass bottles*
- *Part 2: Closures for infusion bottles*
- *Part 3: Aluminium caps for infusion bottles*
- *Part 4: Infusion sets for single use, gravity feed*
- *Part 5: Burette infusion sets for single use, gravity feed*
- *Part 6: Freeze drying closures for infusion bottles*
- *Part 7: Caps made of aluminium-plastics combinations for infusion bottles*
- *Part 8: Infusion equipment for use with pressure infusion apparatus*
- *Part 9: Fluid lines for use with pressure infusion equipment*
- *Part 10: Accessories for fluid lines for use with pressure infusion equipment*
- *Part 11: Infusion filters for use with pressure infusion equipment*
- *Part 12: Check valves*

# Infusion equipment for medical use —

## Part 4:

## Infusion sets for single use, gravity feed

### 1 Scope

This part of ISO 8536 specifies requirements for single use, gravity feed infusion sets for medical use in order to ensure their compatibility with containers for infusion solutions and intravenous equipment.

Secondary aims of this part of ISO 8536 are to provide guidance on specifications relating to the quality and performance of materials used in infusion sets and to present designations for infusion set components.

In some countries, the national pharmacopoeia or other national regulations are legally binding and take precedence over this part of ISO 8536.

**iTeh STANDARD PREVIEW**

### 2 Normative references (standards.iteh.ai)

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 594-1, *Conical fittings with a 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 1: General requirements*

ISO 594-2, *Conical fittings with 6 % (Luer) taper for syringes, needles and certain other medical equipment — Part 2: Lock fittings*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 7864, *Sterile hypodermic needles for single use*

ISO 14644-1, *Cleanrooms and associated controlled environments — Part 1: Classification of air cleanliness<sup>1)</sup>*

ISO 15223-1, *Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements<sup>2)</sup>*

### 3 General requirements

**3.1** The nomenclature to be used for components of infusion sets and of a separate air-inlet device is given in Figures 1, 2 and 3. These figures illustrate examples of the configuration of infusion sets and air-inlet devices; other configurations may be used provided they lead to the same results. Infusion sets as illustrated in Figure 2 should only be used for collapsible plastic containers. Infusion sets as illustrated in Figure 2 used

1) Under preparation. (Revision of ISO 14644-1:1999)

2) To be published. (Revision of ISO 15223-1:2007)