



SLOVENSKI STANDARD SIST EN ISO 11137-3:2017

01-oktober-2017

Nadomešča:

SIST EN ISO 11137-3:2006

Sterilizacija izdelkov za zdravstveno nego - Sevanje - 3. del: Smernice o dozimetričnih vidikih za razvoj, validacijo in rutinski nadzor (ISO 11137-3:2017)

Sterilization of health care products - Radiation - Part 3: Guidance on dosimetric aspects of development, validation and routine control (ISO 11137-3:2017)

Sterilisation von Produkten für die Gesundheitsfürsorge - Strahlen - Teil 3: Anleitung zu dosimetrischen Aspekten der Entwicklung, Validierung und Lenkung der Anwendung (ISO 11137-3:2017)

Stérilisation des produits de santé - Irradiation - Partie 3: Directives relatives aux aspects dosimétriques de développement, la validation et le contrôle de routine (ISO 11137-3:2017)

Ta slovenski standard je istoveten z: EN ISO 11137-3:2017

ICS:

11.080.01 Sterilizacija in dezinfekcija na splošno Sterilization and disinfection in general

SIST EN ISO 11137-3:2017

en

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11137-3:2017](#)

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>

EUROPEAN STANDARD

EN ISO 11137-3

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2017

ICS 11.080.01

Supersedes EN ISO 11137-3:2006

English Version

Sterilization of health care products - Radiation - Part 3: Guidance on dosimetric aspects of development, validation and routine control (ISO 11137-3:2017)

Stérilisation des produits de santé - Irradiation - Partie
3: Directives relatives aux aspects dosimétriques de
développement, la validation et le contrôle de routine
(ISO 11137-3:2017)

Sterilisation von Produkten für die
Gesundheitsfürsorge - Strahlen - Teil 3: Anleitung zu
dosimetrischen Aspekten der Entwicklung, Validierung
und Lenkung der Anwendung (ISO 11137-3:2017)

This European Standard was approved by CEN on 15 March 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
European foreword.....	3

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11137-3:2017](https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017)
<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>

European foreword

This document (EN ISO 11137-3:2017) has been prepared by Technical Committee ISO/TC 198 “Sterilization of health care products” in collaboration with Technical Committee CEN/TC 204 “Sterilization of medical devices” the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2018 and conflicting national standards shall be withdrawn at the latest by January 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11137-3:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

iTeh STANDARD PREVIEW
Endorsement notice
(standards.iteh.ai)

The text of ISO 11137-3:2017 has been approved by CEN as EN ISO 11137-3:2017 without any modification.

[SIST EN ISO 11137-3:2017](https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017)

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11137-3:2017](#)

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>

INTERNATIONAL
STANDARD

ISO
11137-3

Second edition
2017-06

**Sterilization of health care products —
Radiation —**

Part 3:
**Guidance on dosimetric aspects of
development, validation and routine
control**

iTeh STANDARD PREVIEW

(standards.iteh.ai)

Stérilisation des produits de santé — Irradiation —

*Partie 3: Directives relatives aux aspects dosimétriques de
développement, la validation et le contrôle de routine*

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>



Reference number
ISO 11137-3:2017(E)

© ISO 2017

iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 11137-3:2017](https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017)

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms, definitions and symbols	1
3.1 General.....	1
3.2 Symbols.....	3
4 Measurement of dose	4
4.1 General.....	4
4.1.1 Direct and indirect dose measurements.....	4
4.1.2 Dosimetry systems.....	4
4.1.3 Best estimate of dose.....	4
4.2 Dosimetry system selection and calibration.....	5
4.2.1 General.....	5
4.2.2 Selection of dosimetry systems.....	5
4.2.3 Calibration of dosimetry systems.....	5
4.3 Dose measurement uncertainty.....	6
4.3.1 General concepts.....	6
4.3.2 The Guide to the expression of uncertainty in measurement (GUM) methodology.....	6
4.3.3 Radiation sterilization specific aspects of dose measurement uncertainty.....	7
5 Establishing the maximum acceptable dose	8
6 Establishing the sterilization dose	9
7 Installation qualification	10
8 Operational qualification	11
8.1 General.....	11
8.2 Gamma irradiators.....	11
8.3 Electron beam irradiators.....	13
8.4 X-ray irradiators.....	15
9 Performance qualification	17
9.1 General.....	17
9.2 Gamma irradiators.....	18
9.2.1 Loading pattern.....	18
9.2.2 Dosimetry.....	19
9.2.3 Analysis of dose mapping data.....	20
9.3 Electron beam irradiators.....	20
9.3.1 Loading pattern.....	20
9.3.2 Dosimetry.....	22
9.3.3 Analysis of dose mapping data.....	23
9.4 X-ray irradiators.....	23
9.4.1 Loading pattern.....	23
9.4.2 Dosimetry.....	24
9.4.3 Analysis of dose mapping data.....	25
10 Routine monitoring and control	25
10.1 General.....	25
10.2 Frequency of dose measurements.....	26
Annex A (informative) Mathematical modelling	27
Annex B (informative) Tables of references for dosimetry-related testing during IQ/OQ/PQ	30

ISO 11137-3:2017(E)

Annex C (informative) Tolerances associated with doses used in sterilization dose setting/ substantiation in ISO 11137-2 and ISO/TS 13004	33
Annex D (informative) Application of dose measurement uncertainty in setting process target doses	34
Bibliography	40

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[SIST EN ISO 11137-3:2017](https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017)

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical committee ISO/TC 198, *Sterilization of health care products*.

This second edition cancels and replaces the first edition (ISO 11137-3:2006), which has been technically revised.

A list of all parts in the ISO 11137 series can be found on the ISO website.

ISO 11137-3:2017(E)**Introduction**

An integral part of radiation sterilization is the ability to measure dose. Dose is measured during all stages of development, validation and routine monitoring of the sterilization process. It has to be demonstrated that dose measurement is traceable to a national or an International Standard, that the uncertainty of measurement is known, and that the influence of temperature, humidity and other environmental considerations on dosimeter response is known and taken into account. Process parameters are established and applied based on dose measurements. This document provides guidance on the use of dose measurements (dosimetry) during all stages in the development, validation and routine control of the radiation sterilization process.

Requirements in regard to dosimetry are given in ISO 11137-1 and ISO 11137-2 and ISO/TS 13004. This document gives guidance to these requirements. The guidance given is not normative and is not provided as a checklist for auditors. The guidance provides explanations and methods that are regarded as being suitable means for complying with the requirements. Methods other than those given in the guidance may be used, if they are effective in achieving compliance with the requirements of ISO 11137-1, ISO 11137-2 and ISO/TS 13004.

**iTeh STANDARD PREVIEW
(standards.iteh.ai)**

[SIST EN ISO 11137-3:2017](https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017)

<https://standards.iteh.ai/catalog/standards/sist/537e1323-d0b0-4b40-8b94-d6f1eda83a51/sist-en-iso-11137-3-2017>

Sterilization of health care products — Radiation —

Part 3:

Guidance on dosimetric aspects of development, validation and routine control

1 Scope

This document gives guidance on meeting the requirements in ISO 11137-1 and ISO 11137-2 and in ISO/TS 13004 relating to dosimetry and its use in development, validation and routine control of a radiation sterilization process.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11137-1, *Sterilization of health care products — Radiation — Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices*

ISO 11137-2, *Sterilization of health care products — Radiation — Part 2: Establishing the sterilization dose*

ISO/TS 13004, *Sterilization of health care products — Radiation — Substantiation of a selected sterilization dose: Method VD_{max}^{SD}*

ISO 13485, *Medical devices — Quality management systems — Requirements for regulatory purposes*

3 Terms, definitions and symbols

For the purposes of this document, the terms and definitions given in ISO 11137-1 and ISO 11137-2 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1 General

3.1.1

absorbed dose dose

quantity of ionizing radiation energy imparted per unit mass of a specified material

[SOURCE: ISO 11137-1:2006, 3.1, modified]

Note 1 to entry: For the purposes of this document, the term “dose” is used to mean “absorbed dose”.

ISO 11137-3:2017(E)

3.1.2

combined standard measurement uncertainty

standard measurement uncertainty (3.1.13) that is obtained using the individual standard measurement uncertainties associated with the input quantities in a measurement model

[SOURCE: VIM 2012, 2.31]

3.1.3

coverage factor

number larger than one by which a *combined standard measurement uncertainty* (3.1.2) is multiplied to obtain an *expanded measurement uncertainty* (3.1.7)

Note 1 to entry: A coverage factor is usually symbolized as “*k*” (see also the GUM:1995, 2.3.6).

3.1.4

direct dose measurement

measurement of *absorbed dose* (3.1.1) with a dosimeter at the location of interest

Note 1 to entry: For example, a direct measurement of minimum dose is made with a dosimeter at the minimum dose location in an irradiation container.

3.1.5

dose uniformity ratio

ratio of the maximum to the minimum *absorbed dose* (3.1.1) within the irradiation container

3.1.6

dosimetry system

interrelated elements used for determining *absorbed dose* (3.1.1), including dosimeters, instruments, associated reference standards and procedures for their use

[SOURCE: ISO/TS 11139:2006, 2.15]

3.1.7

expanded measurement uncertainty

product of a *combined standard measurement uncertainty* (3.1.2) and a factor larger than the number one

Note 1 to entry: The factor depends on the type of probability distribution of the output quantity in a measurement model and on the selected coverage probability.

Note 2 to entry: The term “factor” in this definition refers to a coverage factor.

3.1.8

indirect dose measurement

measurement of *absorbed dose* (3.1.1) at a location remote from a directly measured dosimeter, calculated by the application of factors

Note 1 to entry: For example, where the minimum dose in an irradiation container cannot easily be measured directly, a dosimeter placed in a remote location may be measured and factors applied to that measurement to calculate the minimum dose.

3.1.9

scan length

dimension of the irradiation zone, perpendicular to the scan width and direction of the electron beam at a specified distance from the accelerator window

Note 1 to entry: ISO/ASTM standards use “beam length” to mean the same thing that “scan length” means in this document. This document uses “scan length” for consistency with ISO 11137-1.

3.1.10 scan width

dimension of the irradiation zone in the direction that the beam is scanned, perpendicular to the scan length and direction of the electron beam at a specified distance from the accelerator window

Note 1 to entry: ISO/ASTM standards use “beam width” to mean the same thing that “scan width” means in this document.

3.1.11 simulated product

material with attenuation and scattering properties similar to those of the product, material or substance to be irradiated

Note 1 to entry: Simulated product is used as a substitute for the actual product, material or substance to be irradiated. When used in routine production runs in order to compensate for the absence of product, simulated product is sometimes referred to as compensating dummy. When used for absorbed dose mapping, simulated product is sometimes referred to as “phantom material”.

Note 2 to entry: In this document, “dose mapping” is used for “absorbed dose mapping.”

3.1.12 spatial resolution

resolution in two dimensions

Note 1 to entry: Ability to detect change in dose in two dimensions.

3.1.13 standard measurement uncertainty

uncertainty of the result of a measurement expressed as a standard deviation

[SOURCE: VIM 2012, 2.30, modified]

3.1.14 uncertainty budget

statement of a measurement uncertainty, of the components of that measurement uncertainty, and of their calculation and combination

Note 1 to entry: An uncertainty budget should include the measurement model, estimates and measurement uncertainties associated with the quantities in the measurement model, covariances, type of applied probability density functions, degrees of freedom, type of evaluation of measurement uncertainty and any coverage factor.

[SOURCE: VIM 2012, 2.33]

3.2 Symbols

Symbol	Meaning
$D_{\max,acc}$	maximum acceptable dose determined in accordance with ISO 11137-1:2006, 8.1
D_{ster}	sterilization dose determined in accordance with ISO 11137-1:2006, 8.2
D_{\max}	direct measurement of maximum dose in a given irradiation container
D_{\min}	direct measurement of minimum dose in a given irradiation container
D_{mon}	direct measurement of dose at the routine monitoring position
$R_{\max/min}$	ratio of maximum to minimum dose (D_{\max}/D_{\min}) determined by dose mapping