

---

---

**Sterilization of health care products —  
Chemical indicators —**

Part 5:

**Class 2 indicators for Bowie and Dick-  
type air removal tests**

*Stérilisation des produits de santé — Indicateurs chimiques —*

*Partie 5: Indicateurs de Classe 2 pour l'essai de Bowie et Dick  
d'enlèvement d'air*

ISO 11140-5:2007

<https://standards.iteh.ai/catalog/standards/iso/cfe94f3d-bb1c-4753-b168-bfc8fb95a0a7/iso-11140-5-2007>



Reference number  
ISO 11140-5:2007(E)

**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

ISO 11140-5:2007

<https://standards.iteh.ai/catalog/standards/iso/cfe94f3d-bb1c-4753-b168-bfc8fb95a0a7/iso-11140-5-2007>

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 General requirements</b> .....	<b>2</b>
<b>5 Indicator system</b> .....	<b>2</b>
<b>5.1 Format</b> .....	<b>2</b>
<b>5.2 Performance</b> .....	<b>2</b>
<b>6 Indicator</b> .....	<b>3</b>
<b>6.1 Format</b> .....	<b>3</b>
<b>6.2 Performance</b> .....	<b>3</b>
<b>7 Packaging and labelling</b> .....	<b>4</b>
<b>8 Quality assurance</b> .....	<b>4</b>
<b>9 Sampling conditioning</b> .....	<b>4</b>
<b>Annex A (normative) Determination of the degree of contrast between the colour of the substrate and the indicator agent</b> .....	<b>5</b>
<b>Annex B (normative) Method of determining uniform colour change on exposure to saturated steam</b> .....	<b>6</b>
<b>Annex C (normative) Method for determining indicator colour change on exposure to dry heat</b> .....	<b>7</b>
<b>Annex D (normative) Method for determining transfer of indicator agent to standard test pack</b> .....	<b>8</b>
<b>Annex E (normative) Standard test pack</b> .....	<b>9</b>
<b>Annex F (normative) Method of determining non-uniform colour change on exposure to a standard fault condition</b> .....	<b>10</b>
<b>Annex G (normative) Steam exposure apparatus</b> .....	<b>11</b>
<b>Bibliography</b> .....	<b>12</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11140-5 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This second edition cancels and replaces the first edition (ISO 11140-5:2000), which has been technically revised.

ISO 11140 consists of the following parts, under the general title *Sterilization of health care products — Chemical indicators*:

- *Part 1: General requirements*
- *Part 3: Class 2 indicator systems for use in the Bowie and Dick-type steam penetration test*
- *Part 4: Class 2 indicators as an alternative to the Bowie and Dick-type test for detection of steam penetration*
- *Part 5: Class 2 indicators for Bowie and Dick-type air removal tests*

## Introduction

The air removal test is used to evaluate the efficacy of air removal during the pre-vacuum phase of a pre-vacuum sterilization cycle or during the pulsing stage of positive pulsing cycles if non-condensable gases were present in the steam. Retention of air due to an inefficient air removal stage or the presence of an air leak or non-condensable gases during the air removal stage are circumstances which can lead to failure of the test. This part of ISO 11140 describes the requirements for Class 2 indicators for Bowie and Dick-type air removal test sheets and packs.

For a description of the classes of chemical indicators, see ISO 11140-1.

The difference between the steam penetration test (ISO 11140-3 and ISO 11140-4) and the air removal test (ISO 11140-5) is described in the chemical indicator guidance document (ISO 15882).

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 11140-5:2007](https://standards.iteh.ai/catalog/standards/iso/cfe94f3d-bb1c-4753-b168-bfc8fb95a0a7/iso-11140-5-2007)

<https://standards.iteh.ai/catalog/standards/iso/cfe94f3d-bb1c-4753-b168-bfc8fb95a0a7/iso-11140-5-2007>



# Sterilization of health care products — Chemical indicators —

## Part 5:

## Class 2 indicators for Bowie and Dick-type air removal tests

### 1 Scope

This part of ISO 11140 specifies the requirements for Class 2 indicators for Bowie and Dick-type air removal tests used to evaluate the effectiveness of air removal during the pre-vacuum phase of pre-vacuum steam sterilization cycles.

Additionally, this part of ISO 11140 includes test methods and equipment used to meet these performance requirements.

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5-4:1995, *Photography — Density measurements — Part 4: Geometric conditions for reflection density*

ISO 5636-3, *Paper and board — Determination of air permeance (medium range) — Part 3: Bendtsen method*

ISO 11140-1:2005, *Sterilization of health care products — Chemical indicators — Part 1: General requirements*

IEC 60584-2:1982, *Thermocouples. Part 2: Tolerances*

IEC 60751:1983, *Industrial platinum resistance thermometer sensors*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11140-1 and the following apply.

#### 3.1

##### **indicator system**

combination of the indicator agent and its substrate subsequently intended to be used in combination with a specific test load

NOTE For the purposes of this document, the specific test load is the standard test pack as defined in Annex E.

#### 3.2

##### **indicator**

combination of the indicator agent and its substrate in the final form in which it is intended to be used

See Annex E of ISO 11140-1:2005.

NOTE The indicator may be user-assembled or pre-assembled. The test load may be disposable, for limited re-use or re-usable.

**3.3**  
**equilibration time**  
period which elapses between the attainment of the sterilization temperature in the sterilizer chamber and the attainment of the sterilization temperature in all parts of the load

## 4 General requirements

Unless otherwise specified in this part of ISO 11140, the general requirements given in ISO 11140-1 apply.

## 5 Indicator system

### 5.1 Format

**5.1.1** The indicator agent shall be uniformly distributed on its substrate to cover not less than 30 % of one side of the substrate.

The pattern of indicator agent distribution should allow easy judgement of the uniformity of the colour change.

**5.1.2** The substrate of the indicator system shall have a uniform ground colour which provides a difference in colour density of not less than 0,3 between ground colour and either the changed or the unchanged indicator agent, as specified by the manufacturer, when the difference in colour density is determined using a reflective densitometer.

Compliance shall be tested in accordance with Annex A.

**5.1.3** The indicator system dimensions shall be  $(200 \pm 20)$  mm  $\times$   $(275 \pm 25)$  mm.

**5.1.4** The indicator system shall have an air porosity not less than  $1,7 \mu\text{m}/(\text{Pa}\cdot\text{s})$  when tested in accordance with ISO 5636-3 at an air pressure of 1,47 kPa.

### 5.2 Performance

**5.2.1** The indicator system shall show a uniform colour change (as specified by the manufacturer) after exposure to saturated steam at  $134^\circ\text{C}$  for  $3,5 \text{ min} \pm 5 \text{ s}$  and/or exposure to saturated steam at  $121^\circ\text{C}$  for  $15 \text{ min} \pm 5 \text{ s}$ , or such other combinations of time and temperature as the manufacturer shall specify for the intended use of product. In all cases, the permitted tolerance on the test temperature shall be  $^{+1,5}_0^\circ\text{C}$  and the time given shall be the time within which the colour change shall occur.

Compliance shall be tested in accordance with Annex B.

**5.2.2** When placed in the centre of the standard test pack as specified in Annex E, the indicator system shall show a non-uniform colour change when the temperature at the centre of the standard test pack is  $(2^{+1}_0)^\circ\text{C}$  lower than the temperature of the chamber drain at the beginning of the final 1 min of a 3,5 min cycle at  $134^\circ\text{C}$ , or at the beginning of the final 5 min of a 15 min cycle at  $121^\circ\text{C}$  of the exposure phase of the steam exposure apparatus. This is the standard fault condition generated by inadequate air removal from the chamber. Any other combination of time and temperature stated by the manufacturer shall exhibit a similar response at the beginning of the final 30 % of the exposure time.

Compliance shall be tested in accordance with Annex F.

**5.2.3** After exposure to dry heat at  $(140 \pm 2)^\circ\text{C}$  for not less than 30 min, the indicator agent shall show either no change or a change that is markedly different from the change occurring after exposure to a steam sterilization process.

Compliance shall be tested in accordance with Annex C.



**5.2.4** Transfer of the indicator agent to the material of the standard test pack in which it is intended to be used shall not compromise the result of the test.

Compliance shall be tested in accordance with Annex D.

NOTE Although some transfer can be possible without adversely affecting the performance of the indicator system or test pack, there are currently no test methods available to verify the acceptable limits of indicator agent transfer.

**5.2.5** The indicator system shall comply with the requirements of this part of ISO 11140 for the duration of the shelf life specified by the manufacturer.

A documented accelerated ageing procedure may be used in demonstrating compliance.

## 6 Indicator

### 6.1 Format

**6.1.1** The indicator agent shall be uniformly distributed on its substrate to cover not less than 30 % of the test area of the substrate.

**6.1.2** The substrate of the indicator system shall have a uniform ground colour which provides a difference in colour density of not less than 0,3 between ground colour and either the changed or the unchanged indicator agent, as specified by the manufacturer, when the difference in colour density is determined using a reflective densitometer.

Compliance shall be tested in accordance with Annex A.

### 6.2 Performance

**6.2.1** The indicator system shall show a uniform colour change complying with 6.1.2 both after exposure to saturated steam at  $(134^{+1,5}_0)$  °C for 3,5 min ± 5 s and/or exposure to saturated steam at  $(121^{+1,5}_0)$  °C for 15 min ± 5 s, or such other combinations of time and temperature as the manufacturer shall specify for the intended use of product. In all cases, the permitted tolerance on the test temperature shall be  $^{+1,5}_0$  °C and the time given shall be the time within which the colour change shall occur.

Compliance shall be tested in accordance with Annex B.

**6.2.2** After exposure to conditions used to produce a standard fault condition (see 5.2.2) in the standard test pack as specified in Annex E, the indicator system shall show a non-uniform colour change.

Compliance shall be tested in accordance with Annex F.

**6.2.3** After exposure to dry heat at  $(140 \pm 2)$  °C for not less than 30 min, the indicator agent shall show either no change or a change that is markedly different from the change occurring after exposure to a steam sterilization process. Compliance shall be tested in accordance with Annex C.

**6.2.4** Transfer of the indicator agent to the test load shall not compromise the result of the test.

**6.2.5** The indicator shall comply with the requirements of this part of ISO 11140 for the duration of the shelf life specified by the manufacturer.

A documented accelerated ageing procedure may be used in demonstrating compliance.

## 7 Packaging and labelling

7.1 The general requirements in ISO 11140-1 apply.

7.2 In addition, each indicator, indicator system and its packaging shall be clearly marked with the following.

AIR REMOVAL
-------------

## 8 Quality assurance

The general requirements in ISO 11140-1 apply.

## 9 Sampling conditioning

Test samples shall be conditioned for at least 1 h immediately prior to testing in an environment between  $(23 \pm 7)$  °C with a relative humidity of 30 % to 70 %.

iTeh Standards  
(<https://standards.iteh.ai>)  
Document Preview

[ISO 11140-5:2007](https://standards.iteh.ai/catalog/standards/iso/cfe94f3d-bb1c-4753-b168-bfc8fb95a0a7/iso-11140-5-2007)

<https://standards.iteh.ai/catalog/standards/iso/cfe94f3d-bb1c-4753-b168-bfc8fb95a0a7/iso-11140-5-2007>