INTERNATIONAL STANDARD



First edition 2007-06-01

Vacuum technology — Standard methods for measuring vacuum-pump performance — General description

Technique du vide — Méthodes normalisées pour mesurer les performances des pompes à vide — Description générale

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 21360:2007 https://standards.iteh.ai/catalog/standards/sist/25b015e4-a9bd-49eb-9c19a2dbd2709ce9/iso-21360-2007



Reference number ISO 21360:2007(E)

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 21360:2007 https://standards.iteh.ai/catalog/standards/sist/25b015e4-a9bd-49eb-9c19a2dbd2709ce9/iso-21360-2007



COPYRIGHT PROTECTED DOCUMENT

© ISO 2007

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.org Web www.iso.org Published in Switzerland

Contents

Forewo	ord	. iv
Introdu	uction	v
1	Scope	1
2	Normative references	1
3	Terms and definitions	2
4	Symbols and abbreviated terms	3
5 5.1 5.2 5.3 5.4 5.5	Test Methods Volume flow rate (pumping speed) measurement with the throughput method Volume flow rate (pumping speed) measurement with the orifice method Volume flow rate (pumping speed) measurement with the pump-down method Measurement of the base pressure Measurement of the compression ratio and the critical backing pressure	4 4 8 13 18 19
Annex	A (informative) Mean free path of some important gases	23
Annex Bibliog	B (informative) Measuring uncertainties Iraphy (standards.iteh.ai)	24 27

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21360 was prepared by Technical Committee ISO/TC 112, Vacuum technology.

iTeh STANDARD PREVIEW (standards.iteh.ai)

Introduction

This International Standard is a basic standard for measuring the performance data of vacuum pumps. The methods described here are well known from existing national and international standards. In developing this International Standard, the aim was to provide one single document containing the measurements of performance data of vacuum pumps and to simplify the future development of specific vacuum pump standards.

Specific vacuum pump standards will contain a suitable selection of measurement methods from this basic standard in order to determine the performance data, limiting values and specific operational conditions on the basis of the specific properties of the particular kind of pump. Whenever a discrepancy exists between the basic standard and the specific standard, it is the specific standard which is valid.

iTeh STANDARD PREVIEW (standards.iteh.ai)

iTeh STANDARD PREVIEW (standards.iteh.ai)

Vacuum technology — Standard methods for measuring vacuum-pump performance — General description

1 Scope

This International Standard is a basic standard which defines three different methods for measuring the volume flow rate and one method each for measuring the base pressure, the compression ratio and the critical backing pressure of a vacuum pump.

The first method for measuring the volume flow rate (the throughput method) is the basic concept, in which a steady gas flow is injected into the pump while the inlet pressure is measured. In practice, the measurement of gas throughput may be complicated or inexact. For this reason, two other methods are described which avoid the direct measurement of throughput.

The second method for measuring the volume flow rate (the orifice method) is used when there is very small throughput occurring at very small inlet pressures (in the high and ultra-high vacuum). It is based on measuring the ratio of pressures in a two-chamber test dome where the two chambers are separated by a wall with a circular orifice.

The third method for measuring the volume flow rate (the pump-down method) is well-suited for automated measurement. It is based on the evacuation of a large vessel. The volume flow rate is calculated from two pressures, before and after a pumping interval, and from the volume of the test dome. Different effects, such as leak and desorption rates das cooling by nearly isentropic expansion during the pumping interval and increasing flow resistance in the connection line between test dome and pump caused by molecular flow at low pressures, influence the results of the pressure measurement and the resulting volume flow rate.

The choice of the required measurement methods depends on the properties of the specific kinds of vacuum pump, e.g. the measurement of the critical backing pressure is only necessary for vacuum pumps which need a backing pump. All data that is measured on a vacuum pump but not described in this International Standard (e.g. measurement of power consumption) is defined in the specific pump standard.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3529-1, Vacuum technology — Vocabulary — Part 1: General terms

ISO 3529-2, Vacuum technology — Vocabulary — Part 2: Vacuum pumps and related terms

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 3529-2 and the following apply.

3.1

volume flow rate

 q_V

volume of gas which, under ideal conditions, flows from the test dome through the pump inlet per unit time

NOTE 1 For practical reasons, however, the volume flow rate of a given pump and for a given gas is conventionally considered to be equal to the quotient of the throughput of this gas and of the equilibrium pressure at a given location. The units adopted for the volume flow rate are cubic metres per hour (m^3/h) or litres per second (l/s).

NOTE 2 The term "pumping speed" and symbol "*S*" are often used instead of "volume flow rate".

3.2

inlet pressure

 p_{1}, p_{d}, p_{e}

pressure at the inlet of the pump, measured at a defined location in the test dome

3.3

base pressure

 p_{b}

pressure obtained in the test dome after conditioning the vacuum pump and the test dome

See 5.4.

iTeh STANDARD PREVIEW

NOTE The ultimate pressure is the value which the pressure in the test dome approaches asymptotically. It is the lowest pressure obtainable with the pump, but there is no practical method of measurement or specification.

ISO 21360:2007

3.4 https://standards.iteh.ai/catalog/standards/sist/25b015e4-a9bd-49eb-9c19maximum working pressure a2dbd2709ce9/iso-21360-2007

p_{1max}

highest pressure on the inlet side that the vacuum pump and the driving device can withstand for a prolonged period of operation time without being damaged

3.5

backing pressure

 p_3

pressure at the outlet of a vacuum pump

3.6

critical backing pressure

 p_{C}

maximum backing pressure for which the conditions are defined in the instruction manual or in a specific standard for the particular vacuum pump

3.7

compression ratio

K_0

ratio of the backing pressure, p_3 , to the inlet pressure, p_1 , of the vacuum pump without throughput, expressed by the following equation:

$$K_0 = \frac{p_3}{p_1}$$

3.8

test dome

special vacuum vessel with precisely defined size, diameter and connection flanges on specified locations, used for standard performance data measurements on vacuum pumps

3.9

throughput *Q*

 \tilde{a} mount of gas flowing through a duct, expressed by the following equation:

$$Q = \frac{p_1 V}{t} = p_1 q_V$$

where

- p_1 is the (high) vacuum pressure on the inlet;
- q_V is the volume flow rate of the test pump;
- t is time;
- *V* is the volume of the test dome.

3.10

standard gas flow rate. Teh STANDARD PREVIEW

volume flow rate at standard reference conditions for gases in accordance with ISO 3529-1

4 Symbols and abbreviated terms SO 21360:2007

https://standards.iteh.ai/catalog/standards/sist/25b015e4-a9bd-49eb-9c19-

	a2dbd2709ce9/iso-21360-2007	
Symbol	Designation	Unit
а	inner diameter of the connection pipe between test pump and quick-acting valve (items 3 and 5 in Figure 6)	m
Α	cross-section of the connection pipe between test pump and quick-acting valve (items 3 and 5 in Figure 6)	m ²
С	conductance	m ³ /s (= 10 ³ l/s)
d	diameter of orifice	m
D	inner diameter of test dome	m
D_{N}	nominal diameter of test dome	m
<i>K</i> ₀	compression ratio of vacuum pump with zero throughput	_
l	length of the connection pipe between test pump and quick-acting valve (items 3 and 5 in Figure 6)	m
ī	mean free path	m
L	thickness of the orifice wall at the orifice diameter	m
М	molar mass of gas	kg/mol
<i>p</i> ₀	101 325 Pa in accordance with ISO 3529-1	Pa
<i>p</i> ₁	(high) vacuum pressure on inlet	Pa (or mbar)
<i>p</i> _{1max}	maximum working pressure on inlet	Pa (or mbar)

ISO 21360:2007(E)

<i>p</i> ₃	vacuum pressure in backing line	Pa (or mbar)
<i>p</i> _{t1} , <i>p</i> _{t2} , <i>p</i> _{t3}	pressures in the test dome for the pump-down method, measured before and after time intervals Δt_1 , Δt_2 , Δt_3	Pa (or mbar)
p_{b1}, p_{b2}, p_{b3}	base pressures	Pa (or mbar)
p_{c}	critical backing pressure	Pa (or mbar)
p_{d}, p_{e}	pressures in the test dome for the orifice method	Pa (or mbar)
Q	gas throughput of vacuum pump	Pa·l/s (or mbar·l/s)
Q_{r}	test gas load	Pa·l/s (or mbar·l/s)
q_V	volume flow rate of test pump	l/s (or m ³ /h)
q_{VB}	volume flow rate of backing pump	l/s (or m ³ /h)
$q_{V m sccm}$	volume flow rate at standard reference conditions for gases in accordance with ISO 3529-1	sccm (or cm ³ /min)
$q_{V \text{std}}$	volume flow rate at standard reference conditions for gases in accordance with ISO 3529-1	l/s (or m ³ /h)
\mathcal{Q}_{max}	maximum gas throughput of vacuum pump which the pump can withstand without damage	Pa·l/s (or mbar·l/s)
R	ideal gas constant	8,314 J/(mol·K)
Т	thermodynamic temperatures TANDARD PREV	κw
T ₀	273,15 K in accordance with ISO 3529-1 ards.iteh.ai)	К
T_{D}	temperature of the test dome	К
T_{f}	temperature of the flow meter https://standards.iteh.ai/catalog/standards/sist/25b015e4-a9bd	-49eb-9c19-
u	measurement uncertainty a2dbd2709ce9/iso-21360-2007	_
V	volume of the test dome	l, m ³
V _i	volume of connection pipe between test pump and quick-acting valve (items 3 and 5 in Figure 6)	l, m ³

5 Test Methods

5.1 Volume flow rate (pumping speed) measurement with the throughput method

5.1.1 General

The throughput method is the one most used for vacuum pumps and is applicable for all pressure ranges and pump sizes where flow meters for the gas throughput measurements are available with sufficient accuracy. The gas flow measuring ranges shall be chosen by multiplying the expected volume flow rate with the maximum and minimum working pressure of the test pump.

All measuring devices shall be calibrated either:

- a) in a traceable way to a vacuum primary or to a national standard, or
- b) by means of instruments of absolute measure which are traceable to the SI units and to which measurement uncertainties can be attributed.

In the case of calibrated measuring instruments, there should exist a calibration certificate in accordance with ISO/IEC 17025.

5.1.2 Test dome for the throughput method

For these measurements, use a test dome as shown in Figure 1 with the same nominal diameter, D_N , as that of the pump inlet. The face of the dome opposite the inlet flange may be flat, conical or slightly curved, with the same average height above the flange as the flat face. Three flanges for pressure measurement at the height of D/2 above the bottom flange are preferable if more than one pressure gauge is used. The diameter of these flanges should be equal to or larger than the flanges of the gauges used, and their mounting dimensions shall be noted. No measuring port shall be located in the angle range \pm 45° next to a gas inlet port. The connection pipes between flange and dome shall not protrude beyond the dome wall on the inside with the exception of the gas inlet pipe.

If necessary for the test pump, the test dome shall be fitted with a device for bake-out that ensures uniform heating of the dome to achieve the base pressure.

The volume of the test dome may depend on the pump type. Refer to the specific pump standard for details.

For pumps with an inlet flange diameter of less than $D_N = 100$ mm, the diameter of the dome shall correspond to $D_N = 100$ mm. The transition to the pump inlet flange shall be made through a 45° conical adaptor, as shown in Figure 1.



Key

- 1 gas inlet pipe and temperature measuring point for T_{D}
- 2 vacuum gauge and mass spectrometer connections

D inner diameter of test dome, in metres



5.1.3 Experimental setup

See Figure 2.

The test dome shall be clean and dry. The cleanness of the pump, seals and other components shall be appropriate for the expected base pressure. All components are mounted together under clean conditions in accordance with Figure 2. Because of the narrow measuring range, flow meters with different ranges may be switched in series. In the case of a flow restriction by a small flow meter, they may be used in parallel with a manifold and a valve between every flow meter and the manifold. Instead of the flow meter and the gas inlet valve, mass flow controllers with programmable throughputs may be used. They shall be combined in parallel on a manifold.

NOTE The leak tightness of large mass flow controllers is not sufficient in many cases. In such cases, it is advisable that valves be used between flow controller and the manifold.

lonization gauges and mass spectrometers shall be installed in such a way that there is no direct geometrical path between them.

CAUTION — Observe the safety instructions of the manufacturer of the vacuum pump.



Key

- 1 test dome
- 2 backing pump
- 3 test pump
- 4 gas inlet valve
- 5 flow meters to measure Q
- 6 vacuum gauge to measure p_1
- 7 heating jacket (optional)
- 8 vacuum gauge to measure p_3
- 9 temperature measuring point for $T_{\rm D}$

NOTE Items 2 and 8 are only used in connection with high vacuum test pumps.

Figure 2 — Arrangement for measuring volume flow rate (pumping speed) with throughput method